

INTERNATIONAL
STANDARD

IEC
CEI

NORME
INTERNATIONALE

60601-2-22

Third edition
Troisième édition
2007-05

Medical electrical equipment –

Part 2-22:

Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment

Appareils électromédicaux –

Partie 2-22:

Règles particulières pour la sécurité de base et les performances essentielles des appareils chirurgicaux, esthétiques, thérapeutiques et de diagnostic à laser



Reference number
Numéro de référence
IEC/CEI 60601-2-22:2007



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2007 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us.

Email: csc@iec.ch
Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch
Tél.: +41 22 919 02 11
Fax: +41 22 919 03 00

INTERNATIONAL
STANDARD

IEC
CEI

NORME
INTERNATIONALE

60601-2-22

Third edition
Troisième édition
2007-05

Medical electrical equipment –

Part 2-22:

Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment

Appareils électromédicaux –

Partie 2-22:

Règles particulières pour la sécurité de base et les performances essentielles des appareils chirurgicaux, esthétiques, thérapeutiques et de diagnostic à laser



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE
CODE PRIX

U

*For price, see current catalogue
Pour prix, voir catalogue en vigueur*

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
201.1 Scope, object and related standards.....	6
201.2 Normative references	8
201.3 Terms and definitions	8
201.4 General requirements.....	10
201.5 General requirements for testing ME EQUIPMENT.....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	10
201.7 ME EQUIPMENT identification, marking and documents.....	10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	12
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	13
201.10 Protection against unwanted and excessive radiation HAZARDS.....	13
201.11 Protection against excessive temperatures and other HAZARDS.....	15
201.12 Accuracy of controls and instruments and protection against hazardous outputs ...	15
201.13 HAZARDOUS SITUATIONS and fault conditions.....	16
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	18
201.15 Construction of ME EQUIPMENT.....	18
201.16 ME SYSTEMS.....	18
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	19
Annexes	19
Annex D (informative) Symbols on marking.....	19
Annex AA (informative) Particular guidance and rationale.....	22
Bibliography.....	24
Index of defined terms used in this particular standard.....	25
Table D.1 – General symbols.....	19

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-22 has been prepared by IEC subcommittee 76: Optical radiation safety and laser equipment.

This third edition cancels and replaces the second edition of IEC 60601-2-22, published in 1995. This edition constitutes a technical revision.

This third edition takes account of the recently published new editions of the General Standard IEC 60601-1 and Group safety publication IEC 60825-1. Additionally, it addresses technical and safety issues which have arisen in the time following the previous second edition.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
76/359/FDIS	76/363/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes Subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard amends and supplements IEC 60601-1 (third edition, 2005: *Medical Electrical Equipment – Part 1: General requirements for basic safety and essential performance*).

This standard also refers to IEC 60825-1 (2007).

The requirements of this standard are the minimum that need to be complied with, in order to achieve a reasonable level of safety and reliability during operation and application of medical laser equipment.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA. Understanding of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revisions necessitated by changes in clinical practice or by developments in technology.

iTech Standards
(<https://standards.iteh.ai>)
Document Preview

<https://standards.iteh.ai/standards/iec/19326332-6c87-41ff-bf6c-04e3c5304f14/iec-60601-2-22-2007>

<https://standards.iteh.ai/standards/iec/19326332-6c87-41ff-bf6c-04e3c5304f14/iec-60601-2-22-2007>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment

201.1 Scope, object and related standards

Clause 1 of the General Standard applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of laser equipment for either surgical, therapeutic, medical diagnostic, cosmetic, or veterinary applications, intended for its use on humans or animals, classified as a CLASS 3B or CLASS 4 LASER PRODUCT as defined by 3.22 and 3.23 in IEC 60825-1, hereafter referred to as LASER EQUIPMENT.

Throughout this International Standard, light emitting diodes (LED) are included whenever the word “laser” is used.

NOTE 1 Refer to Definition 3.49 in IEC 60825-1.

NOTE 2 Laser products for these applications classified as a CLASS 1, 1M, 2, 2M or CLASS 3R LASER PRODUCT, are covered by IEC 60825-1 and IEC 60601-1.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the General Standard.

NOTE See also 4.2 of the General Standard.

This standard can also be applied to surgical, cosmetic, therapeutic and diagnostic laser equipment used for compensation or alleviation of disease, injury or disability.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for the safety of surgical, cosmetic, therapeutic and diagnostic laser equipment.

NOTE Laser classification (IEC 60825-1) must not be confused with electrical classification (IEC 60601-1).

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the General Standard and Clause 2 of this particular standard.

IEC 60601-1-3 does not apply.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in this standard as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the General Standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the General Standard. Collateral standards are referred to by their document number.

The numbering of sections, clauses and subclauses of this particular standard corresponds to that of the General Standard or applicable collateral standard. The changes to the text of the General Standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the General Standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the General Standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the General Standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses or figures which are additional to those of the General Standard are numbered starting from 201.101, additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the General Standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding section, clause or subclause in this particular standard, the section, clause or subclause of the General Standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the General Standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

Concerning LASER RADIATION safety of laser equipment, IEC 60825-1 applies, except that the relevant requirements are specified, changed or amended in this particular standard.

Clauses and subclauses of the General Standard and IEC 60825-1, which are not applicable to laser equipment for medical applications, are not necessarily indicated as "not applicable".

201.2 Normative references

Clause 2 of the General Standard applies, except as follows:

Addition:

IEC 60825-1:2007, *Safety of laser products – Part 1: Equipment classification and requirements*

IEC 60947-3, *Low-voltage switchgear and controlgear – Part 3: Switches, disconnectors, switch-disconnectors and fuse-combination units*

IEC 61010-1, *Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, apply, except as follows:

Addition:

201.3.101

ACCESSIBLE EMISSION LIMIT (AEL)

ACCESSIBLE EMISSION LIMIT for CLASS 1M, 2, 2M, 3R, or 3B lasers (see 3.3 and Tables 4 through 9 of IEC 60825-1)

201.3.102

AIMING BEAM

beam of optical radiation, producing a visible AIMING BEAM SPOT, intended for indication of the anticipated point of impact of the WORKING BEAM

201.3.103

AIMING BEAM SPOT

area of impact of the AIMING BEAM within the WORKING AREA

201.3.104

AIMING LASER

LASER emitting an AIMING BEAM

201.3.105

APERTURE

distal opening of the BEAM DELIVERY SYSTEM (see 3.8 of IEC 60825-1)

201.3.106**BEAM DELIVERY SYSTEM**

optical system which guides the LASER RADIATION from its origin to the WORKING AREA

201.3.107**CLASS 1, 1M, 2, 2M, 3R, 3B, OR 4 LASER PRODUCT**

laser equipment, incorporating a LASER as defined in 3.41 and 3.18 through 3.23 of IEC 60825-1

201.3.108**EMERGENCY LASER STOP**

hand- or foot-actuated device intended to stop the LASER OUTPUT immediately in case of emergency

201.3.109**LASER EMISSION CONTROL SWITCH**

hand- or foot-actuated device intended to initiate and stop WORKING BEAM emission through any APERTURE

201.3.110**LASER EMISSION INDICATOR**

visible and/or audible signal which indicates that the WORKING BEAM is being emitted through any APERTURE

NOTE The LASER EMISSION INDICATOR is different from the LASER RADIATION EMISSION warning requirement 4.7 of IEC 60825-1.

201.3.111**LASER ENERGY**

RADIANT ENERGY of the WORKING BEAM, incident on the WORKING AREA (see 3.70 of IEC 60825-1)

201.3.112**LASER OPERATOR**

the person who handles the laser equipment. In general, the LASER OPERATOR controls the delivery of the laser radiation to the WORKING AREA. The LASER OPERATOR may appoint other person(s), who assist with the selection and/or setting of the parameters

Refer to Definition 3.73 in IEC 60601-1.

NOTE The safety requirements in this standard apply to all above persons.

201.3.113**LASER OUTPUT**

either LASER POWER OR LASER ENERGY

201.3.114**LASER POWER**

RADIANT POWER of the WORKING BEAM, incident on the WORKING AREA, see 3.72 of IEC 60825-1

201.3.115**LASER READY INDICATOR**

visible means which indicate that the laser equipment is in the READY condition, and the purpose of which is to make all persons present in the laser area aware of the need to take precautions against hazardous LASER RADIATION, as detailed in the ACCOMPANYING DOCUMENTS (instructions for use). See 201.7.9.

201.3.116

OPERATOR PROTECTIVE FILTER

a moveable or fixed filter which does not allow radiation in excess of the MAXIMUM PERMISSIBLE EXPOSURE (MPE) to the LASER OPERATOR

NOTE For the definition of MPE, see 3.56 of IEC 60825-1.

201.3.117

SHUTTER

electronic, opto-electronic and/or mechanical means which allows or prevents LASER OUTPUT to be emitted from the APERTURE

201.3.118

STAND-BY/READY

modes of operation when mains supply is connected and the mains switch activated, where the STAND-BY mode means that the laser is not capable of emitting the WORKING BEAM even if the laser control switch is activated, and where the READY mode keeps the laser equipment enabled, so that it is capable of emitting LASER OUTPUT when the control switch is activated

201.3.119

TARGET INDICATING DEVICE

an aiming device which designates the position where the WORKING BEAM will perform its surgical, cosmetic, therapeutic or diagnostic purpose

201.3.120

WORKING AREA

area which is intended to be irradiated with WORKING BEAM

201.3.121

WORKING BEAM

beam of LASER RADIATION emitted by the laser equipment for surgical, cosmetic, therapeutic or diagnostic purposes (other than the AIMING BEAM)

201.4 General requirements

Clause 4 of the General Standard applies.

201.5 General requirements for testing ME EQUIPMENT

Clause 5 of the General Standard applies.

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the General Standard applies.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the General Standard applies, except as follows:

201.7.2 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts

Addition:

201.7.2.101 Additional items

See Clause 5 of IEC 60825-1.

a) General

Laser equipment shall carry labels in accordance with 5.5, 5.6, 5.8, 5.9, 5.10, 5.11 of IEC 60825-1, as applicable.

b)* Aperture label

Laser equipment shall have a label positioned as close as practicable to each laser aperture. The label shall be similar to the laser hazard symbol as specified in IEC 60825-1, Figure 1, with the exception that the size can be adjusted as appropriate or alternatively be similar to the label described in 5.7 of IEC 60825-1. Hand-pieces and other applicators are exempt from these requirements. In this case, a label is to be affixed in a prominent position with either:

- a statement that the laser aperture is on the end of the fibre/applicator, or
- a symbol as detailed in Table D.1, number 113.

NOTE The required information can be combined into one single label if the area where the label is to be affixed is suitable.

201.7.9 ACCOMPANYING DOCUMENTS

201.7.9.1 General

Addition:

The ACCOMPANYING DOCUMENTS shall give adequate instructions for proper operation, including clear warnings concerning precautions to avoid possible exposure to hazardous LASER RADIATION.

201.7.9.2 Instructions for use

201.7.9.2.13 Maintenance

Addition:

The instructions for maintenance shall include clear warnings concerning precautions to avoid possible exposure to hazardous laser radiation.

201.7.9.2.101 Laser specific information for the responsible organization and for the LASER OPERATOR

The instructions for use shall include (as applicable):

- information on the NOMINAL OCULAR HAZARD DISTANCE (NOHD) for the laser equipment in NORMAL USE with each appropriate ACCESSORY;
- a statement in SI units of BEAM DIVERGENCE, PULSE DURATION, maximum LASER OUTPUT of the LASER RADIATION, with the magnitudes of the cumulative measurement uncertainty and any expected increase in the measured quantities at any time after manufacture added to the values measured at the time of manufacture;
- legible reproductions (colour optional) of all required LASER labels and hazard warnings affixed to the laser equipment;

- information and guidance for regular calibration of the LASER OUTPUT in accordance with 201.12.1 of this standard. The information shall include a specification for the measuring equipment and frequency of calibration and clarification requirements concerning regular calibration of LASER OUTPUT;
 - a clear indication of all locations of laser APERTURES;
 - a listing of controls, adjustments and procedures for operation and maintenance by the RESPONSIBLE ORGANISATION, including the warning "Caution – Use of controls or adjustments or performance of procedures other than those specified herein may result in hazardous radiation exposure";
 - a description of the BEAM DELIVERY SYSTEMS including the characteristics of the LASER OUTPUT;
 - A note, saying that laser equipment should be protected against unauthorized use, for example by removal of the key from the key switch;
 - a specification for eye protection;
- NOTE Refer to 8.4.5.2 of IEC/TR 60825-14 (2004), Safety of laser products – Part 14: A user's guide.
- a specification for fume and plume extraction, including a cautionary statement: "Caution – Laser fume and/or plume may contain viable tissue particulates";
 - information about the potential hazards when inserting, sharply bending or improperly securing the fibre optics, stating that not following the recommendations of the manufacturer may lead to damage to the fibre or delivery system and/or harm to the patient or LASER OPERATOR;
 - recommendation, for example as follows: "As the AIMING BEAM passes down the same delivery system as the WORKING BEAM, it provides a good means of checking the integrity of the delivery system. If the AIMING BEAM is not present at the distal end of the delivery system, its intensity is reduced or it looks diffused, this is a possible indication of a damaged or malfunctioning delivery system";
 - a warning, for example as follows: "A risk of fire and/or explosion exists when the LASER OUTPUT is used in the presence of flammable materials, solutions or gases, or in an oxygen enriched environment". The high temperatures produced in NORMAL USE of the laser equipment may ignite some materials, for example cotton wool when saturated with oxygen. The solvents of adhesives and flammable solutions used for cleaning and disinfecting should be allowed to evaporate before the laser equipment is used. Attention should also be drawn to the danger of ignition of endogenous gases.

201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of the General Standard applies except as follows:

201.8.5 Separation of parts

Amendment:

If an OEM (Original Equipment Manufacturer) laser and/or OEM laser power supply is incorporated into the laser product and the OEM product complies with IEC 61010-1, then the OEM product is exempt from the differing requirements of IEC 60601-1. This exemption applies to the means of operator protection (MOOP) and to the requirements of Subclauses 8.6, 8.8, and 8.9.

201.8.7 LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS

This subclause of the General Standard applies except as follows: