

IEC 60746-2  
(Second edition – 2003)

**Expression of performance of electrochemical analyzers –  
Part 2: pH value**

**CORRIGENDUM 2**

Page 18

**Table B.2**

*For buffer solution A, replace the value of the Molality by 0,05.*

*For buffer solution A, replace the value of the Mass by 12,620.*

**iTeh Standards**  
**(<https://standards.itih.ai>)**  
**Document Preview**

[IEC 60746-2:2003/COR2:2003](https://standards.itih.ai/catalog/standards/iec/df3f9ba-bcd9-4583-b0c7-930fd1ecd07d/iec-60746-2-2003-cor2-2003)

<https://standards.itih.ai/catalog/standards/iec/df3f9ba-bcd9-4583-b0c7-930fd1ecd07d/iec-60746-2-2003-cor2-2003>