

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



INTERNATIONAL SPECIAL COMMITTEE ON RADIO INTERFERENCE  
COMITÉ INTERNATIONAL SPÉCIAL DES PERTURBATIONS RADIOÉLECTRIQUES

**Specification for radio disturbance and immunity measuring apparatus and methods –**

**Part 1-5: Radio disturbance and immunity measuring apparatus – Specifications and validation procedures for CALTS and REFTS from 30 MHz to 1 000 MHz**

**Spécifications des méthodes et des appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques –**

**Partie 1-5: Appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques – Spécifications et procédures de validation relatives aux CALTS et aux REFTS dans la plage comprise entre 30 MHz et 1 000 MHz**



**THIS PUBLICATION IS COPYRIGHT PROTECTED**  
**Copyright © 2012 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.  
If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

#### Useful links:

IEC publications search - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The advanced search enables you to find IEC publications by a variety of criteria (reference number, text, technical committee,...).

It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available on-line and also once a month by email.

Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary (IEV) on-line.

Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [csc@iec.ch](mailto:csc@iec.ch).

### A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente. un corrigendum ou amendement peut avoir été publié.

#### Liens utiles:

Recherche de publications CEI - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

La recherche avancée vous permet de trouver des publications CEI en utilisant différents critères (numéro de référence, texte, comité d'études,...).

Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Just Published CEI - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications de la CEI. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (VEI) en ligne.

Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [csc@iec.ch](mailto:csc@iec.ch).

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



INTERNATIONAL SPECIAL COMMITTEE ON RADIO INTERFERENCE  
COMITÉ INTERNATIONAL SPÉCIAL DES PERTURBATIONS RADIOÉLECTRIQUES

**Specification for radio disturbance and immunity measuring apparatus and methods –**

**Part 1-5: Radio disturbance and immunity measuring apparatus – Specifications and validation procedures for CALTS and REFTS from 30 MHz to 1 000 MHz**

**Spécifications des méthodes et des appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques –**

**Partie 1-5: Appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques – Spécifications et procédures de validation relatives aux CALTS et aux REFTS dans la plage comprise entre 30 MHz et 1 000 MHz**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 33.100.10; 33.100.20

ISBN 978-2-8322-0185-5

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
TABLE RECAPITULATING CROSS REFERENCES .....	7
1 Scope.....	8
2 Normative references .....	8
3 Definitions .....	9
4 Specifications and validation procedures for <del>a test site to be used to calibrate antennas in the frequency range of</del> CALTS and REFTS from 30 MHz to 1 000 MHz .....	10
4.1 <del>Introduction</del> General.....	10
4.2 Antenna calibration test site (CALTS) specification.....	11
4.3 Test antenna specification.....	12
4.4 Antenna calibration test site validation procedure.....	14
4.5 Antenna calibration test site compliance criteria .....	19
4.6 The validation report .....	23
4.7 Validation of <del>the CALTS for vertical polarization a</del> REFTS.....	25
Annex A (informative) CALTS requirements.....	28
Annex B (informative) Test antenna considerations .....	31
Annex C (informative) Antenna and site attenuation theory.....	37
Annex D (informative) Application of a fixed length dipole ( $30 \text{ MHz} \leq f \leq 80 \text{ MHz}$ ).....	49
Annex E (informative) Pascal Program used in C.1.3.....	50
Annex F (informative) Checklist validation procedure.....	54
Figure 1 – Schematic diagram of the test antenna .....	13
Figure 2 – Adjustment of a telescopic wire element to the length $L_{we}$ .....	13
Figure 3 – Determination $U_{r1}(f)$ or $U_{r2}(f)$ .....	17
Figure 4 – Determination $U_s(f)$ with the wire antennas in their specified position .....	17
Figure 5 – Relation between the quantities used in the SA compliance criterion.....	22
Figure B.1 – Example of a test antenna .....	32
Figure B.2 – Diagram of the measurement of $S_{11}$ and $S_{12}$ and of $S_{22}$ and $S_{21}$ when generator and load are interchanged (in this figure by putting both switches in their alternative position).....	32
Figure B.3 – Schematic diagram of the determination of the insertion loss $A_1(f)$ .....	35
Figure B.4 – Schematic diagram of the determination of the insertion loss $A_2(f)$ .....	35
Figure C.1 – Network model for SA calculations.....	39
Figure C.2 – Equivalent circuit to the network in figure C.1 .....	39
Figure C.3 – Definition of the mutual couplings, feed-terminal voltages and antenna currents of the antennas above the reflecting plane and their images .....	39

Table 1 – Frequency and fixed receiving antenna height data for SA measurements where $h_t = 2$ m and $d = 10$ m .....	16
Table 2 – Maximum tolerances for $d = 10$ m.....	20
Table 3 – Antenna heights .....	26
Table A.1 – Combinations of fixed-length dipole antenna, swept-frequency range and receiving antenna height .....	29
Table C.1 – Numerical example, calculation of $L_a$ , $SA_C$ (see C.1.3.1).....	43
Table C.2 – Numerical example, calculation of $\Delta SA_t$ (see C.1.3.2).....	44
Table C.3 – Numerical example, calculation of $h_{rc}$ and $\Delta h_{rt}$ (see C.1.3.3).....	45
Table C.4 – Numerical example, calculation of $f_c$ and $\Delta f_t$ (see C.1.3.4) .....	45
Table C.5 – Numerical example calculation of $SA_C$ for vertical polarization, $h_t = 2$ m, except $h_t = 2,75$ m at 30 MHz, 35 MHz and 40 MHz .....	47
Table F.1 – Items to be addressed in the CALTS validation report .....	54

iTech Standards  
(<https://standards.iteh.ai>)  
Document Preview

CISPR 16-1-5:2003

<https://standards.iteh.ai/catalog/standards-iec/22c55acb-0367-45fc-ba8b-3dd820a12840/cispr-16-1-5-2003>

WITHDRAWN

INTERNATIONAL ELECTROTECHNICAL COMMISSION  
INTERNATIONAL SPECIAL COMMITTEE ON RADIO INTERFERENCE

**SPECIFICATION FOR RADIO DISTURBANCE AND IMMUNITY  
MEASURING APPARATUS AND METHODS –**

**Part 1-5: Radio disturbance and immunity measuring apparatus –  
~~Antenna calibration test sites for 30 MHz to 1 000 MHz~~  
Specifications and validation procedures for CALTS and REFTS  
from 30 MHz to 1 000 MHz**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

**This consolidated version of CISPR 16-1-5 consists of the first edition (2003) and its amendment 1 (2012) [documents CISPR/A/994A/FDIS and CISPR/A/1004/RVD]. It bears the edition number 1.1.**

**The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.**

International Standard CISPR 16-1-5 has been prepared by CISPR subcommittee A: Radio interference measurements and statistical methods.

This first edition of CISPR 16-1-5, together with CISPR 16-1-1, CISPR 16-1-2, CISPR 16-1-3 and CISPR 16-1-4, cancels and replaces the second edition of CISPR 16-1, published in 1999, amendment 1 (2002) and amendment 2 (2003). It contains the relevant clauses of CISPR 16-1 without technical changes.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

**IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.**

## INTRODUCTION

CISPR 16-1, CISPR 16-2, CISPR 16-3 and CISPR 16-4 have been reorganised into 14 parts, to accommodate growth and easier maintenance. The new parts have also been renumbered. See the list given below.

Old CISPR 16 publications		New CISPR 16 publications	
CISPR 16-1	Radio disturbance and immunity measuring apparatus	→	CISPR 16-1-1 Measuring apparatus
		→	CISPR 16-1-2 Ancillary equipment – Conducted disturbances
		→	CISPR 16-1-3 Ancillary equipment – Disturbance power
		→	CISPR 16-1-4 Ancillary equipment – Radiated disturbances
		→	CISPR 16-1-5 Antenna calibration test sites for 30 MHz to 1 000 MHz
CISPR 16-2	Methods of measurement of disturbances and immunity	→	CISPR 16-2-1 Conducted disturbance measurements
		→	CISPR 16-2-2 Measurement of disturbance power
		→	CISPR 16-2-3 Radiated disturbance measurements
		→	CISPR 16-2-4 Immunity measurements
CISPR 16-3	Reports and recommendations of CISPR	→	CISPR 16-3 CISPR technical reports
		→	CISPR 16-4-1 Uncertainties in standardised EMC tests
		→	CISPR 16-4-2 Measurement instrumentation uncertainty
		→	CISPR 16-4-3 Statistical considerations in the determination of EMC compliance of mass-produced products
CISPR 16-4	Uncertainty in EMC measurements	→	CISPR 16-4-4 Statistics of complaints and a model for the calculation of limits

More specific information on the relation between the 'old' CISPR 16-1 and the present 'new' CISPR 16-1-5 is given in the table after this introduction (TABLE RECAPITULATING CROSS REFERENCES).

Measurement instrumentation specifications are given in five new parts of CISPR 16-1, while the methods of measurement are covered now in four new parts of CISPR 16-2. Various reports with further information and background on CISPR and radio disturbances in general are given in CISPR 16-3. CISPR 16-4 contains information related to uncertainties, statistics and limit modelling.

CISPR 16-1 consists of the following parts, under the general title *Specification for radio disturbance and immunity measuring apparatus and methods – Radio disturbance and immunity measuring apparatus*:

- Part 1-1: Measuring apparatus,
- Part 1-2: Ancillary equipment – Conducted disturbances,
- Part 1-3: Ancillary equipment – Disturbance power,
- Part 1-4: Ancillary equipment – Radiated disturbances,
- Part 1-5: Antenna calibration test sites for 30 MHz to 1 000 MHz.



## TABLE RECAPITULATING CROSS REFERENCES

Second edition of CISPR 16-1  
Clauses, subclauses

First edition of CISPR 16-1-5  
Clauses, subclauses

1  
2  
3  
5.13

1  
2  
3  
4

Annexes

R  
S  
T  
U  
V  
W

Annexes

A  
B  
C  
D  
E  
F

Figures

55, 56, 57, 58, 59  
S.1, S.2, S.3, S.4  
T.1, T.2, T.3

Figures

1, 2, 3, 4, 5  
B.1, B.2, B.3, B.4  
C.1, C.2, C.3

Tables

19, 20

Tables

1, 2

## SPECIFICATION FOR RADIO DISTURBANCE AND IMMUNITY MEASURING APPARATUS AND METHODS –

### Part 1-5: Radio disturbance and immunity measuring apparatus – ~~Antenna calibration test sites for 30 MHz to 1 000 MHz~~ Specifications and validation procedures for CALTS and REFTS from 30 MHz to 1 000 MHz

#### 1 Scope

~~This part of CISPR 16 is designated a basic standard which specifies the requirements for calibration test sites, used to perform antenna calibrations, as well as the test antenna characteristics, calibration site verification procedure and site compliance criteria. Further information on calibration site requirements, test antenna considerations and the theory of antennas and site attenuation is provided in informative annexes.~~

This part of CISPR 16 specifies the requirements for calibration test sites used to perform antenna calibrations and for reference test sites used to measure the antenna-pair reference site attenuation for compliance test site validations. It describes the test antenna characteristics, calibration and reference test site verification procedures and site compliance criteria. Further information on calibration site requirements, test antenna considerations and the theory of antennas and site attenuation is provided in informative annexes.

Measurement instrumentation specifications are given in CISPR 16-1-1 and CISPR 16-1-4. Further information and background on uncertainties in general is given in CISPR 16-4-1, which may be helpful in establishing uncertainty estimates for the calibration processes of antennas.

#### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CISPR 14-1:2000, *Electromagnetic compatibility – Requirements for household appliances, electric tools and similar apparatus – Part 1: Emission*

CISPR 16-1-1:2003, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus*

CISPR 16-1-4:~~2003~~2010, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-4: Radio disturbance and immunity measuring apparatus – ~~Ancillary equipment – Radiated disturbances~~ Antennas and test sites for radiated disturbance measurements*

CISPR 16-4-1:2003, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 4-1: Uncertainties, statistics and limit modelling - Uncertainties in standardised EMC tests*

CISPR 16-4-2:2003, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 4-2: Uncertainties, statistics and limit modelling – Measurement instrumentation uncertainties*

IEC 60050(161):1990, *International Electrotechnical Vocabulary (IEV) – Chapter 161: Electromagnetic compatibility*

*International Vocabulary of Basic and General Terms in Metrology*, International Organization for Standardization, Geneva, 2nd edition, 1993

### 3 Definitions

For the purpose of this section of CISPR 16, the following definitions apply. Also see IEC 60050(161).

#### 3.1

##### **calibration test site (CALTS)**

open-area test site with metallic ground plane and tightly specified site attenuation performance in horizontal ~~and vertical~~ electric field polarization

~~A CALTS is used for determining the free-space antenna factor of an antenna.~~

~~Site attenuation measurements of a CALTS are used for comparison to corresponding site attenuation measurements of a compliance test site, in order to evaluate the performance of the compliance test site~~

NOTE A CALTS is used for the measurement of height dependent AF, and by the standard site method to measure free-space AF. A CALTS can also be validated for vertical polarization via 4.7. See the related definition of a REFTS.

#### 3.2

##### **compliance test site (COMTS)**

environment which assures valid, repeatable measurement results of disturbance field strength from equipment under test for comparison to a compliance limit

#### 3.3

##### **antenna**

that part of a transmitting or receiving system that is designed to radiate or to receive electromagnetic waves in a specified way

NOTE 1 In the context of this standard, the balun is a part of the antenna.

NOTE 2 See also the term "wire antenna".

#### 3.4

##### **balun**

passive electrical network for the transformation from a balanced to an unbalanced transmission line or device or vice versa

#### 3.5

##### **free-space-resonant dipole**

wire antenna consisting of two straight colinear conductors of equal length, placed end to end, separated by a small gap, with each conductor approximately a quarter-wavelength long such that at the specified frequency the input impedance of the wire antenna measured across the gap is pure real when the dipole is located in the free space

NOTE 1 In the context of this standard, this wire antenna connected to the balun is also called the "test antenna".

NOTE 2 This wire antenna is also referred to as "tuned dipole".

**3.6****site attenuation****SA**

~~site attenuation between two specified positions on a test site is the insertion loss determined by a two-port measurement, when a direct electrical connection between the generator output and receiver input is replaced by transmitting and receiving antennae placed at the specified positions~~

minimum site insertion loss measured between two polarization-matched antennas located on a test site when one antenna is moved vertically over a specified height range and the other is set at a fixed height

**3.7****test antenna**

combination of the free-space-resonant dipole and the specified balun

NOTE For the purpose of this standard only.

**3.8****wire antenna**

a specified structure consisting of one or more metallic wires or rods for radiating or receiving electromagnetic waves

NOTE A wire antenna does not contain a balun.

**3.9****site insertion loss**

~~transmission loss between a pair of antennas placed at specified positions on a test site, when a direct electrical connection between the signal generator output and the measuring receiver input is replaced by transmitting and receiving antennas placed at the specified positions~~

**3.10****reference test site****REFTS**

open-area test site with metallic ground plane and tightly specified site attenuation performance in horizontal and vertical electric field polarizations

NOTE Site attenuation measurements of a REFTS are used for comparison to corresponding site attenuation measurements of a COMTS to evaluate the performance of the COMTS.

## **4 Specifications and validation procedures for ~~a test site to be used to calibrate antennas in the frequency range of CALTS and REFTS~~ from 30 MHz to 1 000 MHz**

Clause 5 of CISPR 16-1-4 specifies the requirements for a test site used to make radio disturbance field strength measurements in the frequency range of 30 MHz to 1 000 MHz. Such a test site may not be suitable for calibrating antennas. This clause specifies the requirements and validation procedure for a test site suitable for the calibration of antennas above a conducting, flat metal plane in the frequency range of 30 MHz to 1 000 MHz. A test site meeting these stringent requirements may also be used as a reference test site for comparison purposes in an alternative validation procedure to 5.2.6 of CISPR 16-1-4.

**4.1 Introduction General**

A test site suitable for performing antenna calibration, referred to herein as CALTS, is intended to provide a suitable environment to calibrate an antenna for its free-space antenna factor. This calibration is performed most conveniently above a reflecting plane by using only horizontal polarization. Subclauses 4.3 through 4.6 specify the characteristics of a CALTS, the characteristics of a calculable test antenna and the CALTS verification (validation) procedure and performance criteria. The CALTS validation procedure given in 4.5 requires the

use of a calculable dipole antenna as specified in 4.4, thus creating the possibility of comparing theoretically predicted site-attenuation to measured CALTS performance. Items to be reported in a CALTS validation report are summarized in 4.7. Annex A provides guidance for constructing a CALTS, which complies with validation criteria specified in 4.6.

~~In order for a CALTS to be used as a reference test site (REFSITE) for validating the performance of test sites according to clause 5 of CISPR 16-1-4, some additional requirements need to be specified. Subclause 4.7 specifies the additional characteristics and performance criteria. Test sites specified in clause 5 of CISPR 16-1-4, which are used for demonstrating compliance with radiated emission limits are referred to herein as a compliance test site (COMTS). Validation of a COMTS may be obtained by comparing it to the theoretical site attenuation given in clause 5 of CISPR 16-1-4 (which takes precedence) or by comparing site attenuation measurements of the REFSITE to corresponding site attenuation measurements of the COMTS, using the same measurement set up and equipment (antennas, cables, generator, receiver, etc.).~~

~~In order for a CALTS to be used as a reference test site for validating the performance of test sites according to Clause 5 of CISPR 16-1-4, the CALTS should be verified using both horizontal and vertical antenna polarizations, as described in 4.7, whereupon it qualifies for use as a REFTS. Test sites specified in Clause 5 of CISPR 16-1-4 that are used for demonstrating compliance with radiated emission limits are referred to herein as compliance test sites (COMTS). Validation of a COMTS may be obtained by comparing it to the theoretical site attenuation given in Clause 5 of CISPR 16-1-4, or by comparing the site attenuation measured on a REFTS to the corresponding site attenuation measured on the COMTS, using the same antennas and the same antenna set-up geometry.~~

The annexes to this standard contain informative specifications of a CALTS and of the calculable, free-space-resonant dipole (tuned dipole) to be used in the CALTS validation procedures. They also give a model for calculating theoretical site attenuation, numerical examples and a checklist for the validation procedure.

## 4.2 Antenna calibration test site (CALTS) specification

### 4.2.1 Introduction

The CALTS comprises the following main components:

- a good-conducting flat metal plane (the reflecting plane);
- an electromagnetically obstruction-free area surrounding the reflecting plane.

In addition, the following ancillary equipment is needed:

- two antenna masts carrying the antennas to be used in either the CALTS validation procedure or the antenna calibration procedure;
- the cables to be connected to these antennas; and
- electronic equipment, such as an RF generator and a measuring receiver.

The normative specification for a CALTS is given in 4.2.2, while annex A contains a number of informative specifications as a guidance to construct and place a CALTS in such a way that the validation criteria will normally be met.

### 4.2.2 Normative specification

~~For the calibration of antennas, the CALTS shall comply with the validation criteria given in 4.5.3, i.e.~~

- ~~a) site attenuation at fixed antenna heights, and~~
- ~~b) antenna heights for maximum site attenuation, or for maximum site attenuation, at all frequencies at which the antennas shall be calibrated.~~

~~NOTE 1 In the CALTS validation procedure, equipment is used which is also subject to normative specifications (see 4.3 and 4.4).~~

~~NOTE 2 The CALTS validation report (4.6) will contain information on how compliance with the requirements is maintained, so that the CALTS is deemed to comply with the requirements during its actual use.~~

For the calibration of antennas, the CALTS shall comply with the validation criteria given in 4.5.3.1 at all frequencies at which the antennas shall be calibrated. The tests in 4.4.3.2 and their associated criteria in 4.5.3.2 and 4.5.3.3 are optional, but are included for those who want to confirm the achievement in 4.5.3.1 of  $T_{SA}(f)$  to less than 0,7 dB.

### 4.3 Test antenna specification

#### 4.3.1 Introduction

To allow (numerical) calculation of the theoretical site attenuation  $SA_c$  needed in the validation procedure, antennas are needed which can be accurately modelled. Therefore, the test antenna shall be a free-space resonant dipole connected to a balun with specified properties. The normative test antenna specifications are given in 4.3.2. An example of the construction of a test antenna is given in annex B.

The test antenna consists of a balun and two colinear wire elements (conductors) each having a diameter  $D_{we}$  and length  $L_{we}$ . These elements are connected to the two feed terminals (A and B in figure 1) at the balun. The gap between these feed terminals has a width  $W_g$ . The tip-to-tip length  $L_a$  of the antenna is given by  $L_a = 2L_{we} + W_g$ . The centre of the test antenna is in the middle of the feed-terminal gap on the centre-line of the two colinear wire elements.

The balun has an unbalanced input/output (transmitting/receiving antenna) port and a balanced port at the two feed terminals A and B. As an example, in figure 1 the purpose of the balun is indicated schematically by the balance/unbalance transformer.

#### 4.3.2 Normative specifications

**4.3.2.1** The test antenna shall have identical wire elements of length  $L_{we}$  which can be disconnected from the balun to enable the balun parameters to be validated, and to allow the balun heads of the two antennas used in site attenuation measurements to be connected together.

**4.3.2.2** The tip-to-tip length  $L_a(f, D_{we})$  of the approximately  $\lambda/2$  wire antenna is determined by the condition that, at the specified frequency  $f$  and in free space, the absolute value of the imaginary part of the input impedance at the feed terminals is smaller than  $1 \Omega$ .

NOTE 1 If the wire elements have a constant diameter and if  $D_{we} \ll L_a$ , then  $L_a(f, D_{we})$  can be calculated from the equation (C.2) in subclause C.1.1. If the diameter is not a constant, e.g. when a telescopic antenna is used,  $L_a(f)$  can only be calculated numerically, see C.2.2.

NOTE 2 When using a telescopic antenna, the telescopic elements should be tuned in such a way that the elements having the largest diameter are used first (see figure 2), and the numerical calculations should account for this approach.