



SLOVENSKI STANDARD

SIST EN 13273:2002

01-januar-2002

Dc j ý]bg_c`U_hj bY'gbcj]!'8 c`c Yj Ub^Y'bY]cbg_\`gi VghUbWj `Ub]cbg_\`dcj fý]bg_c
U_hj b\`gbcj Y`g'hY_c]bg_c`_fca Urc[fU]c`fkD@Ł

Surface active agents - Determination of the content of non-ionic substances in anionic surface active agents by high performance liquid chromatography (HPLC)

Grenzflächenaktive Stoffe - Bestimmung des Gehaltes von nichtionischen Substanzen in anionischen grenzflächenaktiven Stoffen mittels Hochleistungs-Flüssigchromatographie (HPLC)

(standards.iteh.ai)

Agents de surface - Détermination de la teneur en substances non-ioniques dans les agents de surfaces anioniques par chromatographie liquide a haute performance (CLHP)

Ta slovenski standard je istoveten z: EN 13273:2001

ICS:

71.100.40 Površinsko aktivna sredstva Surface active agents

SIST EN 13273:2002

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 13273:2002

<https://standards.iteh.ai/catalog/standards/sist/04460e14-aed0-47a1-833d-0c3a217d8f03/sist-en-13273-2002>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 13273

April 2001

ICS 71.100.40

English version

Surface active agents - Determination of the content of non-ionic substances in anionic surface active agents by high performance liquid chromatography (HPLC)

Agents de surface - Détermination de la teneur en substances non-ioniques dans les agents de surfaces anioniques par chromatographie liquide à haute performance (CLHP)

Grenzflächenaktive Stoffe - Bestimmung des Gehaltes von nichtionischen Substanzen in anionischen grenzflächenaktiven Stoffen mittels Hochleistungs-Flüssigchromatographie (HPLC)

This European Standard was approved by CEN on 20 January 2001.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

	Page
Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents	4
5 Apparatus	4
6 Calibration	5
7 Sampling and preparation of the test sample.....	6
8 Procedure	6
9 Calculation and expression of results	7
10 Precision	7
11 Test report	8
Annex A (informative) Flow diagram for the backflush valve in position 1 and position 2.....	9
Annex B (informative) Example chromatograms	10
Annex C (informative) Ring test results	14

SIST EN 13273:2002

<https://standards.iteh.ai/catalog/standards/sist/04460e14-aed0-47a1-833d-0c3a217d8f03/sist-en-13273-2002>

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 276 "Surface active agents", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2001, and conflicting national standards shall be withdrawn at the latest by October 2001.

Annexes A, B and C are informative.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 13273:2002](https://standards.iteh.ai/catalog/standards/sist/04460e14-aed0-47a1-833d-0c3a217d8f03/sist-en-13273-2002)

<https://standards.iteh.ai/catalog/standards/sist/04460e14-aed0-47a1-833d-0c3a217d8f03/sist-en-13273-2002>

1 Scope

This European Standard specifies a method for the determination of the contents of non-ionic substances in anionic surface active agents (sulfates, sulfonates and ethoxysulfates) by high performance liquid chromatography (HPLC).

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN ISO 3696, *Water for analytical laboratory use – Specification and test methods*. (ISO 3696:1987)

ISO 607, *Surface active agents and detergents – Methods of sample division*.

3 Principle

A specified mass of sample is dissolved and then made up to standard volume with methanol. For sulfonic acids only, the sample is neutralised with aqueous sodium hydroxide solution prior to the dissolution. The solution is then analysed by Reverse Phase High Performance Liquid Chromatography (RP-HPLC). An octadecylsilane bonded stationary phase, methanol/water mobile phase, differential refractometer detector and external means of quantification are employed. A back flush technique is used to "focus" the non-ionic components after all the active matter components and water have eluted from the column.

[SIST EN 13273:2002](https://standards.iteh.ai/catalog/standards/sist/04460e14-aed0-47a1-833d-0c3a217d8f03/sist-en-13273-2002)

<https://standards.iteh.ai/catalog/standards/sist/04460e14-aed0-47a1-833d-0c3a217d8f03/sist-en-13273-2002>

4 Reagents

During the analysis, only use reagents of recognized analytical grade and water complying with grade 3 as defined in EN ISO 3696.

4.1 Methanol, CH₄O HPLC grade.

4.2 Sodium hydroxide NaOH $c(\text{NaOH}) = 1 \text{ mol/l}$ solution.

4.3 Phenolphthalein indicator, ethanolic solution, C₂₀H₁₄O₄, $w(\text{C}_{20}\text{H}_{14}\text{O}_4) = 0,1 \text{ g/100 g}$.

4.4 Calibration substances, based on the feedstock from which the surface active agent/sulfonic acid has been produced, eg. :

- parent linear alkyl benzene (LAB) for alkyl benzene sulfonates ;
- parent alcohols for alcohol sulfates ;
- parent alcohol ethoxylates for alcohol ethoxy sulfates.

5 Apparatus

Ordinary laboratory apparatus and the following.

5.1 High performance liquid chromatograph, suitable for analysis according to the conditions given in 6.3.

5.2 Data logger/plotter giving the chromatographic peak areas.

5.3 Burette, capacity 10 ml with polytetrafluorethylene (PTFE) key, never lubricated with stopcock grease.

5.4 **Syringe**, capacity 1 ml with "Luer" fitting.

5.5 **Needle** with "Luer" fitting to suit sample injection valve, or use an automated injector.

5.6 **Syringe filter or membrane** suitable for use with methanol, 0,45 µm, with "Luer" fittings.

6 Calibration

6.1 General

Always wash glassware thoroughly with water and then acetone (propan-2-one). Dry thoroughly.

6.2 Determination of the optimum backflush time

Prepare and analyse the appropriate calibration substance (4.4) according to clause 7, but keep the backflush valve in position 1 (see Figure A.1) throughout the complete analysis. The chromatogram of the calibration substance alkyl ether sulfate, type 2, used in the ring test is shown in Figure B.1.

Note the elution time just before the first peak of the calibration substance appears. This is the backflush valve switching time.

The above valve switching time needs only be determined if the method is being set up or if it becomes apparent that the retention times have changed during routine use.

6.3 Determination of response factors

Determine the response factors for each type of the surface active agents as follows :

Prepare a calibration solution by weighing $(0,15 \pm 0,02)$ g of the calibration substance (4.4) to the nearest 0,1 mg into a 100 ml volumetric flask. Dissolve with methanol (4.1) and make up to volume.

Load the syringe (5.4) with 1 ml of the calibration solution and mount a new syringe filter fitted with the needle (5.5).

Use the following conditions:

Column :

Type	HPLC column with appropriate fittings. A radial compression type is suitable if used with appropriate conditions.
Length x internal diameter	Dimensions which enable results similar to the chromatograms in annex B. For a radial compression column, 100 mm x 8 mm is suitable. For conventional columns, 250 mm x 4,6 mm is suitable.
Stationary phase	Octadecylsilyl bonded micro-particulate silica of mean particle size not more than 5 µm. This should be of the completely-end-capped or deactivated type.

Mobile Phase :

Composition	A mixture of methanol and water capable of use with the chosen HPLC column so as to give results similar to the examples in annex B. A mixture of methanol : water of 85:15 parts by volume can be suitable
Flow rate	1 ml/min to 2 ml/min to the nearest 0,01 ml/min.

Detector :

Type	Differential refractometer
Sensitivity	A setting which gives a display of peak sizes such that deviations from normal appearance can be readily seen.

Temperatures :

Column	30,0 °C
	NOTE Thermostating of the column gives improved stability, but the method can be used with the column at ambient temperatures.
Detector	40,0 °C

Injection system :

High pressure syringe-loading injection valve or a suitable automated injector, fitted with an injection loop of appropriate volume, for example 50 µl.

Backflush valve :

A 6 port high pressure switching valve for HPLC, fitted with a very low volume by-pass loop.

Valve positioning :

Position 1 at the beginning of the analysis. Switch to position 2 at time A (see Figures A.1 and B.1)

Switch to position 1 on completion of the analysis (typical analysis time : 20 min).

Run the chromatograph in accordance with the manufacturer's instructions. When injecting the blend ensure the sample loop is completely flushed by dispensing the total syringe volume taking care not to introduce any air bubbles. An example chromatogram is given in Figure B.1.

NOTE The calibration peak elutes at a run time twice that of the backflush valve switch time. This time can vary depending on the sample type and is derived in 6.2.

The response factor, F , is calculated using the equation (1) :

$$F = \frac{m_c}{A_c} \quad (1)$$

where

m_c is the mass of the calibration substance used, in grams ;

A_c is the area of the calibration peak, in arbitrary units.

7 Sampling and preparation of the test sample

Sample the material to be tested in accordance with ISO 607.

8 Procedure

8.1 Determine the optimum backflush time from the calibration carried out in 6.2.

8.2 Weigh $(1 \pm 0,1)$ g of the sample to the nearest 0,1 mg into a 25 ml volumetric flask.

8.3 For sulfonic acids only, add one drop of phenolphthalein indicator (4.3) and then add 2,00 ml of sodium hydroxide solution (4.2). Swirl to dissolve. When completely dissolved, titrate dropwise with sodium hydroxide solution (4.2) to the end point characterized by faint pink colour (total titrant addition typically about 3,2 ml). For the other surface active agents, continue to 8.4.

Use a burette (5.3).

8.4 Dissolve and make up to volume with methanol (4.1). Allow to stand for a short time such that any precipitate formed settles to the bottom of the flask.

8.5 Load the syringe (5.4) with 1 ml of supernatant from the sample solution. Mount a new syringe filter or membrane (5.6). Use this to deliver a filtered sample portion for use in the analysis. Where possible, this should be done by direct delivery into the injection valve, via the needle (5.5).

8.6 Analyse the sample according to the conditions given in 6.3. When injecting the sample ensure the sample loop is completely flushed by dispensing the total syringe volume taking care not to introduce any air bubbles.

NOTE 1 An example chromatogram is given in figures B.2 and B.4. If the valve was not used, a chromatogram as shown in Figure B.3 should result.

NOTE 2 The operation of the backflush valve should be automated.

9 Calculation and expression of results

The content of non-ionic substances w_{NIS} of the sample expressed in grams per 100 g is calculated using the equation (2):

$$w_{\text{NIS}} = \frac{A_s \times F \times 100}{m_s \times 4} \quad (2)$$

where

A_s is the area of the peak of non-ionic substances, as used for A_c (see 6.3) in arbitrary units ;

F is the response factor of the calibration solution (6.3) ;

m_s is the mass of the sample (8.2) in grams.

Report the result to the nearest 0,001 g/100 g.

10 Precision

10.1 Repeatability limit

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will not exceed the repeatability limit, r , in more than 5 % of cases.

From the data given in annex C, the relative repeatability limits are :

- 12 % of the mean value at a level of 2 % to 3 % on active matter ;
- 19 % of the mean value at a level below 1 % on active matter.