



SLOVENSKI STANDARD
SIST ISO 7393-1:1996
01-junij-1996

Water quality -- Determination of free chlorine and total chlorine -- Part 1: Titrimetric method using N,N-diethyl-1,4-phenylenediamine

Water quality -- Determination of free chlorine and total chlorine -- Part 1: Titrimetric method using N,N-diethyl-1,4-phenylenediamine

iTeh STANDARD PREVIEW

Qualité de l'eau -- Dosage du chlore libre et du chlore total -- Partie 1: Méthode titrimétrique à la N,N-diéthylphénylène-1,4 diamine

SIST ISO 7393-1:1996

<https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-4f2e-977a-daf4cccc99cf/sist-iso-7393-1-1996>

Ta slovenski standard je istoveten z: ISO 7393-1:1985

ICS:

13.060.50 Ú!^ã \ æ ç ð [å ^ Á æ { ã } ^ Examination of water for chemical substances

SIST ISO 7393-1:1996

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 7393-1:1996

<https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-4f2e-977a-daf4ecec99cf/sist-iso-7393-1-1996>

International Standard



7393/1

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Water quality — Determination of free chlorine and
total chlorine —
Part 1: Titrimetric method using
N,N-diethyl-1,4-phenylenediamine**

iTeh STANDARD PREVIEW

*Qualité de l'eau — Dosage du chlore libre et du chlore total — Partie 1: Méthode titrimétrique
à la *N,N*-diéthylphénylène-1,4 diamine*

(standards.iteh.ai)

First edition — 1985-09-15

[SIST ISO 7393-1:1996](https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-4f2e-977a-daf4eccc99cf/sist-iso-7393-1-1996)

<https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-4f2e-977a-daf4eccc99cf/sist-iso-7393-1-1996>

UDC 543.3 : 543.24 : 546.13

Ref. No. ISO 7393/1-1985 (E)

Descriptors : water, quality, chemical analysis, determination of content, chlorine, volumetric analysis.

Price based on 8 pages

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 7393/1 was prepared by Technical Committee ISO/TC 147, *Water quality*.

iteh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 7393-1:1996](https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-4f2e-977a-daf4eccc99cf/sist-iso-7393-1-1996)

<https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-4f2e-977a-daf4eccc99cf/sist-iso-7393-1-1996>

Water quality — Determination of free chlorine and total chlorine —

Part 1: Titrimetric method using *N,N*-diethyl-1,4-phenylenediamine

0 Introduction

ISO 7393 consists of the following parts:

Part 1: Titrimetric method using *N,N*-diethyl-1,4-phenylenediamine.

Part 2: Colorimetric method using *N,N*-diethyl-1,4-phenylenediamine, for routine control purposes.

Part 3: Iodometric titration method for the determination of total chlorine.¹⁾

1 Scope and field of application

This part of ISO 7393 specifies a titrimetric method for the determination of free chlorine and total chlorine in water.

Sea water and waters containing bromides and iodides comprise a group for which special procedures are required.^[2]

The method is applicable to concentrations, in terms of chlorine (Cl_2), from 0,000 4 to 0,07 mmol/l (0,03 to 5 mg/l) total chlorine and at higher concentrations by dilution of samples. For concentrations above 0,07 mmol/l, ISO 7393/3 can also be used.

In annex A a procedure is presented for the differentiation of combined chlorine of the monochloramine type, combined chlorine of the dichloramine type and combined chlorine in the form of nitrogen trichloride.

Several compounds influence the determination specified in this part of ISO 7393. Interferences are noted in clauses 7 and 9.

2 Definitions (see table 1)

For the purpose of this part of ISO 7393, the following definitions apply.

2.1 free chlorine: Chlorine present in the form of hypochlorous acid, hypochlorite ion or dissolved elemental chlorine.

2.2 combined chlorine: The fraction of total chlorine present in the form of chloramines and organic chloramines.

2.3 total chlorine: Chlorine present in the form of "free chlorine" or "combined chlorine" or both.

2.4 chloramines: Derivatives of ammonia by substitution of one, two or three hydrogen atoms with chlorine atoms (monochloramine NH_2Cl , dichloramine NHCl_2 , nitrogen trichloride NCl_3) and all chlorinated derivatives of organic nitrogen compounds as determined by the method specified in this part of ISO 7393.

3 Principle

3.1 Determination of free chlorine

Direct reaction with the *N,N*-diethyl-1,4-phenylenediamine (DPD) and formation of a red compound at pH 6,2 to 6,5. Titration by means of a standard solution of ammonium iron(II) sulfate to the disappearance of the red colour.

3.2 Determination of total chlorine

Reaction with DPD in the presence of an excess of potassium iodide then titration as in 3.1.

Table 1 — Terms and synonyms in relation to actual compounds in the solution

Term	Synonym		Compounds
Free chlorine	Free chlorine	Active free chlorine	Elemental chlorine, hypochlorous acid
		Potential free chlorine	Hypochlorite
Total chlorine	Total residual chlorine		Elemental chlorine, hypochlorous acid, hypochlorite, chloramines

1) At present at the stage of draft.

4 Reagents

During the analysis, use only reagents of recognized analytical grade, and only water as specified in 4.1.

4.1 Water, free from oxidizing and reducing substances.

Demineralized or distilled water of which the quality is checked as follows.

Into two 250 ml chlorine-demand-free conical flasks (clause 5) place, in order,

- in the first: 100 ml of the water to be checked and about 1 g of potassium iodide (4.4); mix and after 1 min add 5 ml buffer solution (4.2) and 5,0 ml of DPD reagent (4.3);
- in the second: 100 ml of the water to be checked and two drops of sodium hypochlorite solution (4.8); then, after 2 min, 5,0 ml of buffer solution (4.2) and 5 ml of DPD reagent (4.3).

No coloration should appear in the first flask whereas it is essential that a light pink coloration appears in the second flask.

In the case of demineralized or distilled water not having the desired quality it must be chlorinated. After a period of contact followed by dechlorination the quality finally has to be rechecked.

A procedure for the preparation of water free from oxidizing and reducing substances is given in annex B.

4.2 Buffer solution, pH 6,5.

Dissolve in water (4.1) in this order: 24 g anhydrous disodium hydrogen phosphate (Na_2HPO_4) or 60,5 g of the dodecahydrate form ($\text{Na}_2\text{HPO}_4 \cdot 12\text{H}_2\text{O}$) and 46 g of potassium dihydrogen phosphate (KH_2PO_4). Add 100 ml of 8 g/l disodium dihydrogenethylenedinitrilotetraacetate dihydrate (disodium EDTA dihydrate, $\text{C}_{10}\text{H}_{14}\text{N}_2\text{O}_8\text{Na}_2 \cdot 2\text{H}_2\text{O}$) solution (or 0,8 g of the solid form).

If necessary, add 0,020 g of mercury(II) chloride (HgCl_2), to prevent mould growth and interference in the free available chlorine test caused by any trace amounts of iodide in the reagents.

Dilute to 1 000 ml and mix.

NOTE — Solutions containing mercury should be disposed of properly (for example a method is specified in ISO 5790, *Inorganic chemical products for industrial use — General method for determination of chloride content — Mercurimetric method*).

4.3 *N,N*-diethyl-1,4-phenylenediamine sulfate (DPD) $[\text{NH}_2\text{-C}_6\text{H}_4\text{-N}(\text{C}_2\text{H}_5)_2 \cdot \text{H}_2\text{SO}_4]$, solution, 1,1 g/l.

Mix 250 ml water (4.1), 2 ml sulfuric acid ($\rho = 1,84$ g/ml) and 25 ml of 8 g/l disodium EDTA dihydrate solution (or 0,2 g of the solid form). Dissolve in this mixture 1,1 g of anhydrous DPD or 1,5 g of the pentahydrate form, dilute to 1 000 ml and mix.

Store the reagent in a dark bottle protected from heat.

Renew the solution after 1 month or when it becomes discoloured.

4.4 Potassium iodide, crystals.

NOTE — Reagents 4.2, 4.3 and 4.4 may be conveniently replaced by combined reagents commercially available in the form of stable powder or tablets.

4.5 Ammonium iron(II) sulfate, stock solution, $c[(\text{NH}_4)_2\text{Fe}(\text{SO}_4)_2 \cdot 6\text{H}_2\text{O}] = 0,056$ mol/l.

4.5.1 Preparation of the solution

Dissolve 22 g of ammonium iron(II) sulfate hexahydrate (Mohr's salt) in about 250 ml water (4.1) containing 5 ml sulfuric acid ($\rho = 1,84$ g/ml) in a 1 000 ml one-mark volumetric flask. Make up to the mark with water and mix.

Store in a dark bottle.

Standardize this solution by means of the procedure given in 4.5.2, when required for use, or daily if large numbers of determinations have to be done.

4.5.2 Standardization of the solution

Place in a 250 ml conical flask, 50,0 ml of the stock solution (4.5.1), about 50 ml water (4.1), 5 ml orthophosphoric acid ($\rho = 1,71$ g/ml) and 4 drops of barium diphenylamine sulfonate indicator (4.9). Titrate with potassium dichromate solution (4.10). The end-point is reached when one drop produces an intense purple coloration which remains unchanged after further addition of the potassium dichromate solution. The concentration, c_1 , expressed in millimoles of Cl_2 per litre, of this solution is given by the equation

$$c_1 = V_2 \frac{c_2}{V_1}$$

where

c_2 is the concentration of the potassium dichromate standard reference solution (4.10),
 $c(1/6 \text{ K}_2\text{Cr}_2\text{O}_7) = 100$ mmol/l;

V_1 is the volume, in millilitres, of ammonium iron(II) sulfate stock solution (4.5), 50,0 ml;

V_2 is the volume, in millilitres, of the potassium dichromate standard reference solution (4.10) used in the titration.

NOTE — When V_2 becomes less than 22 ml, prepare a fresh stock solution (see 4.5.1).

4.6 Ammonium iron(II) sulfate, standard volumetric solution, $c[(\text{NH}_4)_2\text{Fe}(\text{SO}_4)_2 \cdot 6\text{H}_2\text{O}] = 2,8$ mmol/l.

Place 50,0 ml of the freshly standardized stock solution (4.5.1) in a 1 000 ml one-mark volumetric flask. Make up to the mark with water (4.1) and mix.

Transfer to a dark bottle.

Prepare this solution when required for use, or daily if large numbers of determinations have to be done.

The concentration, c_3 , expressed in millimoles of Cl_2 per litre, of this solution is given by the equation

$$c_3 = \frac{c_1}{20}$$

where c_1 is as defined in 4.5.2.

4.7 Sodium arsenite (NaAsO_2), solution, 2 g/l; or **thioacetamide** (CH_3CSNH_2), solution, 2,5 g/l.

4.8 Sodium hypochlorite, solution [$\rho(\text{Cl}_2)$ about 0,1 g/l].

Prepare by dilution of concentrated commercial sodium hypochlorite solution.

4.9 Barium diphenylamine sulfonate, indicator solution, 3 g/l.

Dissolve 0,3 g barium diphenylamine sulfonate [$(\text{C}_6\text{H}_5\text{-NH-C}_6\text{H}_4\text{-SO}_3)_2\text{Ba}$] in 100 ml water.

4.10 Potassium dichromate, standard reference solution, $c(1/6 \text{ K}_2\text{Cr}_2\text{O}_7) = 100 \text{ mmol/l}$.

Weigh, to the nearest milligram, 4,904 g of anhydrous potassium dichromate. Dissolve in water in a 1 000 ml one-mark volumetric flask. Make up to the mark with water and mix.

5 Apparatus

Ordinary laboratory apparatus, and

Microburette, measuring up to 5 ml and graduated in divisions of 0,02 ml.

NOTE ON THE PREPARATION OF GLASSWARE

Chlorine-demand-free glassware is obtained by filling with sodium hypochlorite solution (4.8) then, after 1 h, rinsing copiously with water (4.1). During the course of the analysis one set of glassware should be kept for the determination of free chlorine and another for the determination of total chlorine in order to avoid contamination of the free chlorine set.

6 Procedure

6.1 Test sample

Start determination immediately after taking samples. At all times avoid bright light, agitation and heat.

6.2 Test portions

Take two test portions, each of 100,0 ml. If the concentration of total chlorine exceeds 70 $\mu\text{mol/l}$ (5 mg/l) it is necessary to take a smaller volume of test sample and to dilute with water (4.1) to 100,0 ml.

6.3 Determination of free chlorine

Place rapidly in a 250 ml conical flask in this order: 5,0 ml of buffer solution (4.2), 5,0 ml of DPD reagent (4.3) and the first test portion (6.2). Mix and titrate immediately to a colourless end-point with ammonium iron(II) sulfate solution (4.6). Note the volume, V_3 , in millilitres, used in the titration.

In the case of an unknown water, possibly being very acid, or very alkaline or with a high concentration of salts, it is advisable to verify that the volume of buffer solution (4.2) added is sufficient to bring the water to pH 6,2 to 6,5. If not, use a greater volume of the buffer solution (4.2).

6.4 Determination of total chlorine

Place rapidly in a 250 ml conical flask in this order: 5,0 ml of buffer solution (4.2), 5,0 ml of DPD reagent (4.3), the second test portion (6.2) and about 1 g of potassium iodide (4.4). Mix and after 2 min, titrate to a colourless end-point with the ammonium iron(II) sulfate solution (4.6). If within 2 min a drift back of colour is observed continue titration to a colourless end-point. Note the volume, V_4 , in millilitres, used in the titration.

In the case of an unknown water, possibly being very acid, or very alkaline or with a high concentration of salts, it is advisable to verify that the volume of buffer solution (4.2) added is sufficient to bring the water to pH 6,2 to 6,5. If not, use a greater volume of the buffer solution (4.2).

7 Correction of interference due to the presence of oxidized manganese

Determine the effect of oxidized manganese by carrying out a supplementary determination on a further test portion (6.2) previously treated with the arsenite or thioacetamide solution (4.7) in order to neutralize all oxidizing compounds other than oxidized manganese.

Place this test portion in a 250 ml conical flask, add 1 ml of sodium arsenite solution (4.7) or thioacetamide solution (4.7) and mix. Again add 5,0 ml of buffer solution (4.2) and 5,0 ml of DPD reagent (4.3). Titrate immediately to a colourless end-point against ammonium iron(II) sulfate solution (4.6). Note the volume, V_5 , in millilitres, corresponding to the oxidized manganese.

8 Expression of results

8.1 Method of calculation

8.1.1 Calculation of the free chlorine concentration

The concentration of free chlorine, $c(\text{Cl}_2)$, expressed in millimoles per litre, is given by the equation

$$c(\text{Cl}_2) = \frac{c_3 (V_3 - V_5)}{V_0}$$

where

c_3 is the concentration, expressed in millimoles of Cl_2 per litre, of ammonium iron(II) sulfate solution;

V_0 is the volume, in millilitres, of test sample in the test portion (6.2);

V_3 is the volume, in millilitres, of the ammonium iron(II) sulfate solution (4.6) used in the titration (6.3);

V_5 is the volume, in millilitres, of the ammonium iron(II) sulfate solution (4.6) used in clause 7 ($V_5 = 0$ ml in the absence of oxidized manganese).

8.1.2 Calculation of the total chlorine concentration

The concentration of total chlorine, $c(\text{Cl}_2)$, expressed in millimoles per litre, is given by the equation

$$c(\text{Cl}_2) = \frac{c_3 (V_4 - V_5)}{V_0}$$

where

c_3 , V_0 and V_5 are as defined in 8.1.1;

V_4 is the volume, in millilitres, of the ammonium iron(II) sulfate solution (4.6) used in the titration (6.4)

8.2 Conversion of amount of substance concentration to mass concentration

The chlorine concentration expressed in moles per litre may be expressed in grams per litre by multiplying by a conversion factor of 70,91.

8.3 Repeatability and reproducibility

To obtain an indication of repeatability and reproducibility, figures are taken from measurements obtained by methods which are the same in principle as that specified in this part of ISO 7393.

The USA-EPA Environmental Monitoring and Support Laboratory^[1] evaluated the titrimetric method with the following results.

For distilled water samples at concentrations of $c(\text{Cl}_2) = 4,79$; $9,17$; and $48,6 \mu\text{mol/l}$ [$\rho(\text{Cl}_2) = 0,34$; $0,65$; and $3,45 \text{ mg/l}$] total chlorine, the relative standard deviations were 5,6 %, 0,5 % and 0,5 %, respectively. Using drinking water containing $c(\text{Cl}_2) = 13,8 \mu\text{mol/l}$ [$\rho(\text{Cl}_2) = 0,98 \text{ mg/l}$], total chlorine, the relative standard deviation was 1,2 %. With more polluted waters almost the same precision was obtained as for drinking water with the exception of raw sewage where, with a total chlorine concentration of $c(\text{Cl}_2) = 11,1 \mu\text{mol/l}$ [$\rho(\text{Cl}_2) = 0,79 \text{ mg/l}$], the relative standard deviation was 3,3 %.

Results published by the British Department of the Environment^[2] showed for total chlorine concentrations of $c(\text{Cl}_2) = 14$ and $71 \mu\text{mol/l}$ [$\rho(\text{Cl}_2) = 1,0$ and $5,0 \text{ mg/l}$], relative standard deviations of 1,4 % and 0,88 %, respectively.

The results presented in the preceding paragraphs relate to replicate determinations in the same laboratory and thus provide a measure of the repeatability of the method. Attempts in the past to quantify the reproducibility of the method by distribution of samples to different laboratories have produced unreliable results because of a general instability of solutions containing free and combined chlorine. More recently it has been found by the Quality Assurance Branch of EMSL-Cincinnati^[5] that a sealed vial of sodium hypochlorite in very pure water is remarkably stable when stored in the dark inside a mailing tube. Subsequent evaluation by various US federal and state laboratories has produced the analytical parameters listed in table 2 for methods in current use.

Table 2 — Analytical parameters from interlaboratory analysis for free residual chlorine

True value $c(\text{Cl}_2)$ $\mu\text{mol/l}$	$[\rho(\text{Cl}_2)]$ (mg/l)	Method code ¹⁾	Number of laboratories (observations)	Mean		Standard deviation	
				$\mu\text{mol/l}$	(mg/l)	$\mu\text{mol/l}$	(mg/l)
7	(0,5)	A	6	6,2	(0,44)	1,3	(0,09)
		B	7	6,8	(0,48)	1,8	(0,13)
11,3	(0,80)	A	10	10,9	(0,77)	1,1	(0,08)
		B	14	11,1	(0,79)	4,1	(0,29)
		C	6	11,6	(0,82)	1,3	(0,09)
15,5	(1,10)	A	10	15,5	(1,10)	2,0	(0,14)
		B	14	16,2	(1,15)	5,5	(0,39)
		C	6	16,5	(1,17)	1,1	(0,08)
18,2	(1,29)	A	6	18,6	(1,32)	1,3	(0,09)
		B	7	19,9	(1,41)	5,4	(0,38)

1) A: Iodometric titration amperometric

B: DPD colorimetric

C: DPD titrimetric

9 Interferences

Two types of interference may be noted.

9.1 Interference by other chlorine compounds

A fraction of any chlorine dioxide that might be present is measured as free chlorine. This interference may be corrected by determining the chlorine dioxide in the water.^[2, 3, 4]

9.2 Interference by compounds other than chlorine compounds

Oxidation of DPD is not specifically caused by chlorine compounds. Depending on the concentration and the chemical oxidation potential, the reaction is effected by other oxidizing agents. The following substances may be mentioned in particular: bromine, iodine, bromamines, iodoamines, ozone, hydrogen peroxide, chromate, oxidized manganese, nitrite, iron(III) ions and copper ions. The interference is suppressed by

the disodium EDTA in reagents 4.2 and 4.3 in the case of copper(II) ions (< 8 mg/l) and iron(III) ions (< 20 mg/l).

Interference by chromate may be eliminated by addition of barium chloride.^[6]

10 Test report

The test report shall include the following information:

- a) a reference to this part of ISO 7393;
- b) all information necessary for complete identification of the sample;
- c) the results and the method of expression used;
- d) details of any operations not included in this part of ISO 7393, or regarded as optional, together with any circumstance that may have affected the results.

Bibliography

- iTeh STANDARD PREVIEW**
standards.iteh.ai
- [1] BENDER, D.F. Comparison of methods for the determination of total available residual chlorine in various sample matrices, Report No. EPA-600/4-78-019. Cincinnati, Ohio 45268, USA, US Environmental Protection Agency, 1978.
- [2] DoE. *Chemical Disinfecting Agents in Waters, and Effluents, and Chlorine Demand, Methods for the Examination of Waters and Associated Materials*. London, UK, HMSO, 1980.
<https://standards.iteh.ai/catalog/standards/sist/cd740c46-6bd6-42e-977a-daf4eccc99cf/sist-iso-7393-1-1996>
- [3] PALIN A.T. Methods for the determination in water of free and combined available chlorine, chlorine, dioxide and chlorite, bromine, iodine and ozone, using diethyl-*p*-phenylenediamine. *J. Inst. Water Eng.* **21** 1967 : 537.
- [4] PALIN A.T. Analytical control of water disinfection with special reference to differential DPD methods for chlorine, chlorine dioxide, bromine, iodine and ozone. *J. Inst. Water Eng.* **28** 1974 : 139.
- [5] Studies WS007 and WS008, Cincinnati, Ohio 45268, USA, Quality Assurance Branch, Environmental Monitoring and Support Laboratory, Office of Research and Development, US Environmental Protection Agency, 1980.
- [6] PALIN, A.T. New correction procedures for chromate interference on the DPD method for residual, free and combined chlorine. *J. Inst. Water Eng. Sci.* **36** 1982: 351.