

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 8891:2000

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 8891

July 2000

ICS 11.060.10

Supersedes EN ISO 8891:1995

English version

Dental casting alloys with noble metal content of at least 25 %
but less than 75 % (ISO 8891:1998)

Alliages dentaires à couler avec une teneur en métaux
précieux supérieure ou égale à 25 % et strictement
inférieure à 75 % (ISO 8891:1998)

Dentalgußlegierungen mit einem Edelmetallanteil von
mindestens 25 % bis unter 75 % (ISO 8891:1998)

This European Standard was approved by CEN on 9 June 2000.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN ISO 8891:2000](https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000)

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Page 2
EN ISO 8891:2000

Foreword

The text of the International Standard from Technical Committee ISO/TC 106 "Dentistry" of the International Organization for Standardization (ISO) has been taken over as an European Standard by Technical Committee CEN/TC 55 "Dentistry", the secretariat of which is held by DIN.

This European Standard replaces EN ISO 8891:1995.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2001, and conflicting national standards shall be withdrawn at the latest by January 2001.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of the International Standard ISO 8891:1998 has been approved by CEN as a European Standard without any modification.

NOTE: Normative references to International Standards are listed in annex ZA (normative).

iteh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 8891:2000

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

Annex ZA (normative)
Normative references to international publications
with their relevant European publications

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN</u>	<u>Year</u>
ISO 3696	1987	Water for analytical laboratory use - Specification and test methods	EN ISO 3696	1995
ISO 6507-1	1997	Metallic materials – Vickers hardness test – Part 1: Test method	EN ISO 6507-1	1997
ISO 9693	1991	Dental ceramic fused to metal restorative materials	EN ISO 9693	1994

iTech STANDARD PREVIEW
 (standards.iteh.ai)

SIST EN ISO 8891:2000
<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 8891:2000

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

INTERNATIONAL STANDARD

**ISO
8891**

Third edition
1998-12-15

Dental casting alloys with noble metal content of at least 25 % but less than 75 %

*Alliages dentaires à couler avec une teneur en métaux précieux supérieure
ou égale à 25 % et strictement inférieure à 75 %*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8891:2000](https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000)

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>



Reference number
ISO 8891:1998(E)

ISO 8891:1998(E)

Contents

1 Scope	1
2 Normative references	1
3 Classification.....	1
4 Requirements	2
5 Sampling.....	3
6 Preparation of test specimens	3
7 Testing	4
8 Information and instructions	5
9 Marking	5
Annex A (normative) Surface corrosion testing — Static immersion test	6
Annex B (normative) Tarnish testing — Sodium sulfide test	8
Annexe C (informative) Bibliography	10

STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 8891:2000

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

© ISO 1998

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 8891 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthetic materials*.

This third edition cancels and replaces the second edition (ISO 8891:1993), of which it constitutes a technical revision.

Annexes A and B form an integral part of this International Standard. Annex C is for information only.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8891:2000](https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000)

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

Introduction

Specific qualitative and quantitative requirements for freedom from biological hazard are not included in this International Standard but it is recommended that, in assessing possible biological hazards, reference should be made to ISO 10993-1 and ISO 7405.

At this time it has not been possible to set requirements for corrosion and tarnish resistance. However it is recommended that the static immersion test given in annex A should be used to provide information on the type and quantity of metal ions which leach from a dental casting alloy and the sodium sulfide tarnish test given in Annex B should be used to provide information on the probability of surface alteration as a result of tarnish.

When an International Standard (e.g. ISO 10271) for corrosion and tarnish testing of dental casting alloys is published, the test procedures given in annexes A and B will be revised to conform with that International Standard, and requirements for maximum permissible corrosion and tarnish will be set. At that time, the implementation of a test on the electrochemical behaviour of dental casting alloys should also be considered.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8891:2000](https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000)

<https://standards.iteh.ai/catalog/standards/sist/3c012ee4-ae1c-4872-8b93-377c382864fc/sist-en-iso-8891-2000>

Dental casting alloys with noble metal content of at least 25 % but less than 75 %

1 Scope

This International Standard gives the classification of, and specifies requirements and test methods for, dental casting alloys with a noble metal content of at least 25 % (mass fraction) up to but not including 75 % (mass fraction).

It applies to casting alloys suitable for the fabrication of dental restorations and appliances.

NOTE 1 For dental casting gold alloys with noble metal content of 75 % and above, see ISO 1562.

NOTE 2 For dental alloys intended as the substructure of a metal-ceramic dental restorative system, see ISO 9693.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3585:1998, *Borosilicate glass 3.3 — Properties.*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods.*

ISO 6507-1:1997, *Metallic materials — Vickers hardness test — Part 1: Test method.*

ISO 6892:1998, *Metallic materials — Tensile testing at ambient temperature.*

ISO 9693:1991, *Dental ceramic fused to metal restorative materials.*

3 Classification

For the purposes of this International Standard, dental casting alloys are classified, according to their mechanical properties and the application for which they are recommended, as follows:

- **Type 1:** low-strength- for castings subject to very slight stress, e.g. inlays;
- **Type 2:** medium-strength- for castings subject to moderate stress, e.g. inlays, onlays, and full crowns;
- **Type 3:** high-strength- for castings subject to high stress, e.g. onlays, thin cast backings, pontics, crowns and saddles;
- **Type 4:** extra-high strength- for castings subject to very high stress and thin cross-section, e.g. saddles, bars, clasps, thimbles, unit castings and partial denture frameworks.