



Standard Test Method for Reaction Induction Time by Thermal Analysis¹

This standard is issued under the fixed designation E 2046; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

1. Scope

1.1 This test method describes the measurement of Reaction Induction Time (RIT) of chemical materials that undergo exothermic reactions with an induction period. The techniques and apparatus described may be used for solids, liquids, or slurries of chemical substances. The temperature range covered by this test method is typically from ambient to 400°C. This range may be extended depending upon the apparatus used.

1.2 The RIT is a relative index value, not an absolute thermodynamic property. As an index value, the RIT value may change depending upon experimental conditions. A comparison of RIT values may be made only for materials tested under similar conditions of apparatus, specimen size, and so forth. Furthermore, the RIT value may not predict behavior of large quantities of material.

1.3 The RIT shall not be used by itself to establish a safe operating temperature. It may be used in conjunction with other test methods (for example, E 487, E 537, and E 1981 as part of a hazard analysis of a particular operation).

1.4 This test method may be used for RIT values greater than 15 min (as relative imprecision increases at shorter periods).

1.5 This test method is used to study catalytic and autocatalytic reactions. These reactions depend upon time as well as temperature. Such reactions are often studied by fixing one experimental parameter (that is, time or temperature) and then measuring the other parameter (that is, temperature or time). This test method measures time to reaction onset detection under isothermal conditions. It is related to Test Method E 487 that measures detected reaction onset temperature under constant time conditions.

1.6 The values stated in SI units are to be regarded as the standard.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to consult and establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

¹ This test method is under the jurisdiction of ASTM Committee E-27 on Hazard Potential of Chemicals and is the direct responsibility of Subcommittee E27.02 on Thermal Stability and Condensed Phases.

Current edition approved Dec. 10, 1999. Published January 2000.

D 3350 Specification for Polyethylene Plastics Pipe and Fittings Materials²

D 3895 Test Method for Oxidative-Induction Time of Polyolefins by Differential Scanning Calorimetry²

D 4565 Test Methods for Physical and Environmental Performance Properties of Insulations and Jackets for Telecommunications Wire and Cable³

D 5483 Test Method for Oxidation Induction Time of Lubricating Greases by Pressure Differential Scanning Calorimetry⁴

D 6186 Test Method for Oxidation Induction Time of Lubricating Oils by Pressure Differential Scanning Calorimetry

E 473 Terminology Relating to Thermal Analysis⁵

E 487 Test Method for Constant-Temperature Stability of Chemical Materials⁵

E 537 Test Method for Assessing the Thermal Stability of Chemicals by Methods of Differential Thermal Analysis⁵

E 967 Practice for Temperature Calibration of Differential Scanning Calorimeters and Differential Thermal Analyzers⁵

E 968 Practice for Heat Flow Calibration of Differential Scanning Calorimeters⁵

E 1445 Terminology Relating to Hazardous Potential of Chemicals⁵

E 1858 Test Method for Determining Oxidative Induction Time of Hydrocarbons by Differential Scanning Calorimetry⁵

E 1860 Test Method for Elapsed Time Calibration of Thermal Analyzers⁵

E 1981 Guide for Assessing the Thermal Stability of Materials by Method of Accelerating Rate Calorimetry⁵

3. Terminology

3.1 The specialized terms used in this test method are described in Terminologies E 473 and E 1445.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *constant temperature stability (CTS) value, n*—the maximum temperature at which a chemical compound or mixture may be held for a minimum of two hours without

² *Annual Book of ASTM Standards*, Vol 08.02.

³ *Annual Book of ASTM Standards*, Vol 10.02.

⁴ *Annual Book of ASTM Standards*, Vol 05.03.

⁵ *Annual Book of ASTM Standards*, Vol 14.02.

exhibiting a measurable exothermic reaction. (See Test Method E 487.)

3.2.2 *reaction induction time (RIT) value, n*—the time a chemical compound or mixture may be held under isothermal conditions until it exhibits a specified exothermic reaction.

4. Summary of Test Method

4.1 A specimen of the chemical compound or mixture is placed in an inert container that is then heated to an operator-selected test temperature of interest. The specimen temperature and the difference in heat flow or temperature between the test specimen and an inert reference are monitored until an exothermic reaction is recorded. The time from the attainment of the isothermal test temperature until the extrapolated onset to the exothermic reaction is taken as the Reaction Induction Time.

4.2 Using fresh specimens measurements at more than one isothermal test temperature may be made.

4.3 The RIT is expressed as time at a specific test temperature. For example:

$$\text{RIT} = 120 \text{ min at } 100^{\circ}\text{C}$$

5. Significance and Use

5.1 This test method measures the time to extrapolated onset of an exothermic reaction under constant temperature (isothermal) conditions for reactions which have an induction period, for example, those which are catalytic or autocatalytic in nature or which contain reaction inhibitors.

5.2 The RIT determined by this test method is to be considered an index measurement that is useful for comparing one material to another at the test temperature of interest and in the same apparatus type only.

5.3 This test method is a useful adjunct to dynamic thermal tests, such as Test Method E 537, which are performed under conditions in which the sample temperature is increased continuously at constant rate. Results obtained under dynamic test conditions may result in higher estimates of temperature at which an exothermic reaction initiates because the detected onset temperature is dependent upon the heating rate and because dynamic methods allow insufficient time for autocatalytic reactions to measurably affect the onset temperature.

5.4 While not described in this test method, RIT values determined under a series of isothermal test conditions may be plotted as their logarithm versus the reciprocal of the absolute temperature to produce a plot, the slope of which is proportional to the activation energy of the reaction.

5.5 This test method may be used in research and development, manufacturing, process and quality control, and regulatory compliance.

5.6 This test method is similar to that for Oxidation Induction Time (OIT) (for example, Specification D 3350 and Test Methods D 3895, D 4565, D 5483, D 6186, and E 1858) where the time to the oxidation reaction under isothermal test conditions is measured. The OIT test method measures the presence of antioxidant packages and is a relative measurement of a material's resistance to oxidation.

6. Apparatus

6.1 The design and complexity of the apparatus required for

this test method depends upon the size of the specimen to be used. In general, observation of an exothermic reaction in small specimens (less than 50 mg) is performed using differential scanning calorimetry or differential thermal analysis equipment and techniques. Large samples (up to 2 g) may be tested using devices such as the Kuhner Micro CTS apparatus.

6.2 The following items are required to obtain the appropriate experimental data.

NOTE 1—Commercially available differential scanning calorimetry apparatus may be used. Alternatively, the apparatus may be assembled or fabricated from commercially available components.

6.2.1 *Test Chamber*, composed of the following:

6.2.1.1 *Furnace(s)*, to provide uniform controlled heating of a specimen and reference to a constant temperature.

6.2.1.2 *Temperature Sensor*, to provide an indication of specimen/furnace temperature to ± 0.1 K.

6.2.1.3 *Differential Sensor*, to detect a difference in heat flow (or temperature) between the specimen and the reference specimen equivalent to 10 μW or 0.01 mK.

NOTE 2—A reference material is used when differential heat flow or differential temperature measurements are made. The reference material should match the physical state and heat capacity of the specimen as closely as practical. Typical reference materials include calcined aluminum oxide, glass beads, silicone oil, or combinations thereof.

6.2.1.4 *Means of Sustaining a Test Chamber Environment*, through the use of an air purge gas at a rate of 10 to 100 ± 5 mL/min.

NOTE 3—Typically, air or inert 99.9+ % pure nitrogen, argon, or helium is employed (when oxidation in air is a concern). Unless effects of moisture are to be studied, use of dry purge gas is recommended.

NOTE 4—For the Kuhner Micro CTS apparatus, the purge gas is provided by operation in a laboratory hood with the door(s) approximately 50 % closed.

6.2.2 *Temperature Controller*, capable of heating from ambient to 400°C at a rate of up to 20°C/min and maintaining an isothermal temperature constant within that range of $\pm 0.4^{\circ}\text{C}$ for the duration of the test, or both.

6.2.3 *Recording Device*, either digital or analog, to record and display any fraction of the differential heat flow or differential temperatures, test specimen temperature, and time, including signal noise, to the sensitivities previously described. Typical temperature ranges are from ambient to 400°C. Typical time ranges are from 0 to 200 min.

6.2.4 *Containers* (pans, crucibles, vials, test tubes, and so forth, and lids), which are inert to the specimen and reference materials at the maximum temperature used and which are of suitable structure, shape, and integrity to contain the specimen and reference in accordance with the temperature and mass requirements as described in this section.

6.3 *Balance*, with a capacity of 100 mg or more to weigh specimens or containers to ± 0.1 mg.

NOTE 5—A balance capacity of 10 g or more with a readability to ± 0.1 g is required for use with the Kuhner Micro CTS device.

7. Hazards

7.1 Dynamic thermal tests, utilizing milligram quantities of materials, such as Test Method E 537, are normally conducted on specimens before the present test is undertaken. The