



SLOVENSKI STANDARD

SIST EN 13657:2004

01-januar-2004

?

Characterization of waste - Digestion for subsequent determination of aqua regia soluble portion of elements

Charakterisierung von Abfällen - Aufschluss zur anschließenden Bestimmung des in Königswasser löslichen Anteils an Elementen in Abfällen

STANDARD PREVIEW

Caractérisation des déchets - Digestion en vue de la détermination ultérieure de la part des éléments solubles dans l'eau régale contenus dans les déchets

[SIST EN 13657:2004](https://standards.iteh.ai/catalog/standards/sist/2d12457-71cd-4f00-b9bb-8ccc95845c43/sist-en-13657-2004)

Ta slovenski standard je istoveten z: **EN 13657:2002**

ICS:

13.030.40	Naprave in oprema za odstranjevanje in obdelavo odpadkov	Installations and equipment for waste disposal and treatment
-----------	--	--

SIST EN 13657:2004

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 13657:2004

<https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 13657

October 2002

ICS 13.030.40

English version

Characterization of waste - Digestion for subsequent determination of aqua regia soluble portion of elements

Caractérisation des déchets - Digestion en vue de la détermination ultérieure de la part des éléments solubles dans l'eau régale contenus dans les déchets

Charakterisierung von Abfällen - Aufschluss zur anschließenden Bestimmung des in Königswasser löslichen Anteils an Elementen in Abfällen

This European Standard was approved by CEN on 19 August 2002.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN 13657:2004](https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004)

<https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

	page
Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Safety remarks	6
5 Sample	6
5.1 Sample pre-treatment	6
5.2 Mass of test portion.....	7
6 Equipment	7
6.1 Closed vessel system	7
6.1.1 Microwave unit.....	7
6.1.2 Digestion vessels.....	7
6.2 Semi-open vessel system.....	8
6.2.1 Microwave unit.....	8
6.2.2 Digestion vessels.....	8
6.3 Apparatus for thermal heating digestion	8
6.3.1 Reaction vessels.....	8
6.3.2 Reflux condenser.....	8
6.3.3 Absorption vessel.....	8
6.3.4 Roughened glass beads.....	8
6.3.5 Temperature-controlled heating apparatus	8
6.4 General equipment.....	8
7 Reagents	9
8 Interferences and sources of error	9
8.1 General informations.....	9
8.2 Closed vessel system	9
8.3 Semi-open vessel system for microwave or thermal heating digestion.....	10
9 Procedure	10
9.1 Blank test.....	10
9.2 Procedure for microwave assisted digestion in closed vessels.....	10
9.3 Procedure for microwave assisted digestion in semi-open vessels.....	11
9.3.1 Amount of samples	11
9.3.2 Digestion	11
9.3.3 Preparation of the solution for analysis	11
9.4 Procedure for thermal heating digestion in reflux systems	11
9.4.1 Amount of samples	11
9.4.2 Digestion	12
9.4.3 Preparation of the solution for analysis	12
10 Test report	12
Annex A (normative) Calibration of the power adjustment	14
Annex B (informative) Validation	15
B.1 Inter-laboratory study.....	15
B.2 Robustness study.....	23
Bibliography	26

Foreword

This document EN 13657:2002 has been prepared by Technical Committee CEN/TC 292 "Characterization of waste", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2003, and conflicting national standards shall be withdrawn at the latest by April 2003.

In this European Standard the annex A is normative and the annexe B is informative.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 13657:2004](https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004)

<https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004>

Introduction

The aim of this European Standard is to describe suitable methods for the extraction of elements in waste, soluble by digestion with aqua regia, i. e. aqua regia soluble elements.

For those types of waste, where the sample material by the described digestion with aqua regia are not brought totally in solution, the obtained results will not be the total amount of the elements in the waste.

The obtained results of aqua regia soluble elements cannot be regarded as available for leaching, as the digestion with aqua regia is too vigorous to represent natural processes.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 13657:2004](#)

<https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004>

1 Scope

This European Standard specifies methods of digestion with aqua regia. Solutions produced by the methods are suitable for analysis e.g. by atomic absorption spectrometry (FLAAS, HGAAS, CVAAS, GFAAS), inductively coupled plasma emission spectrometry (ICP-OES) and inductive coupled plasma mass spectrometry (ICP-MS).

The digestion with aqua regia will not necessarily release all elements completely. However for most environmental application and waste characterization the results fit for the purpose.

The method is applicable to the digestion of waste for example for the following elements: Al, Sb, As, B, Ba, Be, Ca, Cd, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Mo, Ni, P, K, Se, Ag, S, Na, Sr, Sn, Te, Ti, Tl, V, Zn.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text, and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN ISO 3696:1995, *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*.

3 Terms and definitions

For the purposes of this European Standard, the following terms and definitions apply.

3.1

aqua regia

digestion solution obtained by mixing 1 volume of nitric acid (65 % m/m to 70 % m/m) and 3 volumes of hydrochloric acid (35 % m/m to 37 % m/m)

(standards.iteh.ai)
SIST EN 13657:2004
<https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8eec93845c43/sist-en-13657-2004>

3.2

digestion

mineralization of the organic matter of a sample and dissolution of its mineral part, more or less completely, when reacted with a reagent mixture

3.3

sample

portion of material selected from a larger quantity of material

[ENV 12506:2001]

3.4

laboratory sample

sample or subsample(s) sent to or received by the laboratory

[ENV 12506:2001]

3.5

test sample; analytical sample

sample, prepared from the laboratory sample, from which test portions are removed for testing or analysis

[ENV 12506:2001]

EN 13657:2002 (E)**3.6****test portion; analytical portion**

quantity of material of proper size for measurement of the concentration or other properties of interest, removed from the test sample

[ENV 12506:2001]

NOTE 1 The test portion may be taken from the laboratory sample directly if no preparation of sample is required (e. g. with liquids), but usually it is taken from the prepared test sample.

NOTE 2 A unit or increment of proper homogeneity, size and fineness, needing no further preparation, may be a test portion.

3.7**dry residue**

dry matter expressed as a percentage by mass after drying at $105\text{ °C} \pm 5\text{ °C}$ to the constancy of weight

3.8**digestion vessel**

special flask where the test portion and the acid mixture are filled in and the digestion is performed

3.9**microwave unit**

whole microwave digestion system (oven and associated equipment)

3.10**microwave unit cavity**

the inner part of the microwave unit in which the digestion vessel is located and the microwave digestion is performed

3.11

focused microwave unit <https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-4d99e0c11111/sist-en-13657-2004>
microwave unit in which a precise control of the electric field is made by using a wave guide

NOTE Microwaves are focused at the bottom part of the digestion vessel.

4 Safety remarks

All this work has to be performed by skilled persons.

The reagents used within this EN are strongly corrosive and partly very toxic. Safety precautions are absolutely necessary due to strong corrosive reagents, high temperature and high pressure.

All procedures have to be performed in a hood or in closed force-ventilated equipment. By the use of strong oxidising reagents the formation of explosive organic intermediates is possible especially when dealing with samples with a high organic content. Do not open pressurised vessels before they have cooled down. Avoid contact with the chemicals and the gaseous reaction products. Samples and solutions have to be disposed of according to regulations.

5 Sample**5.1 Sample pre-treatment**

The test portion should be transferred into the vessel as it is without any pre-treatment if possible. This is applicable only if the test portion is representative for the laboratory sample and the effectiveness of the digestion is proven.

If these conditions are not met a pre-treatment of the laboratory sample is necessary. This procedure shall not change the concentration of the elements of interest.

Pre-treatment should include drying or grain size reduction below a particle size of 250 µm for solid waste or homogenizing by use of a high speed mixer or sonification for liquid samples.

The mass of laboratory samples shall be sufficient for the multiple digestion procedures and determination of the dry residue.

5.2 Mass of test portion

The mass of test portion for a single digestion has to be selected in a way, that:

- it is representative for the laboratory sample;
- it complies with the specifications of manufacturer of the digestion unit.

NOTE If the representative test portion exceeds the manufacturers specifications the test portion should be divided into smaller quantities and digested separately. The individual digests should be combined prior to analysis.

For representativity reason mass above 200 mg is to be preferred. Unless recommended by the manufacturer the amount of organic carbon shall not exceed 100 mg because of safety reasons in the case of closed digestion vessel.

6 Equipment

iTeh STANDARD PREVIEW
(standards.iteh.ai)

6.1 Closed vessel system

[SIST EN 13657:2004](#)

6.1.1 Microwave unit <https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-8ecc93845c43/sist-en-13657-2004>

The microwave unit shall provide programmable power which can be programmed to within ± 10 W of the required power. Typical units provide a nominal 600 W to 1 200 W of power. If necessary (referring to manufactures specifications) calibration of the microwave unit has to be performed (see annex A).

The microwave unit has to comply to European and national regulations relevant to microwave radiation.

The microwave unit cavity has to be well ventilated. It has to have an exhaust air tube which is connected to a corrosion resistant laboratory air outlet system or the instrument is provided for use in a laboratory hood.

All electronics are sufficiently protected against corrosion for safe operation. All parts which could have contact with acids or their vapours have to be corrosion resistant.

The microwave unit shall be designed in a way that guarantees homogeneous heating of the samples.

The microwave unit cavity has to be built in a way that even in case of leakage or explosion of the vessels the safety of the operators can be guaranteed. Household instruments are not suitable for laboratory use.

NOTE The microwave unit should include a temperature and/or pressure control system.

6.1.2 Digestion vessels

The vessels used in the microwave unit shall be equipped with a pressure relieve valve or another technical equipment which avoids the bursting of the vessels at suddenly occurring excess pressure. The material of the vessels has to be inert to the acids used for digestion. The digestion vessel shall withstand the pressure of at least 8 bar. If the amount of organic carbon exceeds 100 mg it has to be ensured that the digestion vessel is capable of withstanding higher pressures.

EN 13657:2002 (E)**6.2 Semi-open vessel system****6.2.1 Microwave unit**

The microwave unit shall be equipped with power control. Typical unit provides a nominal 200 W or 300 W of released power. If necessary, calibration of the microwave unit has to be performed by the manufacturer.

The microwave unit shall comply with European and national regulations relevant to microwave radiation.

The microwave unit shall be designed for use in a laboratory hood.

Fume extraction equipment, used to extract acid vapours from the reaction vessels during the digestion program, shall have sufficient flow rate to prevent the release of dangerous vapours into the laboratory.

All electronics shall be sufficiently protected against corrosion for safe operation. All parts which can be in contact with acids or their vapours shall be corrosion resistant.

6.2.2 Digestion vessels

The vessel is working at atmospheric pressure and has to be connected with a reflux system to avoid losses of analytes. The vessel shall comply with the manufactures specifications and should have a minimum volume of 50 ml.

The material of the vessels has to be inert to the reagents used for digestion.

6.3 Apparatus for thermal heating digestion**6.3.1 Reaction vessels**

[SIST EN 13657:2004](https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-811111111111)

[https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-](https://standards.iteh.ai/catalog/standards/sist/2db12457-71cd-4f00-b9bb-811111111111)

The reaction vessel shall have a volume of at least 5 times of the volume of the aqua regia used.

6.3.2 Reflux condenser

Water-cooled reflux condenser, minimum assembled length 340 mm with conical ground joints.

6.3.3 Absorption vessel

Non-return type.

6.3.4 Roughened glass beads

Diameter 2 mm to 3 mm (or anti bumping granules), acid washed.

6.3.5 Temperature-controlled heating apparatus

Capable of heating the contents of the reaction vessel to reflux temperature.

6.4 General equipment

The following equipment is used by the systems described in 6.1, 6.2 and 6.3:

- volumetric graduated flasks and pipettes of adequate size;
- filter equipment of adequate chemical resistance and purity or centrifuge;

— analytical balance, with an error limit of $\pm 0,1$ mg.

For the preparation of standards and the treatment and storage of samples for determination of boron the use of borosilicate glass shall be avoided.

7 Reagents

Use reagents of analytical grade quality or better and water of grade 1 according to EN ISO 3696:1995.

— Hydrochloric acid (HCl): a mass fraction of 35 % to 37 %;

— Nitric acid (HNO₃): a mass fraction of 65 % to 70 %;

— Diluted nitric acid: $c(\text{HNO}_3) = 0,5$ mol/l.

8 Interferences and sources of error

8.1 General informations

Due to the volatility of some compounds it is of great importance to take care, that the sample is not heated before the digestion and that the volatile reaction products which might be formed during the digestion are not allowed to escape.

The container in which the sample is delivered and stored can be a source of errors. Its material shall be chosen according to the elements to be determined (e.g. elemental Hg can penetrate polyethylene walls very fast in both directions. Glass can contaminate samples with elements contained: e.g. B, Na, K, Al).

Grinding or milling samples includes a risk of contamination of the sample by the environment (air, dust, wear of milling equipment). Due to elevated temperature losses of volatile compounds are possible.

For the determination of elements forming volatile compounds (e.g. Hg, As, Pb) special care has to be taken at sample pre-treatment.

The use of the described digestion procedures may leave large parts of the sample undissolved. This includes the risk of bad repeatability.

High acid and solute concentrations in the digest may cause interferences at determination.

Depending on the concentration of the element of interest and the wanted precision, a particular caution to the cleaning of the vessels shall be taken. It is recommended to clean the vessels with 10 % nitric acid.

Care shall be taken to ensure that all of the test portion is brought into contact with the acid mixture in the reaction vessel.

Some elements of interest can be lost because of precipitation with some ions of the solution. It is the case for insoluble chlorides, fluorides and sulphates for example. In this case the precipitate can be analysed separately.

In the case of filtration of the digested solution it is necessary to take care that the filtration procedure does not introduce contaminants.

8.2 Closed vessel system

The upper limits of mass of the test portion referring to the manufacturers specifications have to be taken into account.