



SLOVENSKI STANDARD
SIST EN ISO 1797-1:2000/A1:2000
01-januar-2000

Dental rotary instruments - Shanks - Part 1: Shanks made of metals (ISO 1797-1:1992/Amendment 1:1997)

Dental rotary instruments - Shanks - Part 1: Shanks made of metals (ISO 1797-1:1992/Amendment 1:1997)

Rotierende Dentalinstrumente - Schäfte - Teil 1: Schäfte aus Metall (ISO 1797-1:1992/Ergänzung 1:1997)

Instruments rotatifs dentaires - Queues - Partie 1: Queues en matériaux métalliques (ISO 1797-1:1992/Amendement 1:1997)

<https://standards.iteh.ai/catalog/standards/sist/5b08df0e-c994-48df-a841-da20c8bfe4d7/sist-en-iso-1797-1-2000-a1-2000>

Ta slovenski standard je istoveten z: EN ISO 1797-1:1995/A1:1999

ICS:

11.060.25 Zobotehnični instrumenti Dental instruments

SIST EN ISO 1797-1:2000/A1:2000 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 1797-1:2000/A1:2000

<https://standards.iteh.ai/catalog/standards/sist/5b08df0e-c994-48df-a841-da20c8bfe4d7/sist-en-iso-1797-1-2000-a1-2000>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 1797-1:1995/A1

May 1999

ICS 11.060.20

English version

Dental rotary instruments - Shanks - Part 1: Shanks made of metals (ISO 1797-1:1992/Amendment 1:1997)

Instruments rotatifs dentaires - Queues - Partie 1: Queues en matériaux métalliques (ISO 1797-1:1992/Amendement 1:1997)

Rotierende Dentalinstrumente - Schäfte - Teil 1: Schäfte aus Metall (ISO 1797-1:1992/Ergänzung 1:1997)

This amendment A1 modifies the European Standard EN ISO 1797-1:1995; it was approved by CEN on 16 April 1999.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for inclusion of this amendment into the relevant national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This amendment exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

SIST EN ISO 1797-1:2000/A1:2000

<https://standards.iteh.ai/catalog/standards/sist/5b08df0e-c994-48df-a841-da20c8bf4d7/sist-en-iso-1797-1-2000-a1-2000>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Page 2

EN ISO 1797-1:1995/A1:1999

Foreword

The text of the Amendment ISO 1797-1:1995/A1:1999 to the International Standard from Technical Committee ISO/TC 106 "Dentistry" of the International Organization for Standardization (ISO) has been taken over as an Amendment to the European Standard by Technical Committee CEN/TC 55 "Dentistry", the secretariat of which is held by DIN.

This Amendment to the European Standard EN ISO 1797-1:1995 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 1999, and conflicting national standards shall be withdrawn at the latest by November 1999.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of the Amendment to the International Standard ISO 1797-1:1992/Amendment 1:1997 has been approved by CEN as an Amendment to the European Standard without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 1797-1:2000/A1:2000](https://standards.iteh.ai/catalog/standards/sist/5b08df0e-c994-48df-a841-da20c8bfe4d7/sist-en-iso-1797-1-2000-a1-2000)

<https://standards.iteh.ai/catalog/standards/sist/5b08df0e-c994-48df-a841-da20c8bfe4d7/sist-en-iso-1797-1-2000-a1-2000>



INTERNATIONAL STANDARD

ISO
1797-1

First edition
1992-02-15

AMENDMENT 1
1997-12-15

Dental rotary instruments — Shanks —

Part 1: Shanks made of metals

AMENDMENT 1

iTeh STANDARD PREVIEW

Instruments rotatifs dentaires — Queues —
(standards.iteh.ai)

Partie 1:

Queues en matériaux métalliques

<https://standards.iteh.ai/standards/sist/5b08df0e-c994-48df-a841-da20c8bfe4d7/sist-en-iso-1797-1-2000-a1-2000>



Reference number
ISO 1797-1:1992/Amd.1:1997(E)

ISO 1797-1:1992/Amd.1:1997(E)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Amendment 1 to International Standard ISO 1797:1992 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

iTeh STANDARD PREVIEW
(Standards.it.ih.ai)
<https://standards.it.ih.ai/catalog/standards/sist/5b08df0e-c994-48df-a841-da20c8bfe4d7/sist-en-iso-1797-1-2000-a1-2000>

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet central@iso.ch
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

Dental rotary instruments — Shanks —

Part 1: Shanks made of metals

AMENDMENT 1

Page 4

After subclause 5.4 add the following subclause 5.5.

5.5 Marking

Marking, if envisaged, should be outside the fitting length l_1 .

If the marking is applied within the fitting length l_1 , the outside diameter of the marking shall not exceed the effective diameter of the shank.

Marking shall be resistant to disinfection, cleaning and sterilization according to the manufacturer's instructions.