

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-31: Particular requirements for the basic safety and essential performance
of external cardiac pacemakers with internal power source**

**Appareils électromédicaux –
Partie 2-31: Exigences particulières pour la sécurité de base et les performances
essentielles des stimulateurs cardiaques externes à source d'énergie interne**

<https://standards.iteh.ai/catalog/standards/iec/190659af-c8d9-4d13-b6d0-1768e68aa6ce/iec-60601-2-31-2008>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2011 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch

Tel.: +41 22 919 02 11

Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch

Tél.: +41 22 919 02 11

Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-31: Particular requirements for the basic safety and essential performance
of external cardiac pacemakers with internal power source**

**Appareils électromédicaux –
Partie 2-31: Exigences particulières pour la sécurité de base et les performances
essentielle des stimulateurs cardiaques externes à source d'énergie interne**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	8
201.2 Normative references	9
201.3 Terms and definitions	10
201.4 General requirements	11
201.5 General requirements for testing ME EQUIPMENT.....	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	12
201.7 ME EQUIPMENT identification, marking and documents.....	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	16
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	18
201.10 Protection against unwanted and excessive radiation HAZARDS.....	18
201.11 Protection against excessive temperatures and other HAZARDS.....	18
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	19
201.13 HAZARDOUS SITUATIONS and fault conditions.....	23
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	23
201.15 Construction of ME EQUIPMENT.....	23
201.16 ME SYSTEMS	23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	23
202 Electromagnetic compatibility – Requirements and tests.....	23
Annexes	25
Annex AA (informative) Particular guidance and rationale	26
Bibliography	36
Index of defined terms used in this particular standard	37
Figure 201.101 – Measuring circuit for the PATIENT AUXILIARY CURRENT for ME EQUIPMENT with an INTERNAL ELECTRICAL POWER SOURCE	18
Figure 201.102 – Measuring circuit for the MAXIMUM TRACKING RATE	21
Figure 201.103 – Initial oscilloscope display when measuring MAXIMUM TRACKING RATE	21
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements.....	12
Table 201.102 – DUAL CHAMBER connector terminal marking.....	13
Table 201.103 – Measurement method accuracy	20
Table 202.101 – Static discharge requirements	24
Table AA.1 – EXTERNAL PACEMAKER HAZARD inventory	27

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-31: Particular requirements for the basic safety and essential performance of external cardiac pacemakers with internal power source

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of IEC 60601-2-31 consists of the second edition (2008) [documents 62D/603/CDV and 62D/667/RVC] and its amendment 1 (2011) [documents 62D/918/FDIS and 62D/931/RVD]. It bears the edition number 2.1.

The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.

International standard IEC 60601-2-31 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition of IEC 60601-2-31 is aligned with IEC 60601-1:2005, and contains minimal technical revisions from the first edition.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

iTech Standards
(<https://standards.iteh.ai>)
Document Preview

IEC 60601-2-31:2008

<https://standards.iteh.ai/cou/standards/iec/192659af-c8d9-4d13-b6d0-1768e68aa6ce/iec-60601-2-31-2008>

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of external cardiac pacemakers with an internal power source.

Basically, CARDIAC PACEMAKERS treat cardiac arrhythmias. Such arrhythmias reduce cardiac output and can lead to confusion, dizziness, loss of consciousness and death. The objective of pacing is to restore cardiac rhythm and output appropriate to the PATIENT's physiological needs.

There are two distinct families of CARDIAC PACEMAKERS, ~~IMPLANTABLE~~ implantable PACEMAKERS and EXTERNAL PACEMAKERS. EXTERNAL PACEMAKERS are used to pace PATIENTS temporarily prior to implanting an ~~IMPLANTABLE~~ implantable PACEMAKER as well as for temporary pacing related to other medical procedures, e.g. open heart surgery.

CARDIAC PACEMAKERS differ in the various ways in which they maintain and monitor cardiac activity in different circumstances. The simplest model stimulates the atrium or ventricle independently of the cardiac activity; others detect atrial or ventricular activity and stimulate the atrium or ventricle as and when this is necessary, others, more complex, detect the spontaneous heart activity and stimulate appropriately the atrium and/or the ventricle. Certain PACEMAKERS work on preset frequency values, amplitudes and impulse duration. Others can have several values for parameters.

Standards for EXTERNAL PACEMAKERS require attention to information which will aid in selecting and applying these devices. It is through these aspects of standardization that the central role of clinical experience should be, or has been, acknowledged. The ability to predict how a ~~pacemaker~~ PACEMAKER will perform in a specific ~~patient~~ PATIENT based on testing of a device to a set of technical criteria is limited.

This particular standard does not take into consideration the specific safety aspects of EXTERNAL PACEMAKERS that are connected to a SUPPLY MAINS while simultaneously connected to the PATIENT.

This particular standard amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1. General requirements for basic safety and essential performance*, hereinafter referred to as the general standard (see 1.4).

The requirements are followed by specifications for the relevant tests.

Following the decision taken by subcommittee 62D at the meeting in Washington in 1979, a "General guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

Clauses or subclauses for which there are explanatory notes in Annex AA are marked with an asterisk (*).

An inventory of the PATIENT's safety posed by EXTERNAL PACEMAKERS and a rationale for the safety requirements contained in this particular standard are given in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

INTRODUCTION (to amendment 1)

The purpose of this amendment is to address comments received during the process of harmonizing the standard in Europe, update several references to defined terms that were not printed in SMALL CAPS, and improve terminology usage.

Withdrawing

iTech Standards
(<https://standards.iteh.ai>)
Document Preview

<https://standards.iteh.ai/standards/iec/193659af-c8d9-4d13-b6d0-1768e68aa6ce/iec-60601-2-31-2008>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-31: Particular requirements for the basic safety and essential performance of external cardiac pacemakers with internal power source

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of EXTERNAL PACEMAKERS powered by an INTERNAL ELECTRICAL POWER SOURCE, hereafter referred to as ME EQUIPMENT.

~~This standard applies to PATIENT CABLES as defined in 201.3.109.~~

This standard applies to PATIENT CABLES as defined in 201.3.109, but does not apply to LEADS as defined in 201.3.106.

~~If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.~~

HAZARDS inherent in the intended physiological function of ME EQUIPMENT within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

~~This standard does not apply to the implantable parts of active implantable medical devices ACTIVE IMPLANTABLE MEDICAL DEVICES covered by ISO 14708-1 This standard does not apply to EXTERNAL PACEMAKERS which can be connected directly or indirectly to a SUPPLY MAINS.~~

This standard does not apply to transthoracic and oesophageal pacing ME EQUIPMENT and antitachycardia ME EQUIPMENT.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for EXTERNAL PACEMAKERS ~~AS DEFINED IN~~ as defined in 201.3.103.

201.1.3 Collateral standards

Addition:

¹⁾ The general standard is IEC 60601-1:2005.

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 2 of this particular standard.

IEC 60601-1-3 does not apply.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses or figures which are additional to those of the general standard are numbered starting from 201.101, additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility - Requirements and tests*

Addition:

ISO 14708-2:2005, *Implants for surgery – Active implantable medical devices – Part 2: Cardiac pacemakers*

ANSI/AAMI PC69:2007, *Active implantable medical devices – Electromagnetic compatibility – EMC test protocols for implantable cardiac pacemakers and implantable cardioverter defibrillators*

NOTE Informative references are listed in the bibliography on page 36.

201.3 * Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 and ISO 14708-2:2005 apply, except as follows:

NOTE An index of defined terms is found beginning on page 35.

Addition:

201.3.101

ACTIVE IMPLANTABLE MEDICAL DEVICE

active medical device which is intended to be totally or partially introduced, surgically or medically, into the human body or by medical intervention into a natural orifice, and which is intended to remain in place after the procedure

[ISO 14708-1:2000, definition 3.3]

201.3.102

BATTERY DEPLETION INDICATOR

means of indicating when the battery should be replaced

201.3.103

CARDIAC PACEMAKER

ME EQUIPMENT intended to treat bradyarrhythmias

201.3.104

DUAL CHAMBER

relating to both atrium and ventricle

201.3.105

EXTERNAL PACEMAKER

CARDIAC PACEMAKER ~~with~~ consisting of a NON-IMPLANTABLE PULSE GENERATOR and PATIENT CABLE(S) (if used)

201.3.106

LEAD

flexible tube enclosing one or more insulated electrical conductors, intended to transfer electrical energy along its length between the EXTERNAL PACEMAKER and the ~~patient's~~ PATIENT'S heart

[ISO 14708-1:2000, definition 3.5 modified]

201.3.107**MAXIMUM TRACKING RATE**

maximum PULSE RATE at which the NON-IMPLANTABLE PULSE GENERATOR will respond on a 1:1 basis to a triggering signal

[ISO 14708-2:2005, definition 3.3.18 modified]

201.3.108**NON-IMPLANTABLE PULSE GENERATOR**

ME EQUIPMENT with an INTERNAL ELECTRICAL POWER SOURCE which is intended for use outside the body and which produces a periodic electrical ~~pulse~~ PULSE intended to stimulate the heart through a LEAD (or combination of a LEAD and PATIENT CABLE)

201.3.109**PATIENT CABLE**

cable used to extend the distance between the NON-IMPLANTABLE PULSE GENERATOR and the pacing LEAD

201.3.110**POST-VENTRICULAR ATRIAL REFRACTORY PERIOD****PVARP**

atrial refractory period minus the AV delay

201.3.111**PRIMARY BATTERY**

one or more cells, which are not designed to be electrically recharged, that are fitted with devices necessary for use, for example case, terminals, marking and protective devices

[IEC 60050-482:2004, definition 482-01-04 modified]

201.3.112**SINGLE CHAMBER**

relating to either atrium or ventricle

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.3 ESSENTIAL PERFORMANCE

Additional subclause:

201.4.3.101 Additional ESSENTIAL PERFORMANCE requirements

Additional ESSENTIAL PERFORMANCE requirements are found in the subclauses listed in Table 201.101.

Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
BATTERY DEPLETION INDICATOR	201.11.8
ME EQUIPMENT parameter stability	201.12.1.101
PULSE AMPLITUDE stability	201.12.1.102
Disarming runaway rate protection	201.12.4.1
Deliberate action required to change settings	201.12.4.101
Parameter stability at onset of the BATTERY DEPLETION INDICATOR	201.12.4.102
Runaway protection	201.12.4.103
Interference reversion in the presence of sensed electrical interference	201.12.4.104
Limit at which the ventricle is paced in response to sensed atrial activity	201.12.4.105

201.4.10.1 Source of power for ME EQUIPMENT

Replacement:

ME EQUIPMENT shall be powered by a PRIMARY BATTERY.

Compliance is checked by inspection of the ACCOMPANYING DOCUMENTS.

201.4.10.2 ~~Supply mains~~ SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS

This subclause of the general standard does not apply.

201.4.11 * Power input

This subclause of the general standard does not apply.

201.5 General requirements for testing ME EQUIPMENT

Clause 5 of the general standard applies.

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies, except as follows:

201.6.2 * Protection against electric shock

Replacement:

ME EQUIPMENT shall be classified as INTERNALLY POWERED ME EQUIPMENT.

ME EQUIPMENT shall be recognized as INTERNALLY POWERED only if no external connections to an electrical power source are provided.

APPLIED PARTS shall be classified as TYPE CF APPLIED PARTS. APPLIED PARTS shall be classified as DEFIBRILLATION-PROOF APPLIED PARTS.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows: