



**SLOVENSKI STANDARD
SIST EN ISO 4796-3:2000**

01-september-2000

**BUXca Yý U
SIST ISO 4796:1995**

Laboratorijska steklovina - Steklenice - 3. del: Sesalne steklenice (ISO 4796-3:2000)

Laboratory glassware - Bottles - Part 3: Aspirator bottles (ISO 4796-3:2000)

Laborgeräte aus Glas - Flaschen - Teil 3: Tubusflaschen (ISO 4796-3:2000)

Verrerie de laboratoire - Flacons - Partie 3: Flacons à tubulure basse (ISO 4796-3:2000)

Ta slovenski standard je istoveten z: EN ISO 4796-3:2000

[SIST EN ISO 4796-3:2000](https://standards.iteh.ai/SIST/EN-ISO-4796-3-2000)

<https://standards.iteh.ai/SIST/EN-ISO-4796-3-2000>

ICS:

71.040.20	Laboratorijska posoda in aparati	Laboratory ware and related apparatus
-----------	----------------------------------	---------------------------------------

SIST EN ISO 4796-3:2000

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 4796-3:2000](https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000)

<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 4796-3

April 2000

ICS 71.040.20

English version

Laboratory glassware - Bottles - Part 3: Aspirator bottles (ISO 4796-3:2000)

Verrerie de laboratoire - Flacons - Partie 3: Flacons à tubulure basse (ISO 4796-3:2000)

Laborgeräte aus Glas - Flaschen - Teil 3: Tubusflaschen (ISO 4796-3:2000)

This European Standard was approved by CEN on 1 April 2000.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN ISO 4796-3:2000](https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000)

<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Foreword

The text of the International Standard ISO 4796-3:2000 has been prepared by Technical Committee ISO/TC 48 "Laboratory glassware and related apparatus" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2000, and conflicting national standards shall be withdrawn at the latest by October 2000.


According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

NOTE FROM CEN/CS: The foreword is susceptible to be amended on reception of the German language version. The confirmed or amended foreword, and when appropriate, the normative annex ZA for the references to international publications with their relevant European publications will be circulated with the German version.

ITeH STANDARD PREVIEW
(standards.iteh.ai)
Endorsement notice

The text of the International Standard ISO 4796-3:2000 was approved by CEN as a European Standard without any modification.

SEE EN ISO 4796-3:2000
<http://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>



INTERNATIONAL
STANDARD

ISO
4796-3

First edition
2000-04-01

Laboratory glassware — Bottles —

Part 3:
Aspirator bottles

Verrerie de laboratoire — Flacons —

Partie 3: Flacons à tubulure basse

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 4796-3:2000

<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>



Reference number
ISO 4796-3:2000(E)

© ISO 2000

ISO 4796-3:2000(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 4796-3:2000](https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000)

<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 734 10 79
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 4796 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 4796-3 was prepared by Technical Committee ISO/TC 48, *Laboratory glassware and related apparatus*, Subcommittee SC 2, *General laboratory glassware (other than measuring apparatus)*.

ISO 4796 consists of the following parts, under the general title *Laboratory glassware — Bottles*:

— *Part 1: Screw-neck bottles*

— *Part 2: Conical neck bottles*

— *Part 3: Aspirator bottles*

IT-EN STANDARD PREVIEW

(standards.iteh.ai)

[SIST EN ISO 4796-3:2000](https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000)

<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 4796-3:2000

<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-49389d10f86b/sist-en-iso-4796-3-2000>

Laboratory glassware — Bottles —

Part 3: Aspirator bottles

1 Scope

This part of ISO 4796 specifies a series of aspirator bottles with a screw neck or with a conical neck suitable for the delivery of liquid chemicals and reagents in general laboratory use.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 4796. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 4796 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 383:1976, *Laboratory glassware — Interchangeable conical ground joints.*
<https://standards.iteh.ai/catalog/standards/sist/13474201-8310-4ca9-99a0-4938b11086b/sist-en-iso-4796-3-2000>

ISO 4796-1:2000, *Laboratory glassware — Bottles — Part 1: Screw-neck bottles.*

ISO 4796-2:2000, *Laboratory glassware — Bottles — Part 2: Conical neck bottles.*

3 Types and capacities

3.1 There are three types of aspirator bottles specified in this part of ISO 4796:

- type ASN: aspirator bottles with a screw neck;
- type ACS: aspirator bottles with a conical neck socket;
- type ACJ: aspirator bottles with an interchangeable conical ground joint.

3.2 The nominal capacities of aspirator bottles shall be chosen from the following series:

0,5 l — 1 l — 2 l — 5 l — 10 l and 20 l

3.3 The nominal capacity of an aspirator bottle indicates the quantity of liquid which a bottle of average wall thickness shall contain when the bottle is filled to the turn of the shoulder.

3.4 The design of the bottle shall be such that the total capacity to the base of the neck shall be approximately 15 % greater than that to the shoulder.