

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical electrical equipment –  
Part 2-58: Particular requirements for the basic safety and essential performance  
of lens removal devices and vitrectomy devices for ophthalmic surgery**

**Appareils électromédicaux –  
Partie 2-58: Exigences particulières pour la sécurité de base et les performances  
essentielles des dispositifs de retrait du cristallin et des dispositifs de  
vitrectomie pour la chirurgie ophtalmique**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2008 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland  
Email: [inmail@iec.ch](mailto:inmail@iec.ch)  
Web: [www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: [www.iec.ch/online\\_news/justpub](http://www.iec.ch/online_news/justpub)

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: [www.iec.ch/webstore/custserv](http://www.iec.ch/webstore/custserv)

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: [csc@iec.ch](mailto:csc@iec.ch)  
Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00

### A propos de la CEI

La Commission Electrotechnique internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: [www.iec.ch/searchpub/cur\\_fut-f.htm](http://www.iec.ch/searchpub/cur_fut-f.htm)

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: [www.iec.ch/online\\_news/justpub](http://www.iec.ch/online_news/justpub)

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: [www.iec.ch/webstore/custserv/custserv\\_entry-f.htm](http://www.iec.ch/webstore/custserv/custserv_entry-f.htm)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: [csc@iec.ch](mailto:csc@iec.ch)  
Tél.: +41 22 919 02 11  
Fax: +41 22 919 03 00

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Medical electrical equipment –**  
**Part 2-58: Particular requirements for the basic safety and essential performance**  
**of lens removal devices and vitrectomy devices for ophthalmic surgery**

**Appareils électromédicaux –**  
**Partie 2-58: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des dispositifs de retrait du cristallin et des dispositifs de**  
**vitrectomie pour la chirurgie ophtalmique**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

PRICE CODE  
CODE PRIX

U

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
201.1 Scope, object and related standards .....	6
201.2 Normative references .....	7
201.3 Terms and definitions.....	8
201.4 General requirements.....	10
201.5 General requirements for testing of ME EQUIPMENT.....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS .....	11
201.7 ME EQUIPMENT identification, marking and documents.....	11
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	12
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	12
201.10 Protection against unwanted and excessive radiation HAZARDS.....	12
201.11 Protection against excessive temperatures and other HAZARDS.....	12
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	13
201.13 Hazardous situations and fault conditions .....	20
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	20
201.15 Construction of ME EQUIPMENT .....	20
201.16 ME SYSTEMS.....	20
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	20
202 Electromagnetic compatibility – Requirements and tests .....	20
Annexes .....	22
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	22
Annex AA (informative) Particular guidance and rationale.....	23
Bibliography.....	25
Index of defined terms .....	26
Figure 201.101 – Test method for gravity fed IRRIGATION.....	14
Figure 201.102 – Test method for pressurized IRRIGATION.....	15
Figure 201.103 – Test method for ASPIRATION pressure measurement/display accuracy.....	16
Table 201.101 – Legends of symbols for Figure 201.101 to Figure 201.103 .....	16
Table 201.C.101 – ACCOMPANYING DOCUMENTS, instructions for use of LENS REMOVAL DEVICES and VITRECTOMY DEVICES or its parts .....	22

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-58: Particular requirements for the basic safety  
and essential performance of lens removal devices  
and vitrectomy devices for ophthalmic surgery**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-58 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and ISO subcommittee SC 7: Ophthalmic optics and instruments of ISO technical committee 172: Optics and photonics.

It is published as a double logo standard.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/701/FDIS	62D/723/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 14 P-members out of 15 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this particular standard will remain unchanged until the maintenance result date indicated on the IEC web site under “<http://webstore.iec.ch>” in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended

## INTRODUCTION

LENS REMOVAL DEVICES and VITRECTOMY DEVICES are used widely in ophthalmology to perform anterior-segment and posterior-segment surgery on the human eye. Commercial use of these MEDICAL ELECTRICAL EQUIPMENT devices began in the early 1970s. This International Standard defines particular requirements for BASIC SAFETY and ESSENTIAL PERFORMANCE of LENS REMOVAL DEVICES and VITRECTOMY DEVICES, comprising an equipment console, surgical HANDPIECES, and ACCESSORIES connected to this ME EQUIPMENT.

In many parts of the world LENS REMOVAL DEVICES and VITRECTOMY DEVICES are used in combination by ophthalmic surgeons to perform combined anterior-segment (lens removal) and posterior-segment (vitreoretinal) surgical PROCEDURES to maximize surgical outcomes. For this reason both LENS REMOVAL DEVICES and VITRECTOMY DEVICES are covered in this International Standard.

As all particular standards in the IEC 60601-1 series are based on the general standard IEC 60601-1:2005, the user of this standard is reminded that RISK MANAGEMENT plays an important role in the use of this particular standard. Compliance with the requirements of this particular standard should be documented in the RISK MANAGEMENT FILE to ensure the HAZARDS associated with the product have been considered fully.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

IEC 60601-2-58:2008

[https://standards.iteh.ai/catalog/standards/sist/b241d4ce-01a5-486d-be49-f4abe03c7539/iec-](https://standards.iteh.ai/catalog/standards/sist/b241d4ce-01a5-486d-be49-f4abe03c7539/iec-60601-2-58-2008)

60601-2-58-2008



## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-58: Particular requirements for the basic safety and essential performance of lens removal devices and vitrectomy devices for ophthalmic surgery

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1)</sup> applies, except as follows:

##### 201.1.1 Scope

*Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of LENS REMOVAL DEVICES and VITRECTOMY DEVICES for ophthalmic surgery (as defined in 201.3.208 and 201.3.217) and associated ACCESSORIES that can be connected to this MEDICAL ELECTRICAL EQUIPMENT, hereafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for LENS REMOVAL DEVICES and VITRECTOMY DEVICES for ophthalmic surgery (as defined in 201.3.208 and 201.3.217) and associated ACCESSORIES that can be connected to the ME EQUIPMENT and are to be tested together or individually.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1 and Clause 201.2 of this particular standard.

IEC 60601-1-2 applies as modified in Clause 202. All other published collateral standards in the IEC 60601-1 series apply as published.

##### 201.1.4 Particular standards

*Replacement:*

<sup>1)</sup> The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.



In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the “general standard”. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

“Replacement” means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

“Addition” means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

“Amendment” means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses or figures which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this standard” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

## 201.2 Normative references

NOTE Informative references are listed in the bibliography beginning on page 25.

Clause 2 of the general standard applies, except as follows:

### *Addition:*

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility - Requirements and tests*

IEC 60601-2-2, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*<sup>2)</sup>

IEC 60601-2-22, *Medical electrical equipment – Part 2-22: Particular requirements for the basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*

IEC 61847:1998, *Ultrasonics – Surgical systems – Measurement and declaration of the basic output characteristics*

ISO 11607-1:2006, *Packaging for terminally sterilized medical devices – Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2:2006, *Packaging for terminally sterilized medical devices – Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 15752:2000, *Ophthalmic instruments – Endoilluminators – Fundamental requirements and test methods for optical radiation safety*

ISO 17664:2004, *Sterilization of medical devices – Information to be provided by the manufacturer for the processing of resterilizable medical devices*

### **201.3 Terms and definitions**

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, apply, except as follows:

NOTE An index of defined terms is found beginning on page 26.

*Addition:*

#### **201.3.201**

##### **ASPIRATION**

drawing fluid or gas out of the eye by use of suction

#### **201.3.202**

##### **DIATHERMY**

surgical technique using high frequency (HF) electrical currents used for example to coagulate blood or bind tissues together

NOTE The terms “cautery” or “coagulation” have also been used in this context.

#### **201.3.203**

##### **DRAIN CONTAINER**

sealed container (or bag) in which aspirated fluid is collected

#### **201.3.204**

##### **HANDPIECE**

##### **PROBE**

handheld APPLIED PART, an ACCESSORY of LENS REMOVAL DEVICES or VITRECTOMY DEVICES

<sup>2)</sup> 5<sup>th</sup> edition, to be published. The title of the fourth edition is IEC 60601-2-2:2006, *Medical electrical equipment – Part 2-2: Particular requirements for the safety of high frequency surgical equipment*.

**201.3.205****LASER**

any device which can be made to produce or amplify electromagnetic radiation in the wavelength range from 180 nm to 1 mm primarily by the process of controlled stimulated emission

[IEC 60825-1:2007, definition 3.41]

**201.3.206****LASER FRAGMENTATION**

method by which the lens is broken into small fragments using LASER energy

**201.3.207****LENS REMOVAL**

removal of unwanted lens tissue

**201.3.208****LENS REMOVAL DEVICE**

ME EQUIPMENT or ME SYSTEM designed to remove lens material which incorporates an IRRIGATION and ASPIRATION function, and a mechanism for LENS REMOVAL such as PHACOEMULSIFICATION, LIQUEFACTION, or LASER FRAGMENTATION

NOTE These devices may also be used for other ocular surgical purposes

**201.3.209****LIQUEFACTION FRAGMENTATION****LIQUEFACTION**

method by which the lens is broken into small fragments by means of pulses of ophthalmic IRRIGATION solution

**201.3.210****OCULAR IRRIGATION****IRRIGATION**

introduction of a liquid into the eye

NOTE The term "infusion" has also been used in this context

**201.3.211****PHACOFRAGMENTATION**

method by which the lens is broken into small fragments using energy such as from ultrasonic devices

NOTE Refer to the definition of LENS REMOVAL DEVICE in 201.3.208.

Historically PHACOFRAGMENTATION (term is also identified as phacoemulsification) has been a surgical PROCEDURE that uses ultrasonic energy to fragment (or emulsify) a cataractous lens and removes the lens material through a small incision. Recently, other emerging energy modalities, including LASER FRAGMENTATION and LIQUEFACTION, have also been utilized in the removal of the cataractous lens through a small incision.

**201.3.212****PHOTORETINITIS**

retinal injury resulting from a very intense retinal radiant exposure

[ISO 15752:2000, definition 3.7]

**201.3.213****PRIME****PRIMING**

pre-operative setup PROCEDURE to fill TUBING SET (fluid path) with ophthalmic IRRIGATION solution

**201.3.214**

**TIP**

hollow needle-like device that is attached to a HANDPIECE

**201.3.215**

**TUBING SET**

set of tubes to contain fluid, designed to provide IRRIGATION to the eye and ASPIRATION from the eye

**201.3.216**

**VITRECTOMY**

surgical PROCEDURE to remove vitreous humour, membranes, blood, lens tissue and other material from the eye, involving IRRIGATION, ASPIRATION and vitreous cutting

NOTE The PROCEDURE may also include illumination, DIATHERMY, fluid/gas exchanges, and injection of viscous fluids.

**201.3.217**

**VITRECTOMY DEVICE**

ME EQUIPMENT or ME SYSTEM used to perform VITRECTOMY

NOTE These devices may also be used for other ocular surgical purposes.

**201.4 General requirements**

Clause 4 of the general standard applies, except as follows:

**201.4.3 \* ESSENTIAL PERFORMANCE**

*Addition:*

**201.4.3.101 General**

For LENS REMOVAL DEVICES and VITRECTOMY DEVICES no ESSENTIAL PERFORMANCE has been identified in general. If the LENS REMOVAL DEVICES and VITRECTOMY DEVICES have functions other than those specified in Clause 201.12, the MANUFACTURER shall identify which of these functions of the ME EQUIPMENT and ME SYSTEMS is ESSENTIAL PERFORMANCE.

*Compliance is checked by inspection of the RISK MANAGEMENT FILE.*

*Additional subclause:*

**201.4.101 \* Additional functions**

If there is a DIATHERMY function used for the LENS REMOVAL DEVICE and VITRECTOMY DEVICE, that function shall meet the requirements of IEC 60601-2-2.

If the ME EQUIPMENT includes a LASER function, that function shall meet the requirements of IEC 60601-2-22.

If there is an illumination function used to illuminate the eye during surgery that is part of the ME EQUIPMENT or ME SYSTEM then that portion of the ME EQUIPMENT or ME SYSTEM shall meet Clause 4.2 of ISO 15752:2000.

**201.5 General requirements for testing of ME EQUIPMENT**

Clause 5 of the general standard applies.

## 201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies.

## 201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

### 201.7.9.2.2 Warning and safety notices

*Addition:*

The instructions for use shall additionally include the following warning and safety notices:

- a) a warning to use only recommended TUBING SET(s);
- b) if an electrically adjustable ophthalmic IRRIGATION solution support pole is used a warning not to modify pole height or manually force the pole height because this could cause incorrect indication of bottle height and PATIENT injury;
- c) a warning never to intentionally modify HANDPIECES or TIPS (e.g. do not bend, cut, or engrave them) as they could break or malfunction;
- d) a warning to the OPERATOR not to touch an activated ultrasonic HANDPIECE TIP as injuries could occur;
- e) if applicable, warnings related to lamp replacement (e.g. RISK of injury, ratings of lamp, damage to lamp, damage to machine, etc.);
- f) if applicable, a warning to the OPERATOR that care should be taken to avoid concentrating the output of an illumination module on a small area of the retina for unnecessarily prolonged periods of time due to the potential for PHOTORETINITIS and serious permanent PATIENT injury;
- g) if applicable, a warning to the OPERATOR that inadvertent activation of functions that are intended for PRIMING or tuning HANDPIECES while the HANDPIECE is in the eye can create a HAZARDOUS SITUATION that could result in PATIENT injury;
- h) where gravity is relevant to performance, the ophthalmic IRRIGATION solution source shall be at or above the PATIENT's eye level;
- i) a warning to the OPERATOR to ensure sufficient volume of IRRIGATION solution for the PROCEDURE. The level should be monitored during the PROCEDURE;
- j) if applicable, a warning to the OPERATOR to ensure that the maximum capacity of the DRAIN CONTAINER is not exceeded as this could cause a HAZARDOUS SITUATION to the PATIENT.

### 201.7.9.2.8 Start-up PROCEDURE

*Addition:*

The instructions for use shall include instructions to perform functional checks of the system before first use of the day.

### 201.7.9.2.9 Operating instructions

*Addition:*

The operating instructions shall additionally include:

- a) if applicable, instructions regarding loading, PRIMING, changing, and reloading the TUBING SET(s), and the TUBING SET(s) change interval to maintain the specified performance;
- b) if applicable, instructions regarding the use of clamps on a TUBING SET, the avoidance of ophthalmic IRRIGATION solution free flow conditions, and the PROCEDURE to be followed when changing the ophthalmic IRRIGATION solution source;

c) instructions regarding securely attaching plugs, HANDPIECE cables and other connectors.

#### **201.7.9.2.12 Cleaning, disinfection, and sterilization**

For parts that are resterilizable, the information for processing shall be in accordance with ISO 17664:2004. This information shall be provided to the RESPONSIBLE ORGANIZATION or the OPERATOR.

#### **201.7.9.2.13 Maintenance**

*Addition:*

The instructions for use shall provide the OPERATOR or RESPONSIBLE ORGANIZATION with a recommendation to inspect all HANDPIECE cables and any cords on a regular basis and a recommendation as to the action to take if damage (e.g. exposed wire, nicks in the insulation, deformation, etc.) is observed.

#### **201.8 Protection against electrical HAZARDS from ME EQUIPMENT**

Clause 8 of the general standard applies.

#### **201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS**

Clause 9 of the general standard applies.

#### **201.10 Protection against unwanted and excessive radiation HAZARDS**

Clause 10 of the general standard applies.

#### **201.11 Protection against excessive temperatures and other HAZARDS**

Clause 11 of the general standard applies, except as follows:

##### **201.11.1.2 Temperature of APPLIED PARTS**

##### **201.11.1.2.1 APPLIED PARTS intended to supply heat to a PATIENT**

*Replacement:*

HANDPIECES for DIATHERMY, PHACOFAGMENTATION, LASER and LIQUEFACTION are considered to be APPLIED PARTS intended to supply heat to a PATIENT.

The temperature or clinical effects shall be determined and documented in the RISK MANAGEMENT FILE.

##### **201.11.6.7 Sterilization of ME EQUIPMENT and ME SYSTEMS**

*Addition:*

The packaging for terminally sterilized ACCESSORIES for LENS REMOVAL DEVICES and VITRECTOMY DEVICES shall comply with the requirements of ISO 11607-1:2006. Validation requirements for forming, sealing, and assembly processes for this packaging shall be consistent with ISO 11607-2:2006.

## 201.12 Accuracy of controls and instruments and protection against hazardous outputs

Clause 12 of the general standard applies, except as follows:

### 201.12.1 Accuracy of controls and instruments

*Additional subclause*

#### 201.12.1.101 Additional accuracy of controls and instruments requirements

Additional requirements for accuracy of controls and instruments are detailed in subclauses 201.12.1.101.1 to 201.12.1.101.5 and 201.12.1.101.7 to 201.12.1.101.9.

##### 201.12.1.101.1 Accuracy of static IRRIGATION pressure

Static IRRIGATION pressure output shall not deviate from the indicated setting on the LENS REMOVAL DEVICES and VITRECTOMY DEVICES by more than  $\pm 20\%$  or  $\pm 10$  mmHg ( $\pm 1,3$  kPa) whichever is greater for a specific device in a defined configuration (see 201.12.4.101.1 for hazardous output limit).

*Compliance is checked by applying the relevant test method(s) 1 and/or 2:*

*Test method 1 (Gravity fed IRRIGATION)*

- (1) Set the test environment temperature to  $25\text{ °C} \pm 5\text{ °C}$ .
- (2) Install the TUBING SET(s) and PRIME the device in accordance with the MANUFACTURER's instructions for use.
- (3) Zero the pressure meter reading. Connect the pressure meter to the end of the IRRIGATION tubing and position the pressure meter within  $\pm 2,5$  cm of the simulated PATIENT eye level, see Figure 201.101.
- (4) Initiate the flow of fluid in accordance with the MANUFACTURER's instructions for use.
- (5) Set the gravity feed reservoir height to 0 cm or the lowest setting and record the pressure meter reading after 5 s.
- (6) Increase the reservoir height by 20 cm and wait for 5 s and record the pressure meter reading.
- (7) Repeat step 6 until the maximum reservoir height is reached.
- (8) Record the pressure meter reading at the maximum reservoir height.
- (9) Repeat the readings at the heights used in steps 5, 6 and 7 as the height is decreased and wait for 5 s and record the pressure meter reading at each point.
- (10) Confirm that all the readings are within the stated range.