SLOVENSKI STANDARD

SIST EN 60645-5:2005

december 2005

Elektroakustika – Avdiometrična oprema – 5. del: Instrumenti za merjenje ušesne akustične impedance/admitance (IEC 60645-5:2004)

Electroacoustics – Audiometric equipment – Part 5: Instruments for the measurement of aural acoustic impedance/admittance (IEC 60645-5:2004)

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 60645-5:2005 https://standards.iteh.ai/catalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-7831a85fe898/sist-en-60645-5-2005

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 60645-5:2005</u> https://standards.iteh.ai/catalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-7831a85fe898/sist-en-60645-5-2005

EUROPEAN STANDARD

EN 60645-5

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2005

ICS 17.140.50

Supersedes EN 61027:1993

English version

Electroacoustics – Audiometric equipment Part 5: Instruments for the measurement of aural acoustic impedance/admittance

(IEC 60645-5:2004)

Electroacoustique –
Appareillage audiométrique
Partie 5: Instruments pour la mesure
de l'impédance ou de l'admittance aurale
(impédancemètres ou admittancemètres)
(CEI 60645-5:2004)

Akustik –
Audiometer
Teil 5: Geräte zur Messung der
akustischen Impedanzen/Admittanzen
des Gehörs
(IEC 60645-5:2004)

(standards.iteh.ai)

SIST EN 60645-5:2005

https://standards.iteh.ai/catalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-

This European Standard was approved by CENELEC on 2004-12-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 29/563/FDIS, future edition 1 of IEC 60645-5, prepared by IEC TC 29, Electroacoustics, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60645-5 on 2004-12-01.

This European Standard supersedes EN 61027:1993.

The following dates were fixed:

 latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement

(dop) 2005-09-01

 latest date by which the national standards conflicting with the EN have to be withdrawn

(dow) 2007-12-01

Annex ZA has been added by CENELEC.

Endorsement notice iTeh STANDARD PREVIEW

The text of the International Standard IEC 60645-5:2004 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

ISO 389-1 NOTE Harmonized as EN ISO 389-1:2000 (not modified).

ISO 389-2 NOTE Harmonized as EN ISO 389-2:1996 (not modified).

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE Where an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	EN/HD	<u>Year</u>
IEC 60126	- 1)	IEC reference coupler for the measurement of hearing aids using earphones coupled to the ear by means of ear inserts	HD 305 S1	1977 2)
IEC 60601-1	- 1)	Medical electrical equipment Part 1: General requirements for safety	EN 60601-1	1990 ²⁾
IEC 60601-1-2	- ¹⁾	Medical electrical equipment Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2	2001 ²⁾
IEC 60601-1-4	_ 1) https://st	Medical electrical equipment Part 1-4: General requirements for safety Collateral standard: Programmable electrical medical systems	EN 60601-1-4 se-bf49-	1996 ²⁾
IEC 60645-1	2001	Electroacoustics - Audiological equipment Part 1: Pure-tone audiometers	EN 60645-1	2001
BIPM/IEC/IFCC/ ISO/IUPAP/OIML	- 1)	Guide to the expression of uncertainty in measurement (GUM)	-	-

¹⁾ Undated reference.

²⁾ Valid edition at date of issue.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 60645-5:2005</u> https://standards.iteh.ai/catalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-7831a85fe898/sist-en-60645-5-2005

NORME INTERNATIONALE INTERNATIONAL STANDARD

CEI IEC 60645-5

> Première édition First edition 2004-11

Electroacoustique – Appareillage audiométrique –

Partie 5:

Instruments pour la mesure de l'impédance pou de l'admittance aurale (impédancemètres ou admittancemètres) (standards.iten.ai)

Electroacoustics -5:2005 https://standards.iteh.avcatalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-Audiometriceguipment

Part 5: Instruments for the measurement of aural acoustic impedance/admittance

© IEC 2004 Droits de reproduction réservés — Copyright - all rights reserved

Aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'éditeur.

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



CODE PRIX PRICE CODE



CONTENTS

FO	REWC	PRD	5			
INT	INTRODUCTION9					
1	Scop	e	.11			
2	Norm	ative references	.11			
3	Terms and definitions					
4	Requ	irements for specific instruments	. 19			
5	General specifications					
	5.1	Measuring system	.21			
	5.2	Pneumatic system				
	5.3	Acoustic reflex activating stimulus system	. 25			
6	Demo	onstration of conformity with specifications	. 29			
	6.1	General	. 29			
	6.2	Impedance/admittance measuring system	. 29			
	6.3	Probe signal	. 29			
	6.4	Pneumatic system				
	6.5	Acoustic reflex activating stimulus signals				
	6.6	Maximum permitted expanded uncertainty of measurements	. 33			
7	Calib	ration cavities Teh STANDARD PREVIEW				
	7.1	General	. 35			
	7.2					
	7.3	Material of calibration cavity walls 160645-52005				
•	7.4	Connection of probe to a calibration cavity/a3d9dcfa-23a2-4c6c-bf49				
8		ral requirements				
	8.1	Marking				
	8.2	Instruction manual				
	8.3	Safety requirements				
	8.4	Warm-up time				
	8.5 8.6	Supply variation and environmental conditions				
9		pols, forms and formats for acoustic impedance/admittance data				
9	9.1	Symbols for indicating measured quantities				
	9.1	Tympanogram format				
	9.3	Acoustic reflex test format				
	9.4	Eustachian tube function test format				
10		ional characteristics to be specified by the manufacturer				
		Aural impedance/admittance measurement system				
		Pneumatic system				
		Acoustic reflex activating stimulus system				
		Analogue electrical output				
Ann		(informative) Routine calibration				
	Annex B (informative) Units and terms49					
Annex C (informative) Overall temporal characteristics						
	Bibliography57					
-	- 0 - 1	· <i>J</i>	_			

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTROACOUSTICS – AUDIOMETRIC EQUIPMENT –

Part 5: Instruments for the measurement of aural acoustic impedance/admittance

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.

 7831a85fe898/sist-en-60645-5-2005
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60645-5 has been prepared by IEC technical committee 29: Electroacoustics.

This first edition of IEC 60645-5 cancels and replaces the first edition of IEC 61027, published in 1991, and constitutes a technical revision.

The text of this standard is based on the following documents:

FDIS	Report on voting
29/563/FDIS	29/567/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

IEC 60645 consists of the following parts:

Part 1: Pure tone audiometers

Part 2: Equipment for speech audiometry

Part 3: Auditory test signals of short duration for audiometric and neuro-otological purposes

Part 4: Equipment for extended high-frequency audiometry

Part 5: Instruments for the measurement of aural acoustic impedance/admittance¹

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

reconfirmed: iTeh STANDARD PREVIEW

withdrawn; (standards.iteh.ai)

replaced by a revised edition, or

amended. <u>SIST EN 60645-5:2005</u>

https://standards.iteh.ai/catalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-7831a85fe898/sist-en-60645-5-2005

Future editions of all parts of this series will appear under the new general title Electroacoustics – Audiometric equipment.

INTRODUCTION

Developments in the field of diagnostic hearing measurement have resulted in a number of instruments designed to evaluate the acoustic impedance/admittance of the human ear by means of acoustic probe signals having different frequencies and temporal characteristics. The practical use of such instruments concerns to a large extent the measurement of changes in acoustic impedance/admittance caused either by varying the air pressure in the external acoustic meatus or by activating the middle ear muscle reflex.

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 60645-5:2005 https://standards.iteh.ai/catalog/standards/sist/a3d9dcfa-23a2-4e6e-bf49-7831a85fe898/sist-en-60645-5-2005

ELECTROACOUSTICS -AUDIOMETRIC EQUIPMENT -

Part 5: Instruments for the measurement of aural acoustic impedance/admittance

1 Scope

This part of IEC 60645 applies to instruments designed primarily for the measurement of acoustic impedance/admittance in the human external acoustic meatus using a stated probe tone. It is recognized that other probe signals may also be used. The standard defines the characteristics to be specified by the manufacturer, lays down performance specifications for three types of instruments and specifies the facilities to be provided on these types. This standard describes methods of test to be used for approval testing and guidance on methods for undertaking routine calibration.

The purpose of this standard is to ensure that measurements made under comparable test conditions with different instruments complying with the standard will be consistent. The standard is not intended to restrict development or incorporation of new features, nor to discourage innovative approaches.

iTeh STANDARD PREVIEW

Normative references (standards.iteh.ai)

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60126, IEC reference coupler for the measurement of hearing aids using earphones coupled to the ear by means of ear inserts 2

IEC 60601-1, Medical electrical equipment – Part 1: General requirements for safety

IEC 60601-1-2, Medical electrical equipment - Part 1-2: General requirements for safety -Collateral standard: Electromagnetic compatibility - Requirements and tests

IEC 60601-1-4, Medical electrical equipment - Part 1: General requirements for safety -Collateral standard: Programmable electrical medical systems

IEC 60645-1:2001, Electroacoustics - Audiological equipment - Part 1: Pure-tone audiometers

BIPM/IEC/ISO/IUPAC/IUPAP/OIML, Guide to the expression of uncertainty in measurement (GUM)

² To be revised as IEC 60318-5.