

**CEI 60601-1**  
(Troisième édition - 2005)

**Appareils électromédicaux –  
Partie 1: Exigences générales pour la sécurité de base  
et les performances essentielles**

**CORRIGENDUM 1**

Page 104

**7.2.3 Consultation des documents d'accompagnement**

*Dans la deuxième phrase:*

*au lieu de:*

« CEI 60878 Sécurité 01 »

*lire:*

« ISO 7010-M002 »

Page 136

**iTeh Standards**

**(<https://standards.iteh.ai>)**

**8.2.2 Connexion à une source externe en courant continu**

*Dans la première phrase:* **Document Preview**

*au lieu de:*

[IEC 60601-1:2005/COR1:2006](#)

« PERFORMANCES ESSENTIELLES »

<https://standards.iteh.ai/catalog/standards/iec/e0dfc349-b6cb-4e7e-9d87-3f1e0fbc7ffc/iec-60601-1-2005-cor1-2006>

*lire :*

« sa fonction prévue »

**IEC 60601-1**  
(Third edition - 2005)

**Medical electrical equipment –  
Part 1: General requirements for basic safety  
and essential performance**

**CORRIGENDUM 1**

Page 105

**7.2.3 Consult accompanying documents**

*In the second sentence:*

*instead of:*

“IEC 60878 Safety 01”

*read:*

“ISO 7010-M002”

Page 137

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

**8.2.2 Connection to an external d.c. power source**

*In the first sentence:* [IEC 60601-1:2005/COR1:2006](#)

<https://standards.iteh.ai/catalog/standards/iec/e0dfc349-b6cb-4e7e-9d87-3f1e0fbc7ffc/iec-60601-1-2005-cor1-2006>

*instead of:*

“ESSENTIAL PERFORMANCE”

*read:*

“its intended function”