
**Small-bore connectors for liquids and
gases in healthcare applications —**

**Part 1:
General requirements**

*Raccords de petite taille pour liquides et gaz utilisés dans le domaine
de la santé —*

Partie 1: Exigences générales

[ISO 80369-1:2010](https://standards.iteh.ai/catalog/standards/sist/80369-1-2010)

<https://standards.iteh.ai/catalog/standards/sist/80369-1-2010>

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

Withhold
iTeh STANDARD PREVIEW
(standards.itih.ai)
ISO 80369-1:2010
<https://standards.itih.ai/standards/standards/iso/80369-1-2010>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	2
3 Terms and definitions	2
4 Materials used for SMALL-BORE CONNECTORS.....	3
5 Requirements for SMALL-BORE CONNECTORS for specific APPLICATIONS.....	4
5.1 SMALL-BORE CONNECTOR incompatibility.....	4
5.2 BREATHING SYSTEMS and driving gases APPLICATIONS.....	4
5.3 Enteral and gastric APPLICATIONS.....	4
5.4 Urethral and urinary APPLICATIONS.....	4
5.5 Limb cuff inflation APPLICATIONS.....	4
5.6 Neuraxial APPLICATIONS.....	5
5.7 Intravascular or hypodermic APPLICATIONS.....	5
5.8 Alternative SMALL-BORE CONNECTORS.....	5
6 Additional SMALL-BORE CONNECTOR APPLICATIONS.....	5
7 PROCEDURE to assess a proposed new design of SMALL-BORE CONNECTOR for inclusion in this series of standards	6
7.1 General	6
7.2 Proposal initiation	6
7.3 PROCEDURE to assess acceptability and NON-INTERCONNECTABLE characteristics.....	6
7.3.1 Design.....	6
7.3.2 Design realization.....	6
7.3.3 Design VERIFICATION	7
7.3.4 Design validation.....	7
7.4 Design review	7
7.5 Subsequent parts of this series of standards	7
Annex A (informative) Rationale.....	8
Annex B (normative) Mechanical tests for verifying NON-INTERCONNECTABLE characteristics.....	11
Annex C (informative) APPLICATIONS of SMALL-BORE CONNECTORS	12
Annex D (informative) Reference to the Essential Principles	14
Bibliography.....	16
Terminology – Alphabetized index of defined terms.....	17

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 80369-1 was prepared by a Joint Working Group of Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*, IEC/TC 62, *Electrical equipment*, Subcommittee SC D, *Electrical equipment in medical practice* and CEN/CENELEC TC 3/WG 2, *Small-bore connectors*.

ISO 80369 consists of the following parts, under the general title, *Small-bore connectors for liquids and gases in healthcare applications*:

— *Part 1: General requirements*

The following parts are under preparation:

— *Part 2: Connectors for breathing systems and driving gases applications*

— *Part 3: Connectors for enteral applications*

— *Part 4: Connectors for urethral and urinary applications*

— *Part 5: Connectors for limb cuff inflation applications*

— *Part 6: Connectors for neuraxial applications*

— *Part 7: Connectors for intravascular or hypodermic applications*

Introduction

In the 1990s concern grew regarding the proliferation of MEDICAL DEVICES fitted with Luer CONNECTORS and the reports of PATIENT death or injury arising from misconnections that resulted in the inappropriate delivery of enteral solutions, intrathecal medication or compressed gases.

Concerns regarding the use of Luer CONNECTORS with enteral feeding tubes and gas sampling and gas delivery systems were raised with CEN/BT and the European Commission. In November 1997 the newly created CHeF steering group set up a Forum Task Group (FTG) to consider the problem.

The FTG produced CEN Report CR 13825, in which they concluded that there is a problem arising from the use of a single CONNECTOR design to a number of incompatible APPLICATIONS. In a coronary care unit there are as many as 40 Luer CONNECTORS on the MEDICAL DEVICES used with a single PATIENT. Therefore it is not surprising that misconnections are made.

MEDICAL DEVICES have for many years followed the established principle of "safety under single fault conditions". Simply stated this means that a single fault should not result in an unacceptable RISK. This principle is embodied in the requirements of numerous MEDICAL DEVICE standards. Extending this principle to the application of Luer CONNECTORS, i.e. that misconnection should not result in an unacceptable RISK to a PATIENT, the FTG recommended that the Luer CONNECTOR should be restricted to MEDICAL DEVICES intended to be connected to the vascular system or a hypodermic syringe. In addition, new designs of SMALL-BORE CONNECTORS should be developed for other APPLICATIONS, and these should be NON-INTERCONNECTABLE with Luer CONNECTORS and each other.

ISO/TR 16142:2006 addresses this type of problem in Essential Principle A.1.2:

The solutions adopted by the manufacturer for the design and construction of the devices should conform to safety principles, taking into account the generally acknowledged state of the art.

In selecting the most appropriate solutions, the manufacturer should apply the following principles in the following order:

- identify hazards and the associated risks arising from the intended use and foreseeable misuse;
- eliminate or reduce risks as far as possible (inherently safe design and construction);

It is understood that SMALL-BORE CONNECTOR systems cannot be designed to overcome all chances of misconnection or to eliminate deliberate misuse. However, a number of steps that would improve the current situation and lead to greater PATIENT safety can be taken. This will only be achieved through a long-term commitment involving industry, healthcare professionals, MEDICAL DEVICE purchasers and MEDICAL DEVICE regulatory authorities.

This is the first edition of ISO 80369-1 and it cancels and replaces EN 15546-1:2008 which has been editorially revised.

Part 1 of this International Standard and its parts are intended to be the reference documents in which the necessary measures and PROCEDURES to prevent misconnection between SMALL-BORE CONNECTORS used in different APPLICATIONS and designs of SMALL-BORE CONNECTORS for APPLICATIONS are listed. The JWG of ISO/TC 210 – IEC 62D and CEN/CENELEC TC 3/WG 2 is developing this series of standards in such a way that ISO 80369-1 includes general requirements to prevent misconnections between SMALL-BORE CONNECTORS used in different APPLICATIONS.

This part 1 of this International Standard contains general requirements to ensure the prevention of misconnection between SMALL-BORE CONNECTORS used in different APPLICATIONS. Subsequent parts of this series of standards are expected to include requirements with regard to the CONNECTORS used in different APPLICATION categories.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS TYPE.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

The attention of Member Bodies and National Committees is drawn to the fact that equipment MANUFACTURERS and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication for equipment newly designed and not earlier than 5 years from the date of publication for equipment already in production.

Small-bore connectors for liquids and gases in healthcare applications —

Part 1: General requirements

1 Scope

This part of ISO 80369 specifies general requirements for SMALL-BORE CONNECTORS, which convey liquids or gases in healthcare APPLICATIONS. These SMALL-BORE CONNECTORS are used in MEDICAL DEVICES or ACCESSORIES intended for use with a PATIENT.

This International Standard also specifies the healthcare fields in which these SMALL-BORE CONNECTORS are intended to be used.

These healthcare fields of use include, but are not limited to, APPLICATIONS for:

- BREATHING SYSTEMS and driving gases,
- enteral and gastric,
- urethral and urinary,
- limb cuff inflation,
- neuraxial devices, and
- intravascular or hypodermic.

SMALL-BORE CONNECTORS as specified in this International Standard are NON-INTERCONNECTABLE with:

- the cones and sockets of ISO 5356-1:2004 and ISO 5356-2:2006;
- the temperature sensor CONNECTOR and mating ports specified in Annex DD of ISO 8185:2007; and
- the nipples of EN 13544-2:2002.

This International Standard provides the methodology to assess NON-INTERCONNECTABLE characteristics of SMALL-BORE CONNECTORS based on their inherent design and dimensions in order to reduce the RISK of misconnections between MEDICAL DEVICES or between ACCESSORIES for different APPLICATIONS and to reduce the RISK of misconnections between MEDICAL DEVICES with 6 % Luer CONNECTORS, and all other non-Luer CONNECTORS that will be developed under future parts of this series of standards.

It does not specify requirements for the MEDICAL DEVICES or ACCESSORIES that use these SMALL-BORE CONNECTORS. Such requirements are given in particular International Standards for specific MEDICAL DEVICES or ACCESSORIES.

NOTE 1 It is intended that new designs of SMALL-BORE CONNECTORS will be included in this series of standards after they have been assessed according to the PROCEDURE given in Clause 6.

NOTE 2 MANUFACTURERS are encouraged to incorporate the SMALL-BORE CONNECTORS specified in this series of standards into MEDICAL DEVICES, medical systems or ACCESSORIES, even if currently not required by the relevant particular MEDICAL DEVICE standards. It is expected that when the relevant particular MEDICAL DEVICE standards are revised, requirements for SMALL-BORE CONNECTORS as specified in the series of standards will be included.

NOTE 3 MANUFACTURERS and RESPONSIBLE ORGANIZATIONS are encouraged to report their experience with the SMALL-BORE CONNECTORS specified in this series of standards to the Secretariat of ISO/TC 210 to consider this feedback during the revision of the relevant part of this series of standards.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5356-1:2004, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*

ISO 5356-2:2006, *Anaesthetic and respiratory equipment — Conical connectors — Part 2: Screw-threaded weight-bearing connectors*

ISO 14971:2007, *Medical devices — Application of risk management to medical devices*

EN 13544-2:2002, *Respiratory therapy equipment — Part 2: Tubing and connectors*

IEC 62366:2007, *Medical devices — Application of usability engineering to medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions specified in ISO 14971:2007, IEC 62366:2007 and the following apply. For convenience, the sources of all defined terms used in this document are given in an index on page 17.

3.1

ACCESSORY

additional part(s) for use with MEDICAL DEVICE in order to:

- achieve the INTENDED USE,
- adapt it to some special use,
- facilitate its use,
- enhance its performance, or
- enable its functions to be integrated with those of other MEDICAL DEVICES

[Modified from IEC 60601-1:2005, definition 3.3]

3.2

APPLICATION

specific healthcare field in which a SMALL-BORE CONNECTOR is intended to be used

NOTE Annex C lists SMALL-BORE CONNECTOR APPLICATIONS.

3.3

BREATHING SYSTEM

inspiratory and expiratory pathways through which gas flows at respiratory pressures and bounded by the port through which fresh gas enters, the PATIENT CONNECTION port and the exhaust port

3.4**CONNECTION**

union or joining of mating halves of a CONNECTOR

3.5**CONNECTOR**

mechanical device, consisting of one of two mating halves and designed to join a conduit to convey liquids or gases

3.6**NON-INTERCONNECTABLE**

having characteristics which incorporate geometries or other characteristics that prevent different CONNECTORS from being connected

3.7**PATIENT**

person undergoing a medical, surgical or dental PROCEDURE

[Modified from IEC 60601-1:2005, definition 3.76]

3.8**RESPONSIBLE ORGANIZATION**

entity accountable for the use and maintenance of a MEDICAL DEVICE

NOTE 1 The accountable entity can be, for example, a hospital, an individual clinician or a layperson. In home use applications, the PATIENT, USER and RESPONSIBLE ORGANIZATION can be one and the same person.

NOTE 2 Education and training is included in "use."

[Modified from IEC 60601-1:2005, definition 3.101]

3.9**RIGID MATERIAL**

material with a modulus of elasticity either in flexure or in tension greater than 35 000 kg/cm² (3 433 MPa)

3.10**SEMI-RIGID MATERIAL**

material with a modulus of elasticity either in flexure or in tension, between 700 kg/cm² and 35 000 kg/cm² (69 MPa and 3 433 MPa)

3.11**SMALL-BORE**

inner-fluid pathway of a CONNECTION with a diameter less than 8,5 mm

NOTE For the purposes of this standard, the 8,5 mm cone and socket of ISO 5356-1 is not considered a SMALL-BORE CONNECTOR.

4 Materials used for SMALL-BORE CONNECTORS

A SMALL-BORE CONNECTOR shall be made of RIGID MATERIAL or SEMI-RIGID MATERIAL.

Check compliance by application of the tests of ASTM D747 or ASTM D790 at (23 ± 2) °C and (50 ± 5) % relative humidity.

5 Requirements for SMALL-BORE CONNECTORS for specific APPLICATIONS

5.1 SMALL-BORE CONNECTOR incompatibility

SMALL-BORE CONNECTORS of each APPLICATION category specified in this International Standard shall be NON-INTERCONNECTABLE with any of the SMALL-BORE CONNECTORS of every other APPLICATION category for RISKS to be acceptable, unless otherwise indicated.

SMALL-BORE CONNECTORS for APPLICATION categories specified in this International Standard shall be NON-INTERCONNECTABLE with:

- the cones and sockets of ISO 5356-1:2004 and ISO 5356-2:2006; and
- the nipples of EN 13544-2:2002;
- the temperature sensor CONNECTOR and mating ports specified in Annex DD of ISO 8185:2007;

unless otherwise indicated.

Check compliance by confirming that OBJECTIVE EVIDENCE verifies that RISKS have been reduced to acceptable levels for the acceptability criteria specified in Annex B and other acceptability criteria established by the MANUFACTURER for NON-INTERCONNECTABLE characteristics. Verify that the SMALL-BORE CONNECTOR is NON-INTERCONNECTABLE.

5.2 BREATHING SYSTEMS and driving gases APPLICATIONS

SMALL-BORE CONNECTORS intended to be used either as an ancillary port CONNECTION in a BREATHING SYSTEM or a respirable driving gas APPLICATION of MEDICAL DEVICES and ACCESSORIES shall comply with ISO 80369-2:—, unless the use of these CONNECTORS creates an unacceptable RISK for a specific MEDICAL DEVICE, or shall comply with 5.8.

Check compliance by inspection of the RISK MANAGEMENT FILE for a specific MEDICAL DEVICE or application of the tests of ISO 80369-2:—. See also 5.8 for alternative methods of compliance.

5.3 Enteral and gastric APPLICATIONS

SMALL-BORE CONNECTORS intended to be used for CONNECTIONS in enteral and gastric APPLICATIONS shall comply with ISO 80369-3:—, unless the use of these CONNECTORS creates an unacceptable RISK for a specific MEDICAL DEVICE or shall comply with 5.8.

Check compliance by inspection of the RISK MANAGEMENT FILE for a specific MEDICAL DEVICE or application of the tests of ISO 80369-3:—. See also 5.8 for alternative methods of compliance.

5.4 Urethral and urinary APPLICATIONS

SMALL-BORE CONNECTORS intended to be used for CONNECTIONS in urethral and urinary APPLICATIONS shall comply with ISO 80369-4:— or 5.8, unless the use of these CONNECTORS creates an unacceptable RISK for a specific MEDICAL DEVICE or shall comply with 5.8.

Check compliance by inspection of the RISK MANAGEMENT FILE for a specific MEDICAL DEVICE or application of the tests of ISO 80369-4:—. See also 5.8 for alternative methods of compliance.

5.5 Limb cuff inflation APPLICATIONS

SMALL-BORE CONNECTORS intended to be used for CONNECTIONS in limb cuff inflation APPLICATIONS shall comply with ISO 80369-5:—, unless the use of these CONNECTORS creates an unacceptable RISK for a specific MEDICAL DEVICE or shall comply with 5.8.

Check compliance by inspection of the RISK MANAGEMENT FILE for a specific MEDICAL DEVICE or application of the tests of ISO 80369-5:—. See also 5.8 for alternative methods of compliance.

5.6 Neuraxial APPLICATIONS

SMALL-BORE CONNECTORS intended to be used for CONNECTIONS in neuraxial APPLICATIONS shall comply with ISO 80369-6:—, unless the use of these CONNECTORS creates an unacceptable RISK for a specific MEDICAL DEVICE or shall comply with 5.8.

Check compliance by inspection of the RISK MANAGEMENT FILE for a specific MEDICAL DEVICE or application of the tests of ISO 80369-6:—. See also 5.8 for alternative methods of compliance.

5.7 Intravascular or hypodermic APPLICATIONS

SMALL-BORE CONNECTORS intended to be used for CONNECTIONS in intravascular or hypodermic APPLICATIONS shall comply with ISO 80369-7:—, unless the use of these CONNECTORS creates an unacceptable RISK for a specific MEDICAL DEVICE or shall comply with 5.8.

NOTE Upon its publication, ISO 80369-7:— will replace ISO 594-1:1986 and ISO 594-2:1998.

Check compliance by inspection of the RISK MANAGEMENT FILE for a specific MEDICAL DEVICE or application of the tests of ISO 80369-7:—. See also 5.8 for alternative methods of compliance.

5.8 Alternative SMALL-BORE CONNECTORS

Alternative designs of SMALL-BORE CONNECTORS to those specified in 5.2 to 5.7 may be used in a MEDICAL DEVICE or ACCESSORY, and if used, they shall:

- a) be NON-INTERCONNECTABLE with the CONNECTORS specified in 5.2 through 5.7, inclusive;
- b) not create an unacceptable RISK for a specific MEDICAL DEVICE or ACCESSORY;
- c) be evaluated according to Annex B; and
- d) comply with Clause 4.

Check compliance by confirming that OBJECTIVE EVIDENCE verifies that RISKS have been reduced to acceptable levels for the acceptability criteria specified in Annex B and other acceptability criteria established by the MANUFACTURER for NON-INTERCONNECTABLE characteristics for the MEDICAL DEVICE or ACCESSORY.

6 Additional SMALL-BORE CONNECTOR APPLICATIONS

Alternative designs of SMALL-BORE CONNECTORS to those specified in 5.2 to 5.7, which convey liquids or gases, and which are used in MEDICAL DEVICES and ACCESSORIES intended for use with a PATIENT for other APPLICATIONS, may be used, and if used, they shall:

- a) be NON-INTERCONNECTABLE with the CONNECTORS specified in 5.2 through 5.7, inclusive;
- b) not create an unacceptable RISK for a specific MEDICAL DEVICE or ACCESSORY; and
- c) be evaluated according to Annex B.

Check compliance by confirming that OBJECTIVE EVIDENCE verifies that RISKS have been reduced to acceptable levels for the acceptability criteria specified in Annex B and other acceptability criteria established by the MANUFACTURER for NON-INTERCONNECTABLE characteristics.