



Standard Terminology Relating to Analytical Chemistry for Metals, Ores, and Related Materials¹

This standard is issued under the fixed designation E 135; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This is a compilation of terms commonly used in analytical chemistry for metals, ores, and related materials. Terms that are generally understood or defined adequately in other readily available sources are either not included or their sources are identified.

1.2 A definition is a single sentence with additional information included in a Discussion.

1.3 Definitions identical to those published by another standards organization or ASTM committee are identified with the name of the organization or the identifying document and ASTM committee.

1.4 Definitions specific to a particular field (such as *emission spectrometry*) are identified with an italicized introductory phrase.

2. Referenced Documents

2.1 ASTM Standards:

E 180 Practice for Determining the Precision of ASTM Methods for Analysis and Testing of Industrial Chemicals²

E 876 Practice for Use of Statistics in the Evaluation of Spectrometric Data³

E 1914 Practice for Use of Terms Relating to the Development and Evaluation of Methods for Chemical Analysis³

2.2 ISO Standard:

ISO Guide 30—Terms and Definitions Used in Connection with Reference Materials⁴

3. Significance and Use

3.1 Definitions given in Section 4 are intended for use in all standards on analytical chemistry for metals, ores, and related materials. The definitions should be used uniformly and consistently. The purpose of this terminology is to promote clear understanding and interpretation of the standards in which definitions are used.

¹ This terminology is under the jurisdiction of ASTM Committee E-1 on Analytical Chemistry for Metals, Ores, and Related Materials and is the direct responsibility of Subcommittee E01.23 on Terminology.

Current edition approved Dec. 10, 1999. Published February 2000. Originally published as E 135 – 58. Last previous edition E 135 – 98.

² *Annual Book of ASTM Standards*, Vol 15.05.

³ *Annual Book of ASTM Standards*, Vol 03.06.

⁴ Available from American Standards Institute, 11 West 42nd St., 13th Floor, New York, NY 10036.

4. Terminology Definitions

absorbance, *n*—*in spectrometry*, the logarithm to the base 10 of the reciprocal of the transmittance, *T*.

$$A = \log_{10}(1/T) = -\log_{10} T \quad (1)$$

absorption (of electromagnetic radiation), *n*—a decrease in radiant energy when passing through matter, resulting in a corresponding increase in the energy of the absorbing system.

absorption spectrometry, *n*—the branch of spectroscopy pertaining to the absorption of electromagnetic radiation by atoms, ions, radicals, and molecules.

analytical curve—see **calibration curve**.

analyte, *n*—*in methods of chemical analysis*, the component determined by a method.

analytical gap, *n*—*in optical emission spectrometry*, the region between two electrodes in which the specimen is excited and from which radiant energy is used for analysis.

analytical line, *n*—the particular wavelength of an element used in determining the presence or concentration of that element.

anneal, *vt*—*in fire assay*, to heat and then gradually cool a metal to remove internal stresses and make the material less brittle.

arc, condensed—see **triggered capacitor discharge**.

arc, continuous dc, *n*—a self-maintaining dc discharge.

arc line—not recommended, see **atom line**.

arc, noncapacitive ac, *n*—*in optical emission spectrometry*, a series of separate electrical discharges, individually self-initiating or initiated separately by another means, in which each current pulse has a polarity that is reversed from the previous one.

arc, noncondensed, intermittent dc—see **arc, noncapacitive, intermittent dc**.

atom line, *n*—a spectral line resulting from radiation emitted during electron transition as an excited atom returns to a lower energy level.

atomic absorption spectrometry, *n*—the branch of spectroscopy pertaining to the absorption of electromagnetic radiation by free atoms.

buffer, *n*—*in spectrometric analysis*, a substance that tends to minimize the effects of one or more elements on the emission of other elements.

burn, *n*—*in emission spectrometry*, (*I*) that portion of a solid

specimen from which atoms were volatilized;

(2) the act of burning.

burn, *vt*—*in emission spectrometry*, to vaporize and excite a specimen with sufficient energy to generate spectral radiation.

calibrant, *n*—a reference material used for a calibration.

calibrate, *vt*—(1) to establish the relationship between the response of an instrument and the concentration or mass of the analyte; (2) to establish a table of corrections to improve the accuracy of equipment used to measure physical properties such as mass, volume, temperature, etc.

calibration, *n*—the act, process, or result of establishing: (1) the relationship between the response of an instrument and the concentration or mass of the substance determined; (2) a table of corrections to improve the accuracy of equipment used to measure physical properties such as mass, volume, temperature, etc.

calibration curve, *n*—the graphical or mathematical representation of the relationship between the response of an instrument and the concentration or mass of the analyte.

certified reference material (CRM), *n*—a reference material, the composition or properties of which are certified by a recognized standardizing agency or group.

DISCUSSION—A standard reference material, SRM, is a certified reference material issued by the National Institute of Standards and Technology.

characteristic emulsion curve—see **emulsion calibration curve**.

characteristic radiation, *n*—of *X rays*, a unique set of *X rays* emitted by an element.

combine, *v*—*in sampling*, to join together two or more increments without prior division or comminution.

comminution, *n*—*in sample preparation*, processes which improve the microscale homogeneity of a sample by reducing the particle size.

concentration, (*c*), *n*—the quantity of a substance contained in a unit quantity of sample.

concentration index, *n*—the concentration of an element at which the intensities of the analytical and internal standard lines are equal.

concentration range, *n*—*in an analytical method*, the concentrations within which a method has been tested and found suitable for use.

confidence interval, *n*—the range of values that may be expected to encompass the true value, generally stated at some probability, the confidence level.

confidence level, *n*—the probability that the true value lies within a stated range (the confidence interval).

counter electrode, *n*—*in optical emission spectrometry*, the electrode in an analytical pair that does not contain the specimen being analyzed.

cupel, *vt*—to refine precious metals in a cupel by exposure to high temperature in an oxidizing atmosphere.

cupel, *n*—a small, shallow, porous cup, used in assaying to separate precious metals from lead and other base metals.

densitometer—not recommended, see **microphotometer**.

detection limit—a stated limiting value that designates the lowest concentration or mass that can be estimated or

determined with confidence and that is specific to the analytical procedure used (see Practice E 876).

division, *n*—*in sample preparation*, processes which divide a sample into two or more subsamples of equal mass and composition.

doré bead, *n*—a gold and silver alloy bead which results from cupellation.

drift, *n*—*in instrumental methods of quantitative analysis*, a gradual change in instrument response from start to completion of a set of determinations.

electrode, *n*—*in emission spectrometry*, either of two terminals between which an electrical discharge occurs.

electrode gap—not recommended, see **analytical gap**.

emission spectrometry, *n*—the branch of spectroscopy pertaining to the emission of electromagnetic radiation by atoms, ions, radicals, and molecules.

emulsion calibration curve, *n*—*in photographic optical emission spectrometry*, the plot of the degree of blackening of the developed photographic emulsion as a function of the intensity of the spectral line to which it has been exposed.

excitation potential (X-ray), *n*—the potential required to produce characteristic radiation from an element.

exposure, *n*—the irradiance of a receiver integrated over the exposure time.

exposure time, *n*—the time during which a receiver is irradiated.

fatigue, *n*—the decrease in response of a photoelectric radiant energy receiver caused by the accumulated exposure of the receiver to radiant energy.

filter, *n*—a substance that attenuates the radiant power reaching the detector in a definite manner with respect to spectral distribution.

filter, neutral, *n*—a filter that attenuates the radiant power reaching the detector by the same factor at all wavelengths within a prescribed wavelength region.

filter, nonselective—not recommended, see **filter, neutral**.

goniometer, *n*—*in X-ray spectrometry*, a device used to adjust the angular relationships among a sample, crystal, and detector in an X-ray spectrometer.

grating, concave, *n*—a diffraction grating on a concave mirror surface.

grating, diffraction, *n*—a series of a large number of narrow, close, equally spaced, diffracting slits or grooves capable of dispersing light into its spectrum.

grating, plane, *n*—a transmission or reflecting grating whose surface is flat.

grating, reflection, *n*—a diffraction grating from which the incident light is reflected to form a spectrum.

grating, transmission, *n*—a transparent diffraction grating through which light is transmitted.

homologous lines, *n*—*in optical emission spectrometry*, spectral lines that exhibit minimal change in their intensity ratios with variations in excitation conditions.

increment, *n*—*in sampling*, a portion of material removed from a lot by a single operation.

inquartation, *vt*—the addition of silver to an assay sample to facilitate parting.