



Designation: E 578 – 83 (Reapproved 1998)

Standard Test Method for Linearity of Fluorescence Measuring Systems¹

This standard is issued under the fixed designation E 578; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers a procedure for evaluating the limits of linearity of fluorescence-measuring systems under operating conditions. Particular attention is given to slit widths, filters, and sample containers. This test method can be used to test the overall linearity under a wide variety of instrumental and sampling conditions. The results obtained apply only to the tested combination of slit width and filters, and the size, type and illumination of the sample cuvette, all of which must be stated in the report. The sources of nonlinearity may be the measuring electronics, excessive absorption of either the exciting or emitted radiation, or both, and the sample handling technique, particularly at low concentrations.

1.2 This test method has been applied to fluorescence-measuring systems utilizing continuous and low-energy excitation sources (for example, an excitation source of 150-W electrical input or less). There is no assurance that extremely intense illumination will not cause photodecomposition of the compounds suggested in this test method.² For this reason it is recommended that this test method not be indiscriminately employed with high-intensity light sources. It is not a test method to determine the linearity of response of other materials. If this test method is extended to employ other chemical substances, the user should be aware of the possibility that these other substances may undergo decomposition, or adsorption onto containers.

1.3 *This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Summary of Test Method

2.1 This procedure is used for testing the linearity of fluorescence-measuring systems by using solutions of quinine sulfate dihydrate in sulfuric acid as standard test solutions. Other stable solutions which may be more suitable to the user

can be employed (Note 1). The standard used to determine linearity should be stated in the report. The fluorescence of the test solution is measured in the measuring system with the cuvettes, slits, or filters that are to be employed in projected use.

NOTE 1—A substitute standard should have the following properties: (1) It should have a large quantum yield at very high dilution; (2) it should be stable to the exciting radiation during spectral measurements; (3) its fluorescence and its absorption spectra overlap should be small; (4) its quantum yield should not be strongly concentration dependent; and (5) it should have a broad emission spectrum, so that little error is introduced when wide slits are used.³

2.2 *Upper Limit of Linearity*—The fluorescence intensity of a series of standard solutions is measured, the resultant instrument readings are plotted against concentration on log-log graph paper, and a smooth curve is drawn through the data points. The point (concentration) at which the upper end of the curve deviates by more than 5 % of the signal from the straight line (defined by the center region of the curve) is taken as the upper limit of linearity. The limit is expressed in micrograms per milliliter of quinine sulfate dihydrate.

NOTE 2—Absorption of the exciting radiation at high solute concentrations is dependent on instrument geometry, and can result in fluorescence signal nonlinearity.

2.3 *Lower Limit of Linearity*—The lower limit of linearity is taken as the point (concentration) at which the lower end of the curve deviates from the straight line defined by the central portion of the curve by more than twice the average percent deviation of the points that determine the straight line.

3. Significance and Use

3.1 The range of concentration of a fluorescing substance in solution over which the fluorescence varies linearly with the concentration is the range most useful for quantitative analysis. This range is affected by properties of the solution under analysis and by features of the measuring system. This test method provides a means of testing the performance of a fluorescence measuring system and of determining the concentration range over which the system is suitable for making a given quantitative analysis.

3.2 This test method is not meant for comparing the

¹ This test method is under the jurisdiction of ASTM Committee E-13 on Molecular Spectroscopy and is the direct responsibility of Subcommittee E13.06 on Molecular Luminescence.

Current edition approved Aug. 26, 1983. Published October 1983. Originally published as E 578 – 76. Last previous edition E 578 – 76.

² Lukasiewicz, R. J., and Fitzgerald, J. M., *Analytical Chemistry*, ANCHA, Vol 45, 1973, p. 511.

³ Gill, J. E., *Photochemistry and Photobiology*, PHCBA, Vol 9, 1969, p. 313.