

# INTERNATIONAL STANDARD

**ISO**  
**8754**

First edition  
1992-05-01

---

---

## **Petroleum products — Determination of sulfur content — Energy-dispersive X-ray fluorescence method**

**iTeh STANDARD PREVIEW**

**(standards.iteh.ai)**

*Produits pétroliers — Détermination de la teneur en soufre — Méthode  
par fluorescence X dispersive d'énergie*

ISO 8754:1992

<https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992>

INTERNATIONAL

ISO



Reference number  
ISO 8754:1992(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 8754 was prepared by Technical Committee ISO/TC 28, *Petroleum products and lubricants*.

[ISO 8754:1992](https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992)

<https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992>

© ISO 1992

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

# Petroleum products — Determination of sulfur content — Energy-dispersive X-ray fluorescence method

## 1 Scope

This International Standard specifies a method for the determination of the sulfur content of hydrocarbons such as naphthas, distillates, fuel oils, residues, lubricating base oils only, unleaded gasolines and components. The method is applicable to products having sulfur contents in the range 0,01 % (*m/m*) to 5 % (*m/m*).

Samples containing heavy-metal additives, lead alkyls, etc., interfere with the method. Elements such as silicon, phosphorus, calcium, potassium and halides interfere if present in concentrations of more than a few hundred milligrams per kilogram.

## 2 Principle

The sample is placed in the beam emitted from a suitable low-energy radioactive source, for example <sup>55</sup>Fe source (typical radiation of 740 MBq). The resultant excited characteristic X-radiation is measured and the accumulated count is compared with counts from previously calibrated blends to obtain the sulfur concentration as a percentage by mass. Three groups of calibration samples are required to span the concentration range 0,01 % (*m/m*) to 5 % (*m/m*).

## 3 Reagents

**3.1 Di-*n*-butyl sulfide**, sulfur content 21,91 % (*m/m*)

**3.2 White oil**, high-purity grade, containing less than 20 mg of sulfur per kilogram.

## 4 Apparatus

Ordinary laboratory apparatus and

**4.1 Energy-dispersive X-ray fluorescence analyser:** any suitable model can be used, set up according to the manufacturer's instructions.<sup>1)</sup>

**4.2 Analytical balance**, accurate to 0,1 mg.

## 5 Procedure

### 5.1 Safety precautions

The X-ray equipment and manner of use shall comply with the regulations governing the use of ionizing radiation and/or recommendations of the International Commission on Radiological Protection. The radiation source shall be checked for radiation leakage at intervals as required by the regulations.

Servicing of the source shall only be carried out by a fully trained and competent person using the correct shielding techniques.

**NOTE 1** At least one commonly used instrument has a thin beryllium window above the radioactive source. This window is very fragile and is easily broken during cleaning of the sample area to remove oily residues.

### 5.2 Sample cell preparation

The cells shall be thoroughly clean and dry before use. Handling of the film shall be kept to an absolute minimum. The window shall be renewed for the measurement of each sample.

#### NOTES

2 Window material is usually 6 μm polyester or polycarbonate film. Normal commercially available polyester window material contains small but variable amounts of calcium which may introduce errors.

3 Samples having a high aromatic content are likely to dissolve polycarbonate films.

1) Details of suppliers of suitable equipment may be obtained from the Secretariat of ISO/TC 28.

### 5.3 Preparation of apparatus

Set up the apparatus according to the manufacturer's instructions. Where regulations allow, the instrument shall be continuously switched on, to maintain optimum stability. Check the calibration of the apparatus at intervals in accordance with the manufacturer's instructions.

Some instruments require a helium purge to obtain accurate measurements. Good-quality helium, at a steady pressure, as specified by the instrument manufacturer, shall be used to avoid background interference which could give variable results.

### 5.4 Determination

Fill the sample cell to the required depth [3 mm to 20 mm (approximately half full)]. Ensure that the thin-film window does not bow. Ensure also that there are no air bubbles between the window and the liquid. For viscous samples, it may be necessary to apply heat so that they can easily be poured into the cell. Obtain two consecutive counts using the recommended counting time for the instrument on a portion of the sample. Calculate the average count for the sample (typically acceptable counting times are 50 s to 200 s).

The analysis of a sample shall be carried out at the same period of time after its preparation in the cell as that used in the preparation of the calibration curve.

## 6 Calibration

### 6.1 Preparation of standards

#### 6.1.1 Primary standards

Prepare primary standards having sulfur contents of 5 % (m/m), 2,5 % (m/m) and 1 % (m/m). Prepare each standard separately as follows:

Weigh, to the nearest 0,1 mg, the appropriate quantity of white oil (3.2), shown in table 1, into a suitable, narrow-necked container and then weigh in the appropriate quantity of di-*n*-butyl sulfide. Mix thoroughly (a glass-coated magnetic stirrer is advisable) at room temperature.

Calculate the sulfur content of each standard to three decimal places.

Table 1 — Composition of primary standards

Sulfur content % (m/m)	Mass of white oil g	Mass of di- <i>n</i> -butyl sulfide g
5	48,6	14,4
2,5	44,7	5,7
1	47,7	2,3

#### 6.1.2 Calibration blends

Make up the calibration standards in three ranges by diluting primary standards with white oil (table 2).

Table 2 — Calibration standards

Range	Expected sulfur content % (m/m)	Sulfur content of standards % (m/m)			
		0,0 <sup>1)</sup>	0,3	0,7	1,0
1	0,01 to 1,0	1,0 <sup>2)</sup>	1,5	2,0	2,5
2	1,0 to 2,5	2,0 <sup>3)</sup>	3,0	4,0	5,0
3	2,0 to 5,0				

1) White oil (3.2), assumed to be 0 % (m/m) sulfur.  
2) 1,0 % (m/m) standard from range 1 can be used.  
3) 2,0 % (m/m) standard from range 2 can be used.

### 6.2 Storage of standards

Store the standards in dark glass-stoppered bottles in a cool, dark place until required. As soon as any sediment or change of concentration is observed, discard the standard.

### 6.3 Preparation of calibration curve

Proceed as described in clause 4. Obtain four readings on each calibration blend in the selected range, using the recommended counting time for the instrument. Immediately repeat the procedure using freshly prepared cells and fresh portions of each calibration blend. From the data obtained, calculate the average reading for each sulfur concentration. Prepare a calibration graph from the averaged results.

Some automated equipment ejects the test portion after two counts. Additional test portions shall then be used to obtain the required number of counts for each calibration blend.

## 7 Expression of results

### 7.1 Calculation

Read the concentration of sulfur in the sample from the calibration curve, using the averaged counts for each oil. Record the results to two decimal places and report the result as the total sulfur content as a percentage by mass.

### 7.2 Precision

The precision of the method, as obtained by statistical examination of inter-laboratory test results, is as follows:

#### 7.2.1 Repeatability

The difference between successive test results, obtained by the same operator with the same apparatus under constant operating conditions on identical test material, would, in the long run, in the normal and correct operation of the test method, exceed the value  $0,017 (X + 0,8)$ , where  $X$  is the mean sulfur content, only in one case in 20.

#### 7.2.2 Reproducibility

The difference between two single and independent results, obtained by different operators working in different laboratories on identical test material, would, in the long run, in the normal and correct operation of the test method, exceed the value  $0,055 (X + 0,8)$ , where  $X$  is the mean sulfur content, only in one case in 20.

NOTE 4 The types of sample used in the correlations were naphtha, kerosine, gas oil, lubricating oil base stock and fuel oils.

## 8 Test report

The test report shall contain at least the following information:

- a) a reference to this International Standard;
- b) the type and identification of the product tested;
- c) the result of the test to two decimal places;
- d) any deviation, by agreement or otherwise, from the procedure specified;
- e) the date of the test.

ISO 8754:1992

<https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

**This page intentionally left blank**

[ISO 8754:1992](#)

<https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992>

## **iTeh STANDARD PREVIEW (standards.iteh.ai)**

**This page intentionally left blank**

[ISO 8754:1992](#)

<https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 8754:1992

<https://standards.iteh.ai/catalog/standards/sist/bad704a7-4b21-4c0c-b29c-a8c2c0a58907/iso-8754-1992>

---

---

**UDC 665.7:543.845:543.426**

**Descriptors:** petroleum products, hydrocarbons, determination of content, chemical analysis, sulphur, X-ray fluorescence spectrometry.

Price based on 3 pages

---

---