

IEC 61157
(Second edition – 2007)

Standard means for the reporting of the acoustic output
of medical diagnostic ultrasonic equipment

CORRIGENDUM 1

Page 6

3 Terms, definitions and symbols

**3.7
beam axis**

In the bracketed note on the origin of the definition, instead of:

[IEV 62127-1, definition 3.8 modified]

read:

[IEC 62127-1, definition 3.8 modified]

**3.9
beamwidth midpoint**

Delete the right bracket at the end of the note.

**3.12
external transducer aperture**

Replace “NOTE 1” with “NOTE”.

[IEC 61157:2007/COR1:2008](https://standards.iteh.ai/catalog/standards/iec/407a5931-e692-4d4e-b291-7a94fdc4a95d/iec-61157-2007-cor1-2008)

**3.43
ultrasonic transducer element group dimensions**

Delete “NOTE 2” and replace “NOTE 1” with “NOTE”.

Page 15

Table 1 – List of symbols

Add the following symbol, term and reference:

f_{awf}	zero-crossing acoustic-working frequency	IEC 62127-1
-----------	--	-------------

Page 16

4 Requirements

4.1 General

After the 2nd bullet point, replace “NOTE” with “NOTE 1”.

After the 5th bullet point, replace “NOTE” with “NOTE 2”.

August 2008