IEC 61326-3-1 (First edition – 2008)

Electrical equipment for measurement, control and laboratory use – EMC requirements –

Part 3-1: Immunity requirements for safety-related systems and for equipment intended to perform safety-related functions (functional safety) – General industrial applications

CORRIGENDUM 1

Page 22

Replace the existing Tables 2 and 3 with the following corrected tables:

Table 2 - Selected frequencies for electromagnetic field tests

Centre frequency	Frequency range	Purpose		
MHz	MHz	The state of		
84,000	83,996 – 84,004	ISM UK only		
	137 – 74	Div. Mobile and SRD		
168,000	167,992 – 168,008	ISM UK only		
	390 – 430	TETRA		
(https://	430 – 470	AMATEUR		
433,920	433,05 –434,79	ISM Region 1only		
896,000	886,000 - 906,000	ISM UK only		
897,500	880 - 915	GSM		
915,000	902 – 928 CORI	2 ISM Region 2 only		
log/staxogrds/10/6e.	925 – 960 47-87	5-e0e3b GSM 3a6/iec-	61326-3-1-2008-cor1-2008	
1 745,750	1 710 – 1 785	GSM		
	1 805 – 1 880	GSM		
	1 900 – 2 025	UMTS		
	2 110 – 2 200	UMTS		
2 450	2 400 – 2 500	ISM		
\ \ \ \	2 500 – 2 690	UMTS		

Table 3 - Selected frequencies for conducted r.f. tests

_	<u> </u>				
Centre frequency MHz		Frequency range	Burnooo		
		MHz	Purpose		
	3,39	3,370 – 3,410	ISM Netherlands only		
	6,780	6,765 – 6,795	ISM		
	13,560	13,553 – 13,567	ISM		
	27,120	26,957 – 27,283	ISM/CB/SRD		
	40,680	40,66 - 40,70	ISM/SRD		