



Standard Practice for Electronic Interchange of Color and Appearance Data¹

This standard is issued under the fixed designation E 1708; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers procedures to be used in the electronic exchange of color and appearance data between users, by either modem or the physical transfer of electronic media. It is intended for use by manufacturers of color-measuring instruments and developers of software so that any instrument may acquire data for its use that may have been measured on an instrument of another manufacturer, at another place, or at another time.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

E 284 Terminology of Appearance²

2.2 ANSI Standard:

CGATS.5 Graphic Technology—Spectral Measurement and Colorimetric Computation for Graphic Arts Images³

3. Terminology

3.1 *Definitions*—Definitions of terms in Terminology E 284 are applicable to this practice.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *data value identifier, n*—a keyword that identifies the contents of a data field and, by definition, signals the data type of the field as string, integer, or floating point.

3.2.2 *keyword, n*—in *electronic interchange*, an alphanumeric designator that defines the nature or number of following data fields.

3.2.2.1 *Discussion*—The word is used in its ordinary dictionary sense in Section 7.

3.2.3 *white space, n*—a valid character representation that does not generate a printable character in a file.

3.2.3.1 *Discussion*—Valid ASCII white space characters are *space* (decimal 32), *carriage return* (decimal 13), *form feed*

(decimal 12), *vertical tab* (decimal 11), *newline* (decimal 10), and *horizontal tab* (decimal 9).

4. Summary of Practice

4.1 Users of this practice are given a mandatory list of keywords. The user fills out the list with the data values associated with each keyword. The user may then select from an optional list of keywords. The user then associates with these keywords the data values required to be transmitted.

4.2 The keywords are used at the receiving end of the transmission to decode the format and number of data values present in the transmission.

5. Significance and Use

5.1 This practice should be used by manufacturers of color-measuring instruments and developers of software when the transmission of color and appearance data is desired between instruments or computers.

5.2 The method of transmission may be via direct connection, modem, or the transfer of electronic media, for example, by floppy disk transfer.

5.3 The practice lends itself to the transmission of either a single record of data or multiple record files.

5.4 Examples of files generated in this practice are contained in Fig. X1.1 and Fig. X2.1 of the appendixes.

6. Procedure

6.1 Prepare a file using ASCII code that consists of the mandatory keyword list followed by the associated data values. Each keyword and each data value is delimited by white space. The record may be organized into logical units associated with each keyword by using the carriage return character for the white space delimiter at the end of the data values associated with any keyword.

6.1.1 Every keyword in the record must be followed by at least one character of white space as a delimiter. Do not precede keywords on a line by other than white space. The exception to this recommendation, which is made to assist human readability, is the use of keywords as data value identifiers, which are listed in order on one line, if possible, so that they may be interpreted by the human reader as column headings.

6.1.2 The specification of data value identifiers is entirely optional regarding the choice and number of identifiers. Data

¹ This practice is under the jurisdiction of ASTM Committee E-12 on Appearance and is the direct responsibility of Subcommittee E12.04 on Color and Appearance.

Current edition approved May 15, 1995. Published July 1995.

² *Annual Book of ASTM Standards*, Vol 06.01.

³ Available from NPES, The Association for Suppliers of Printing and Publishing Technologies, 1899 Preston White Drive, Reston, VA 22091-4367.

value identifiers are placed between the BEGIN_DATA_FORMAT keyword and the END_DATA_FORMAT keyword. Each data value identifier must be separated by at least one character of white space.

6.2 *Mandatory Keywords*—The following keywords are mandatory in a record:

6.2.1 *E1708YY*—The first seven characters in the file must be this keyword, where the digits YY are numeric for the year of revision of this practice that was used to formulate the record.

6.2.2 *ORIGINATOR*—The alphanumeric data value associated with this keyword is the specific system, organization, or individual originating the record.

6.2.3 *DESCRIPTOR*—The alphanumeric data value associated with this keyword contains a description of the purpose or content of the record.

6.2.4 *CREATED*—The alphanumeric data value associated with this keyword contains the date on which the record was created or the data measured.

6.2.5 *NUMBER_OF_FIELDS*—The integer data value associated with this keyword contains the number of data format identifiers prescribed in the data format definition that follows.

6.2.6 *BEGIN_DATA_FORMAT*—No data value is associated with this keyword. The keyword identifies the beginning of the list of data value identifiers.

6.2.7 *END_DATA_FORMAT*—No data value is associated with this keyword. The keyword identifies the end of the list of data value identifiers.

6.2.8 *NUMBER_OF_SETS*—The integer data value associated with this keyword denotes the number of data sets with the format described by the list of data value identifiers.

6.2.9 *BEGIN_DATA*—No data value is associated with this keyword. The keyword marks the beginning of the data value list defined by the data value identifier list and the NUMBER_OF_SETS of data keyword.

6.2.10 *END_DATA*—No data value is associated with this keyword. The keyword marks the end of the data list.

6.3 *Data Value Identifiers*—The following default data value identifiers are recognized by automated readers. The list contains the data value identifier, followed by its data type, and finally a description of the contents of the data field associated with the descriptor. The alphanumeric data type is indicated by CS. The integer data type is indicated by I. The floating point data type is indicated by F.

6.3.1 *SPECIMEN_ID*, CS—This identifies the specimen that the data represent.

NOTE 1—This keyword represents an occurrence of the incompatibility of ASTM keywords with keywords defined by CGATS.5. That standard uses SAMPLE_ID for this keyword. The term “sample” refers to a part taken to represent the whole. A specimen is an individual, or instance, taken from the sample that represents the whole.

6.3.2 *STRING*, CS—This indicates that the field contains any alphanumeric value it may be necessary to transmit. This keyword is used to identify label information, lot number information, or other string values associated with the record.

6.3.3 *SPECTRAL_NM*, I—This identifies the wavelength, in nm, of an associated data value.

6.3.4 *SPECTRAL_PC*, F—This indicates that the associ-

ated data value is in units of percent reflectance or percent transmittance.

6.3.5 *SPECTRAL_RT*, F—This indicates that the associated data value is a spectral reflectance factor or spectral transmittance factor.

NOTE 2—This keyword represents an occurrence of the incompatibility of ASTM keywords with keywords defined by CGATS.5. That standard uses SPECTRAL_FR for reflectance and transmittance factors. The term “F(R)” is accepted routinely as representing Kubelka-Munk K/S in the European color community and in the analytical chemistry community worldwide. The keyword SPECTRAL_FR is thus adopted by ASTM to refer to that parameter (see 6.5.5.5).

6.3.6 *SPECTRAL_RM*, F—This indicates that the associated value is a spectroradiometric quantity.

6.3.7 *XYZ_X*, F—This indicates that the associated data value is a CIE X tristimulus value.

6.3.8 *XYZ_Y*, F—This is a CIE Y tristimulus value.

6.3.9 *XYZ_Z*, F—This is a CIE Z tristimulus value.

6.3.10 *XYY_CAPY*, F—This is a CIE Y tristimulus value. This keyword is used as an alternative to XYZ_Y when the data are furnished in chromaticity coordinates and luminance format.

6.3.11 *XYY_X*, F—This is a CIE x chromaticity coordinate.

6.3.12 *XYY_Y*, F—This is a CIE y chromaticity coordinate.

6.3.13 *LAB_L*, F—This is a CIE 1976 metric lightness value, L^* .

6.3.14 *LAB_A*, F—This is a CIE 1976 redness-greenness value, a^* .

6.3.15 *LAB_B*, F—This is a CIE 1976 yellowness-blueness value, b^* .

6.3.16 *LAB_U*, F—This is a CIE 1976 redness-greenness value, u^* .

6.3.17 *LAB_V*, F—This is a CIE 1976 yellowness-blueness value, v^* .

6.3.18 *LAB_C*, F—This is a CIE 1976 metric chroma value, C^* .

6.3.19 *LAB_H*, F—This is a CIE 1976 a, b hue angle, h_{ab} .

6.3.20 *LAB_DE*, F—This is a CIE 1976 metric color difference value, ΔE^*_{ab} .

6.4 *Data Values*—Data values must be formatted to correspond to the type of data called for by the keyword definition.

6.4.1 Alphanumeric data must be enclosed by the character “double-quote” (decimal 34) at beginning and ending. The double-quote character embedded in a string is designated by two double-quotes (“”) in a row, as is required by C language syntax. In this syntax, one double-quote is an element of the data, and the other is an indicator to the automated reader that the occurrence of the character *double-quote* does not terminate the string in which the double-quote is embedded.

6.4.2 It is not mandatory that the data value associated with the keyword SPECIMEN_ID be enclosed by double-quotes unless the specimen data value contains white space.

6.4.3 The integer data type is formed by combinations of the characters 0 through 9 inclusive, without decimal point or alphanumeric characters.

6.4.4 The floating point data type should contain a decimal point. A decimal point will be assumed at the end of the number represented if a floating point number fails to contain an explicit decimal point.