

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-6: Particular requirements for the basic safety and essential performance
of microwave therapy equipment**

**Appareils électromédicaux –
Partie 2-6: Exigences particulières pour la sécurité de base et les performances
essentielles des appareils de thérapie à micro-ondes**

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2016 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 15 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.



IEC 60601-2-6

Edition 2.1 2016-04
CONSOLIDATED VERSION

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-6: Particular requirements for the basic safety and essential performance
of microwave therapy equipment**

**Appareils électromédicaux –
Partie 2-6: Exigences particulières pour la sécurité de base et les performances
essentiels des appareils de thérapie à micro-ondes**

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.60

ISBN 978-2-8322-3378-8

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

REDLINE VERSION

VERSION REDLINE



**Medical electrical equipment –
Part 2-6: Particular requirements for the basic safety and essential performance
of microwave therapy equipment**

**Appareils électromédicaux –
Partie 2-6: Exigences particulières pour la sécurité de base et les performances
essentielles des appareils de thérapie à micro-ondes**

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	7
201.2 Normative references	8
201.3 Terms and definitions	8
201.4 General requirements	9
201.5 General requirements for testing of ME EQUIPMENT.....	9
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	9
201.7 ME EQUIPMENT identification, marking and documents.....	10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	12
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	12
201.10 Protection against unwanted and excessive radiation HAZARDS	12
201.11 Protection against excessive temperatures and other HAZARDS	13
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	13
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	15
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	15
201.15 Construction of ME EQUIPMENT.....	15
201.16 ME SYSTEMS	15
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	15
202 Electromagnetic compatibility disturbances – requirements and tests	15
Annexes	16
ANNEX C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	17
Annex AA-(informative) Particular guidance and rationale	18
Index of defined terms used in this particular standard	21
Table 201.101 – Additional ESSENTIAL PERFORMANCE requirements	9
Table 201.C.101 – Marking on the outside of MICROWAVE THERAPY EQUIPMENT or its parts.....	17
Table 201.C.102 – Marking on the inside of MICROWAVE THERAPY EQUIPMENT or its parts.....	17

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-6: Particular requirements for the basic safety and essential performance of microwave therapy equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 60601-2-6 edition 2.1 contains the second edition (2012-04) [documents 62D/985/FDIS and 62D/1008/RVD] and its amendment 1 (2016-04) [documents 62D/1331/FDIS and 62D/1351/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

International standard IEC 60601-2-6 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition of IEC 60601-2-6, published in 1984. This edition constitutes a technical revision and has been aligned ~~to the third edition of~~ with IEC 60601-1:2005+A1:2012.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC 60601-2-6:2012](#)

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of microwave therapy equipment.

This particular standard amends and supplements IEC 60601-1 ~~(third edition, 2005 and amendment 1, 2012)~~; *Medical electrical equipment – Part 1: General requirements for safety and essential performance*, hereinafter referred to as the general standard (see 201.1.4).

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

Clauses or subclauses for which there are explanatory notes in Annex AA are marked with an asterisk (*).

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC 60601-2-6:2012](#)

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-6: Particular requirements for the basic safety and essential performance of microwave therapy equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard specifies requirements for the safety of MICROWAVE THERAPY EQUIPMENT used in medical practice, as defined in 201.3.204.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MICROWAVE THERAPY EQUIPMENT as defined in 201.3.204.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard.

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x”

¹ The general standard is IEC 60601-1:2005+A1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

“Replacement” means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

“Addition” means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

“Amendment” means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this standard” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies.

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:~~2005+A1:2012~~², apply, except as follows:

Addition:

201.3.201

APPLICATOR

microwave radiator for local application of microwave energy to the PATIENT

Note 1 to entry: Some examples are dipoles, dipoles with reflectors, modified dipoles, dipole arrays, open waveguides, and dielectric radiators.

² ~~A1:2012~~ To be published

201.3.202

* CONTACT APPLICATOR

APPLICATOR that contacts the PATIENT and is thus an APPLIED PART

201.3.203

MATCHED LOAD

complex load which, when connected, results in the maximum power being delivered from the MICROWAVE THERAPY EQUIPMENT into the load

201.3.204

MICROWAVE THERAPY EQUIPMENT

MEDICAL ELECTRICAL EQUIPMENT for the treatment of the PATIENT by means of a propagated electromagnetic field in the frequency range of more than 300 MHz but not exceeding 30 GHz

201.3.205

* NON-CONTACT APPLICATOR

an APPLICATOR that does not contact or touch the PATIENT

201.3.206

PHANTOM

device which receives the radiated microwave energy and is intended to simulate the PATIENT for test purposes

201.3.207

* RATED OUTPUT POWER

value of the maximum high-frequency power which can be fed into a MATCHED LOAD

201.3.208

* UNWANTED RADIATION

microwave radiation which is not incident on or in the PATIENT for treatment purposes

201.4 General requirements

[IEC 60601-2-6:2012](https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012)

<https://standards.iteh.ai/catalog/standards/iec/dd21b1c9-6ce9-4e88-b729-13635b826582/iec-60601-2-6-2012>

Clause 4 of the general standard applies, except as follows:

201.4.3 ESSENTIAL PERFORMANCE

Addition:

201.4.3.101 * Additional ESSENTIAL PERFORMANCE requirements

Additional ESSENTIAL PERFORMANCE requirements are found in the subclauses listed in Table 201.101.

Table 201.101 – Additional ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
Free from the display of incorrect numerical values associated with the therapy to be performed.	201.12.1

201.5 General requirements for testing of ME EQUIPMENT

Clause 5 of the general standard applies.

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.2 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts

Additional subclause:

201.7.2.101 Output

MICROWAVE THERAPY EQUIPMENT shall be marked with the following information:

- RATED OUTPUT POWER in watts;
- MATCHED LOAD in ohms;
- operating frequency in megahertz or gigahertz;
- symbol number 5140 (non-ionizing electromagnetic radiation) of IEC 60878.

Compliance is checked by inspection.

201.7.3 * Marking on the inside of ME EQUIPMENT or ME EQUIPMENT parts

Additional subclauses:

201.7.3.101

Symbol number 5140 (non-ionizing electromagnetic radiation) of IEC 60878 shall be applied to any internal ACCESS COVER if the removal of that cover might cause the ME EQUIPMENT to fail the requirement of 201.10.3.102.

Compliance is checked by the test of 201.10.3.102 with any internal ACCESS COVER removed if it is not marked with the above symbol and also with any external ACCESS COVER not bearing this symbol removed.

201.7.3.102

Symbols number 2 (general warning sign) and number 10 (follow operating instructions) of Table D.2 in Annex D of the general standard shall be displayed on or near components or on panels giving access to components if adjustment or replacement of these components might cause the ME EQUIPMENT to fail to comply with IEC 60601-1-2 and Clause 202.

Compliance is checked by inspection.

201.7.4.2 * Control devices

Addition:

The output control shall have a scale and/or associated indicator representing the microwave energy output. The numeral "0" shall not be used unless any microwave energy delivered in this position is less than 10 mW. If the output scale or indicator represents watts of output power, it shall be so marked.