



SLOVENSKI STANDARD

SIST EN ISO 6817:1997

01-september-1997

Measurement of conductive liquid flow in closed conduits - Method using electromagnetic flowmeters

Measurement of conductive liquid flow in closed conduits - Methods using electromagnetic flow-meters (ISO 6817:1992)

Durchflußmessung von leitfähigen Flüssigkeiten in geschlossenen Leitungen - Verfahren mit magnetisch-induktiven Durchflußmeßgeräten (ISO 6817:1992)

Mesure de débit d'un fluide conducteur dans les conduites fermées - Méthode par débitmètres électromagnétiques (ISO 6817:1992)

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

Ta slovenski standard je istoveten z: EN ISO 6817:1995

ICS:

17.120.10	Pretok v zaprtih vodih	Flow in closed conduits
-----------	------------------------	-------------------------

SIST EN ISO 6817:1997

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

EUROPEAN STANDARD

EN ISO 6817

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 1995

ICS 17.120.10

Descriptors: liquid flow, pipe flow, flow measurements, flowmeters, electromagnetic equipments, installation, specifications, marking

English version

**Measurement of conductive liquid flow in closed
conduits - Method using electromagnetic
flowmeters (ISO 6817:1992)**

Mesure de débit d'un fluide conducteur dans les conduites fermées - Méthode par débitmètres électromagnétiques (ISO 6817:1992) (standards.iteh.ai)

Durchflußmessung von leitfähigen Flüssigkeiten in geschlossenen Leitungen - Verfahren mit magnetisch-induktiven Durchflußmeßgeräten (ISO 6817:1992)

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

This European Standard was approved by CEN on 1995-08-31. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

© 1995

All rights of reproduction and communication in any form and by any means reserved in all countries to CEN and its members.

Ref. No. EN ISO 6817:1995 E

Foreword

This European Standard was taken over by CEN from the work of ISO/TC 30 "Measurement of fluid flow in closed conduits" of the International Standards Organization (ISO).

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 1996, and conflicting national standards shall be withdrawn at the latest by March 1996.

According to the CEN/CENELEC Internal Regulations, the following countries are bound to implement this European Standard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, United Kingdom.

Endorsement notice

The text of the International Standard ISO 6817:1992 was approved by CEN as a European Standard without any modification.

NOTE: Normative references to International publications are listed in annex ZA (normative).

Itch STANDARD PREVIEW
(standards.itech.ai)
SIST EN ISO 6817:1997
<https://standards.itech.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

Annex ZA (normative)
Normative references to international publications
with their relevant European publications

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN</u>	<u>Year</u>
ISO 4006	1991	Measurement of fluid flow in closed conduits - Vocabulary and symbols	EN 24006	1993
ISO 9104	1991	Measurement of fluid flow in closed conduits - Methods of evaluating the performance of electromagnetic flow-meters for liquids	EN 29104	1993

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

INTERNATIONAL STANDARD

**ISO
6817**

First edition
1992-12-01

Measurement of conductive liquid flow in closed conduits — Method using electromagnetic flowmeters

iTeh STANDARD PREVIEW

(standards.iteh.ai)

*Mesure de débit d'un fluide conducteur dans les conduites fermées —
Méthode par débitmètres électromagnétiques*

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>



Reference number
ISO 6817:1992(E)

ISO 6817:1992(E)

Contents

Page

1	Scope	1
2	Normative references	1
3	Definitions	1
4	Symbols and units	2
5	Theoretical requirements	2
6	Construction and principle of operation	3
7	Installation design and practice	7
8	Equipment marking	11
9	Calibration and test conditions	12
10	Uncertainty analysis	12

Annexes

A	Materials for construction of primary devices	16
B	Bibliography	18

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

© ISO 1992

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 6817 was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Sub-Committee SC 5, *Electromagnetic flowmeters*.

The first edition cancels and replaces ISO/TR 6817:1980, of which it constitutes a technical revision.

Annexes A and B of this International Standard are for information only.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

This page intentionally left blank

SIST EN ISO 6817:1997

<https://standards.iteh.ai/catalog/standards/sist/c09e4e05-a076-4b45-824a-84a97141ba6c/sist-en-iso-6817-1997>

Measurement of conductive liquid flow in closed conduits — Method using electromagnetic flowmeters

1 Scope

This International Standard describes the principle and main design features of industrial electromagnetic flowmeters for the measurement of flow-rate of a conductive liquid in a closed conduit running full. It covers their installation, operation, performance and calibration.

This International Standard does not specify safety requirements in relation to hazardous environmental usage of the meter, nor does it apply to the measurement of magnetically permeable slurries, liquid metals nor usage in medical applications.

This International Standard covers flowmeter types in both a.c. and pulsed d.c. versions.

ISO 7066-2:1988, *Assessment of uncertainty in the calibration and use of flow measurement devices — Part 2: Non-linear calibration relationships.*

ISO 9104:1991, *Measurement of fluid flow in closed conduits — Methods of evaluating the performance of electromagnetic flow-meters for liquids.*

3 Definitions

For the purposes of this International Standard, the definitions given in ISO 4006 and the following definitions apply. Many of these are extracted from ISO 4006 for ease of reference.

3.1 electromagnetic flowmeter: Flowmeter which creates a magnetic field perpendicular to the flow, so enabling the flow-rate to be deduced from the induced electromotive force (e.m.f.) produced by the motion of a conducting liquid¹⁾ in the magnetic field. The electromagnetic flowmeter consists of a primary device and one or more secondary devices.

3.1.1 primary device: Device containing the following elements:

- an electrically insulated meter tube through which the conductive liquid to be metered flows,
- one or more pairs of electrodes, diametrically opposed, across which the signal generated in the liquid is measured,
- an electromagnet for producing a magnetic field in the meter tube.

The primary device develops a signal proportional to the flow-rate and in some cases the reference signal.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 4006:1991, *Measurement of fluid flow in closed conduits — Vocabulary and symbols.*

ISO 5168:1978, *Measurement of fluid flow — Estimation of uncertainty of a flow-rate measurement.*

ISO 7066-1:1989, *Assessment of uncertainty in the calibration and use of flow measurement devices — Part 1: Linear calibration relationships.*

1) In the present International Standard, for electromagnetic flowmeters, the more correct term "liquid" replaces the word "fluid" (covering liquids and gases) of the general definition in ISO 4006. This usage also aligns with that in ISO 9104.