



# SLOVENSKI STANDARD

## SIST ISO 10155:1996

01-oktober-1996

---

### Emisije nepremičnih virov - Avtomatski monitoring masne koncentracije delcev - Delovne karakteristike, preskusne metode in specifikacije

Stationary source emissions - Automated monitoring of mass concentrations of particles  
- Performance characteristics, test methods and specifications

## iTeh STANDARD PREVIEW

Émissions de sources fixes - Contrôle automatique des concentrations en masse de  
particules - Caractéristiques de fonctionnement, modes opératoires d'essai et  
spécifications

[SIST ISO 10155:1996](https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996)

[https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-](https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996)

[870f92e2d7dc/sist-iso-10155-1996](https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996)

Ta slovenski standard je istoveten z: **ISO 10155:1995**

---

#### **ICS:**

13.040.40      Emisije nepremičnih virov      Stationary source emissions

**SIST ISO 10155:1996**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 10155:1996

<https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996>

INTERNATIONAL  
STANDARD

**ISO**  
**10155**

First edition  
1995-04-01

---

---

**Stationary source emissions — Automated  
monitoring of mass concentrations of  
particles — Performance characteristics,  
test methods and specifications**

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

*Émissions de sources fixes — Contrôle automatique des concentrations  
en masse de particules — Caractéristiques de fonctionnement, modes  
opératoires, d'essai et spécifications*

<https://standards.iteh.ai/catalog/standards/sist/1995-04-01/iso-10155-1996>  
434a-a891-870f92e2d7dc/sist-iso-10155-1996



Reference number  
ISO 10155:1995(E)

## ISO 10155:1995(E)

## Contents

	Page
<b>1</b> Scope .....	<b>1</b>
<b>2</b> Normative references .....	<b>1</b>
<b>3</b> Definitions .....	<b>2</b>
<b>4</b> Measurement system components .....	<b>3</b>
<b>4.1</b> Sampling and analytical components .....	<b>3</b>
<b>4.2</b> Data recorder .....	<b>3</b>
<b>5</b> Installation criteria .....	<b>3</b>
<b>5.1</b> Sampling location .....	<b>3</b>
<b>5.2</b> Environmental conditions .....	<b>4</b>
<b>6</b> System performance specifications .....	<b>4</b>
<b>6.1</b> Response time .....	<b>4</b>
<b>6.2</b> Zero drift .....	<b>4</b>
<b>6.3</b> Span .....	<b>5</b>
<b>6.4</b> Span drift .....	<b>5</b>
<b>6.5</b> Calibration specifications .....	<b>5</b>
<b>7</b> Test procedures to validate specifications .....	<b>5</b>
<b>7.1</b> Response time test .....	<b>5</b>
<b>7.2</b> Zero and span setting in the field .....	<b>6</b>
<b>7.3</b> System calibration .....	<b>6</b>
 <b>Annexes</b>	
<b>A</b> Statistical treatment of data .....	<b>8</b>
<b>B</b> Possible sampling configurations .....	<b>13</b>
<b>C</b> Test data forms .....	<b>14</b>
<b>D</b> Example of calculations .....	<b>16</b>

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

SIST ISO 10155:1996  
<https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996>

© ISO 1995

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 10155 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 1, *Stationary source emissions*.

[SIST ISO 10155:1996](https://standards.iteh.ai/catalog/standards/sist/10155-1996)

<https://standards.iteh.ai/catalog/standards/sist/10155-1996> Annex A forms an integral part of this International Standard. Annexes B, C and D are for information only.

## iTeh STANDARD PREVIEW

This page intentionally left blank  
(standards.iteh.ai)

SIST ISO 10155:1996

<https://standards.iteh.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996>

# Stationary source emissions — Automated monitoring of mass concentrations of particles — Performance characteristics, test methods and specifications

## 1 Scope

This International Standard specifies conditions and criteria for the automated monitoring of the mass concentration of particulate matter in stationary source gas streams. This includes performance characteristics and test procedures.

This International Standard provides the field evaluation test program and its application to automated monitoring systems. Its approach is general and not limited to a specific measurement principle or instrument system. The scope includes system operation characteristics, calibration, test procedures and data treatment.

This International Standard is applicable only on a site-specific basis by direct correlation with the manual testing method in ISO 9096. If site conditions change (i.e. change in emission controls, change in fuel type), the calibration needs to be repeated. The mass concentration range of application is only confined to those automated methods that can meet the calibration specifications. Consequently, the actual range will vary according to the measurement technique of the automated method.

Changes in the physical properties (i.e. size, shape, colour, etc.) and chemical composition of the particulate matter may exist, to the extent that the integrity of the calibration cannot be maintained for the measurement system used. In such instances this technique ceases to be applicable. Reported and suspected limitations of various measurement methods need to be evaluated on a site-by-site basis.

## 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 4225:1994, *Air quality — General aspects — Vocabulary*.

ISO 6879:1983, *Air quality — Performance characteristics and related concepts for air quality measuring methods*.

ISO 7504:1984, *Gas analysis — Vocabulary*.

ISO 9096:1992, *Stationary source emissions — Determination of concentration and mass flow rate of particulate material in gas-carrying ducts — Manual gravimetric method*.

ISO 9169:1994, *Air quality — Determination of performance characteristics of measurement methods*.

### 3 Definitions

For the purposes of this International Standard, the following definitions apply.

#### 3.1 monitoring

- (1) In the wide sense of the term, repeated measurement to follow changes over a period of time.
- (2) In the restricted sense of the term, regular measurement of pollutant levels in relation to some standard or in order to assess the effectiveness of the system of regulation and control. [ISO 4225]

#### 3.2 analyser: An assembly comprising:

- a) lines permitting the introduction and removal of the gas to be analysed and/or of calibration gas;
- b) a measuring cell which, from the physical or chemical properties of the components of the gas to be analysed, gives signals allowing their identification or measurement;
- c) signal processing devices (amplification, recording) or, if need be, data processing devices. [ISO 7504]

#### 3.3 System operating characteristics

**3.3.1 response time:** Time taken for an instrument to respond to a rapid change in value of the air quality characteristic. It can be divided into two parts. [ISO 6879]

**3.3.1.1 lag time:** Time taken to reach 10 % of the total change in instrument response.

**3.3.1.2 rise time (fall time):** Time taken to pass from 10 % to 90 % of the total change in instrument response.

**3.3.2 zero instability (drift):** Change in instrument reading in response to a zero sample over a stated period of unattended operation. [ISO 6879]

**3.3.3 span:** Difference between the instrument readings for a stated value of air quality characteristic and a zero sample. By convention, this value of air quality characteristic is selected to be 95 % of the upper limit of measurement. [ISO 6879]

**3.3.4 span instability (drift):** Change which takes place in instrument span over a stated period of unattended operation. [ISO 6879]

#### 3.4 System calibration characteristics

**3.4.1 calibration function:** Instrument reading as a function of some measurable property of the particulate mass concentration under investigation and represented by the reference manual method (see ISO 9096) with all the interferents remaining constant.

**3.4.2 linear function:** A calibration function wherein the instrument reading is linear with respect to the particulate mass concentration measured by the reference manual method (see ISO 9096).

**3.4.3 nonlinear function:** A calibration function wherein the instrument reading is not linear with respect to the particulate mass concentration measured by the reference manual method (see ISO 9096). Nonlinearity may be expressed by higher order regression coefficients.

**3.4.4 confidence interval:** The interval with lower and upper limits within which the mean values of the regression line lie with a given level of confidence.

**3.4.5 tolerance interval:** The interval with lower and upper limits within which a specified percentage of the population lies with a given level of confidence.



## 4 Measurement system components

The system consists of the total equipment required for the determination of particulate mass concentration in a source effluent. Systems shall consist of major components as follows (see also figure D.1).

### 4.1 Sampling and analytical components

#### 4.1.1 Sampling components

This is the portion of the system which samples the source effluent. The sampling is carried out by extractive or nonextractive (*in situ*) means, either on a continuous or noncontinuous basis.

##### 4.1.1.1 Extractive sampling

This requires the transport of a volume of the gas stream outside the source environment into the instrument system for subsequent analysis. The interface includes all the elements necessary to extract, transport, and maintain the particulate mass concentration.

##### 4.1.1.2 Nonextractive (*in situ*) sampling

This does not involve extraction and transport. Sampling consists of defining and limiting the sample volume to be analysed, which is usually done by physical methods, such as electromagnetic radiation techniques.

##### 4.1.1.3 Continuous sampling

This is associated with both extractive and nonextractive sampling techniques and, primarily, with continuous analysis.

##### 4.1.1.4 Sequential sampling

This is most often associated with extractive sampling techniques having discrete sampling volumes and sampling time intervals.

#### 4.1.2 Analyser

This portion of the monitoring system shall sense the particulate quality characteristic of the gas sample and generate a detection signal output that can be correlated to mass concentration.

### 4.2 Data recorder

This portion of the monitoring system shall process the analyser (4.1.2) output and provide a record of the output signal in terms of particulate mass concentration.

## 5 Installation criteria

### 5.1 Sampling location

The place at which the sample is taken for the automated measurement of the mass concentration of particulate matter emissions in stationary source gas streams shall be designated as the sampling location. This is the location where there is no interference between the reference method (ISO 9096) and automated method. This arrangement ensures that both methods are operating on gases with the same characteristics. Openings shall be provided in the gas flow duct for the sampling probes of the reference measurement method (see ISO 9096) so that reference measurements can be made in the same vicinity as the continuous measurements.

The sampling location may require one of a number of sampling configurations which can vary in complexity depending upon the design of the measurement system and the spatial and time dependent distribution of the mass concentration of particulate matter in the cross-section of the gas stream. The objective ideally shall be to obtain,

with the simplest configuration, a sample representative of the cross-sectional mass concentration of the particulate matter.

The sampling location shall be defined in order to minimize spatial and time dependent distribution of the mean concentration of particulate matter in the cross-section of the gas stream. The representativeness of the sampling procedure and location is verified by compliance with the calibration specifications (see 6.5 and 7.3.4).

## 5.2 Environmental conditions

A number of conditions shall be considered in the installation of emission measurement systems both with respect to the sampling environment and the sampling technique. In particular, the following conditions are to be met.

- a) Compliance with the operational temperatures specified by the manufacturer.
- b) Provision for adequate protection against weather influences.
- c) Avoidance of exposure to corrosive and damaging gases on the measurement system.
- d) Avoidance of exposure to interfering electric and magnetic fields in the immediate vicinity.
- e) Avoidance of locations where vibration affects the instrument's stability.

Working platforms for conducting reference measurements and for checking and maintaining the continuous measurement systems shall be required to provide an ongoing and reliable measurement operation. Access to these platforms, to the measurement systems, and to the sampling locations are to be easily and safely achievable.

iTech STANDARD PREVIEW  
(standards.itech.ai)

### 5.2.1 In-stack

Nonextractive techniques usually involve electro-optical measurement methods which require that special precautions be taken against stray light.

SIST ISO 10155:1996  
<http://standards.itech.ai/catalog/standards/sist/f2062e23-d956-434a-a891-870f92e2d7dc/sist-iso-10155-1996>

### 5.2.2 Out-of-stack

Extractive techniques impose a number of precautions. These include the following:

- a) keeping sampling trains as short as possible to avoid long response times, chemical conversions, and particulate matter entrainment in sampling lines;
- b) maintaining isokinetic sampling;
- c) using heated sampling lines to avoid condensation and loss of particulate matter.

## 6 System performance specifications

The following specifications shall be met when measurements are made according to clause 7.

### 6.1 Response time

The response shall be within the manual reference sampling time.

### 6.2 Zero drift

The zero drift shall be within  $\pm 2\%$  of the full working range when determining system calibration and within a period of one month of unattended operation.

For automated zero adjustments, the accumulated corrections shall be within 6 % of the working range.

### 6.3 Span

The span shall be between two and three times the allowed emission concentration.

### 6.4 Span drift

The span drift shall be within  $\pm 2\%$  of the full working range when determining system calibration and within a period of one month of unattended operation.

For automated span adjustments, the accumulated corrections shall be within 6 % of the working range.

### 6.5 Calibration specifications

The following specifications shall be met.

**6.5.1** The correlation coefficient shall be  $\geq 0,95$ .

**6.5.2** The confidence interval (95 %) shall be within  $\pm 10\%$  of a site emission standard for a particulate mass concentration.

**6.5.3** The tolerance interval shall have 95 % confidence that 75 % of all possible values are within  $\pm 25\%$  of a given emission standard.

**6.5.4** In cases where any of these specifications are not met, the system performance shall be examined with regard to the following:

- a) testing procedures for the manual reference method;
- b) sampling location;
- c) sampling strategy;
- d) applicability of analytical principle;
- e) change in plant operating conditions;
- f) change in particulate matter composition;
- g) number or level of calibration data.

## 7 Test procedures to validate specifications

### 7.1 Response time test

#### 7.1.1 For automated monitoring/nonextractive sampling

Perform this test prior to the installation of the system on the stack. It may be performed at the monitoring site. Set up the measurement system and operate as specified by the manufacturer's written instructions for the monitor path length to be used at the installation. The analyser is spanned as specified in 7.2.1.1 and 7.3.3.

**Response test.** Insert a simulated reference particulate mass standard in the mass monitor five times, and record the time required for the system to respond to 95 % of the final zero and high-range particulate mass values. (See Form C-1 in annex C.)