INTERNATIONAL STANDARD



First edition 1992-11-15

Water quality — Determination of dissolved fluoride, chloride, nitrite, orthophosphate, bromide, nitrate and sulfate ions, using liquid iTeh Standard versew

(**Part 1: ards.iteh.ai**) Method for water with low contamination

<u>ISO 10304-1:1992</u>

https://standards.iteh.ai/catalog/standards/sist/c9277918-24f7-4c62-a1e2-

Qualité de l'éau ¹⁰³ Dosage des ions fluorure, chlorure, nitrite, orthophosphate, bromure, nitrate et sulfate dissous, par chromatographie des ions en phase liquide —

Partie 1: Méthode applicable pour les eaux faiblement contaminées



Reference number ISO 10304-1:1992(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member VIEW bodies casting a vote.

International Standard ISO 10304-1 was prepared by Technical Committee ISO/TC 147, Water quality, Sub-Committee SC 2, Physical, chemical, biochemical methods.

https://standards.iteh.ai/catalog/standards/sist/c9277918-24f7-4c62-a1e2-ISO 10304 consists of the following parts, under the general title Water quality — Determination of dissolved fluoride, chloride, nitrite, orthophosphate, bromide, nitrate and sulfate ions, using liquid chromatography of ions:

- Part 1: Method for water with low contamination

Part 2: Method for waste water

— Part 3: Determination of chromate, thiocyanate and thiosulfate

Annex A of this part of ISO 10304 is for information only.

© ISO 1992

International Organization for Standardization

Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Introduction

Many different ion-exchange chromatography systems can be considered. It is therefore not appropriate to indicate the type of column, mobile phase, detector type etc. that is to be used. Thus detailed information is not given at any stage during the method, although guidance is provided. However, the quality parameters which should be met by the chromatography conditions chosen are defined.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO 10304-1:1992</u> https://standards.iteh.ai/catalog/standards/sist/c9277918-24f7-4c62-a1e2-9eef742af31b/iso-10304-1-1992

iTeh STANDARD PREVIEW (standards iteh ai) This page intentionally left blank

<u>ISO 10304-1:1992</u> https://standards.iteh.ai/catalog/standards/sist/c9277918-24f7-4c62-a1e2-9eef742af31b/iso-10304-1-1992

Water quality — Determination of dissolved fluoride, chloride. nitrite, orthophosphate, bromide, nitrate and sulfate ions, using liquid chromatography of ions -

Part 1:

Method for water with low contamination

1 Scope

General

1.1

1.2.2 The determination of fluoride in particular is subject to interference by formic acid, acetic acid and carbonate, even at low concentrations. iTeh STANDARD KĽV

This part of ISO 10304 specifies a method for the site of the case of laws life fluoride, determination of chloride, nitrite. between the anions determined (F, CI, NO₂, PO₄, Br, orthophosphate, bromide, nitrate and sulfatesin wath-1:19NO3, SO4).

ter with low contamination (e.g. drinking water, rain rds/sist/c9277918-24f7-4c62-a1c2water, ground water and surface water) in the fol-water, ground water and surface water) in the following ranges:

Fluoride (F):	0,01 mg/l to 10 mg/l
Chloride (Cl):	0,1 mg/l to 50 mg/l
Nitrite (NO ₂):	0,05 mg/l to 20 mg/l
Orthophosphate (PO ₄):	0,1 mg/l to 20 mg/l
Bromide (Br):	0,05 mg/l to 20 mg/l
Nitrate (NO ₃):	0,1 mg/l to 50 mg/l
Sulfate (SO ₄):	0,1 mg/l to 100 mg/l

NOTE 1 Anion symbols are used without charges throughout the text.

In certain cases, the range of application may be changed by variations in the working conditions (e.g. sample volume, dilution, separating columns, preconcentration techniques, sensitivity ranges of detectors, etc.).

1.2 Interferences

1.2.1 Some organic acids, such as malonic acid, maleic acid and malic acid, may interfere in the determination of inorganic anions if they are present in high concentrations.

interfere in the working range specified, unless otherwise stated.

1.2.5 In a buffered eluent (e.g. carbonate/hydrogencarbonate), the determination will not be influenced by the sample pH in the range of pH 2 to pH 9.

1.2.6 The concentration ratios given in table 1 were checked experimentally for various representative conditions. No interferences could be observed when 50 μ l of sample volume was used for the chromatographic analysis.

1.2.7 The information in 1.2.4 to 1.2.6 is valid only as long as the quality requirements of the separating column are fulfilled (see clause 6) and the electrical conductivity of the sample is less than 1 000 μ S/cm (except fluoride: < 500 μ S/cm). For natural samples, the peak resolution (R) needs to be better than 1,3 (see figure 3).

1.2.8 Solid material and organic compounds (e.g. mineral oils, detergents and humic acids) shorten the life-time of the separating column. They should therefore be removed from the sample prior to analysis (see clause 7).

1.2.9 Inorganic acids, such as fluoroborate acid or chlorite, may interfere with the determination.

		5 Sensitivity Of		
Ratio of the mass concentrations solute/interfering ion		Maximum tole concentration ior	of interfering	
,	.	mç	mg/l	
F/CI	1:500	Sum of all ions	400	
CI/NO2	1:50	NO2	5	
CI/NO3	1:500	NO3	500	
CI/SO4	1:500	SO4	500	
	1:250	CI	100	
NO ₂ /PO ₄	1:50	PO4	20	
NO ₂ /NO ₃	1:500	NO ₃	500	
NO ₂ /SO ₄	1:500	SO4	500	
PO ₄ /Cl	1:500	CI	500	
PO₄/NO₃	1:500	NO ₃	400	
°O₄/Br	1:100	Br	100	
O ₄ /NO ₂	1:100	NO2	100	
O ₄ /SO ₄	1:500	^{so₄} iTe	h STM	
Br/Cl	1:500	СІ	(s ⁵⁰⁰ n	
Br/PO ₄	1:100	PO4	100	
Br/NO3	1:50	NO ₃	100	
3r/SO ₄	1:500	SO ₄ https://stan	dards.iteh.ai/cata	
103/CI	1:500	СІ	900f74	
lO₃/Br	1:100	Br	100	
10 ₃ /SO ₄	1:500	SO4	500	
60 ₄ /Cl	1:500	CI	500	
SO₄/NO₃	1:500	NO₃	400	

Table 1 — Cross sensitivity of anions

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 10304. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 10304 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 5667-1:1980, Water quality — Sampling — Part 1: Guidance on the design of sampling programmes. ISO 5667-2:1991, Water quality — Sampling — Part 2: Guidance on sampling techniques.

ISO 5667-3:—¹⁾, Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples.

ISO 8466-1:1990, Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function.

3 Principle

Liquid chromatographical separation of ions by means of a separating column. Use of a low capacity anion exchanger as the stationary phase, and usually aqueous solutions of salts of weak monobasic and dibasic acids as mobile phases (eluent, see 4.11). Detection by a conductivity detector (most widely used); sometimes employed in combination with a suppressor device (e.g. a cation exchanger) which decreases the conductivity of the eluent and converts the separated anions into their corresponding acids.

DANOTE 2 If no suppressor device is used, the conductivity of the eluent must be as low as possible. References covering this analytical techniques are summarized in annex A.

> $^{3}4^{-1}$ Reagents dards/str/92/1918-24f7-4c62-ale2- $^{1}03047^{-1}03047^{-1}092$ Use only reagents of recognized analytical grade. The water shall have an electrical conductivity of $< 0,1 \ \mu$ S/cm and shall be free of particulate matter with diameter $> 0,45 \ \mu$ m. The increase of the electrical conductivity due to an uptake of carbon dioxide does not interfere with the determination.

- 4.1 Sodium hydrogencarbonate (NaHCO₃).
- 4.2 Sodium carbonate (Na₂CO₃).
- 4.3 Potassium hydrogenphthalate (C₈H₅O₄K).
- 4.4 Sodium fluoride (NaF).
- 4.5 Sodium chloride (NaCl).
- **4.6 Sodium nitrite** (NaNO₂).
- 4.7 Potassium dihydrogenphosphate (KH₂PO₄).
- 4.8 Sodium bromide (NaBr).
- 4.9 Sodium nitrate (NaNO₃).

¹⁾ To be published. (Revision of ISO 5667-3:1985)

4.10 Sodium sulfate (Na_2SO_4) .

4.11 Eluents.

Different eluents are used, their choice depending on the type of separating column and detector. Therefore, follow the column manufacturer's instructions for the exact composition of the eluent. The eluent compositions described in 4.11.1.1 and 4.11.2.2 are examples only.

Prepare the eluents using degassed water. Avoid any renewed gas pick-up during operation (e.g. by helium superposition). In order to avoid the growth of bacteria or algae, store the eluent in the dark and renew every 2 d to 3 d.

4.11.1 Examples of eluents for ion chromatography using the suppressor technique.

For the application of the suppressor technique, sodium hydroxide and salt solutions of weakly dissociated acids, such as sodium carbonate/ sodium hydrogencarbonate, sodium hydrogencarbonate and sodium tetraborate are used.

4.11.1.1 Sodium carbonate/sodium hydrogencarbonate concentrate. (standards.iteor the determination of F, Cl, NO₂, PO₄, Br, NO₃,

The addition of the following eluent concentrate to $and SO_4$ in a single run, the following eluent, as an the sample has proved to be successful for <u>sample4-1:199e</u> xample, has proved to be successful. pretreatment and eluent preparation (see 4.11/1:2) dards/sist/Place 50 ml of the concentrate (4.11.2.1) in a gradu-

Place 25,4 g of sodium carbonate (4.2) and 25,2 g of sodium hydrogencarbonate (4.1) in a graduated flask of nominal capacity 1 000 ml, and dilute to volume with water.

The solution contains 0,24 mol/l of sodium carbonate and 0,3 mol/l of sodium hydrogencarbonate and is stable for several months if stored at 4 $^{\circ}$ C to 6 $^{\circ}$ C.

4.11.1.2 Sodium carbonate/sodium hydrogencarbonate eluent.

The following eluent has proved to be applicable for the determination of F, Cl, NO_2 , PO_4 , Br, NO_3 and SO_4 in a single determination.

Place 50 ml of the concentrate (4.11.1.1) in a graduated flask of nominal capacity 5 000 ml, and dilute to volume with water.

The solution contains 0,002 4 mol/l of sodium carbonate and 0,003 mol/l of sodium hydrogencarbonate.

4.11.2 Examples of eluents for ion chromatography without using the suppressor technique.

For ion chromatography techniques which do not utilize suppressor devices, salt solutions (e.g. potassium hydrogenphthalate, *p*-hydroxybenzoic acid, sodium borate/gluconate and sodium benzoate) are used. The concentration of the salts is usually in the range of 0,000 5 mol/l to 0,01 mol/l. Concentrate and eluent solutions are prepared as described in 4.11.2.1. Note that some alkaline concentrate solutions of the cited salts are not stable. Adjust the pH of the eluent after the dilution of the concentrate.

4.11.2.1 Potassium hydrogenphthalate concentrate.

The addition of the following eluent concentrate to the sample has proved to be helpful for sample pretreatment and eluent preparation (4.11.2.2).

Place 20,5 g of potassium hydrogenphthalate (4.3) in a graduated flask of nominal capacity 1 000 ml, and dilute to volume with water.

The solution contains 0,1 mol/l of potassium hydrogenphthalate and is stable for a longer period of time if stored at 4 °C to 6 °C.

¹⁵¹ Place 50 ml of the concentrate (4.11.2.1) in a graduated flask of nominal capacity 5 000 ml, dilute to volume with water and adjust to pH 8,5.

The solution contains 0,001 mol/l of potassium hydrogenphthalate.

4.12 Stock solutions.

Prepare stock solutions of concentration 1 000 mg/l for each of the anions: F, Cl, NO_2 , PO_4 , Br, NO_3 and SO_4 .

Dissolve the appropriate mass of each of the substances, prepared as stated in table 2, in a small quantity of water in graduated flasks of nominal capacity 1 000 ml. Dilute to volume with water. The solutions are stable for several months if stored at 4 °C to 6 °C.

Nitrite is easily oxidized to nitrate and therefore care should be taken to ensure the stability of its concentration.

Alternatively, use commercially available stock solutions of the required concentration.

Pretreatment by drying ¹⁾				
Anion	Salt	Duration h	Temperature °C	Mass of portion g
Fluoride	NaF	1	105	2,210 0
Chloride	NaCl	2	105	1,648 4
Nitrite	NaNO2	1	105	1,499 8
Orthophosphate	КН ₂ РО ₄	1	105	1,433 0
Bromide	NaBr	6	105	1,287 7
Nitrate	NaNO3	24	105	1,370 7
Sulfate	Na ₂ SO ₄	1	105	1,479 0

Table 2 — M	lass portion and	pretreatment fo	or stock solutions
-------------	------------------	-----------------	--------------------

4.12.1 Mixed standard solutions.

Table 3 — Volumes of stock solutions for the Depending upon requirements, prepare standard preparation of the mixed standard solution I solutions of different anion composition and concentration from the stock solutions (4.12). The risk Stock solution Anion of changes in concentration caused by interaction ml with the vessel material increases with decreasing anion concentration. Thus, the use of polytetrafluoroethylene (PTFE) or polyethylene vessels has proved to be satisfactory for the storage of fluoride D and chloride standard solutions. Experience, indicates that nitrate standard solutions are more stable al in borosilicate bottles. Rr

To avoid cross contamination, always use the same ISO 10 04-1:19 No. 10 lards/sist/s92 7918-4f7-4c62 ale2 atalog/stan vessels for the same anions and concentrations. so-10304-992 eef742af31b/i

4.12.2 Mixed standard solution I.

The mass concentrations of this solution are as follows:

 ρ (F, NO₂, PO₄, Br) = 10 mg/l

 ρ (CI, NO₃, SO₄) = 100 mg/I

Pipette the volumes given in table 3 into a graduated flask of nominal capacity 100 ml, and dilute to volume with water.

Store the solution in a polyethylene vessel. If stored at 4 °C to 6 °C, the solution is stable for about one week.

4.12.3 Mixed standard solution II.

The mass concentrations of this solution are as follows:

 ρ (F, NO₂, PO₄, Br) = 1 mg/l

$$\rho$$
 (CI, NO₃, SO₄) = 10 mg/I

Pipette 10 ml of the mixed standard solution I (4.12.2) into a graduated flask of nominal capacity 100 ml, and dilute to volume with water.

1

1

1

Anion

concentration

mg/l

10

100

10

10

10

100

100

Store the solution in a polyethylene vessel. The solution is stable for only 1 d to 2 d, even if stored at 4 °C to 6 °C.

4.12.4 Mixed standard solution III.

The mass concentrations of this solution are as follows:

$$\rho$$
 (F, NO₂, PO₄, Br) = 0,1 mg/I

 ρ (CI, NO₃, SO₄) = 1,0 mg/I

Pipette, 1 ml of the mixed standard solution I (4.12.2) into a graduated flask of nominal capacity 100 ml, and dilute to volume with water.

Store the solution in a PTFE vessel. Prepare the solution on the day of use.

4.13 Anion calibration solution.

Depending on the anion concentration expected, use the stock solution (4.12) or the mixed standard solutions I and II (4.12.2 and 4.12.3 respectively) to prepare five to ten calibration solutions covering the expected working range as evenly as possible.

For example, proceed as follows for the range of 0,1 mg/l to 1,0 mg/l for F, NO₂, PO₄, Br, and 1 mg/l to 10 mg/l for Cl, NO₃, SO₄. Pipette into a series of graduated flasks of nominal capacity 100 ml, 1 ml, 2 ml, 3 ml, 4 ml, 5 ml, 6 ml, 7 ml, 8 ml, 9 ml and 10 ml of the mixed standard solution I (4.12.2), dilute to volume with water and add 1 ml of the eluent.

The concentrations of the calibration solutions are listed in table 4.

4.14 Blank solutions.

Fill a graduated flask of nominal capacity 100 ml, up to volume with water and add 1 ml of the eluent concentrate (4.11.1.1 or 4.11.2.1).

5 Apparatus

5.1 Ion chromatographic system, complying with the quality requirements of clause 6. In general, it shall consist of the following components (see figure 1).

5.1.1 Ion exchange chromatographic apparatus, comprising the following:

- eluent reservoir;
- pump having a very low pulsation effect;
- sample injection system (e.g. sample loop, 50 µl);
- precolumn (see clause 7);
- separating column with the required separating performance (see clause 6);
- conductivity detector. with without a or suppressor device;
- recording device (e.g. integrator with plotter).

Usual laboratory apparatus and STANDARD PREVIEW (standards.iteh.ai)

ISO 10304-1:1992

Anion	Concentrations of the calibration solutions mg/l
	Working range 0,1 mg/l to 1,0 mg/l
F	0,1; 0,2; 0,3; 0,4; 0,5; 0,6; 0,7; 0,8; 0,9; 1,0
NO2	0,1; 0,2; 0,3; 0,4; 0,5; 0,6; 0,7; 0,8; 0,9; 1,0
PO4	0,1; 0,2; 0,3; 0,4; 0,5; 0,6; 0,7; 0,8; 0,9; 1,0
Br	0,1; 0,2; 0,3; 0,4; 0,5; 0,6; 0,7; 0,8; 0,9; 1,0
	Working range 1,0 mg/l to 10,0 mg/l
CI	1; 2; 3; 4; 5; 6; 7; 8; 9; 10
NO3	1; 2; 3; 4; 5; 6; 7; 8; 9; 10
SO₄	1; 2; 3; 4; 5; 6; 7; 8; 9; 10

https://standards.iteh.ai/catalog/standards/sist/c9277918-24f7-4c62-a1e2-

NOTE - The addition of 1ml of the eluent concentrate solution decreases the concentration of the calibration solution. However, this deviation is compensated by the equal treatment of the sample (see clause 7).




Figure 1 — Schematic diagram of an ion chromatography system

- 5.2 Additional equipment, including the following:
- drying oven;
- desiccator;

where graduated flasks, of nominal capacities 100 ml, is the retention time, in seconds, of the iTeh STANDARD P 1st peak; 1 000 ml, and 5 000 ml;

- graduated flasks, of nominal capacity 100 ml, dards the retention time, in seconds, of the plastics, to be used for low concentrations (e.g. 2nd peak; < 0,1 mg/l;
- ISO 10304-11/192 is the peak width, in seconds, of the graduated pipettes, of nominal/capacityteh.meatoog/standards/sist/c92719/jpeak7-4c62-a1e2-9eef742af31b/iso-10304-1-1992 W_2 is the peak width, in seconds, of the 10 ml or microlitre syringes;
- membrane filtering apparatus with membrane filters of mean pore size 0,45 µm.

Quality requirements of the separating 6 column

The separating column is the essential part of the ion chromatographic system. Its separating performance depends on several operating factors, such as column material and type of eluents. Within the scope of this part of ISO 10304, use only those separating columns that yield a baseline resolved separation of all the components of the injected standard solution containing seven anions (F, Cl, NO₂, PO₄, Br, NO₃, and SO₄), at a concentration level of 1 mg/l, (see figure 2). If only some of the anions shown in figure 2 have to be determined, this requirement is applicable for those anions. Therefore, the peak resolution shall not fall below R = 1.3 (see equation 1 and figure 3).

2nd peak.

Calculate the peak resolution R using the equation


...(1)

 $R = \frac{2(t_{R2} - t_{R1})}{W_1 + W_2}$

7 Sampling and sample pretreatment


Use vessels made of polytetrafluoroethylene (PTFE) or polyethylene for sampling. Use only new vessels or those that have been rinsed thoroughly with water (see clause 4). To prevent the risk of contamination of the sample, do not use strong mineral acids or alkaline detergent solutions for cleaning the vessels.

After the arrival of the sample in the laboratory, filter it through a membrane filter (of pore size $0.45 \,\mu\text{m}$) to prevent adsorption of the anions on particulate matter or conversion of anions by bacterial growth. Prevent the risk of contamination of the sample by the membrane (e.g. filter the sample and discard the first portion of the filtrate).



NOTE – Elution sequences and retention times t_R may vary, depending on the type of column and the eluent composition.

(standards.iteh.ai)



https://standards.iteh.ai/catalog/standards/sist/c9277918-24f7-4c62-a1e2-9eef742af31b/iso-10304-1-1992




Figure 3 – Idealized representation of the chromatographic separation