



SLOVENSKI STANDARD

SIST EN 30012-1:1997

01-december-1997

Zahteve zagotavljanja kakovosti za merilno opremo - 1. del: Sistem meroslovnega potrjevanja merilne opreme (ISO 10012-1:1992)

Quality assurance requirements for measuring equipment - Part 1: Metrological confirmation system for measuring equipment (ISO 10012-1:1992)

Forderungen an die Qualitätssicherung für Meßmittel - Bestätigungssystem für Meßmittel (ISO 10012-1:1992)

Exigences d'assurance de la qualité des équipements de mesure - Partie 1: Confirmation métrologique de l'équipement de mesure (ISO 10012-1:1992)

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4aec-9dae-296d9231cdc4/sist-en-30012-1-1997>

Ta slovenski standard je istoveten z: **EN 30012-1:1993**

ICS:

03.120.10	Vodenje in zagotavljanje kakovosti	Quality management and quality assurance
17.020	Meroslovje in merjenje na splošno	Metrology and measurement in general

SIST EN 30012-1:1997

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 30012-1:1997](https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4aee-9dae-296d9231cdc4/sist-en-30012-1-1997)

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4aee-9dae-296d9231cdc4/sist-en-30012-1-1997>

EUROPEAN STANDARD

EN 30012-1:1993

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 1993

UDC 658.56:681.2

Descriptors: Measuring instruments, standard measures, reference materials, test equipment, quality assurance, quality control, definitions, specifications

English version

**Quality assurance requirements for measuring
equipment - Part 1: Metrological confirmation
system for measuring equipment
(ISO 10012-1:1992)**

Exigences d'assurance de la qualité des
équipements de mesure - Partie 1: Confirmation
métrologique de l'équipement de mesure
(ISO 10012-1:1992)

Forderungen an die Qualitätssicherung für
Meßmittel - Bestätigungssystem für Meßmittel
(ISO 10012-1:1992)

STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 30012-1:1997

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4acc-9dae-296d9231cdc4/sist-en-30012-1-1997>

This European Standard was approved by CEN on 1993-10-28. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels



iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 30012-1:1997](https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4aec-9dae-296d9231cdc4/sist-en-30012-1-1997)

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4aec-9dae-296d9231cdc4/sist-en-30012-1-1997>

INTERNATIONAL STANDARD

ISO 10012-1

First edition
1992-01-15

Corrected and reprinted
1993-05-01

Quality assurance requirements for measuring equipment —

Part 1:

Metrological confirmation system for
measuring equipment

<https://standards.iteh.ai/catalog/standards/sist/en-30012-1-1997/iso-10012-1-1992>
*Exigences d'assurance de la qualité des équipements de mesure —
Partie 1. Confirmation métrologique de l'équipement de mesure*



Reference number
ISO 10012-1:1992(E)

ISO 10012-1:1992(E)**Contents**

Page

1	Scope	1
2	Normative references	1
3	Definitions	2
4	Requirements	4
4.1	General	4
4.2	Measuring equipment	4
4.3	Confirmation system	4
4.4	Periodic audit and review of the confirmation system	5
4.5	Planning	5
4.6	Uncertainty of measurement	6
4.7	Documented confirmation procedures	6
4.8	Records	6
4.9	Nonconforming measuring equipment	7
4.10	Confirmation labelling	7
4.11	Intervals of confirmation	8
4.12	Sealing for integrity	8
4.13	Use of outside products and services	9
4.14	Storage and handling	9
4.15	Traceability	9
4.16	Cumulative effect of uncertainties	10
4.17	Environmental conditions	10
4.18	Personnel	10

ITeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 30012-1:1997](https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-196d9231cdc4/sist-en-30012-1-1997)

[https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-](https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-196d9231cdc4/sist-en-30012-1-1997)

[196d9231cdc4/sist-en-30012-1-1997](https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-196d9231cdc4/sist-en-30012-1-1997)

© ISO 1992

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Annexes

A	Guidelines for the determination of confirmation intervals for measuring equipment	11
A.1	Introduction	11
A.2	Initial choice of confirmation intervals	11
A.3	Methods of reviewing confirmation intervals	11
B	Bibliography	14

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 30012-1:1997](https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-296d9231cdc4/sist-en-30012-1-1997)

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-296d9231cdc4/sist-en-30012-1-1997>

ISO 10012-1:1992(E)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 10012-1 was prepared by Technical Committee ISO/TC 176, *Quality management and quality assurance*, Sub-Committee SC 3, *Supporting technologies*.

ISO 10012 consists of the following parts, under the general title *Quality assurance requirements for measuring equipment*:

- *Part 1: Metrological confirmation system for measuring equipment*
- *Part 2: Measurement assurance*

Annex A is based on Organisation Internationale de Métrologie Légale (OIML) International Document No. 10, *Guidelines for the determination of recalibration intervals of measuring equipment used in testing laboratories*.

Annexes A and B of this part of ISO 10012 are for information only.

Introduction

This part of ISO 10012 is written in the context of a Purchaser and a Supplier, both terms being interpreted in the broadest sense. The "Supplier" may be a manufacturer, an installer or a servicing organization responsible for providing a product or a service. The "Purchaser" may be a procurement authority or a customer using a product or service. Suppliers become Purchasers when procuring supplies and services from vendors or other outside sources. The subject of the negotiations relating to this part of ISO 10012 may be a design, an artefact, a product or a service. This part of ISO 10012 may be applied, by agreement, to other situations.

Reference to this part of ISO 10012 may be made:

- by a Purchaser when specifying products or services required;
- by a Supplier when specifying products or services offered;
- by consumer or employee interests, or by legislative or regulatory bodies;

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4aee-9dae-296d9231cdc4/sist-en-30012-1-1997>

— in assessment and audit of laboratories.

This part of ISO 10012 includes both requirements and (in clause 4) guidance on the implementation of the requirements.

In order to distinguish clearly between requirements and guidance, in clause 4 the latter appears in italic type-face, in a box, after each corresponding paragraph under the heading "GUIDANCE".

The text under "GUIDANCE" is for information only and contains no requirements. Statements given there are not to be construed as adding to, limiting or modifying any requirement.

NOTE 1 Use of the masculine gender in this part of ISO 10012 is not meant to exclude the feminine gender where applied to persons. Similarly, use of the singular does not exclude the plural (and vice versa) when the sense allows.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

This page intentionally left blank

[SIST EN 30012-1:1997](#)

<https://standards.iteh.ai/catalog/standards/sist/ac329248-d0af-4ace-9dae-296d9231cdc4/sist-en-30012-1-1997>

Quality assurance requirements for measuring equipment —

Part 1:

Metrological confirmation system for measuring equipment

1 Scope

1.1 This part of ISO 10012 contains quality assurance requirements for a Supplier to ensure that measurements are made with the intended accuracy. It also contains guidance on the implementation of the requirements.

1.2 This part of ISO 10012 specifies the main features of the confirmation system to be used for a Supplier's measuring equipment.

1.3 This part of ISO 10012 is applicable to measuring equipment used in the demonstration of compliance with a specification: it does not apply to other items of measuring equipment. This part of ISO 10012 does not deal extensively with other elements that may affect measurement results such as methods of measurement, competence of personnel etc.; these are dealt with more specifically in other International Standards, such as those referred to in 1.4.

1.4 This part of ISO 10012 is applicable:

- to testing laboratories, including those providing a calibration service; this includes laboratories operating a quality system in accordance with ISO/IEC Guide 25;
- to Suppliers of products or services who operate a quality system in which measurement results are used to demonstrate compliance with specified requirements; this includes operating systems that meet the requirements of ISO 9001, ISO 9002 and ISO 9003. The guidance given in ISO 9004 is also relevant;

— to other organizations where measurement is used to demonstrate compliance with specified requirements.

1.5 The role of the Purchaser in monitoring a Supplier's compliance with the requirements of this part of ISO 10012 may be fulfilled by a third party, such as an accreditation or certification body.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 10012. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 10012 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 8402:1986, *Quality — Vocabulary*.

ISO 9001:1987, *Quality systems — Model for quality assurance in design/development, production, installation and servicing*.

ISO 9002:1987, *Quality systems — Model for quality assurance in production and installation*.

ISO 9003:1987, *Quality systems — Model for quality assurance in final inspection and test*.

ISO 9004:1987, *Quality management and quality system elements — Guidelines*.