



SLOVENSKI STANDARD SIST EN ISO 19739:2006

01-februar-2006

NYa Y'g]d`]b`E`8 c`c Yj Ub`Y`yj Yd`cj] `gdc`b`g`d`]bg_c`_fca Uhc[fUz`c`fIGC
% + ' - . & \$\$ (£

Natural gas - Determination of sulfur compounds using gas chromatography (ISO 19739:2004)

Erdgas - Bestimmung von Schwefelverbindungen mittels Gaschromatographie (ISO 19739:2004)

Gaz naturel - Détermination des composés soufrés par chromatographie en phase gazeuse (ISO 19739:2004)

Ta slovenski standard je istoveten z: EN ISO 19739:2005

ICS:

75.060

Zemeljski plin

Natural gas

SIST EN ISO 19739:2006

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 19739:2006

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>

EUROPEAN STANDARD

EN ISO 19739

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2005

ICS 75.060

English Version

Natural gas - Determination of sulfur compounds using gas chromatography (ISO 19739:2004)

Gaz naturel - Détermination des composés soufrés par chromatographie en phase gazeuse (ISO 19739:2004)

Erdgas - Bestimmung von Schwefelverbindungen mittels Gaschromatographie (ISO 19739:2004)

This European Standard was approved by CEN on 3 November 2005.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

SIST EN ISO 19739:2006

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

EN ISO 19739:2005 (E)**Foreword**

The text of ISO 19739:2004 has been prepared by Technical Committee ISO/TC 193 "Natural gas" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 19739:2005 by CMC.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2006, and conflicting national standards shall be withdrawn at the latest by May 2006.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 19739:2004 has been approved by CEN as EN ISO 19739:2005 without any modifications.

(standards.iteh.ai)

[SIST EN ISO 19739:2006](https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006)

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>

INTERNATIONAL STANDARD

ISO 19739

First edition
2004-05-01

Corrected version
2005-11-15

Natural gas — Determination of sulfur compounds using gas chromatography

*Gaz naturel — Détermination des composés soufrés par
chromatographie en phase gazeuse*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 19739:2006](https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006)

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>



Reference number
ISO 19739:2004(E)

© ISO 2004

ISO 19739:2004(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 19739:2006](https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006)

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 Principle.....	3
5 Apparatus	3
6 Sampling.....	4
6.1 Safety precautions.....	4
6.2 Temperature control.....	4
6.3 Construction materials.....	5
6.4 Cleanness	5
6.5 Installation of the calibration gas cylinder.....	5
6.6 Pressure control	5
6.7 Purging of reduction valve and transfer lines.....	5
6.8 Flow control.....	6
6.9 Diffusion control	6
6.10 Automation and sequences of sampling	6
7 Calibration	6
8 Analysis	6
9 Performance characteristics required for sulfur analysis.....	7
10 Test report.....	8
Annex A (informative) Columns mostly used in sulfur analysis (with internal phase and dimensions).....	9
Annex B (informative) Types of detectors used in sulfur analysis	10
Annex C (informative) GC method using capillary column and FPD.....	13
Annex D (informative) GC method using ED	19
Annex E (informative) GC method using MSD	25
Annex F (informative) GC method using AED.....	28
Annex G (informative) GC methods using column switching and FPD.....	31
Annex H (informative) GC method using capillary column and SCD	43
Annex I (informative) GC method using capillary column and PFPD.....	50
Bibliography	55

ISO 19739:2004(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 19739 was prepared by Technical Committee ISO/TC 193, *Natural gas*, Subcommittee SC 1, *Analysis of natural gas*.

This first edition of ISO 19739 cancels and replaces ISO 6326-2:1981 and ISO 6326-4:1994, of which it constitutes a technical revision.

This corrected version of ISO 19739:2004 incorporates the following corrections:

- page 2, 3.2: a prime has been added to the second R in the formula and the description in parenthesis;
- page 2, 3.3: a prime has been added to the second R in the formula and the description in parenthesis;
- page 7, 9 b): the name of the molecule has been corrected to 2-methylpropane-2-thiol;
- pages 20, 31, 43 and 44: the footnotes have been corrected to read "...this International Standard...";
- page 22, D.3.4.2: the temperature has been changed to 293 K;
- page 43: the name of the molecule has been corrected to 2-methylpropane-2-thiol;
- other minor editorial rectifications.

Introduction

Sulfur compounds may occur naturally in natural gas and remain as traces after treatment, or they may have been injected deliberately to allow subsequent olfactory detection for safety reasons.

The standardization of several methods for the determination of sulfur compounds in natural gas is necessary in view of the diversity of these compounds (hydrogen sulfide, carbonyl sulfide, tetrahydrothiophene, etc.) and the requirements of the determinations (e.g. required uncertainty, measurement at the drilling head, clean-up plant or in transmission pipes).

In order to enable its user to choose the most appropriate method for his/her particular needs and perform the measurements under the best conditions, this International Standard gives the requirements needed to perform a sulfur analysis.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 19739:2006](https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006)

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 19739:2006

<https://standards.iteh.ai/catalog/standards/sist/f4217a1b-f3c8-4c5f-ae20-9fc7f18aed67/sist-en-iso-19739-2006>

Natural gas — Determination of sulfur compounds using gas chromatography

WARNING — Some sulfur compounds can constitute a serious health hazard.

1 Scope

This International Standard specifies the determination of hydrogen sulfide, carbonyl sulfide, C₁ to C₄ thiols, sulfides and tetrahydrothiophene (THT) using gas chromatography (GC). Depending on the method chosen from those given in the annexes, the application ranges for the determination of sulfur compounds can vary, but whichever of the methods is used, the requirements of this International Standard apply.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

ISO 6141, *Gas analysis — Requirements for certificates for calibration gases and gas mixtures*

ISO 6143, *Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures*

ISO 6145-10, *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods — Part 10: Permeation method*

ISO 10715:1997, *Natural gas — Sampling guidelines*

ISO 14532:2001, *Natural gas — Vocabulary*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

alkane thiol

alkyl mercaptan

organic sulfur compound with the general formula R-SH (where R is the alkyl group), either naturally present or added as an odorant to natural gas

[ISO 14532:2001, definition 2.5.3.3.1]

EXAMPLE Methanethiol (MeSH), ethanethiol (EtSH), 2-methylpropane-2-thiol (tert-butylmercaptan TBM).

ISO 19739:2004(E)**3.2****alkyl disulfide**

organic sulfur compound with the general formula R-S-S-R' (where R and R' are alkyl groups)

[ISO 14532:2001, definition 3.5.3.3.3]

3.3**alkyl sulfide
thioether**

organic sulfur compound with the general formula R-S-R' (where R and R' are alkyl groups), either naturally present or added as an odorant to natural gas

[ISO 14532:2001, definition 3.5.3.3.2]

EXAMPLE Dimethyl sulfide (DMS), diethyl sulfide (DES).

3.4**carbonyl sulfide
COS**

sulfur compound found in natural gas, which contributes to the total sulfur content

[ISO 14532:2001, definition 3.5.3.3.4]

3.5**chromatographic resolution**

column efficiency characteristic describing the degree of separation of two adjacent peaks in gas chromatography

[ISO 14532:2001, definition 2.5.3.4.6]

NOTE The resolution is measured as twice the distance between the maximum of the named peaks divided by the sum of the intercepts on the baseline made by tangents drawn to the peaks at half the height.

3.6**cyclic sulfide
thioether**

cyclic organic sulfur compound with one sulfur atom incorporated in a saturated hydrocarbon ring

EXAMPLE Tetrahydrothiophene (thiophane or thiacyclopentane, THT), i.e. C₄H₈S, which is added as an odorant to natural gas.

[ISO 14532:2001, definition 2.5.3.3.6]

3.7**hydrogen sulfide
H₂S**

colourless, toxic gas with an odour similar to rotten eggs

[ISO 14532:2001, definition 2.5.3.3.8]

3.8**normal reference conditions**

reference conditions of pressure, temperature and humidity (state of saturation) equal to 101,325 kPa and 273,15 K for a dry, real gas

[ISO 14532:2001, definition 2.6.1.3]

3.9**standard reference conditions**

reference conditions of pressure, temperature and humidity (state of saturation) equal to 101,325 kPa and 288,15 K for a dry, real gas

[ISO 14532:2001, definition 2.6.1.4]

3.10**total sulfur**

total amount of sulfur found in natural gas

[ISO 14532:2001, definition 2.5.3.3.17]

NOTE The total amount of sulfur may be determined by an analytical method not differentiating between individual sulfur compounds.

3.11**working reference gas mixture****WRM**

working standard gas mixture

gas mixture whose component quantity levels have been validated by direct comparison with a secondary standard gas mixture (CRM)

[ISO 14532:2001, definition 2.5.3.5.2.3]

3.12**secondary standard gas mixture**

gas mixture whose component quantity levels have been validated by direct comparison with a PSM

[ISO 14532:2001, definition 2.5.3.5.2.2]

3.13**primary standard gas mixture****PSM**

gas mixture whose component quantity levels have been determined with the utmost accuracy and can be used as a reference gas for determining the component quantity levels of other gas mixtures

[ISO 14532:2001, definition 2.5.3.5.2.1]

4 Principle

All significant components or groups of components to be determined in a gaseous sample are physically separated by means of gas chromatography (GC) and measured by comparison with calibration or reference gases. The gas being used for calibration and the sample gas shall be analysed with the same measuring system under the same set of conditions

5 Apparatus

5.1 Gas chromatograph, containing injection device, oven, regulation system for temperature control and pressure, detector.

5.2 Chromatographic columns, with column tubing made of a material inert to sulfur compounds (see 6.4), and a stationary phase able to separate the sulfur compounds to be analysed in order that the resolution between two adjacent peaks shall not be less than 1,5.

NOTE 1 See Annex A for a list of chromatographic columns mostly used in sulfur analysis.

NOTE 2 The absence of chromatographic separation between COS and H₂S will lead to an error in total sulfur amount calculation.

ISO 19739:2004(E)**5.3 Detectors** for detecting sulfur compounds:

- sulfur-specific,
- multi-specific (respond to halogen, sulfur), and/or
- general detectors.

See Annex B for descriptions of suitable detectors.

NOTE 1 Separation problems on a column could be solved by using a sulfur-specific detector, as hydrocarbons will not be seen by it.

NOTE 2 Matrix effects can occur in sulfur analysis with certain methods/detectors.

NOTE 3 Sulfur response can be affected by quenching effects produced by hydrocarbons.

NOTE 4 Many detectors use an excited state of a molecule or atom to detect sulfur. An atom or molecule with one electron shifted from its normal orbit to another (more energetic) is said to be excited. When it relaxes, returning to its normal state, the electron falls back to its normal orbit emitting a photon. The energy of this photon is relative to the difference in energy between the orbits. The wavelength of the photon is specific for each excited state. So, if photons of different wavelength are separated (by a filter, monochromator, diffraction prism, etc.), the amount of specific photons can be measured.

6 Sampling**iTeh STANDARD PREVIEW**

The sampling procedures are very important in the analysis of sulfur compounds. Sulfur compounds have a strong tendency to adsorb on to, or to chemically react with, different materials of construction. Low contents of sulfur compounds in samples and calibration gas mixtures place demands on the sampling procedure to ensure that the sulfur compounds in correct quantity reach the analytical detector.

Carry out representative sampling in such a way that the sample represents the bulk of the gas at the time of sampling. Sampling and sample transfer shall be in accordance with ISO 10715.

Purge time should be long enough to have replicate stable analytical results within the acceptable standard deviation of the analyser. Purge time needed will depend on the type and concentration of the sulfur compound, materials of construction in gas contact and the gas flow through the sample loop.

6.1 Safety precautions

Safety precautions required in handling gas cylinders with pressurised flammable gas mixtures are described in the ISO 10715. If a pressure regulator is to be connected to the cylinder, always use a regulator with materials of construction recommended by the producer of the calibration gas.

6.2 Temperature control

When a cylinder of a calibration or sample gas mixture arrives at the place of use, ensure that the cylinder temperature is kept above the condensation temperature (as stated on the certificate). If condensation may have occurred during transportation or storage, store the cylinder at ambient temperature in a horizontal position for at least 7 days. Rolling of the cylinder will lower the homogenisation time.

Always store both calibration and sample gases at the same suitable temperature.

To reduce any adsorption of low concentration levels of sulfur compounds when using the calibration gas or a sample, the transfer lines from the cylinder and the bypass injection valve should be heated (to, for example, 90 °C).

6.3 Construction materials

The presence of sulfur compounds in the calibration or sample gas makes the choice of materials of construction in the pressure reduction device, the transfer line, the sample loop and the separation column very important. The general considerations of ISO 10715 should always be followed.

6.4 Cleanness

When a calibration or sample gas cylinder is to be connected to a gas system, always inspect visually the connection on the cylinder valve outlet. Carefully clean out any dirt, dust or particles with a dust-free cloth. Any trace of humidity is to be purged out with dry inert gas.

Make sure that all transfer lines are free of dirt, rust, grease or other particles. Change all tubing/fittings if there is any suspicion of impurities. Particle filters may be helpful, but they shall only contain material proposed in ISO 10715.

6.5 Installation of the calibration gas cylinder

The installation of a calibration gas cylinder and use of the certified gas mixture is dependent on the method by which a gas sample is taken and is to be analysed/compared. To minimise the surface in gas contact, it is important to connect the calibration gas as near as possible to the injection point. One principle for the connection of a calibration gas cylinder in direct sampling is shown in ISO 10715:1997, Annex A.

6.6 Pressure control

As described for the sample handling in ISO 10715, very often a pressure reduction device is required in order to feed the calibration and sample gas to an analyser. Normally, this is a reduction valve connected directly or close to the calibration and sample gas cylinder. Only use a pressure regulator made of the material approved by the producer of the calibration gas mixture.

To further minimise any adsorption effects, a fine regulating needle valve (made in approved material) could be connected directly to the cylinder valve. Be sure that the certified pressure range of this valve suits that of the total system and that no local or national safety regulation prohibits such an arrangement.

Never use a calibration gas mixture with a total pressure lower than that recommended on the certificate. If no recommendation is stated, stop using the mixture if the total pressure is lower than 10 % of the certified filling pressure.

Always use the same reduced pressure when injecting the calibration mixture and the natural gas sample. Control the purge flow by a needle valve, not by adjusting the reduction pressure valve.

If several calibration gases with different concentrations of the sulfur compounds are used, it is very important to always use the same needle valve for the same calibration gas mixture. Be aware of the need to change needle valves to different concentration levels.

6.7 Purging of reduction valve and transfer lines

Due to the strong tendency of sulfur compounds to adsorb to different materials of construction, it is important to purge all surfaces (which are in contact with the gas) from the cylinder valve to the injection point. Using a pressure-reducing valve mounted directly onto the cylinder valve connection, the purging should include a number of "fill and empty" cycles as described in ISO 10715. A good practice is also to connect the total transfer line from the reducing valve to vent and include the purge all the way through the sample loop.

When analysing calibration gases with different concentration levels, always flush the transfer lines and the valves with dry N₂ in order to avoid memory effects.