

## SLOVENSKI STANDARD SIST EN ISO 6326-5:2000

01-december-2000

# Zemeljski plin - Določevanje žveplovih spojin - 5. del: Metoda sežiga po Lingenerju (ISO 6326-5:1989)

Natural gas - Determination of sulfur compounds - Part 5: Lingener combustion method (ISO 6326-5:1989)

Erdgas - Bestimmung von Schwefelverbindungen - Teil 5: Verbrennung nach dem Lingener-Verfahren (ISO 6326 5:1989) DARD PREVIEW

Gaz naturel - Détermination des composés soufrés - Partie 5: Méthode de combustion Lingener (ISO 6326-5:1989) SIST EN ISO 6326-5:2000

https://standards.iteh.ai/catalog/standards/sist/e59e4252-5ab3-4c54-b614-

Ta slovenski standard je istoveten z: EN ISO 6326-5-2000 EN ISO 6326-5:1997

ICS:

75.060 Zemeljski plin

Natural gas

SIST EN ISO 6326-5:2000

en

SIST EN ISO 6326-5:2000

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6326-5:2000 https://standards.iteh.ai/catalog/standards/sist/e59e4252-5ab3-4c54-b614-26c1d7df5b96/sist-en-iso-6326-5-2000

### SIST EN ISO 6326-5:2000

## EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

## EN ISO 6326-5

December 1997

ICS 75.060

Descriptors: See ISO document

English version

### Natural gas - Determination of sulfur compounds - Part 5: Lingener combustion method (ISO 6326-5:1989)

Gaz naturel - Détermination des composés soufrés - Partie 5: Méthode de combustion Lingener (ISO 6326-5:1989) Erdgas - Bestimmung von Schwefelverbindungen - Teil 5: Verbrennung nach dem Lingener-Verfahren (ISO 6326-5:1989)

This European Standard was approved by CEN on 6 November 1997.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

SIST EN ISO 6326-5:2000

https://standards.iteh.ai/catalog/standards/sist/e59e4252-5ab3-4c54-b614-26c1d7df5b96/sist-en-iso-6326-5-2000



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

© 1997 CEN

 All rights of exploitation in any form and by any means reserved worldwide for CEN national Members. Ref. No. EN ISO 6326-5:1997 E

Page 2 EN ISO 6326-5:1997

1.1

#### Foreword

The text of the International Standard from Technical Committee ISO/TC 193 "Natural gas" of the International Organization for Standardization (ISO) has been taken over as an European Standard by CEN/CS.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 1998, and conflicting national standards shall be withdrawn at the latest by June 1998.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

#### **Endorsement notice**

The text of the International Standard ISO 6326-5:1989 has been approved by CEN as a European Standard without any modification.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6326-5:2000 https://standards.iteh.ai/catalog/standards/sist/e59e4252-5ab3-4c54-b614-26c1d7df5b96/sist-en-iso-6326-5-2000 SIST EN ISO 6326-5:2000

## INTERNATIONAL STANDARD



First edition 1989-07-01

# Natural gas — Determination of sulfur compounds —

Part 5 : Lingener combustion method iTeh STANDARD PREVIEW

#### (standards.iteh.ai) Gaz naturel – Détermination des composés soufrés –

Partie 5 : Méthode de combustion Lingener

https://standards.iteh.ai/catalog/standards/sist/e59e4252-5ab3-4c54-b614-26c1d7df5b96/sist-en-iso-6326-5-2000



#### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 6326-5 was prepared by Technical Committee ISO/TC 158, Analysis of gases.

#### SIST EN ISO 6326-5:2000

ISO 6326 consists of the followingppartsnumder thei/generalstittlea Watural 5gas 2-5-2-5ab3-4c54-b614-Determination of sulfur compounds: 26c1d7df5b96/sist-en-iso-6326-5-2000

- Part 1: General introduction

- Part 2: Gas chromatographic method using an electrochemical detector for the determination of odoriferous sulfur compounds

- Part 3: Determination of hydrogen sulfide, mercaptan sulfur and carbonyl sulfide sulfur by potentiometry

 Part 4: Determination of individual sulfur compounds by gas chromatography with a flame photometric detector

- Part 5: Lingener combustion method

Annex A of this part of ISO 6326 is for information only.

© ISO 1989

Printed in Switzerland

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization

Case postale 56 • CH-1211 Genève 20 • Switzerland

#### Introduction

The standardization of several methods for the determination of sulfur compounds in natural gas is necessary in view of the diversity of these compounds [hydrogen sulfide, carbonyl sulfide, thiols (mercaptans), tetrahydrothiophene (THT), etc.] and the purposes of the determinations (required accuracy, measurement at the drilling head or in the transmission pipes, etc.).

In order to enable the user to choose the method most appropriate to his needs and to perform the measurements under the best conditions, ISO 6326 has been prepared in several parts.

ISO 6326-1 gives a rapid comparison of standardized methods and therefore provides information for the choice of the method.

The other parts of ISO 6326, including this part, describe in detail the various stan-(dardized methods s.iteh.ai)

The determination of total sulfur is specified in ISO 4260 : 1987, Petroleum products and <u>Mydrocarbons</u> <u>326</u>-Determination of sulfur content – Wickbold combustion https://standards.it.method.log/standards/sist/e59e4252-5ab3-4c54-b614-26c1d7df5b96/sist-en-iso-6326-5-2000

## iTeh STANDARD PREVIEW (Shis page intentionally left blank

SIST EN ISO 6326-5:2000 https://standards.iteh.ai/catalog/standards/sist/e59e4252-5ab3-4c54-b614-26c1d7df5b96/sist-en-iso-6326-5-2000

### Natural gas — Determination of sulfur compounds —

Part 5 : Lingener combustion method

## iTeh STANDARD PREVIEW (standards.iteh.ai)

#### 1 Scope

This part of ISO 6326 specifies a method for the determination 6326A5 measured volume of gas is burnt with air at atmospheric of total sulfur in natural gas. The method is applicable to gases and sipressure in a glassicombustion apparatus. The resulting sulfur with sulfur contents between 0,5 mg/m<sup>3</sup> and 1,000 mg/m<sup>3</sup>-en-iso oxides 5 are 0 converted into sulfuric acid by absorption in With a total sulfur content of more than 0,1 mg sulfur in the absorption solution, visual titration with an indicator can be chosen, whereas for lower contents turbidimetric titration is preferable.

NOTE - In all parts of ISO 6326, 1 m<sup>3</sup> of gas is expressed at normal conditions (0 °C; 101,325 kPa).

#### Normative references 2

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 6326. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 6326 are encouraged to investigate the possibility of applying the most recent editions of the standards listed below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 385-1 : 1984, Laboratory glassware - Burettes - Part 1 : General requirements.

ISO 648 : 1977, Laboratory glassware - One-mark pipettes.

ISO 1042 : 1983, Laboratory glassware One-mark volumetric flasks.

ISO 3585 : 1976, Glass plant, pipeline and fittings - Properties of borosilicate glass 3.3.

hydrogen peroxide solution. Depending on the sulfur content of the test gas, the sulfate ions in the absorption solution are determined using either visual titration with an indicator or turbidimetric titration.

#### 4 Reagents and materials

During the analysis, use only reagents of recognized analytical grade and only distilled water or water of equivalent purity.

**4.1** Hydrogen peroxide, 10 % (m/m) aqueous solution, sulfur-free, as absorbing liquid for the sulfur oxides.

4.2 Activated carbon, for adsorption of sulfur impurities from the combustion air.

Absorbing liquid : 30 % (m/m) aqueous solution of 4.3 potassium hydroxide, for the purification of the combustion air.

#### 5 Apparatus

The schematic layout of the apparatus is shown in figure 1.

Ordinary laboratory apparatus and