



SLOVENSKI STANDARD
SIST ISO 11048:1996
01-september-1996

Kakovost tal - Ugotavljanje vodotopnega in kislinsko topnega sulfata

Soil quality -- Determination of water-soluble and acid-soluble sulfates

Qualité du sol -- Dosage du sulfate soluble dans l'eau et dans l'acide

Ta slovenski standard je istoveten z: ISO 11048:1995

[SIST ISO 11048:1996](https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-b128-a4ee63166fa1/sist-iso-11048-1996)

<https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-b128-a4ee63166fa1/sist-iso-11048-1996>

ICS:

13.080.10 S^{\ } \ ^{ } } æ ã [• çãç Chemical characteristics of soils

SIST ISO 11048:1996

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 11048:1996

<https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-b128-a4ee63166fa1/sist-iso-11048-1996>

INTERNATIONAL STANDARD

ISO
11048

First edition
1995-05-01

Soil quality — Determination of water-soluble and acid-soluble sulfate

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Qualité du sol — Dosage du sulfate soluble dans l'eau et dans l'acide

SIST ISO 11048:1996

<https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-b128-a4ee63166fa1/sist-iso-11048-1996>



Reference number
ISO 11048:1995(E)

ISO 11048:1995(E)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 11048 was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

Annex A of this International Standard is for information only.

ITeH STANDARD PREVIEW
(standards.iteh.ai)
SIST ISO 11048:1996
<https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-d128-666666666666/sist-iso-11048-1996>

© ISO 1995

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Soil quality — Determination of water-soluble and acid-soluble sulfate

Section 1: General

1.1 Scope

This International Standard specifies procedures for the preparation of water and acid extracts of air-dried soils and soil-like materials. The sulfate content of these extracts is determined by a gravimetric method in which barium chloride is added to the water or acid extract and the precipitate of barium sulfate is dried and weighed. The sulfate content is then calculated from the mass of the material used in the analysis and the mass of barium sulfate precipitated.

This International Standard is applicable to all types of air-dried soils, for example pretreated according to ISO 11464.

This International Standard consists of seven sections:

Section 1: General

Section 2: Pretreatment of samples and determination of dry matter contents

Section 3: Extraction of soil with water at a mass:volume ratio of 1 soil:5 water

Section 4: Extraction of soil with water at a mass:volume ratio of 1 soil:2 water

Section 5: Extraction of soil with dilute hydrochloric acid

Section 6: Determination of sulfate in solution by a gravimetric method using barium chloride

Section 7: Precision of method and test report

1.2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*.

ISO 9280:1990, *Water quality — Determination of sulfate — Gravimetric method using barium chloride*.

ISO 11464:1994, *Soil quality — Pretreatment of samples for physico-chemical analyses*.

ISO 11465:1993, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*.

1.3 Principle

Samples of air-dried soil are extracted with:

- dilute hydrochloric acid; or
- water in a mass:volume ratio of soil:water of 1:2 or 1:5.

The sulfate content of these extracts is determined by a gravimetric method in which barium chloride is

added to the water or acid extract and the precipitate of barium sulfate is dried and weighed. The sulfate content is then calculated from the mass of soil used in the analysis and the mass of barium sulfate precipitated.

Section 6 is similar to ISO 9280.

It is recognized that other acceptable methods for the determination of sulfate are available but these are not covered by this International Standard. Suitable methods may include: direct barium titrimetry, inductively coupled plasma emission spectrometry, flow injection analysis using a turbidimetric method, continuous flow indirect spectrophotometry using 2-aminoperimidine, air-segmented continuous flow colorimetry using methylthymol blue, and ion chromatography. Reference should be made to the appropriate international and national standards for information regarding the applicability and limitations of such methods.

1.4 Applicability and limitations

The procedures in sections 3 and 4 are considered to be relevant to the extraction of sulfate from soils and may be relevant to the assessment of the "availability" of sulfate to plants (a variety of extractants have been used for this purpose and none were shown to be generally applicable to all soil types and climates). It should be noted that there is no verified relation between water-extractable sulfate and the need for sulfur fertilization. The procedure in section 4 is also relevant to the potential for the attack of sulfates on building materials, especially concrete.

The use of the procedures described in sections 4, 5 and 6 should enable a distinction to be made between the presence of sulfates that are highly soluble in water and those of limited solubility such as gypsum ($\text{CaSO}_4 \cdot 2\text{H}_2\text{O}$).

Section 6 is applicable to concentrations of sulfate ion (SO_4^{2-}) in solution of 10 mg/l and above, i.e. in relation to the methods in sections 3 and 4 to minimum water-soluble concentrations in air-dried soil of about 50 mg/kg and 20 mg/kg [0,005 % (*m/m*) and 0,002 % (*m/m*)] respectively and, in relation to the method in section 5, to a minimum acid-soluble concentration in air-dried soil of about 500 mg/kg [0,05 % (*m/m*)].

If concentrations in solution are lower than this, a concentration step will be required, or a more sensitive analytical procedure than that described in section 6 should be used [e.g. ion chromatography, inductively coupled plasma (ICP) spectrometry or atomic emission spectrometry (AES)].

In most instances, complete characterization of a soil, in terms of the concentrations and nature of sulfates present, will require that determinations are made following the procedure in section 5 and in either section 3 or section 4 or both, whichever is the more appropriate for the purpose of investigation of the soil.

The methods described should be suitable for a range of "soils" including agricultural soils and materials from polluted sites. However, no method can be expected to be universally applicable, and users must recognize this fact and be prepared to deal with the analytical difficulties that may sometimes be encountered. Soils containing sulfides may require special attention (see note 7 to 5.1).

"Soils" from some polluted sites, for example former gasworks sites and colliery spoil tips, may contain a variety of sulfur species (e.g. elemental sulfur, sulfate, sulfite, thiosulfate, thiocyanate) that may be transformed from one to another under changing chemical conditions and/or microbial action. The presence of other ions may affect solubilities. Thus, in these circumstances, particular care may be required in sample handling and preparation, and in the interpretation of the results.

The limitations of the method in section 6, in terms of acceptable concentrations of other ions, are defined.

The tests described are strictly empirical and thus cannot be considered to measure the true sulfate content in soil, although reproducibility should be good.

Because of the limited solubility of some sulfates (e.g. calcium sulfate) and possible interferences from other ions, a single extraction following the procedure described in section 3 may not give a true measure of the "total water extractable" sulfate. This potential problem can be tackled either by repeating the extraction one or more times on the same sample, or by substantially increasing the water:soil ratio. Another option is to use a "buffered" extraction solution which will enhance the solubility of the sulfates present, but such procedures are beyond the scope of this International Standard.

If calcium is the only cation present, sulfate ion in solution will not exceed 1 441 mg SO_4^{2-} /l at 20 °C. This means that the maximum amount of sulfate that can be extracted from such a soil by the procedure in section 3 is about 7 205 mg SO_4^{2-} /kg (10 g soil:50 ml water) and by the procedure in section 4 is about 2 882 mg SO_4^{2-} /kg (50 g soil:100 ml water).

Section 2: Pretreatment of samples and determination of dry matter contents

2.1 Scope

In this section, procedures are described for the preparation of air-dried test samples for use in the extraction procedures described in sections 3 to 5 of this International Standard.

2.2 Determination of dry matter content

Determine the dry matter content of the field-moist and air-dried materials in accordance with ISO 11465, but at a temperature not exceeding 75 °C to 80 °C (there is a danger if gypsum is present that water of crystallization will be lost at higher temperatures).

2.3 Preparation of soil samples for extraction with water in the mass:volume ratio 1 soil:5 water, or for extraction with dilute acid

2.3.1 Apparatus

The following items of apparatus, additional to those specified in ISO 11465, allow the preparation of a single test sample.

2.3.1.1 Desiccator, containing anhydrous silica gel.

2.3.1.2 Glass container, capable of holding approximately 10 g of soil and fitted with a ground glass stopper.

2.3.2 Preparation of test sample

Prepare each test sample for analysis from the laboratory sample as follows.

Dry and prepare a laboratory sample passing a 2 mm sieve of sufficient size, for all the tests to be performed following the procedures described in ISO 11464.

Divide the material passing the 2 mm sieve by successive riffing or another approved means (see ISO 11464) to produce a sample weighing approximately 100 g.

Mill this fraction to pass a 250 µm sieve.

Subdivide the sample by successive riffing to obtain a test sample weighing approximately 10 g (additional test samples can be taken at this stage if required).

Place the test sample in a suitable glass container (2.3.1.2) and dry at a temperature of not more than 40 °C. The test sample shall be deemed to be dry when the differences in successive weighings at intervals of 4 h do not exceed 0,1 % (*m/m*) of the test sample.

Allow the test sample(s) to cool in a desiccator (2.3.1.1).

NOTES

1 It is assumed that any material retained on the 2 mm test sieve will not contain sulfates. This is generally true, but certain soils may contain lumps of gypsum larger than 2 mm in diameter and in such cases the gypsum should be removed by hand, crushed to pass a 2 mm test sieve and incorporated into the fraction passing the sieve.

2 Other suitable mechanical sample dividers may be used instead of the riffle boxes, provided that their size is suitable for the sizes and gradings of the samples to be obtained, or, alternatively, manual coning and quartering may be used.

2.4 Preparation of soil samples for extraction with water in the mass:volume ratio of 1 soil:2 water

2.4.1 Apparatus

The following items of apparatus, additional to those specified in ISO 11465, allow the preparation of a single test sample:

2.4.1.1 Desiccator, containing anhydrous silica gel.

2.4.1.2 Glass container, capable of holding approximately 125 g of soil and fitted with a ground glass stopper.

2.4.1.3 Glass container, capable of holding approximately 50 g of soil and fitted with a ground glass stopper.

2.4.2 Preparation of air-dried soil test sample

Prepare each test sample for analysis from the laboratory sample as follows.

Dry and prepare a laboratory sample passing a 2 mm sieve of sufficient size, for all the tests to be performed following the procedures described in ISO 11464.

Divide the material passing the 2 mm sieve by successive riffing to produce a sample weighing approximately 125 g.

Mill this fraction to pass a 250 µm sieve.

Subdivide the sample by successive riffing to obtain a test sample weighing approximately 50 g (additional test samples can be taken at this stage if required).

Place the test sample in a suitable glass container (2.4.1.3) and dry at a temperature of not more than 40 °C. The test sample shall be deemed to be dry when the differences in successive weighings at intervals of 4 h do not exceed 0,1 % (*m/m*) of the test sample.

Allow the test sample(s) to cool in a desiccator (2.4.1.1).

NOTES

3 It is assumed that any material retained on the 2 mm test sieve will not contain sulfates. This is generally true, but certain soils may contain lumps of gypsum larger than 2 mm in diameter and in such cases the gypsum should be removed by hand, crushed to pass a 2 mm test sieve and incorporated into the fraction passing the sieve.

4 Other suitable mechanical sample dividers may be used instead of the riffle boxes, provided that their size is suitable for the sizes and gradings of the samples to be obtained, or, alternatively, manual coning and quartering may be used.

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 11048:1996

<https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-b128-a4ee63166fa1/sist-iso-11048-1996>

Section 3: Extraction of soil with water at a mass:volume ratio of 1 soil:5 water

3.1 Principle

A procedure is described for the determination of the water-soluble sulfate content of soil, for which a 1 part by mass soil:5 parts by volume water extract is first prepared.

Determination of sulfate ions in solution is subsequently made using the gravimetric method described in section 6 of this International Standard.

NOTE 5 If calcium is the only cation present, the sulfate content of the aqueous extract will not exceed 1 441 mg SO_4^{2-} /l at 20 °C. Sulfate contents in excess of this figure in the soil-water extract, or in the groundwater, as determined in this test, therefore indicate the presence of other more soluble sulfate salts.

3.2 Preparation of 1:5 water extract

3.2.1 Reagent

The following reagent is required for preparing the water extract.

3.2.1.1 Distilled water or water of equivalent purity, in accordance with grade 3 of ISO 3696.

3.2.2 Apparatus

The following items of apparatus are required for preparing the water-soluble sulfate extract from a single test sample.

3.2.2.1 Extraction bottle, of capacity approximately 100 ml.

3.2.2.2 Buchner funnel, of diameter about 100 mm.

3.2.2.3 Vacuum filtration flask, of capacity approximately 500 ml, to take the funnel.

3.2.2.4 Filter papers, of medium grade, porosity = 8 μm .

3.2.2.5 Source of vacuum, for example a filter pump.

3.2.2.6 Length of rubber tubing, to fit the vacuum pump and filter flask.

3.2.2.7 Mechanical shaker or stirrer, capable of keeping 10 g of soil in continuous suspension in 50 ml of water.

3.2.2.8 Balance, capable of weighing to an accuracy of 0,001 g.

3.2.3 Preparation of air-dried soil test sample

Prepare an air-dried test sample following the procedure described in 2.3 of this International Standard.

3.2.4 Preparation of extract

Carry out the extraction at a temperature in the range 20 °C to 25 °C.

The 1 soil:5 water extract for the determination of the water-soluble sulfate content is obtained from a single air-dried test sample as follows.

Transfer a test portion of mass 10 g \pm 0,1 g to the extraction bottle (3.2.2.1).

Add 50 ml \pm 0,5 ml of distilled water (3.2.1.1) to the extraction bottle which shall then be stoppered tightly, placed in the shaker (3.2.2.7) and agitated for 16 h.

Filter the soil suspension into a clean and dry flask (3.2.2.3) through a suitable filter paper (3.2.2.4) placed in the Buchner funnel (3.2.2.2).

Retain the filtrate for measurement of the sulfate content.

Record the volume of the filtrate (V_E).

3.3 Determination of dry matter content

Determine the dry matter content of the air-dried material following the procedure described in 2.2 of this International Standard.

Section 4: Extraction of soil with water at a mass:volume ratio of 1 soil:2 water

4.1 Principle

A procedure is described for the determination of the water-soluble sulfate content of soil, for which a 1 soil:2 water extract is first prepared (see note 5 to 3.1).

Determination of sulfate ions in solution is made using the gravimetric method described in section 6 of this International Standard.

4.2 Preparation of 1:2 water extract

4.2.1 Reagent

The following reagent is required for preparing the water extract.

4.2.1.1 Distilled water or water of equivalent purity, in accordance with grade 3 of ISO 3696.

4.2.2 Apparatus

The following items of apparatus are required for preparing the water-soluble sulfate extract from a single test sample.

4.2.2.1 Beaker, of capacity 500 ml, preferably conical, with glass rods and cover glasses.

4.2.2.2 Buchner funnel, of diameter about 100 mm.

4.2.2.3 Vacuum filtration flask, of capacity approximately 500 ml, to take the funnel.

4.2.2.4 Filter papers, of medium grade, porosity = 8 μm .

4.2.2.5 Source of vacuum, for example a filter pump.

4.2.2.6 Length of rubber tubing, to fit the vacuum pump and filter flask.

4.2.2.7 Mechanical shaker or stirrer, capable of keeping 50 g of soil in continuous suspension in 100 ml of water.

4.2.2.8 Extraction bottle, of capacity approximately 250 ml.

4.2.2.9 Pipette, of capacity 100 ml.

4.2.2.10 Balance, capable of weighing to an accuracy of 0,001 g.

4.2.2.11 Watch glass, capable of holding 50 g of soil.

4.2.3 Preparation of air-dried soil test sample

Prepare an air-dried test sample weighing approximately 50 g following the procedure described in 2.4 of this International Standard.

4.2.4 Preparation of extract

Carry out the extraction at a temperature of 20 °C to 25 °C.

The 1 soil:2 water extract for the determination of the water-soluble sulfate content is obtained from each of the prepared test samples as follows.

Weigh an air-dried test sample of 50 g \pm 0,5 g onto a watch glass (4.2.2.11) and transfer it to a clean and dry extraction bottle (4.2.2.8).

Add 100 ml \pm 1 ml of distilled water (4.2.1.1) to the extraction bottle (see note 6), which shall then be stoppered tightly, placed in the shaker (4.2.2.7) and agitated for 16 h.

Filter the soil suspension into a clean and dry beaker (4.2.2.1) through a suitable filter paper (4.2.2.4) placed in the Buchner funnel (4.2.2.2).

Retain the filtrate for measurement of the sulfate content.

Measure the volume of the filtrate (V_E).

NOTE 6 If a centrifuge is available, it will probably prove more convenient to use a centrifuging tube for the extraction. The suspension can then be centrifuged instead of filtered and 25 ml of the clear supernatant liquid can be used for analysis after appropriate dilution, if required.

4.3 Determination of dry matter content

Determine the dry matter content of the air-dried

material following the procedure described in 2.2 of this International Standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 11048:1996

<https://standards.iteh.ai/catalog/standards/sist/480d404d-44be-4433-b128-a4ee63166fa1/sist-iso-11048-1996>