

# SLOVENSKI STANDARD SIST ISO 11292:2003

01-september-2003

## =bghUbh\_UjU'Ë'8 c`c YjUb^Y`dfcgh]\`]b`WY`chb]\`c[`1\_cj]\`\]XfUhcj`!`A YhcXU i dcfUVY`j]gc\_c``c `1]jY`Ub]cbg\_Y`]na Yb^YjU'bY`\_fcaUhc[fU2]^Y

Instant coffee -- Determination of free and total carbohydrate contents -- Method using high-performance anion-exchange chromatography

# iTeh STANDARD PREVIEW

Café soluble -- Détermination des teneurs en hydrates de carbone libres et totaux --Méthode par chromatographie d'échange d'anions à haute performance

SIST ISO 11292:2003

Ta slovenski standard je istoveten z: df696/sist-iso-11292:1995

ICS:

67.140.20 Kava in kavni nadomestki

Coffee and coffee substitutes

SIST ISO 11292:2003

en



# iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 11292:2003</u> https://standards.iteh.ai/catalog/standards/sist/fc5d7570-fa9a-46da-a865-084a9c8df696/sist-iso-11292-2003

### SIST ISO 11292:2003

# INTERNATIONAL STANDARD

ISO 11292

First edition 1995-06-15

Corrected and reprinted 1997-02-01

## Instant coffee — Determination of free and total carbohydrate contents — Method using high-performance iTeh STANDARD exchange chromatography

(standards.iteh.ai) Café soluble — Détermination des teneurs en hydrates de carbone libres et totaux ---- Méthode par chromatographie d'échange d'anions à haute https://standards.iten.avcatalog/standards/sist/fc5d7570-fa9a-46da-a865-

084a9c8df696/sist-iso-11292-2003



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting VIEW a vote.

International Standard ISO 11292 was prepared by Technical Committee ISO/TC 34, *Agricultural food products*, Subcommittee SC 15, *Coffee*. SIST ISO 11292:2003

Annexes A and B of this International Standard area for information only7570-fa9a-46da-a865-084a9c8df696/sist-iso-11292-2003

© ISO 1995

Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization

#### INTERNATIONAL STANDARD © ISO

## Instant coffee — Determination of free and total carbohydrate contents — Method using high-performance anion-exchange chromatography

#### Scope 1

For the purposes of this International Standard, the This International Standard specifies a method for the definitions given in ISO 3509 and the following defidetermination of free and total carbohydrate contents nitions apply. in instant coffee using high-performance anioneh.ai exchange chromatography. In particular, it determines S.IT 3.1 free carbohydrate content: Content of each the content of individual monosaccharides, sucrose SIST ISO 11292.20individual monosaccharide (arabinose, and mannitol.

https://standards.iteh.ai/catalog/standards/sist/1931actose.oglucose.6mannose), and the sucrose and 084a9c8df696/sist-iso-11299anoitol contents, determined under the conditions

3 Definitions

#### 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 1042:1983, Laboratory glassware — One-mark volumetric flasks.

ISO 3509:1989, Coffee and products its Vocabulary.

ISO 3726:1983, Instant coffee - Determination of loss in mass at 70 °C under reduced pressure.

described (method A). Content is expressed as a percentage by mass on a dry basis.

3.2 total carbohydrate content: Content of each individual monosaccharide (arabinose, galactose, glucose, mannose, xylose) and the mannitol content, determined under the conditions described, which includes a strong hydrolysis step (method B). Content is expressed as a percentage by mass on a dry basis.

#### Principle

#### 4.1 Method A

Dissolution of a test portion in water. Separation of the carbohydrates present in the filtered extract by ion chromatography on a high-performance anionexchange column (HPAEC) using pure water as eluent. Electrochemical detection of the eluted compounds by means of a pulsed amperometric detector (PAD) and quantification by comparison with peak areas given by standard solutions.

fructose.

#### 4.2 Method B

Hydrolysis of a test portion with aqueous hydrochloric acid. Analysis of the carbohydrates present in the filtered hydrolysed solution as described in method A.

#### Reagents 5

Use only reagents of recognized analytical grade, unless otherwise specified, and distilled or demineralized water or water of equivalent purity.

**5.1 Sodium hydroxide** (NaOH), 50 % (*m*/*m*) aqueous solution.

The reagent should contain the minimum amount of sodium carbonate and mercury. Do not shake or stir the solution before use.

Weigh, to the nearest 0,1 mg, approximately 100 mg of each carbohydrate into separate 100 ml volumetric flasks (6.2) and dilute to the mark with water (stock standard solutions of 1 000 mg/l).

Mixed standard solutions can also be prepared from separate stock solutions once the retention time of each carbohydrate is known under the prevailing chromatographic conditions.

Further dilute the standard solutions to reach carbohydrate concentrations similar to those found in the non-hydrolysed or hydrolysed instant coffee sample solutions.

The resolution of rhamnose from arabinose is difficult to achieve. If these two monosaccharides coelute, do not add rhamnose in a mixed standard solution.

Usual laboratory apparatus and, in particular, the fol-

### 6 Apparatus

5.2 Hydrochloric acid (HCI), 1,00 mol/l standard volumetric solution.

iTeh STANDAlowing. PREVIEW (standard, s. Analytical balance, capable of weighing to an

5.3 Eluent 1 (S1), demineralized water accuracy of  $\pm$  0,1 mg.

(18 MΩ·cm).

SIST ISO 11292:2003

https://standards.iteh.ai/catalog/standards/sist/fc5d7570-fa9a-46da-a865-

brane filters. Degas by sparging with helium for between 20 min and 30 min.

5.4 Eluent 2 (S2), sodium hydroxide (NaOH), 300 mmol/l solution.

To 985 ml of degassed water (5.3), pipette 15,6 ml of the sodium hydroxide solution (5.1).

CAUTION - It is extremely important to remove dissolved carbon dioxide from the eluents prior to use. Carbonate will act as a strong "pusher" on the column, resulting in a drastic reduction in resolution and efficiency. Prepare the solution the day before the analysis.

#### 5.5 Carbohydrate standard solutions.

Prepare fresh solutions of arabinose, fructose, galactose, glucose, mannose, sucrose and mannitol.

Filter the demineralized water through 0,2 µmmemdi696/si6.20- One-mark volumetric flasks, of capacity 100 ml (in accordance with class A of ISO 1042).

> 6.3 Graduated cylinders, of capacities 1 000 ml and 50 ml. tall form.

6.4 Vacuum filtration system.

**6.5** Folded filter papers, medium fast, gualitative.

6.6 Disposable C18 filter cartridges<sup>1</sup>, to be used according to the manufacturer's recommendations.

6.7 Disposable membrane filters, 0,2 µm pore size.

6.8 Water bath, capable of being maintained at 100 °C + 5 °C.

<sup>1)</sup> Sep-Pack C18 (Waters) and Supelclean LC-18 (Supelco) are examples of suitable products available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.

**6.9 Metal-free liquid chromatograph**<sup>2</sup>), with a high-performance anion-exchange column<sup>3</sup> filled with pellicular polystyrene-divinylbenzene resin and precolumn (guard column)<sup>4</sup>) and postcolumn delivery system.

6.10 Pulsed amperometric detector (PAD) with gold electrode<sup>5)</sup>.

**6.11** Integrator chromatography data station<sup>6)</sup>.

**6.12 Disposable cartridges**<sup>7</sup>, to be used according to the manufacturer's recommendations.

### 7 Sampling

8

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 6670:1983, *Instant coffee in cases with liners* — *Sampling*. **IIICH STANDARD** 

complete. Dilute to the mark with water. Filter 5 ml to 10 ml of this solution through a cartridge (6.6). Discard the first few millilitres.

# 8.3 Preparation of sample for analysis — Method B

Weigh, to the nearest 0,1 mg, approximately 300 mg of the laboratory sample directly into a 100 ml volumetric flask (6.2). Add 50 ml of the hydrochloric acid (5.2) and swirl. Place the flask in a boiling water bath (6.8) for 150 min.

Keep the level of the sample solution always below that of the water in the bath. Swirl the solution by hand every 30 min. Cool to room temperature by passing the flask under tap water. Dilute to the mark with water and filter the solution through a folded filter paper (6.5). Pass 3 ml of the filtrate through a disposable cartridge (6.12). Discard the first millilitre.

#### 8.4 Chromatographic analysis

Set up the chromatograph (6.9), detector (6.10) and integrator (6.14).

When the instant coffee is not in cases with liners, take a well-mixed representative sample from single **S.it Alow the** chromatograph to equilibrate. packed units. Filter the standard solutions (5.5) and the test sol-

Sist iso 11292:2001Procedurehttps://standards.iteh.ai/catalog/standards/sist/frée/7570-fa9a-46da-a865-<br/>084a9c8df696/sist-iso-11292-2003

#### 8.1 Determination of dry matter

Calculate the dry matter determined on a portion of the laboratory sample in accordance with ISO 3726.

# 8.2 Preparation of sample for analysis — Method A

Weigh, to the nearest 0,1 mg, approximately 300 mg of the laboratory sample directly into a 100 ml volumetric flask (6.2). Add, using a graduated cylinder (6.3), 70 ml of water and shake until dissolution is

Inject the same volume of filtered standard and test solutions into the chromatograph and separate carbohydrates under the conditions given in tables 1 and 2.

Identify and quantify carbohydrates in the sample solution by comparison with retention times and areas of corresponding peaks obtained using the standard solution.

Inject a standard solution every four injections, in order to account for any changes in retention times or peak integrations.

This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.

<sup>2)</sup> The BioLC system (Dionex) consisting of a model GPM-II quarternary gradient pump (Dionex) with a model SP8875 autosampler (Spectra - Physics) filled with a 20  $\mu$ I loop, model EDM-II eluent degas module (Dionex) and reagent reservoir for NaOH postcolumn addition (Dionex) are examples of suitable equipment available commercially.

<sup>3)</sup> CarboPac PA1 (10  $\mu$ m, 250 mm × 4 mm) (Dionex) is an example of a suitable analytical column available commercially.

<sup>4)</sup> CarboPac PA (Dionex) is an example of a suitable precolumn available commercially.

<sup>5)</sup> Model PAD-II (Dionex) is an example of suitable equipment available commercially.

<sup>6)</sup> Model Autolon Al-450 is an example of suitable equipment available commercially.

<sup>7)</sup> On Guard-AG (Dionex) is an example of a suitable cartridge commercially available.

Eluent	Time	Eluent S1	Eluent S2	Procedure
	min	ml	ml	
Isocratic	0	100	0	Start data acquisition
	50,0	100	0	Stop data acquisition
	50,1	0	100	Start clean-up
	65,0	0	100	Stop clean-up
	65,1	100	0	Start re-equilibrium
	80,0	100	0	Stop re-equilibrium

Table	1	 Pren	paration	of	column
anic	8	 110		01	corariin

NOTES

1) Retention times tend to vary from one column to another. Start clean-up only when the last monosaccharide (ribose) has been eluted.

2) It may be necessary to perform two or three injections of standard solution or to increase the reequilibrium time in order to achieve a good separation of sucrose and xylose.

## **iTeh STANDARD PREVIEW**

## Table 254 Conditions for analysis

Injection	20 μl SIST ISO 11292:2003
Flowrate ht	pt; (5 tmt/hmids.iteh.ai/catalog/standards/sist/fc5d7570-fa9a-46da-a865-
Postcolumn addition	Eluent S2 (5.4) at a flowrate of 0,6 ml/min
Temperature	Ambient
Detector	Fill up the reference cell with Eluent S2 (5.4). Use the optimum conditions given by the manufacturer.

#### Calculation 9

The carbohydrate content,  $\omega$ , expressed as a percentage by mass, is equal to

$$\omega = \frac{A \cdot m_0 \cdot V}{A_0 \cdot m \cdot V_0} \times 100$$

where

- is the peak area of the individual carbo-Α hydrate in the test solution (8.4);
- is the peak area of the individual carbo- $A_0$ hydrate in the standard solution (8.4);

- is the mass, in grams, of the test portion т in the test solution (8.2 or 8.3), expressed on a dry basis;
- is the mass, in grams, of the carbohydrate  $m_0$ in the standard solution (5.5);
- is the volume, in millilitres, of the test sol-Vution (8.2 or 8.3);
- is the volume, in millilitres, of the standard  $V_0$ solution (5.5).

Take as the result the arithmetic mean of the two determinations. Express the result either as free (method A) or total (method B) carbohydrate content to the nearest 0,01 % (m/m) for each carbohydrate of interest, or the total of all carbohydrates detected.

#### Precision 10

Details of an interlaboratory test on the precision of the method are summarized in annex A. The values derived from this interlaboratory test may not be applicable to concentration ranges and matrices other than those given.

The values of repeatability and reproducibility for each carbohydrate show a marked relationship with the wide percentage concentration of each carbohydrate, which in many cases in the samples examined was very low. This is reflected especially in some poor reproducibility figures, as might be expected. However, with contents at higher concentrations, both precision values are acceptable, as is also indicated in annex A.

#### 10.1 Repeatability

Repeatability is defined as the absolute difference between two independent single test results obtained on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time. il'eh S'l'ANDAR laboratories with different operators using different equipment.

Again it is not possible to give an exact reproducibility figure for each and every carbohydrate, both free and total, for a full range of possible carbohydrate contents

Provided, however, that the content of a given carbohydrate is greater than 0.3 % (m/m), the data show that the relative standard deviation averages 14,3 % (except for fructose data). At a content of less than 0.3 % (m/m), the coefficient of variation increases very sharply.

For each carbohydrate, the reproducibility and repeatability ranges as found in the interlaboratory test and the applied mean content range are given in table 3.

#### 11 **Test report**

**K**E

The test report shall specify

 the method in accordance with which sampling was carried out, if known,

the method used, It is not possible therefore to give an exact repeateh.aı) ability figure for each and every carbohydrate, both - the result obtained for each carbohydrate of free and total, for a full range of possible percentage 2003 interest. contents.

https://standards.iteh.ai/catalog/standards/sist/fc5d7570-fa9a-46da-a865-

Provided, however, that the content of a given carbo iso-11292-200 the repeatability has been checked, the final quoted result obtained. hydrate is greater than 0.3 % (m/m), the data show that the relative standard deviation averages 4,5 %.

#### 10.2 Reproducibility

Reproducibility is defined as the absolute difference between two single test results, obtained using the same method on identical test material in different

It shall also mention all operating details not specified in this International Standard, or regarded as optional, together with details of any incidents which may have influenced the test result.

The test report shall include all information necessary for the complete identification of the sample.

#### SIST ISO 11292:2003

Carbohydrate	Mean content range applied in test $\% (m/m)$ d.b.	Range of coefficient of variation of reproducibility, %	Range of coefficient of variation of repeatability, %	
Free mannitol	0,02 to 1,6	59,5 to 9,9	0,9 to 8,2	
Free arabinose	0,46 to 1,3	13,8 to 5,1	1,6 to 7,3	
Total arabinose	3,5 to 4,8	21,1 to 4,9	6,6 to 3,0	
Free galactose	0,19 to 0,56	13,0 to 4,1	9,8 to 3,0	
Total galactose	8,1 to 18,5	7,5 to 12,9	8,1 to 1,7	
Free glucose	0,04 to 2,0	23,8 to 6,1	10,2 to 2,5	
Total glucose	0,68 to 16,6	12,5 to 24,3	8,7 to 3,8	
Free mannose	0,16 to 1,0	40,0 to 16,9	8,2 to 3,8	
Total mannose	2,6 to 19,1	10,6 to 21,7	2,0 to 5,8	
Total xylose	0,1 to 1,9	37,7 to 20,2	22,9 to 3,7	
Free fructose	0,05 to 3,6	45,2 to 15,5	21,0 to 0,2	
Sucrose	0,15 to 1,3	41,6 to 10,0	15,1 to 1,8	
	(stanua)	ius.iteii.al)		

Table 3 — Coefficient of variation of reproducibility and repeatability	Table	3 —	Coefficient	of variation	of reproducibility	and repeatabilit
---	-------	-----	-------------	--------------	--------------------	------------------

<u>SIST ISO 11292:2003</u> https://standards.iteh.ai/catalog/standards/sist/fc5d7570-fa9a-46da-a865-084a9c8df696/sist-iso-11292-2003 SIST ISO 11292:2003

ISO 11292:1995(E)

### Annex A

(informative)

## **Results of an interlaboratory test**

An interlaboratory test, carried out at the international level in 1991 under the auspices of ISO/TC 34, *Agricultural food products*, SC 15, *Coffee*, and executed by the technical committee of AFCASOLE, in which 11 laboratories participated, each of which carried out 2 determinations on each sample of 6 different commercial coffees, gave the statistical results (evaluated in accordance with ISO 5725<sup>8</sup>) shown in tables A.1 to A.8.

# iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 11292:2003</u> https://standards.iteh.ai/catalog/standards/sist/fc5d7570-fa9a-46da-a865-084a9c8df696/sist-iso-11292-2003

7

<sup>8)</sup> ISO 5725:1986, Precision of test methods — Determination of repeatability and reproducibility for a standard test method by interlaboratory tests.