



SLOVENSKI STANDARD
SIST-TS ISO/TS 11370:2010
01-september-2010

**Kakovost vode - Določevanje izbranih organskih sredstev za zaščito rastlin -
Avtomatizirana večrazvojna tehnika (AMD)**

Water quality - Determination of selected organic plant-treatment agents - Automated multiple development (AMD) technique

iTeh STANDARD PREVIEW

Qualité de l'eau - Dosage de certains agents organiques de traitement des plantes -
Méthode automatisée par développement multiple (ADM)

[SIST-TS ISO/TS 11370:2010](https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c51ccc4451f8/sist-ts-iso-ts-11370-2010)

Ta slovenski standard je istoveten z: ISO/TS 11370:2000

ICS:

13.060.50	Preiskava vode na kemične snovi	Examination of water for chemical substances
-----------	---------------------------------	--

SIST-TS ISO/TS 11370:2010

en,fr

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TS ISO/TS 11370:2010](https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c31ccc4431f8/sist-ts-iso-ts-11370-2010)

<https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c31ccc4431f8/sist-ts-iso-ts-11370-2010>

TECHNICAL SPECIFICATION

ISO/TS 11370

First edition
2000-11-15

Corrected and reprinted
2001-06-01

Water quality — Determination of selected organic plant-treatment agents — Automated multiple development (AMD) technique

*Qualité de l'eau — Dosage de certains agents organiques de traitement des
plantes — Méthode automatisée par développement multiple (ADM)*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TS ISO/TS 11370:2010](https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c31ccc4431f8/sist-ts-iso-ts-11370-2010)

[https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-
c31ccc4431f8/sist-ts-iso-ts-11370-2010](https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c31ccc4431f8/sist-ts-iso-ts-11370-2010)



Reference number
ISO/TS 11370:2000(E)

© ISO 2000

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c31ccc4431f8/sist-ts-iso-ts-11370-2010>

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

	Page
1 Scope	1
2 Normative references	2
3 Interferences	2
4 Principle	2
5 Reagents	3
6 Apparatus	4
7 Sampling and samples	5
8 Procedure	5
9 Calibration	15
10 Evaluation	18
11 Expression of results	18
12 Test report	19
13 Precision data	19

Annexes

A Interlaboratory trial.....	20
A.1 Recovery summary	20
A.2 Information about the interlaboratory trial	20
B Compilation of recovery rates	22
Bibliography.....	23

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST-TS ISO/TS 11370:2010
<https://standards.iteh.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-c31ccc443118/sist-ts-iso-ts-11370-2010>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In other circumstances, particularly when there is an urgent market requirement for such documents, a technical committee may decide to publish other types of normative document:

- an ISO Publicly Available Specification (ISO/PAS) represents an agreement between technical experts in an ISO working group and is accepted for publication if it is approved by more than 50 % of the members of the parent committee casting a vote;
- an ISO Technical Specification (ISO/TS) represents an agreement between the members of a technical committee and is accepted for publication if it is approved by 2/3 of the members of the committee casting a vote.

An ISO/PAS or ISO/TS is reviewed every three years with a view to deciding whether it can be transformed into an International Standard.

[SIST-TS ISO/TS 11370:2010](https://standards.itih.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07-)
<https://standards.itih.ai/catalog/standards/sist/3fc6025c-f5a9-4794-8d07->

Attention is drawn to the possibility that some of the elements of this Technical Specification may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TS 11370 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical, biochemical methods*.

Annexes A and B of this Technical Specification are for information only.

Water quality — Determination of selected organic plant-treatment agents — Automated multiple development (AMD) technique

1 Scope

The method described in this Technical Specification is applicable to the determination of selected plant-treatment agents and some of their main degradation products (metabolites) in drinking water, with a validated reporting limit of about $> 0,05 \mu\text{g/l}$ (see examples in Table 1). The method may be extended to include additional substances and ground water, provided the method is validated for each individual case.

The selection of the plant-treatment agents and main degradation products in Table 1 and Table A.2 has been made according to the knowledge at the time of the interlaboratory trial (1992). Data for some other substances are given in annex A.

Table 1 — Plant-treatment agents determinable by this method

Name	Molecular formula	CAS No. ^a	Molar mass g/mol	Peak in Figure No.						
				1	2	3	4	5	6	7
Alachlor ^b	$\text{C}_{14}\text{H}_{20}\text{ClNO}_2$	015972-60-8	269,8	6				6		
Atrazine	$\text{C}_8\text{H}_{14}\text{ClN}_5$	001912-24-9	215,7	2				4		
Chlorfenvinphos ^b	$\text{C}_{12}\text{H}_{14}\text{Cl}_3\text{O}_4\text{P}$	000470-90-6	359,6	5				3		
Chlortoluron ^b	$\text{C}_{10}\text{H}_{13}\text{ClN}_2\text{O}$	015545-48-9	212,7	1						3
Cyanazine ^b	$\text{C}_9\text{H}_{12}\text{ClN}_6$	021725-46-2	240,7				1			4
2,4-D	$\text{C}_8\text{H}_6\text{Cl}_2\text{O}_3$	000094-75-7	221,0	4				1		
MCPA ^b	$\text{C}_9\text{H}_9\text{ClO}_3$	000094-74-6	200,6				2	2		
Metazachlor	$\text{C}_{14}\text{H}_{16}\text{ClN}_3\text{O}$	067129-08-2	277,8			3				5
Metobromuron	$\text{C}_9\text{H}_{11}\text{BrN}_2\text{O}_2$	003060-89-7	259,1			5				6
Metolachlor ^b	$\text{C}_{15}\text{H}_{22}\text{ClNO}_2$	051218-45-2	283,8			4				7
Metoxuron	$\text{C}_{10}\text{H}_{13}\text{ClN}_2\text{O}_2$	019937-59-8	228,7			1				1
Monuron ^b	$\text{C}_9\text{H}_{11}\text{ClN}_2\text{O}$	000150-68-5	198,7			2				2
Parathion ^b	$\text{C}_{10}\text{H}_{14}\text{NO}_5\text{PS}$	000056-38-2	291,3	7				7		
Pendimethalin	$\text{C}_{13}\text{H}_{19}\text{N}_3\text{O}_4$	040487-42-1	281,3		6					6
Propazine ^b	$\text{C}_9\text{H}_{16}\text{ClN}_5$	000139-40-2	229,7	3				5		
Sebuthylazine ^b	$\text{C}_9\text{H}_{16}\text{ClN}_5$	007286-69-3	229,7		2					3
Simazine	$\text{C}_7\text{H}_{12}\text{ClN}_5$	000122-34-9	201,7		1					2
2,4,5-T ^b	$\text{C}_8\text{H}_5\text{Cl}_3\text{O}_3$	000093-76-5	255,5		4					1
Terbutylazine ^b	$\text{C}_9\text{H}_{16}\text{ClN}_5$	005915-41-3	229,7		3					4
Trifluralin ^b	$\text{C}_{13}\text{H}_{16}\text{F}_3\text{N}_3\text{O}_4$	001582-09-8	335,3	8				8		
Vinclozoline ^b	$\text{C}_{12}\text{H}_9\text{Cl}_2\text{NO}_3$	050471-44-8	286,1		5					5

^a CAS No.: Chemical abstracts system.

^b Not included in the precision data (Table A.2).

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this Technical Specification. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this Technical Specification are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 5667-1:1980, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes.*

ISO 5667-2:1991, *Water quality — Sampling — Part 2: Guidance on sampling techniques.*

ISO 5667-3:1994, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples.*

ISO 8466-1:1990, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function.*

ISO 8466-2:1993, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 2: Calibration strategy for non-linear second order calibration functions.*

ISO/TR 13530:1997, *Water quality — Guide to analytical quality control for water analysis.*

3 Interferences

iTeh STANDARD PREVIEW
(standards.iteh.ai)

3.1 Interferences with the extraction

The commercially available RP-C18 materials are often of varying quality. Considerable batch-to-batch differences regarding quality and selectivity of this material, even from one manufacturer, are possible. The recovery may vary with the concentration. Co-extractants eluted from the sorbent material can affect the blank and the recovery. Therefore the calibration and analysis shall be performed on exactly the same batch of sorbent. Also any UV-absorbing material occurring in the water which passes through the procedure and has a similar migration distance to that of the reference standard will interfere. Suspended matter in the water sample may clog the packing. In this case the water sample shall be filtered through a glass fibre filter prior to the extraction.

If the water sample has been acidified to pH 2, humic substances will also be extracted. They may interfere with the determination.

3.2 Interferences with the HPTLC measurement

A contaminated laboratory atmosphere may lead to interferences due to an uncontrolled contamination of the HPTLC-layer. Extremely concentrated solutions may crystallize during sample application, leading to incorrect quantification. Failure of the AMD vacuum will result in poor resolution.

Substances which absorb at the wavelengths of detection and have migration distances similar to those of the compounds to be investigated will interfere with the determination. This shall be taken into account especially when examining samples other than ground- and drinking water.

4 Principle

The substances in the water sample are extracted by solid-liquid extraction on RP-C18 material (RP = reversed phase), eluted with a solvent and then separated by high performance thin layer chromatography (HPTLC), using the Automated Multiple Development (AMD) technique. The detection and determination is performed by diffuse *in-situ* reflection measurement at different UV-wavelengths.

5 Reagents

Water, solvents and reagents shall be of sufficient purity (e.g. residue grade, HPLC grade or AMD grade) and as far as possible shall not contain any measurable UV-absorbing substances which could interfere with the compounds of interest.

5.1 Hydrochloric acid, $c(\text{HCl}) = 1 \text{ mol/l}$ (for example).

5.2 Sodium hydroxide solution, $c(\text{NaOH}) = 1 \text{ mol/l}$ (for example).

5.3 Ammonia solution, $w(\text{NH}_3) = 25 \%$.

5.4 Gases, for drying, conditioning of the HPTLC plate in the AMD system and for evaporation, e.g. high-purity nitrogen.

5.5 Formic acid, $w(\text{HCOOH}) = 98 \%$ to 100% .

5.6 Solvents, e.g. acetone, $\text{C}_3\text{H}_6\text{O}$; acetonitrile, CH_3CN ; dichloromethane, CH_2Cl_2 ; 3,3-dimethyl-2-oxabutane, $\text{C}_6\text{H}_{12}\text{O}$; ethyl acetate, $\text{C}_4\text{H}_8\text{O}_2$; hexane, C_6H_{14} ; methanol, CH_3OH ; 2-propanol, $\text{C}_3\text{H}_8\text{O}$.

WARNING — These solvents are toxic agents. Caution shall be exercised when handling.

5.7 RP-C18 sorbent, for solid-phase extraction. For quality and selectivity of the material, see 3.1.

NOTE Other solid-phase adsorbents may be used, if the performance is comparable to this material and if it has been proved suitable according to 3.1.

5.8 Reference standards (see Table 1), of high purity or certified material.

5.9 Solutions of the individual reference standards.

Place, for example, 50 mg of the reference standards (5.8) in a 100 ml volumetric flask. Dissolve it in methanol or in another solvent (5.6) and make up to volume with the solvent.

NOTE Simazine is poorly soluble in acetonitrile.

Store the solutions at about 4°C , protected from light. They are stable for at least one month depending on the compound of interest. For longer use, check regularly by comparison with an independent, preferably certified standard solution.

The solution of vinclozoline shall be prepared freshly every second day.

5.10 Stock solution.

As an example, pipette 1 ml each of the solution of the individual standards (5.9) into a 100 ml volumetric flask, and make up to volume with methanol or another solvent (5.6).

Store the solutions at about 4°C , protected from light. They are stable for at least one month depending on the compound of interest.

5.11 Reference solution for the multipoint calibration.

Prepare the solution by an adequate dilution of the stock solution (5.10), e.g. $\rho_i = 10 \text{ ng}/\mu\text{l}$.

Store the solution at about 4°C , protected from light. It is stable for at least one week.

5.12 Reference solutions for the determination of the recovery.

Prepare the solution by an adequate dilution of the stock solution (5.10), e.g. $\rho_i = 20 \text{ ng/ml}$ to 200 ng/ml .

ISO/TS 11370:2000(E)

Store the solution at about 4 °C, protected from light. It is stable for at least one week.

5.13 Precoated HPTLC plates, with silica gel 60, 20 cm × 10 cm, preferably 100 µm layer thickness, with fluorescence indicator.

The plate shall be prewashed before use (see 8.5.2).

6 Apparatus

Equipment or parts of it which may come into contact with the sample or its extract shall be free from residues that could cause unacceptable interference in blanks. It is recommended to use glass, stainless steel or polytetrafluoroethene (PTFE) and, for cartridges, also polypropylene.

6.1 Flat-bottom flasks or bottles for sampling, preferably brown glass, e.g. 1 000 ml and 2 000 ml, stoppered with ground glass stoppers or with polytetrafluoroethene-lined screw caps,

6.2 Graduated cylinders, e.g. 500 ml and 1 000 ml.

6.3 Volumetric flasks, e.g. 10 ml, 25 ml, 50 ml and 100 ml.

6.4 pH meter.

6.5 Vacuum- or overpressure assembly, for sample enrichment and extract concentration.

6.6 Cartridges, from polypropylene or glass, filled with RP-C18 (e.g. internal diameter 9 mm, length 8 cm) or commercially available prefilled cartridges.

6.7 Glass vessels with stopper, for the collection and evaporation of the eluates (e.g. 5 ml graduated flasks or sampling vessels with conical bottom).

6.8 Borosilicate glass fibre filter, diameter 0,75 µm to 1,5 µm, with inorganic binding material.

6.9 Equipment for the evaporation of the eluates, e.g. a rotary evaporator with vacuum stabilizer and temperature-controlled water bath, or equipment for the evaporation of solvent with nitrogen.

6.10 Microlitre syringes, e.g. 100 µl, 250 µl and 1 000 µl, for the preparation of the reference solution and for adding the solvent to redissolve the residue of the evaporated eluates.

6.11 TLC chamber, to wash the HPTLC plates.

6.12 Heating device, suitable to dry the HPTLC plates, e.g. temperature-controlled hot plate.

A drying oven with forced air recirculation is not suitable for this purpose.

6.13 Desiccator, to store the pre-washed and dried HPTLC plates.

6.14 Applicator, for the band-shaped application of the solution onto the HPTLC plates.

6.15 AMD system, with vacuum pump.

6.16 UV scanner, for measurement and evaluation of thin-layer chromatograms, preferably computer-controlled and interfaced to a multicolour plotter.

Measurements below $\lambda = 200$ nm require an instrument equipped with a photomultiplier, specifically suitable for measurements at $\lambda = 190$ nm. Flush the monochromator housing with nitrogen.

7 Sampling and samples

To avoid interferences collect samples as stated below and according to ISO 5667-1, ISO 5667-2 or ISO 5667-3.

Use thoroughly cleaned, preferably brown, flat-bottom flasks (6.1). Rinse the flasks with the water to be sampled; treat the ground glass stoppers or the lined caps in the same way.

Fill the bottles to the brim with the water to be examined.

Transport the cooled sample, protected from light.

Extract substances from the water samples as soon as possible after sample collection.

If storage is unavoidable, keep the water samples at about + 4 °C in the dark or freeze them in suitable containers to approximately –18 °C. Half-filled glass bottles, welded into polyethylene bags, have proved to be appropriate.

NOTE If the samples are stored at about + 4 °C for longer than 3 days, analyte losses may occur.

8 Procedure

8.1 General requirements

The same conditions (e.g. amount of adsorbent, type of cartridge, conditioning, sample volume and flow, eluting steps and volumes) shall be used for all samples within one batch, including the procedure recovery samples.

Low recovery rates can occur from an insufficient amount of RP-C18 sorbent or an insufficient volume of methanol for the conditioning or elution step. Before analysing, these conditions should be checked and optimized in each laboratory. For typical recoveries, see annex A.

8.2 Conditioning of the RP-C18 material

For a water sample of 1 000 ml, place 1 g to 2 g of RP-C18 material (5.7) into a cartridge or glass column or use an adequate commercial device.

NOTE For more polar substances, e.g. metabolites, bad recoveries arise when using 1 g of RP-C18 for a 1 000 ml sample.

Rinse the RP-C18 material in the cartridge or glass column with five times its bed volume of eluting solvent (see 5.6).

Rewash with water (see clause 5) (five times its volume) and use the moist carrier material for the extraction.

The sorbent shall remain moist.

8.3 Extraction

If necessary remove suspended matter by filtration through a glass fibre filter and record this in the final report.

If filtration is carried out, use spiked samples in order to verify that the recovery is not influenced by this additional step.

Measure the water sample to be examined, e.g. 1 000 ml, adjust the pH from 6 to 8 with either hydrochloric acid (5.1) or sodium hydroxide solution (5.2).

If acidic compounds are to be determined, adjust the pH to between 1,5 and 2 with hydrochloric acid (5.1) immediately before the enrichment step.

NOTE Some plant-treatment agents may be destroyed by pH adjustment (e.g. vinclozoline, parathion).