

## SLOVENSKI STANDARD SIST EN 10136:1997

01-december-1997

Kemična analiza železovih zlitin - Določevanje niklja v jeklih in litinah - Plamenska atomska absorpcijska spektrometrična metoda

Chemical analysis of ferrous materials - Determination of nickel in steels and irons - Flame atomic absorption spectrometric method

Chemische Analyse von Eisenwerkstoffen - Bestimmung von Nickel in Stahl - Flammenatomabsorptionsspektrometrisches Verfahren FV

Analyse chimique des matériaux sidérurgiques - Dosage du nickel dans les aciers et les fontes - Méthode par spectrométrie d'absorption atomique dans la flamme

https://standards.iteh.ai/catalog/standards/sist/27d30aac-b157-4c0d-ad24-

Ta slovenski standard je istoveten z: EN 10136-1997

ICS:

77.040.30 Kemijska analiza kovin Chemical analysis of metals

SIST EN 10136:1997 en

SIST EN 10136:1997

# iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 10136:1997

https://standards.iteh.ai/catalog/standards/sist/27d30aac-b157-4c0d-ad24-c39aac9ee4a9/sist-en-10136-1997

## EUROPEAN STANDARD NORME EUROPÉENNE EUROPAISCHE NORM

**EN** 10 136

January 1989

UDC 669.14:543.422:546.74

Key words: Iron- and steel products, steels, cast iron, chemical analysis,

determination of content, nickel, atomic absorption

spectrophotometry, flame photometry.

English version

Chemical analysis of ferrous materials Determination of nickel in steels and irons Flame atomic absorption spectrometric method

Analyse chimique des matériaux Chemische Analyse von Eisenwerksidérurgiques - Dosage du nickel DARD Pstoffen - Bestimmung von Nickel
dans les aciers et les fontes in Stahl
Méthode par spectrométrie d'absorption des items en atomabsorptions spektroatomique dans la flamme metrisches Verfahren

#### SIST EN 10136:1997

This European Standard and size depths standard to comply with the requirements of 19 then CEN CENELEC Rules which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Central Secretariat or to any CEN member.

This European Standard exists in the official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to CEN Central Secretariat has the same status as the official versions.

CEN members are the national standards organizations of Austria, Belgium, Denmark. Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxemburg, Netherlands, Norway, Portugal, Spain. Sweden, Switzerland and United Kingdom.

## CEN

European Committee for Standardization Comité Européen de Normalisation Europäisches Komitee für Normung

Central Secretariat : Rue Bréderode 2, B-1000 Brussels

#### Brief Mistory

This European Standard takes over the content of EURONORM 136-83 "Chemical analysis of ferrous materials - Determination of nickel in steels and irons - flame atomic absorption spectrometric method", prepared by ECISS/TC 10 "Methods of chemical analysis"; the Secretariat of which is allocated to the Dansk Standardiseringsrad (DS).

It has been submitted to the CEN Formal Vote following the decision of the Coordinating Commission (COCCR) of the European Committee for Iron and Steel Standardization on 1987-11-24/25.

It has been adopted inc hatried by CIN BT on 1988-11-05.

According to the Common tath dansies Rules, alfollowing countries are bound to implement this European Standard:

SIST EN 10136:1997

Austria, Belgum/stablenmanha/cabbgandlardrance30@GermanyodGneece, Iceland, Ireland, Italy, Lukemburgac9eeepherdands6-19Norway, Portugal, Spain, Sweden, Switzerland and United Mingdom.

Mote in clauses 1 and 9 EUROMORM shall read EUROPEAN STANDARD.

EN 10 136 Page 3

## Chemical analysis of ferrous materials Determination of nickel in steels and irons Flame atomic absorption spectrometric method

#### CONTENTS

- SCOPE AND FIELD OF APPLICATION
- REFERENCE
- PRINCIPLE
- REAGENTS
- **APPARATUS**
- SAMPLING

- 7. PROCEDURE
  - 7.1 Test portion
  - 7.2 Blank test
  - 7.3 Determination
  - 7.4 Establishment of the calibration graph
- 8. EXPRESSION OF RESULTS
- 9. TEST REPORT

ANNEX: Precision data

## 1. SCOPE AND FIELD OF APPLICATION

iTeh STANDARD PREVIE

The method is applicable to steels and irons with nickel This EURONORM specifies a method for the determination contents of 0.003 to 2.0% (m/m). of nickei in steels and irons by means of flame atomic absorption spectrometry.

https://standards.iteh.ai/catalog/standards 30aac-b157-4c0d-ad24c39aachee4REFERENCE1997

EURONORM 18 - Selection and preparation of samples and test pieces for steel and iron and steel products.

General guidelines for the application of flame atomic absorption spectrometric methods are in course of preparation.

#### 3. PRINCIPLE

Dissolution of a test portion in a mixture of appropriate acids and furning with perchloric acid.

Spraying of the solution into an air-acetylene flame. Determination of the nickel by means of the spectrometric measurement of the atomic absorption of the 232.0 nm or 352.5 nm line emitted by a nickel hollow cathode lamp.

The instrument is calibrated by addition of a nickel standard solution to a similar matrix to that of the test solution.

NOTE - At the wavelength of 352.5 nm the signal-tonoise ratio is higher than at a wavelength of 232.0 nm. Generally, use of the 352.5 nm line will lead to a better reproducibility.

However, as the sensitivity at 352.5 nm is less than the sensitivity at 232.0 nm. with some instruments the use of the longer wavelength will be impossible when analysing low nickel contents.

#### 4. REAGENTS

During the analysis use only reagents of recognized analytical reagent quality and having a very low nickel content, and only distilled water or water of equivalent purity.

Carefully check the nickel content of all reagents.

If possible, use only freshly prepared distilled or deionized water.

- 4.1 Iron of high purity, with a nickel content < 0.0005% (m/m)
- 4.2 Hydrochloric acid-nitric acid mixture

Mix three volumes of hydrochloric acid,  $\rho$  1.19 g/ml approximately, one volume of nitric acid. p 1.40 g/ml approximately -EN 10 136 Page 4

and two volumes of water. This mixture is to be prepared immediately before use.

#### 4.3 Nitric acid—perchloric acid mixture

Mix 100 mi of nitric acid, p1.40 g/mi approximately with 300 ml of perchloric acid. p 1.54 g/ml approximately. Dilute to II with water and mix.

NOTE - Perchloric acid (p 1.67 g/ml approximately) may also be used, 100 mi of perchloric acid (p 1.54 g/mi approximateiv) is equivalent to 79 ml of perchione acid a 1.67 g/ml approximately).

#### 4.4 Nickel stock solution, corresponding to 1 mg of nickel per ml approximately

Weigh, to the nearest 0.001 g, about 0.5 g of high purity nickel ( 99.9% pure). Transfer the weighed mass to a 400 ml beaker and dissolve in 25 ml of nitric acid (p 1.40 g/ml approximately diluted 1 + 1 (v/v)). Boil to remove oxides of nitrogen. Cool and transfer the solution to a 500 ml volumetric flask, dilute to the mark with water and mix. Calculate the concentration of nickel in this stock solution, in mg/ml.

#### 4.5 Nickel reference solution, corresponding to 40 ug of nickel per mi approximately

Transfer 10.0 ml of nickel stock solution (4.4) to a 250 ml volumetric flask, dilute to the mark with water and mix. Calculate the concentration of nickel in the reference solution. in ug/ml.

#### 5. APPARATUS

Ordinary laboratory equipment and

pion PRE iTeh STANDA

5.1 Atomic absorption spectrometer: a nickel hollow cathode lamp; supplies of air and acetylene sufficiently pure to give a steady clear fuel-lean flame, free from water and oil, and free from EN 10136:199 nickel

For example, if the top and bottom calibration solutions represent 0.1% and 0.01% nickel in the sample, the precision called for (as two standard deviations) would be 0.002% and 0.001% nickel respectively, assuming curve linearity.

mean absorbency of the most concentrated calibration solu-

https://standards.iteh.ai/catalog/standards/sint/Additionalberrotmance/equirements

c39aac9ee4a9/sist-en-10136-199

The atomic absorption spectrometer used will be satisfactory if after optimization according to 7.3.4 the limit of detection and characteristic concentration are in reasonable agreement with the values given by the manufacturer and it meets the following performance requirements.

#### 5.1.1 Minimum precision

The standard deviation of 10 measurements of the absorbency of the most concentrated solution shall not exceed 1.0% of the mean absorbance.

The standard deviation of 10 measurements of the absorbance of the least concentrated calibration solution (excluding the zero calibration solution) shall not exceed 0.5% of the

It is also desirable that the instrument should conform to the following additional performance requirements.

- (a) Characteristic concentration The characteristic concentration for nickel in a matrix similar to the final test solution shall be better than 0.10 µg/mi.
- (b) Limit of detection The limit of detection of nickel in a matrix similar to the final test solution shall be better than 0.15 µg/mi.

#### 5.2 Ancillary equipment

A strip chart recorder and/or digital readout device is recommended to evaluate the criteria in 5.1 and for all subsequent measurements.

#### 6. SAMPLING

Sampling shall be carried out in accordance with EURONORM 18.

#### PROCEDURE

WARNING - Perchloric acid vapours may cause explosions in the presence of ammonia, nitrous fumes or organic matter in general. Always use a specially designed fume-hood.

#### 7.1 Test portion

Weigh the mass (m) indicated below to the nearest 0.001 g:  $m = 1 g \pm 5\%.$ 

#### EN 10 136 Page 5

#### 7.2 Blank test

With each analytical run, carry out an analysis on a reagent blank in parallel with the test portion analysis, using identical reagents, conditions, analytical procedures and dilutions throughout

#### 7.3 Determination

#### 7.3.1 Preparation of the test solution

Place the test portion (7.1) in a 250 ml beaker. Add 10 ml of hydrochlone acid-nitric acid mixture (4.2). Heat gently until the reaction ceases.

Add 15 ml of nitric acid-perchloric acid mixture (4.3) and heat until dense-wnite fumes of perchloric acid appear. Continue heating for one minute and allow to cool.

NOTE - If the test portion is readily soluble in the nitric acid-perchloric acid mixture (4.3), the addition of hydrochloric acid-nitric acid mixture (4.2) can be omitted. In that case the test portion is dissolved in the nitric acid-perchloric acid mixture (4.3) and the solution thus obtained is heated as described above.

Dissolve in 25 ml of water by heating gently. Cool again and transfer the solution to a 100 mi volumetric flask, dilute to the mark with water and mix. This is the test solution. If some residue has been left in the solution due to silicon, tungsten. mobium or tantaium, filter the solution through a dry, medium textured filter paper, discarding the first runnings. [ 2]

If the expected nickel content of the test sample exceeds 0.1% (m/m), dilute the solution as follows: SIST EN 10136:19

dilute to the mark with water and mix.

NOTE - If the solution has to be diluted to give the test solution, the blank test solution (7.2) must be diluted in the same way.

#### 7.3.2 Preparation of the calibration solutions

Place 10 g of iron (4.1) in a 800 ml beaker. Add 100 ml of hydrochloric acid—nitric acid mixture (4.2) and heat gently to dissoive the from.

Add 150 mi of nitric acid-perenionic acid mixture (4.3) and heat until dense white fumes of perculoric acid appear. Conunue heating for one minute and allow to cool. Dissoive in 100 mi of water by heating gently. Cool again and transfer the solution to a 250 ml volumetric flask, dilute to the mark with water and mix.

#### 7.3.2.1 Nickei content < 0.1% (m/m)

Transfer a series of 25.0 ml aliquots of the iron solution to 100 ml volumetric flasks, add to each flask by means of a burette or pipette respectively 0, 2.5, 5.0, 10.0, 15.0, 20.0 and 25.0 ml of nickel reference solution (4.5), dilute to the mark with water and mix.

#### 7.3.2.2 Nickel content 0.1 to 2.0% (m/m)

Transfer 25.0 ml of the iron solution to a 500 ml volumetric flask, dilute to the mark with water and mix. From this solution transfer 25.0 mi aliquots to 100 mi volumetric flasks.

Add to each flask by means of a burette or pipette respectively 0. 2.5, 5.0, 10.0, 15.0, 20.0 and 25.0 ml of nickel reference solution (4.5), dilute to the mark with water and

NOTE - 1 ml of nickel reference solution (4.5) is approximately equivalent to 0.004% (m/m) in the sample in the case of 7.3.2.1 and 0.08% (m/m) in the sample in the case of 7.3.2.2.

#### 7.3.3 Adjustment of atomic absorption spectrometer (5.1)

Type of lamp: Wavelength:

Flame:

Nickel hollow cathode. 232.0 nm or 352.5 nm. Air-acetylene clear fuel-lean

Lamn current -

flame adjusted for maximum nickel response

Slit width

Follow manufacturer's recommendations.

NOTE - The manufacturer's recommendations should be closely followed and particular attention is drawn to the following safety points:

- (a) the explosive nature of acetylene and regulations concerning its use:
- (b) the need to shield the eyes of the operator from UV radiation by means of tinted glass:
- (c) the need to keep burner head clear of deposits. A badly clogged burner may cause a flash back:
- (d) make sure that the liquid trap is filled with water.

Transfer 25.0 ml of the solution topa: 500 ml wotumetric alasks/standards/st settings

c39aac9ee4a9/sist-en-10136-1597
Follow the manufacturers instructions for preparing the instrument for use.

> When the current to the lamp, the wavelength and flow of gas have been adjusted and the burner lit, spray water until the indication has stabilized. Set the absorbance value to zero. Choose a damping setting or integration time to give a signal steady enough to fulfil the precision requirements (5.1).

> Adjust the flame to be clear fuel-lean and the burner height to about 1 cm below the light path. Spraying, alternately, the high and zero calibration solutions, adjust the gas flow and burner position (horizontally, vertically and rotationally) until the difference in absorpance between the calibration solutions is at a maximum

> Evaluate the criteria of 5.1.1 to ensure that the instrument is suitable for the determination.

#### 7.3.5 Atomic absorption measurements

Set the scale expansion so that the top calibration solution gives nearly full scale deflection. Aspirate the calibration solutions in ascending order repetitively until each gives the acceptable precision (see 5.1.1), thus showing that the instrument has achieved stability. Select two calibration solutions, one having an absorbency just lower than the test solution and one just higher. Aspirate these first in ascending order, then in descending order, with the test solution as the middle solution in each case, measuring the absorbency in relation to water. Aspirate the complete range of calibration solutions again.

It is recognized that these procedures cannot be followed with automatic instruments which accept two calibration solutions

#### EN 10 136 Page 6

only. In this case, it is suggested that the two 'sandwiching' solutions should not be used for the primary calibration but should be analysed alternately with the test solution.

Obtain the net mean absorbance of each calibration solution by subtracting the mean absorbance of the zero calibration

Obtain the mean absorbance of the test solution and the mean absorbance of the reagent blank solution.

Convert the mean absorbances of the test solution and of the reagent blank solution to micrograms of nickel per millilitre by means of the calibration graph (7.4).

### 7.4 Establishment of the calibration graph

It is necessary to draw up a new calibration graph for each series of determinations, and for the range of nickel contents expected.

Prepare a calibration graph by plotting the net mean absorbance values of the calibration solutions against micrograms of nickel per millilitre, calculated according to 4.5. Refer the net absorbances of the two adjacent calibration solutions to the graph. If these two calibration readings do not deviate from the graph by more than the permitted precision criteria then the test solution readings are acceptable.

#### EXPRESSION OF RESULTS

The percentage by mass of nickel (Ni) is given by the expression

Ni % (m/m) = 
$$\frac{(b-c) \cdot D \cdot 100}{10^{\circ}} \cdot \frac{100}{m} = \frac{(b-c) \cdot D}{100 m}$$

where:

b is the concentration of nickel in the test solution in  $\mu g/ml$ ;

- c is the concentration of nickel in the reagent blank solution in ug/ml:
- D is the dilution factor in 7.3.1;
  - (D = 1 for samples with an expected nickel content of 0.1% or less.
  - D = 20 for samples with an expected nickel content of more than 0.1%);

m is the mass of the test portion, in g.

## iTeh STANDARD PREVIEW

(spantestreporch.ai)

The test report shall contain the following particulars:

SIST EN 1((d)) any operations not specified in this EURONORMor any (a) The method of analysis used by reference to this EURO standards/soptional operations which could have had an influence on c39aac9ee4a9/sist-en-10136-1997 NORM:

- (b) the results obtained, as well as the form in which they are
- (c) any particular details which may have been noted during the determination:
- (e) all indications necessary for the identification of the sampie:
- (f) the laboratory and the date of analysis.

#### **ANNEX**

#### Precision data

Planned trials of this method were carried out by analysts from 6 to 19 laboratories: 2-5 determinations were carried out by each analyst on each sample. From the results

obtained the 95% (2s) confidence limits have been calculated in accordance with ISO 5725 and are tabulated as follows: