

SLOVENSKI STANDARD SIST EN 10178:1997

01-december-1997

Kemična analiza železovih zlitin - Določevanje niobija v jeklih - Spektrofotometrična metoda

Chemical analysis of ferrous materials - Determination of niobium in steels - Spectrophotometric method

Chemische Analyse von Eisenwerkstoffen - Bestimmung von Niob in Stählen - Photometrisches Verfahrenh STANDARD PREVIEW

(standards.iteh.ai)
Analyse chimique des matériaux sidérurgiques - Dosage du niobium dans les aciers - Méthode spectrophotométrique

https://standards.iteh.ai/catalog/standards/sist/8d39a1a4-e62b-479b-b8ce-

Ta slovenski standard je istoveten z: EN 10178-1997

ICS:

77.040.30 Kemijska analiza kovin Chemical analysis of metals

SIST EN 10178:1997 en

SIST EN 10178:1997

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 10178:1997</u> https://standards.iteh.ai/catalog/standards/sist/8d39a1a4-e62b-479b-b8ce-00427d069338/sist-en-10178-1997

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 10 178

January 1989

UDC 543.42.062:546.882:669.14

English version

Chemical analysis of ferrous materials
Determination of niobium in steels
Spectrophotometric method

Analyse chimique sidérurgiques - Dosage du niobium dans les aciers -Méthode spectrophotométrique Chemische analyse von Eisenwerk stoffen - Bestimmung von Niob in Stählen - Photometrisches Verfahren

iTeh STANDARD PREVIEW (standards.iteh.ai)

This European Standard was accepted by CEN on 1989-01-15. CEN members are bound to comply with the requirements of the CEN/CENELEC Rules which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Central Secretariat or to any CEN member.

This European Standard exists in the official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to CEN Central Secretariat has the same status as the official versions.

CEN members are the national standards organizations of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxemburg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization Comité Européen de Normalisation Europäisches Komitee für Normung

Central Secretariat : Rue Bréderode 2, B-1000 Brussels

Brief History

This European Standard takes over the content of EURONORM 178-85 "Chemical analysis of ferrous materials - Determination of niobium in steels - Spectrophotometric method", prepared by ECISS/TC 20 "Methods of chemical analysis"; the Secretariat of which is allocated to the Dansk Standardiseringsrad (DS).

It has been submitted to the CEN Formal Vote following the decision of the Coordinating Commission (COCOR) of the European Committee for Iron and Steel Standardization on 1987-11-24/25.

It has been adopted and ratified by CEN BT on 1988-11-05.

According to the Common CEN/CENELEC Rules, following countries are bound to implement this European Standard:

Austria, Belgium, Denmark, Finland, France; Germany, Greece, Iceland, Ireland, Italy, Luxemburg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom 1997

https://standards.iteh.ai/catalog/standards/sist/8d39a1a4-e62b-479b-b8ce-00427d069338/sist-en-10178-1997

Note in clauses 1 and 9 EURONORM shall read EUROPEAN STANDARD.

Chemical analysis of ferrous materials Determination of niobium in steels Spectrophotometric method

CONTENTS

- 1. SCOPE AND FIELD OF APPLICATION
- 2. REFERENCE
- 3. PRINCIPLE
- 4. REAGENTS
- 5. APPARATUS
- 6. SAMPLING

- 7. PROCEDURE
 - 7.1 Test portion
 - 7.2 Blank test
 - 7.3 Determination
 - 7.4 Establishment of the calibration graph
- 8. EXPRESSION OF RESULTS
- 9. TEST REPORT

ANNEX: Precision data

1. SCOPE AND FIELD OF APPLICATION

This EURONORM specifies a method for the spectrophotometric determination of niobium in steels.

The method is applicable to all types of steels with niobium contents up to 1.3% (m/m), with a lower detection limit of 0.002% (m/m).

iTeh STANDARD PKEV

(standards.iteh.ai

EURONORM 18 — Selection and preparation of samples and test pieces for steel and iron and steel products.

00427d069338/sist-en-10178-1997

3. PRINCIPLE

Dissolution of a test portion with hydrochloric acid followed by oxidation with hydrogen peroxide.

Precipitation of niobium and tantalum with phenylarsonic acid using zirconium as a carrier.

Formation of a complex of niobium with 4-(2-pyridylazo)-resorcinol (PAR) in a buffered sodium tartrate medium.

Spectrophotometric measurement of the coloured compound at a wavelength of 550 nm.

REAGENTS

During the analysis, use only reagents of recognized analytical quality and only distilled water or water of equivalent purity.

- 4.1 Iron, of high purity, free from niobium
- 4.2 Potassium hydrogen sulphate (KHSO₄)
- 4.3 Hydrochloric acid, ρ 1.19 g/ml approximately, (12 mol/l approximately)
- 4.4 Hydrochloric acid, ρ 1.19 g/ml approximately, diluted 1 + 4 (V/V), (2.4 mol/l approximately)
- 4.5 Hydrochloric acid, ρ 1.19 g/ml approximately, diluted 1 + 9 (V/V), (1.2 mol/l approximately)
- 4.6 Sulphuric acid, ρ 1.84 g/ml approximately, diluted 1 + 1 (V/V), (9 mol/l approximately)
- 4.7 Sulphuric acid, ρ 1.84 g/ml approximately, diluted 1 + 4 (V/V), (3.6 mol/l approximately)
- 4.8 Ethylenediaminetetra-acetic acid di-sodium N 10178:1997 salt, (EDTA Na2), 115 sg/lasolution hai/catalog/standards/4117/3 Niobium reference solution, corresponding to

Store in a polyethylene bottle.

- 4.9 Hydrogen peroxide, 30% w/v (100 vol.)
- 4.10 Phenylarsonic acid, 40 g/l solution

4.11 Phenylarsonic acid, 0.5 g/l solution

4.12 4-(2-pyridylazo)-resorcinol mono-sodium salt, (PAR), 0.6 g/l solution

NOTE - The di-sodium salt may also be used but the identical salt must be used for both calibration and tests.

4.13 Sodium acetate buffer, 350 g/l solution

Dissolve 350 g of sodium acetate trihydrate in 700 ml of water, add 5.5 ml of glacial acetic acid, ρ 1.05 g/ml, dilute to 1 000 ml and mix. Adjust the pH value to 6.3 with small additions of acetic acid or sodium hydroxide solution (4.14) using a pH meter for measurement.

4.14 Sodium hydroxide, 120 g/l solution

Store in a polyethylene bottle.

- 4.15 Tartaric acid, 100 g/l solution
- 4.16 Zirconium nitrate, 3 g/l solution in hydro-Dehloric acid medium

Dissolve 0.3 g of zirconium nitrate in 50 ml of hydrochloric acid solution (4.4). Filter through a fine filter paper, dilute to 100 ml with water and mix.

00427d069338/sist-en-101780.29mg of niobium per millilitre

Weigh 0.1431 g of niobium pentoxide and transfer to a platinum dish. Fuse with 3.5 g of potassium hydrogen sulphate (4.2). Cool and dissolve in 40 ml of tartaric acid solution (4.15). Add a further 160 ml of tartaric acid solution (4.15). Transfer to a 500 ml volumetric flask, dilute to the mark with water and mix.

APPARATUS

Ordinary laboratory equipment and a spectrophotometer suitable for measuring the absorbance of the solution at a wavelength of 550 nm, together with 4 or 1 cm cells.

SAMPLING

Sampling shall be carried out in accordance with EURONORM 18.

PROCEDURE

7.1 Test portion

Weigh the mass (m) indicated to the nearest 0.001 g: $m = 1 g \pm 5\%$

7.2 Blank test

With each analytical run, carry out an analysis on a 1 g portion of pure iron (4.1) in parallel with the test portion

_EN 10 178 Page 5

analysis, using identical reagents, conditions and dilutions throughout.

7.3 Determination

7.3.1 Preparation of the test solution

Transfer the test portion (7.1) to a 400 ml squat beaker, add 40 ml of hydrochloric acid solution (4.3), cover the beaker with a clock glass and heat until solvent action ceases. Cool slightly and add with caution 5 ml of hydrogen peroxide solution (4.9). Boil the solution for 1 min., dilute to approximately 200 ml with warm water and add 5 ml of zirconium nitrate solution (4.16).

7.3.2 Separation of the niobium

Heat the solution to boiling and add 25 ml of a boiling solution of phenylarsonic acid (4.10). Boil for 5 min, add a small amount of filter paper pulp, mix well and allow to stand for 10 min.

Filter through a pulp pad prepared from macerated filter paper and remove adhering particles from the beaker with a rubber-tipped glass rod. Wash the filter alternately with hot dilute hydrochloric acid solution (4.5) and cold phenylarsonic acid solution (4.11) until freed from iron salts. Finally wash several times with cold phenylarsonic acid solution (4.11). Transfer the filter and precipitate to a silica crucible. Dry, ignite at as low a temperature as possible until all carbonaceous matter is removed, and finally at 800 °C for at least 15 min. Cool in a desiccator, add a few drops of sulphuric acid solution (4.6) and evaporate to dryness very carefully. Heat to remove sulphur trioxide.

https://standards.iteh.ai/catalog/standards 00427d069338/sist-e

7.3.3 Preparation of the solution for spectrophotometry

Add 2 g of potassium hydrogen sulphate (4.2) and fuse carefully until a clear melt is obtained. Cool and dissolve the fusion products with 50 ml of a warm solution of tartaric acid (4.15) and transfer to a 400 ml beaker. Add 50 ml of water and mix.

Add 25 ml of sodium hydroxide solution (4.14) and cool. By means of a pH meter adjust the pH of the solution to approximately 6.0 with either sulphuric acid solution (4.7) or sodium hydroxide solution (4.14) as required. Cool to room temperature, transfer to a 250 ml volumetric flask, dilute to the mark with water and mix.

7.3.4 Development of the colour

Take an aliquot volume of the test solution (7.3.3) according to the following table:

Niobium % (m/m)	Volume of aliquot (ml) 25.0 10.0		
Less than 0.26			
0.26-0.65			
0.65-1.30	5.0		

Transfer the aliquot to a 100 ml volumetric flask. By means of a pipette add 10 ml of EDTA Na₂ solution (4.8), 10 ml of

PAR solution (4.12) and 10 ml of buffer solution (4.13), mixing well after each addition. Allow to stand for 15 min. at approximately 20°C, then dilute to the mark with water and mix. Allow to stand for a further 30 min.

7.3.5 Spectrophotometric measurement

Carry out the spectrophotometric measurement at a wavelength of 550 nm after having set the spectrophotometer to zero absorbance in relation to water. Use 4 cm cells for niobium contents up to 0.06% and 1 cm cells for contents greater than 0.06%.

Convert the readings for the test portion solution and for the blank test solution to milligrams of niobium by reference to the calibration graph (7.4).

7.4 Establishment of the calibration graph

7.4.1 Preparation of the calibration solutions

Weigh 1.0 g portions of pure iron (4.1) into a series of 400 ml beakers. Add from a burette, volumes of niobium reference solution (4.17) as shown in the table below:

Niobium Solution MI	Niobium added mg	Cell size
itehai)	0	l and 4 cm
1.0	0.2	4 cm
2.0	0.4	4 cm
78:1997 3.0	0.6	1 and 4 cm
sist/8d39 5.9 a4-e62	h-479h-hR 0 e-	i cm
en-10178-1 9 97	1.4	1 cm
9.0	1.8	1 cm
11.0	2.2	1 cm
13.0	2.6	1 cm

Continue as described in clauses 7.3.1 to 7.3.4 but in all cases taking a 25 ml aliquot in 7.3.4.

7.4.2 Spectrophotometric measurements

Carry out spectrophotometric measurements according to the method described in paragraph 1 of clause 7.3.5, after having adjusted the spectrophotometer to zero absorbance in relation to water.

7.4.3 Plotting the calibration graph

From each of the absorbance readings, subtract the reading obtained on the test portion with no added niobium. Prepare a calibration graph by plotting the net absorbance readings against milligrams of niobium.

NOTE: For test samples containing more than 0.26% of niobium and for which 10 ml or 5 ml aliquots are used the relationship with the calibration curve based on 25 ml aliquots is as follows:

Cell	Equivalent % Nb on 1 g sample		
size	5 ml aliquot	10 ml aliquot	25 mi aliquot
l and 4 cm	0	0	0
4 cm	-	_	0.02
4 cm	_	_	0.04
l and 4 cm	_	_	0.06
1 cm	_	0.25	0.10
· 1 cm	0.70	0.35	0.14
l cm	0.90	0.45	0.18
l cm	1.10	0.55	0.22
1 cm	1.30	0.65	0.26

EXPRESSION OF RESULTS

 m_0

٧

The percentage by mass of niobium (Nb) is given by the expression

Nb% (m/m) =
$$\frac{m_1 - m_0}{m \cdot 1000} \cdot \frac{250}{V} \cdot 100 = \frac{m_1 - m_0}{m} \cdot \frac{25}{V}$$

where:

m

is the mass, in grams, of the test portion;

is the mass, in milligrams, of niobium found in the m. aliquot of the test portion solution;

> is the mass in milligrams of niobium found in the blank test solution;

> is the aliquot volume, in millilitres, taken from the test portion solution.

TEST REPORT

The test report shall contain the following particulars:

- (a) the method of analysis used by reference to this EURO-
- (b) the results obtained, as well as the form in which they are expressed;
- (c) any particular details which may have been noted during the determination:
- (d) any operations not specified in this EURONORM or any optional operations which could have had an influence on the result;
- (e) all indications necessary for the identification of the sam-

https://standards.iteh.ai/catalog/standards

(f) the laboratory and the date of analysis.

d39a1a4-e62b-479b-b8c

SISTENANNEX

Precision data

Planned trials of this method were carried out by 10 analysts from different laboratories; 5 determinations were carried out

by each analyst on each 4 samples; 5 determinations were 00427d069338/sist-en-1 made by 7 of the analysts on a fifth sample.* From the results obtained the 95% (2s) confidence limits have been calculated in accordance with ISO 5725, and are tabulated as follows:

Alloy type	Niobium	Repeatability	Reproducibility
	% (m/m)	r	R
Non-alloy steel Non-alloy steel Magnet alloy, 13% Ni, 8% Al, 3% Cu, 25% Co Niobium stabilized steel 18% Cr, 9.5% Ni Niobium stabilized steel 17.5% Cr, 13% Ni *	0.029	0.0020	0.0024
	0.099	0.0051	0.0076
	0.589	- 0.0127	0.0212
	1.035	0.0229	0.0408
	0.906	0.0175	0.0416

Repeatability

The difference between two single results found on identical material by one analyst using the same apparatus within a short time interval will exceed the repeatability, r, not more than once in 20 cases in the normal and correct operation of the method.

Reproducibility

The difference between two single and independent results found by two operators working in different laboratories on identical test material will exceed the reproducibility, R. on average, not more than once in 20 cases in the normal and correct operation of the method.