



## Dentistry — Resin-based filling materials

### TECHNICAL CORRIGENDUM 1

*Art dentaire — Produits d'obturation à base de résines synthétiques*  
*RECTIFICATIF TECHNIQUE 1*

Technical corrigendum 1 to International Standard ISO 4049 : 1988 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Sub-Committee SC 1, *Filling and restorative materials*.

ITeH STANDARD PREVIEW  
(standards.iteh.ai)

[ISO 4049:1988/Cor 1:1992](https://standards.iteh.ai/catalog/standards/sist/e25bedd3-c4b1-4619-9098-31660cff0fe6/iso-4049-1988-cor-1-1992)

<https://standards.iteh.ai/catalog/standards/sist/e25bedd3-c4b1-4619-9098-31660cff0fe6/iso-4049-1988-cor-1-1992>

Page 2

#### Subclause 4.2.7

Last line, replace "5 µg/mm<sup>3</sup>" by "7,5 µg/mm<sup>3</sup>".

Page 9

#### Subclause 7.9.4.4

First, second and third paragraphs, replace "5 µg/mm<sup>3</sup>" by "7,5 µg/mm<sup>3</sup>" (four occurrences in total).

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

This page intentionally left blank

[ISO 4049:1988/Cor 1:1992](#)

<https://standards.iteh.ai/catalog/standards/sist/e25bedd3-c4b1-4619-9098-31660cff0fe6/iso-4049-1988-cor-1-1992>