

SLOVENSKI STANDARD SIST EN ISO/IEC 17025:2005

01-julij-2005

BUXca Yý U. SIST EN ISO/IEC 17025:2002

Splošne zahteve za usposobljenost preskuševalnih in kalibracijskih laboratorijev (ISO/IEC 17025:2005)

General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)

Allgemeine Anforderungen an die Kompetenz von Prüf und Kalibrierlaboratorien (ISO/IEC 17025:2005) (standards.iteh.ai)

Ta slovenski standard je istoveten z: EN ISO/IEC 17025:2005

ICS:

03.120.20 Certificiranje proizvodov in Product and company

podjetij. Ugotavljanje certification. Conformity

skladnosti assessment

19.020 Preskuševalni pogoji in Test conditions and

postopki na splošno procedures in general

SIST EN ISO/IEC 17025:2005 sl,en

SIST EN ISO/IEC 17025:2005

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO/IEC 17025:2005

https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005

EUROPEAN STANDARD NORME EUROPÉENNE **EUROPÄISCHE NORM**

EN ISO/IEC 17025

May 2005

ICS 03.120.20

Supersedes EN ISO/IEC 17025:2000

English version

General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)

Exigences générales concernant la compétence des laboratoires d'étalonnages et d'essais (ISO/IEC 17025:2005)

Allgemeine Anforderungen an die Kompetenz von Prüf- und Kalibrierlaboratorien (ISO/IEC 17025:2005)

This European Standard was approved by CEN and CENELEC on 15 March 2005.

CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN or CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN or CENELEC member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees, respectively, of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

<u>SIST EN ISO/IEC 17025:2005</u>

https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005





CEN Management Centre: rue de Stassart, 36 B-1050 Brussels

CENELEC Central Secretariat: rue de Stassart, 35 B-1050 Brussels

Foreword

This document (EN ISO/IEC 17025:2005) has been prepared by Technical Committee ISO/CASCO "Committee on conformity assessment" in collaboration with Technical Committee CEN/CLC/TC 1 "Criteria for conformity assessment bodies", the secretariat of which is held by SN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2005, and conflicting national standards shall be withdrawn at the latest by November 2005.

This document supersedes EN ISO/IEC 17025:2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO/IEC 17025:2005 has been approved by CEN and CENELEC as EN ISO/IEC 17025:2005 without any modifications. dards.iteh.ai)

SIST EN ISO/IEC 17025:2005 https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005 **SIST EN ISO/IEC 17025:2005**

INTERNATIONAL STANDARD

17025

Second edition 2005-05-15

General requirements for the competence of testing and calibration laboratories

Exigences générales concernant la compétence des laboratoires d'étalonnages et d'essais

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO/IEC 17025:2005 https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005



Reference number ISO/IEC 17025:2005(E)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO/IEC 17025:2005 https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

| Forewo | ord | ٠. ٧ |
|----------------------------|---|--------------|
| Introdu | iction | . v i |
| 1 | Scope | 1 |
| 2 | Normative references | 2 |
| 3 | Terms and definitions | |
| _ | | |
| 4 4.1 | Management requirements | |
| 4.1 4.2 | Organization | |
| 4.2 | Document control | |
| 4.3.1 | General | |
| 4.3.2 | Document approval and issue | |
| 4.3.3 | Document changes | |
| 4.4 | Review of requests, tenders and contracts | 5 |
| 4.5 | Subcontracting of tests and calibrations | |
| 4.6 | Purchasing services and supplies | 6 |
| 4.7 | Service to the customers T.A. N.D.A.R.D. D.R.E.V.E.V.Complaints | 6 |
| 4.8 | Complaints | 7 |
| 4.9 | Control of nonconforming testing and/or calibration work | 1 |
| 4.10 4.11 | Corrective action | <i>1</i> |
| 4.11.1 | General SIST EN ISO/IEC 17025:2005 | C |
| 4.11.2 | Corrective action | ر |
| 4.11.3 | Selection and implementation of corrective actions 5-2005 | 6 |
| 4.11.4 | Monitoring of corrective actions | |
| | Additional audits | |
| 4.12 | Preventive action | |
| 4.13 | Control of records | |
| 4.13.1 | General | |
| 4.13.2 | Technical records | |
| 4.14 | Internal audits | |
| 4.15 | Management reviews | 10 |
| 5 | Technical requirements | |
| 5.1 | General | |
| 5.2 | Personnel | |
| 5.3 | Accommodation and environmental conditions | |
| 5.4 | Test and calibration methods and method validation | |
| 5.4.1 | General | |
| 5.4.2 5.4.3 | Selection of methodsLaboratory-developed methods | |
| 5.4.3 5.4.4 | Non-standard methods | |
| 5.4. 4 5.4.5 | Validation of methods | |
| 5.4.6 | Estimation of uncertainty of measurement | |
| 5.4.7 | Control of data | |
| 5.5 | Equipment | |
| 5.6 | Measurement traceability | |
| 5.6.1 | General | |
| 5.6.2 | Specific requirements | 17 |
| 5.6.3 | Reference standards and reference materials | |
| 5.7 | Sampling | 19 |

SIST EN ISO/IEC 17025:2005

ISO/IEC 17025:2005(E)

| 5.9 Assuring the quality of test and calibration results | 20 20 20 |
|--|----------------|
| 5.10 Reporting the results | 20 20 20 |
| 5.10.2 Test reports and calibration certificates | 20 |
| | |
| 5 10 3 Test reports | |
| v: i v:v = i vot i vpvi to: | 21 |
| 5.10.4 Calibration certificates | 22 |
| 5.10.5 Opinions and interpretations | 22 |
| 5.10.6 Testing and calibration results obtained from subcontractors | 23 |
| 5.10.7 Electronic transmission of results | 23 |
| 5.10.8 Format of reports and certificates | 23 |
| 5.10.9 Amendments to test reports and calibration certificates | 23 |
| Annex A (informative) Nominal cross-references to ISO 9001:2000 | 24 |
| Annex B (informative) Guidelines for establishing applications for specific fields | 26 |
| Bibliography | 27 |

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO/IEC 17025:2005 https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work. In the field of conformity assessment, the ISO Committee on conformity assessment (CASCO) is responsible for the development of International Standards and Guides.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

Draft International Standards are circulated to the national bodies for voting. Publication as an International Standard requires approval by at least 75 % of the national bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/IEC 17025 was prepared by the ISO Committee on conformity assessment (CASCO).

It was circulated for voting to the national bodies of both ISO and IEC, and was approved by both organizations.

(standards.iteh.ai)

This second edition cancels and replaces the first edition (ISO/IEC 17025:1999), which has been technically revised.

https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-0ef4349d5b83/sist-en-iso-iec-17025-2005

Introduction

The first edition (1999) of this International Standard was produced as the result of extensive experience in the implementation of ISO/IEC Guide 25 and EN 45001, both of which it replaced. It contained all of the requirements that testing and calibration laboratories have to meet if they wish to demonstrate that they operate a management system, are technically competent, and are able to generate technically valid results.

The first edition referred to ISO 9001:1994 and ISO 9002:1994. These standards have been superseded by ISO 9001:2000, which made an alignment of ISO/IEC 17025 necessary. In this second edition, clauses have been amended or added only when considered necessary in the light of ISO 9001:2000.

Accreditation bodies that recognize the competence of testing and calibration laboratories should use this International Standard as the basis for their accreditation. Clause 4 specifies the requirements for sound management. Clause 5 specifies the requirements for technical competence for the type of tests and/or calibrations the laboratory undertakes.

Growth in the use of management systems generally has increased the need to ensure that laboratories which form part of larger organizations or offer other services can operate to a quality management system that is seen as compliant with ISO 9001 as well as with this International Standard. Care has been taken, therefore, to incorporate all those requirements of ISO 9001 that are relevant to the scope of testing and calibration services that are covered by the laboratory's management system.

Testing and calibration laboratories that comply with this international Standard will therefore also operate in accordance with ISO 9001.

Conformity of the quality management system within which the laboratory operates to the requirements of ISO 9001 does not of itself demonstrate the competence of the laboratory to produce technically valid data and results. Nor does demonstrated conformity to this international Standard imply conformity of the quality management system within which the laboratory operates to all the requirements of ISO 9001.

The acceptance of testing and calibration results between countries should be facilitated if laboratories comply with this International Standard and if they obtain accreditation from bodies which have entered into mutual recognition agreements with equivalent bodies in other countries using this International Standard.

The use of this International Standard will facilitate cooperation between laboratories and other bodies, and assist in the exchange of information and experience, and in the harmonization of standards and procedures.

General requirements for the competence of testing and calibration laboratories

1 Scope

- **1.1** This International Standard specifies the general requirements for the competence to carry out tests and/or calibrations, including sampling. It covers testing and calibration performed using standard methods, non-standard methods, and laboratory-developed methods.
- **1.2** This International Standard is applicable to all organizations performing tests and/or calibrations. These include, for example, first-, second- and third-party laboratories, and laboratories where testing and/or calibration forms part of inspection and product certification.

This International Standard is applicable to all laboratories regardless of the number of personnel or the extent of the scope of testing and/or calibration activities. When a laboratory does not undertake one or more of the activities covered by this International Standard, such as sampling and the design/development of new methods, the requirements of those clauses do not apply. PREVIEW

- **1.3** The notes given provide clarification of the text, examples and guidance. They do not contain requirements and do not form an integral part of this International Standard.
- **1.4** This International Standard is for use by laboratories in developing their management system for quality, administrative and technical operations. Laboratory customers, regulatory authorities and accreditation bodies may also use it in confirming or recognizing the competence of laboratories. This International Standard is not intended to be used as the basis for certification of laboratories.
- NOTE 1 The term 'management system' in this International Standard means the quality, administrative and technical systems that govern the operations of a laboratory.
- NOTE 2 Certification of a management system is sometimes also called registration.
- **1.5** Compliance with regulatory and safety requirements on the operation of laboratories is not covered by this International Standard.
- **1.6** If testing and calibration laboratories comply with the requirements of this International Standard, they will operate a quality management system for their testing and calibration activities that also meets the principles of ISO 9001. Annex A provides nominal cross-references between this International Standard and ISO 9001. This International Standard covers technical competence requirements that are not covered by ISO 9001.
- NOTE 1 It might be necessary to explain or interpret certain requirements in this International Standard to ensure that the requirements are applied in a consistent manner. Guidance for establishing applications for specific fields, especially for accreditation bodies (see ISO/IEC 17011) is given in Annex B.
- NOTE 2 If a laboratory wishes accreditation for part or all of its testing and calibration activities, it should select an accreditation body that operates in accordance with ISO/IEC 17011.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC 17000, Conformity assessment — Vocabulary and general principles

VIM, International vocabulary of basic and general terms in metrology, issued by BIPM, IEC, IFCC, ISO, IUPAC, IUPAP and OIML

NOTE Further related standards, guides, etc. on subjects included in this International Standard are given in the Bibliography.

3 Terms and definitions

For the purposes of this document, the relevant terms and definitions given in ISO/IEC 17000 and VIM apply.

NOTE General definitions related to quality are given in ISO 9000, whereas ISO/IEC 17000 gives definitions specifically related to certification and laboratory accreditation. Where different definitions are given in ISO 9000, the definitions in ISO/IEC 17000 and VIM are preferred.

4 Management requirements STANDARD PREVIEW

4.1 Organization

(standards.iteh.ai)

- **4.1.1** The laboratory or the organization of which it is part shall be an entity that can be held legally responsible.

 https://standards.iteh.ai/catalog/standards/sist/579b4b8e-8c6d-4786-bc2f-
- **4.1.2** It is the responsibility of the laboratory to carry out its testing and calibration activities in such a way as to meet the requirements of this International Standard and to satisfy the needs of the customer, the regulatory authorities or organizations providing recognition.
- **4.1.3** The management system shall cover work carried out in the laboratory's permanent facilities, at sites away from its permanent facilities, or in associated temporary or mobile facilities.
- **4.1.4** If the laboratory is part of an organization performing activities other than testing and/or calibration, the responsibilities of key personnel in the organization that have an involvement or influence on the testing and/or calibration activities of the laboratory shall be defined in order to identify potential conflicts of interest.
- NOTE 1 Where a laboratory is part of a larger organization, the organizational arrangements should be such that departments having conflicting interests, such as production, commercial marketing or financing do not adversely influence the laboratory's compliance with the requirements of this International Standard.
- NOTE 2 If the laboratory wishes to be recognized as a third-party laboratory, it should be able to demonstrate that it is impartial and that it and its personnel are free from any undue commercial, financial and other pressures which might influence their technical judgement. The third-party testing or calibration laboratory should not engage in any activities that may endanger the trust in its independence of judgement and integrity in relation to its testing or calibration activities.

4.1.5 The laboratory shall

 a) have managerial and technical personnel who, irrespective of other responsibilities, have the authority and resources needed to carry out their duties, including the implementation, maintenance and improvement of the management system, and to identify the occurrence of departures from the management system or from the procedures for performing tests and/or calibrations, and to initiate actions to prevent or minimize such departures (see also 5.2);