



Designation: D 6822 – 02

An American National Standard



Designation: Manual of Petroleum Measurement Standards (MPMS), Chapter 9.3

## Standard Test Method for Density, Relative Density, and API Gravity of Crude Petroleum and Liquid Petroleum Products by Thermohydrometer Method<sup>1</sup>

This standard is issued under the fixed designation D 6822; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reappraisal.

### 1. Scope

1.1 This test method covers the determination, using a glass thermohydrometer, of the density, relative density (specific gravity), or API gravity of crude petroleum and liquid petroleum products with Reid vapor pressures of 101.325 kPa (14.696 psi) or less. Values are determined at existing temperatures and corrected to 15°C or 60°F by means of international standard tables.

1.2 The values stated in either SI units or inch pound-units are to be regarded separately as standard. The values stated in each system may not be exact equivalents; therefore, each system shall be used independently of the other. Combining values from the two systems may result in nonconformance with the standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

### 2. Referenced Documents

#### 2.1 ASTM Standards:

D 1298 Test Method for Density, Relative Density (Specific Gravity), or API Gravity of Crude Petroleum and Liquid Petroleum Products by Hydrometer Method (API MPMS Chapter 9.1)<sup>2</sup>

D 6300 Practice for Determination of Precision and Bias Data for Use in Test Methods for Petroleum Products and Lubricants<sup>2</sup>

E 100 Specification for ASTM Hydrometers<sup>3</sup>

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and the API Committee on Petroleum Measurement, and is the direct responsibility of Subcommittee D02.02/COMQ, the joint ASTM-API Committee on Static Petroleum Measurement.

Current edition approved Aug. 10, 2002. Published November 2002.

<sup>2</sup> Annual Book of ASTM Standards, Vol 05.04.

<sup>3</sup> Annual Book of ASTM Standards, Vol 14.03.

#### 2.2 Adjunct:

Adjunct to D 1250 Guide for Petroleum Measurement Tables (API MPMS Chapter 11.1)<sup>4</sup>

#### 2.3 API Standards:<sup>5</sup>

MPMS Chapter 9.1 Hydrometer Test Method for Density, Relative Density (Specific Gravity) or API Gravity of Crude Petroleum and Liquid Petroleum Products (ASTM Test Method D 1298)

MPMS Chapter 11.1 Volume Correction Factors

### 3. Terminology

#### 3.1 Definitions of Terms Specific to This Standard:

3.1.1 *API gravity* ( $^{\circ}$ API),  $n$ —a special function of relative density 60/60°F, represented by:

$$^{\circ}\text{API} = [141.5 / (\text{relative density } 60/60^{\circ}\text{F})] - 131.5 \quad (1)$$

3.1.1.1 *Discussion*—No statement of reference temperature is required, as 60°F is included in the definition.

3.1.2 *density*,  $n$ —the mass of liquid per unit volume at 15°C and 101.325 kPa with the standard unit of measurement being kilograms per cubic metre (kg/m<sup>3</sup>).

3.1.2.1 *Discussion*—Other reference temperatures, such as 20°C may be used for some products or in some locations. Less preferred units of measurement; for example, kg/L or g/mL, are still in use.

3.1.3 *observed values*,  $n$ —hydrometer readings observed at a temperature other than the defined reference temperature.

3.1.3.1 *Discussion*—These values are only hydrometer readings and not density, relative density, or API gravity at the temperature.

3.1.4 *relative density*,  $n$ —the ratio of the mass of a given volume of liquid at a specific temperature to the mass of an equal volume of pure water at the same or different temperature. Both reference temperatures shall be explicitly stated.

<sup>4</sup> Annual Book of ASTM Standards, Vol 05.01.

<sup>5</sup> Available from American Petroleum Institute, 1220 L St., NW, Washington, DC 20005.

3.1.4.1 *Discussion*—Common reference temperatures include 15/15°C, 60/60°F, 20/20°C, and 20/4°C. The historic term specific gravity may still be found.

3.1.5 *thermohydrometer, n*—a glass hydrometer with a self-contained thermometer.

#### 4. Summary of Test Method

4.1 The density or API gravity, after temperature equilibrium has been reached, is read by observing the freely floating thermohydrometer and noting the graduation nearest to the apparent intersection of the horizontal plane surface of the liquid with the vertical scale of the hydrometer after temperature equilibrium has been reached. The observed thermohydrometer reading is reduced to the reference temperature value by means of the Petroleum Measurement Tables (the appropriate adjunct to Guide D 1250/API MPMS Chapter 11.1) and observed temperature from the enclosed thermometer.

#### 5. Significance and Use

5.1 Density and API gravity are used in custody transfer quantity calculations and to satisfy transportation, storage, and regulatory requirements. Accurate determination of density or API gravity of crude petroleum and liquid petroleum products is necessary for the conversion of measured volumes to volumes at the standard temperatures of 15°C or 60°F.

5.2 Density and API gravity are also factors that indicate the quality of crude petroleum. Crude petroleum prices are frequently posted against values in kg/m<sup>3</sup> or in degrees API. However, this property of petroleum is an uncertain indication of its quality unless correlated with other properties.

5.3 *Field of Application*—Because the thermohydrometer incorporates both the hydrometer and thermometer in one device, it is more applicable in field operations for determining density or API gravity of crude petroleum and other liquid petroleum products. The procedure is convenient for gathering main trunk pipelines and other field applications where limited laboratory facilities are available. The thermohydrometer method may have limitations in some petroleum density determinations. When this is the case, other methods such as Test Method D 1298 (API MPMS Chapter 9.1) may be used.

5.4 This test method is suitable for determining the density, relative density, or API gravity of low viscosity transparent or opaque liquids, or both. This test method, when used for opaque liquids, requires the use of a meniscus correction (see 7.2).

#### 6. Apparatus

6.1 *Glass Thermohydrometers*, as specified in Specification E 100 (shown in Fig. 1), and graduated in:

6.1.1 Kilograms/cubic metre (kg/m<sup>3</sup>) and degrees Celsius for density hydrometers, as shown in Table 1.

6.1.2 Degrees API (°API) and degrees Fahrenheit for hydrometers measuring in API Gravity, as shown in Table 2.

6.2 *Hydrometer Cylinders*, clear glass, plastic, or metal. For convenience of pouring, the cylinder may have a pouring lip. The inside diameter shall be at least 25 mm (1 in.) greater than the outside diameter of the thermohydrometer used. The height of the cylinder shall be such that the bottom of the thermohy-

TABLE 1 Density Thermohydrometers

ASTM Hydrometer No.	Density, Range, kg/m <sup>3</sup>
300H	600 to 650
301H	650 to 700
302H	700 to 750
303H	750 to 800
304H	800 to 850
305H	850 to 900
306H	900 to 950
307H	950 to 1000
308H	1000 to 1050
309H	1050 to 1100
Hydrometer	
Total length, mm	374 to 387
Body diameter, mm	18 to 25
Stem diameter, mm, min	4.0
Hydrometer Scale	
Standard temperature, °C	15
Subdivisions, kg/m <sup>3</sup>	0.5
Short intermediate lines at, kg/m <sup>3</sup>	1
Long intermediate lines at, kg/m <sup>3</sup>	5
Main (numbered) lines at, kg/m <sup>3</sup>	10
Scale error at any point not to exceed, kg/m <sup>3</sup>	0.5
Length of nominal scale, mm	125 to 145
Scale extension beyond nominal range limits, kg/m <sup>3</sup>	2.5
Thermometer Scale	
Range, °C	
Designation L	−20 to +65
Designation M	0 to +85
Designation H	+20 to +105
Immersion	total
Subdivisions, °C	1.0
Intermediate lines at, °C	5
Main (numbered) lines at, °C	10
Scale error at any point not to exceed, °C	1.0
Scale length, mm	80 to 100

drometer clears the bottom of the cylinder by at least 25 mm (1 in.) when suspended in the sample test portion.

6.2.1 For field testing, a sample thief of suitable dimensions may be more convenient than a hydrometer cylinder. The liquid level shall be level with the top of the thief.

6.3 *Temperature Bath*, to control temperature close to the bulk hydrocarbon temperature or to control temperature close to the reference temperature of 15°C or 60°F.

#### 7. Procedure

##### 7.1 Effect of Test Temperature:

7.1.1 The density or API gravity determined by the thermohydrometer method is most accurate at or near the reference temperature of 15°C or 60°F. Other temperatures within the range of the enclosed thermometer may be used, if consistent with the type of sample and the necessary limiting conditions shown in Table 3.

7.1.2 If the test temperature is significantly different from the reference temperature of 15°C or 60°F, the expansion or contraction of the glass may affect the calibration of the thermohydrometer. A hydrometer correction factor (HYC) may be applied to the measured density value to provide a corrected reading. Historically, the following equations have been used within volume correction factor tables.

$$\rho_{\text{corrected}} = \rho_{\text{test}} \times \text{HYC} \tag{2}$$

**TABLE 2 API Gravity Thermohydrometers**

NOTE—For petroleum products and other liquids of similar surface tensions (33 dynes/cm or less).

Thermometer Scale in Body		Thermometer Scale in Stem	
ASTM Hydrometer No.	Nominal API Gravity Range, degrees	ASTM Hydrometer No.	Nominal API Gravity Range, degrees
41H-66	15 to 23	71H-62	-1 to +11
42H-66	22 to 30	72H-62	9 to 21
43H-66	29 to 37	73H-62	19 to 31
44H-66	36 to 44	74H-62	29 to 41
45H-66	43 to 51		
51H-62	-1 to +11		
52H-62	9 to 21		
53H-62	19 to 31		
54H-62	29 to 41		
55H-62	39 to 51		
56H-62	49 to 61		
57H-62	59 to 71		
58H-62	69 to 81		
59H-62	79 to 91		
60H-62	89 to 101		

  

Hydrometer		
	Thermometer Scale in Body	Thermometer Scale in Stem
Total length, mm	374 to 387	374 to 387
Body diameter, mm	18 to 25	23 to 27
Stem diameter, mm, min	4.0	6.0

  

Hydrometer Scale	
Standard temperature, °F	60
Subdivisions, °API	0.1
Intermediate lines at, °API	0.5
Main (numbered) lines at, °API	1.0
Scale error at any point not to exceed, °API	0.1
Length of nominal scale, mm	125 to 145

  

Thermometer Scale		
	Thermometer Scale in Body	Thermometer Scale in Stem
Range, °F <sup>A</sup>		
Designation L	0 to 150	
Designation M	30 to 180	30 to 220
Designation H	60 to 220	
Immersion	Total	Total
Subdivisions, °F	2	2
Intermediate lines at, °F	10	10
Main (numbered) lines at, °F	20	20
Scale error at any point not to exceed, °F	1	1
Scale length, mm	80 to 110	105 to 145

<sup>A</sup> Indication of the thermometer range is made by the use of the listed designation used as a suffix to the ASTM hydrometer number. For example, 54HL is an instrument with an API gravity range of 29 to 41°API and a thermometer range of 0 to 150°F. An instrument with the same gravity range but a thermometer range of 60 to 220°F would be designated 54HH. The number 57HM would identify an instrument with an API gravity range of 59 to 71°API and a thermometer range of 30 to 180°F.

**TABLE 3 Limiting Conditions and Test Temperatures**

Sample Type	Initial Boiling Point	Other Limits	Test Temperature
Volatile	120°C (250°F) or lower		Cool in original closed container to 18°C (65°F) or lower
Volatile and viscous	120°C (250°F) or lower	Viscosity too high at 18°C (65°F)	Heat to minimum temperature to obtain sufficient fluidity
Non-volatile	Above 120°C (250°F)		Use any temperature between -18°C and 90°C (0 and 195°F) as convenient
Mixture with non-petroleum products	...		Test at 15 ± 0.2°C or 60 ± 0.5°F