



# SLOVENSKI STANDARD SIST EN 301 839-1 V1.2.1:2007

01-december-2007

9`Y\_fca U[ bYfbUnXfi y`^j cgh]b`nUXYj Yj`nj Yn]`nfUX]`g\_`ja `gdY\_fca `fØFAŁ!  
BUdfUj Y\_fUh\_Y[ UXcgY[ UfGF8Ł!`5\_hj b]`a YX]W]bg\_]j gUX\_]i `fUa U\ b] `a c ]  
fl @!5 AŁ]b`df]dUXU`c Y`dYf]ZYfbY`bUdfUj Y`fl @!5 AŁDŁž\_]`XYi `Y`c`j `ZY\_j Yb bYa  
cVa c `f`cX`(\$&A<n`Xc`(\$)`A<n!`%`XY.`HY b] bY`\_fU`hf]gh\_]`b`dfYg\_i gbY  
a YtcXY

Electromagnetic compatibility and Radio spectrum Matters (ERM) - Short Range Devices (SRD) - Ultra Low Power Active Medical Implants (ULP-AMI) and Peripherals (ULP-AMI-P) operating in the frequency range 402 MHz to 405 MHz - Part 1: Technical characteristics and test methods  
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/3429f0c3-778a-4dbd-bb33-c8dd68b44dde/sist-en-301-839-1-v1-2-1-2007>

Ta slovenski standard je istoveten z: EN 301 839-1 Version 1.2.1

**ICS:**

11.040.40	Implantanti za kirurgijo, protetiko in ortetiko	Implants for surgery, prosthetics and orthotics
33.060.20	Sprejemna in oddajna oprema	Receiving and transmitting equipment
33.100.01	Elektromagnetna združljivost na splošno	Electromagnetic compatibility in general

**SIST EN 301 839-1 V1.2.1:2007** en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 301 839-1 V1.2.1:2007

<https://standards.iteh.ai/catalog/standards/sist/3429f0c3-778a-4dbd-bb33-c8dd68b44dde/sist-en-301-839-1-v1-2-1-2007>

# ETSI EN 301 839-1 V1.2.1 (2007-07)

---

*European Standard (Telecommunications series)*

**Electromagnetic compatibility  
and Radio spectrum Matters (ERM);  
Short Range Devices (SRD);  
Ultra Low Power Active Medical Implants (ULP-AMI)  
and Peripherals (ULP-AMI-P)  
operating in the frequency range 402 MHz to 405 MHz;  
Part 1: Technical characteristics and test methods**

---

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 301 839-1 V1.2.1:2007](https://standards.iteh.ai/catalog/standards/sist/3429f0c3-778a-4dbd-bb33-c8dd68b44dde/sist-en-301-839-1-v1-2-1-2007)

<https://standards.iteh.ai/catalog/standards/sist/3429f0c3-778a-4dbd-bb33-c8dd68b44dde/sist-en-301-839-1-v1-2-1-2007>



## Reference

---

REN/ERM-TG30-002-1

## Keywords

---

radio, SRD, testing**ETSI**

---

650 Route des Lucioles  
F-06921 Sophia Antipolis Cedex - FRANCE

Tel.: +33 4 92 94 42 00 Fax: +33 4 93 65 47 16

Siret N° 348 623 562 00017 - NAF 742 C  
Association à but non lucratif enregistrée à la  
Sous-Préfecture de Grasse (06) N° 7803/88

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 301 839-1 V1.2.1:2007<https://standards.iteh.ai/catalog/standards/sist/3429f0c3-778a-4dbd-bb33-c8dd68b44d4d/etsi-en-301-839-1-v1-2-1-2007>**Important notice**

Individual copies of the present document can be downloaded from:

<http://www.etsi.org>

The present document may be made available in more than one electronic version or in print. In any case of existing or perceived difference in contents between such versions, the reference version is the Portable Document Format (PDF). In case of dispute, the reference shall be the printing on ETSI printers of the PDF version kept on a specific network drive within ETSI Secretariat.

Users of the present document should be aware that the document may be subject to revision or change of status. Information on the current status of this and other ETSI documents is available at

<http://portal.etsi.org/tb/status/status.asp>

If you find errors in the present document, please send your comment to one of the following services:

[http://portal.etsi.org/chaicor/ETSI\\_support.asp](http://portal.etsi.org/chaicor/ETSI_support.asp)

---

**Copyright Notification**

No part may be reproduced except as authorized by written permission.  
The copyright and the foregoing restriction extend to reproduction in all media.

© European Telecommunications Standards Institute 2007.  
All rights reserved.

**DECT™**, **PLUGTESTS™** and **UMTS™** are Trade Marks of ETSI registered for the benefit of its Members.  
**TIPHON™** and the **TIPHON logo** are Trade Marks currently being registered by ETSI for the benefit of its Members.  
**3GPP™** is a Trade Mark of ETSI registered for the benefit of its Members and of the 3GPP Organizational Partners.

# Contents

Intellectual Property Rights .....	6
Foreword.....	6
Introduction .....	7
1 Scope .....	8
2 References .....	8
3 Definitions, symbols and abbreviations .....	9
3.1 Definitions .....	9
3.2 Symbols.....	11
3.3 Abbreviations .....	11
4 Technical requirements and specifications.....	12
4.1 General requirements .....	12
4.1.1 Transmitter requirements .....	12
4.1.2 Receiver requirements .....	12
4.2 Presentation of equipment for testing purposes.....	12
4.2.1 Choice of model for testing .....	12
4.2.2 Testing of equipment with alternative power levels .....	12
4.2.3 Presentation of equipment that does not have an external RF connector (integral antenna equipment) .....	13
4.2.3.1 Equipment with an internal permanent or temporary antenna connector.....	13
4.2.3.2 Equipment with a temporary antenna connector .....	13
4.2.3.3 Equipment intended to be implanted in a human body .....	13
4.3 Mechanical and electrical design.....	13
4.3.1 General.....	13
4.3.2 Controls .....	13
4.3.3 Transmitter shut-off facility.....	13
4.3.4 Marking .....	13
4.3.5 Equipment identification.....	13
4.4 Declarations by the Applicant .....	14
4.5 Auxiliary test equipment .....	14
4.6 Interpretation of the measurement results .....	14
5 Test conditions, power sources and ambient temperatures .....	14
5.1 Normal and extreme test conditions .....	14
5.2 Test power source.....	14
5.2.1 External test power source.....	14
5.2.2 Internal test power source .....	15
5.3 Normal test conditions.....	15
5.3.1 Normal temperature and humidity.....	15
5.3.2 Normal test power source .....	15
5.3.2.1 Mains voltage.....	15
5.3.2.2 Regulated lead-acid battery power sources .....	15
5.3.2.3 Other power sources.....	15
5.4 Extreme test conditions .....	16
5.4.1 Extreme temperatures .....	16
5.4.1.1 Procedure for tests at extreme temperatures.....	16
5.4.1.1.1 Procedure for equipment designed for continuous operation .....	16
5.4.1.1.2 Procedure for equipment designed for intermittent operation .....	16
5.4.1.2 Extreme temperature ranges.....	17
5.4.2 Extreme test source voltages.....	17
5.4.2.1 Mains voltage.....	17
5.4.2.2 Regulated lead-acid battery power sources .....	17
5.4.2.3 Power sources using other types of batteries.....	17
5.4.2.4 Other power sources.....	18
6 General conditions.....	18

6.1	Normal test signals and test modulation.....	18
6.1.1	Normal modulation test signals for data .....	18
6.2	Antennas.....	18
6.3	Artificial antenna.....	18
6.3.1	Artificial antenna for transmitters with 50 $\Omega$ impedance connector.....	18
6.4	Test fixture for ULP-AMI-P.....	19
6.5	Test fixture for ULP-AMI .....	19
6.6	Test sites and general arrangements for radiated measurements .....	19
6.7	Modes of operation of the transmitter .....	19
6.8	Measuring receiver .....	20
7	Measurement uncertainty .....	20
8	Methods of measurement and limits for transmitter parameters .....	21
8A	Introduction .....	21
8.1	Frequency error .....	21
8.1.1	Definition.....	22
8.1.1.1	Method of measurement for systems with an unmodulated carrier frequency operating mode.....	22
8.1.1.2	Method of measurement for systems with a modulated carrier frequency.....	22
8.1.2	Limit .....	22
8.2	Emission bandwidth measurement.....	23
8.2.1	Definition.....	23
8.2.1.1	Method of measurement.....	23
8.2.2	Limits.....	23
8.3	Effective radiated power of the fundamental emission .....	23
8.3.1	Definition.....	23
8.3.1.1	Methods of measurement .....	24
8.3.2	Limits.....	24
8.4	Spurious emissions .....	25
8.4.1	Definition.....	25
8.4.1.1	Method of measuring the effective radiated power of spurious emissions.....	25
8.4.2	Limits.....	26
8.5	Frequency stability under low voltage conditions.....	26
8.5.1	Definition.....	26
8.5.1.1	Method of measurement.....	26
8.5.2	Limits.....	27
8.6	ULP-AMI with restricted power and duty cycle .....	27
8.6.1	Definitions .....	27
8.6.2	Declaration of Duty Cycle .....	27
8.6.3	Limit for duty cycle and maximum number of transmissions .....	27
9	Methods of measurement and limits for receiver parameters.....	27
9.1	Spurious radiation.....	28
9.1.1	Definition.....	28
9.1.1.1	Method of measuring the effective radiated power of spurious radiations.....	28
9.1.2	Limits.....	29
10	Requirements and Measuring Methods for Monitoring Systems .....	29
10A	Purpose .....	29
10B	General Remarks on the Measurement Configuration .....	30
10.1	LBT threshold power level.....	30
10.1.1	Measurement method using out-of-operating-region disturbance .....	31
10.1.2	Measurement method using frequency administration commands .....	31
10.1.3	Results based on above test method.....	31
10.1.4	Limit .....	32
10.2	Monitoring system bandwidth.....	32
10.2.1	Measurement method using out-of-operating-region disturbance .....	32
10.2.2	Measurement method using frequency administration commands .....	33
10.2.3	Results based on above test method.....	33
10.3	Monitoring system scan cycle time and minimum channel monitoring period.....	33
10.3.1	Measurement method using out-of-operating-region disturbance .....	33
10.3.1.1	Scan cycle time .....	33
10.3.1.2	Minimum channel monitoring period .....	34

10.3.2	Measurement method using frequency administration commands .....	34
10.3.3	Results based on above test method.....	34
10.3.3.1	Scan cycle time .....	34
10.3.3.2	Minimum Channel Monitoring Period.....	34
10.4	Channel access based on ambient levels relative to the calculated access LBT threshold level, $P_{Th}$ .....	35
10.4.1	Access based on lowest ambient level above $P_{Th}$ using out-of-operating-region disturbance.....	35
10.4.2	Access based on lowest ambient level above $P_{Th}$ using frequency administration commands .....	35
10.4.3	Results based on above test method.....	36
10.5	Discontinuation of MICS session if a silent period greater than or equal to 5 s occurs .....	36
10.5.1	Measurement method.....	36
10.5.2	Results based on above test method.....	36
10.6	Use of pre-scanned alternate channel .....	36
10.6.1	Measurement method for alternate channel selection using out-of-operating-region disturbance.....	37
10.6.2	Measurement method for alternate channel selection using frequency administration commands .....	38
10.6.3	Results based on above test method.....	38
<b>Annex A (normative): Radiated measurements .....</b>		<b>39</b>
A.1	Test sites and general arrangements for measurements involving the use of radiated fields .....	39
A.1.1	Outdoor test site .....	39
A.1.1.1	Standard position .....	39
A.1.1.2	Equipment in close proximity to the human body but external to it .....	40
A.1.1.3	Human torso simulator for ULP-AMI .....	40
A.1.2	Test antenna.....	41
A.1.3	Substitution antenna .....	41
A.1.4	Optional additional indoor site .....	42
A.2	Guidance on the use of radiation test sites .....	43
A.2.1	Measuring distance.....	43
A.2.2	Test antenna.....	43
A.2.3	Substitution antenna .....	43
A.2.4	Artificial antenna.....	43
A.2.5	Auxiliary cables.....	43
A.3	Further optional alternative indoor test site using an anechoic chamber .....	44
A.3.1	Example of the construction of a shielded anechoic chamber .....	44
A.3.2	Influence of parasitic reflections in anechoic chambers.....	44
A.3.3	Calibration of the shielded RF anechoic chamber .....	45
<b>Annex B (normative): Technical performance of the spectrum analyser.....</b>		<b>47</b>
<b>Annex C (informative): Strategy for ULP-AMI/ULP-AMI-P in the band 402 MHz to 405 MHz (source ERM_TG30 see URL: <a href="http://www.etsi.org/">http://www.etsi.org/</a>) .....</b>		<b>48</b>
<b>Annex D (informative): Bibliography .....</b>		<b>49</b>
History .....		50

---

## Intellectual Property Rights

IPRs essential or potentially essential to the present document may have been declared to ETSI. The information pertaining to these essential IPRs, if any, is publicly available for **ETSI members and non-members**, and can be found in ETSI SR 000 314: "*Intellectual Property Rights (IPRs); Essential, or potentially Essential, IPRs notified to ETSI in respect of ETSI standards*", which is available from the ETSI Secretariat. Latest updates are available on the ETSI Web server (<http://webapp.etsi.org/IPR/home.asp>).

Pursuant to the ETSI IPR Policy, no investigation, including IPR searches, has been carried out by ETSI. No guarantee can be given as to the existence of other IPRs not referenced in ETSI SR 000 314 (or the updates on the ETSI Web server) which are, or may be, or may become, essential to the present document.

---

## Foreword

This European Standard (Telecommunications series) has been produced by ETSI Technical Committee Electromagnetic compatibility and Radio spectrum Matters (ERM).

For non EU countries the present document may be used for regulatory purposes.

The present document is part 1 of a multi-part deliverable covering Short Range Devices (SRD); Ultra Low Power Active Medical Implants (ULP-AMI) and Peripherals (ULP-AMI-P) operating in the frequency range 402 MHz to 405 MHz, as identified below:

**Part 1: "Technical characteristics and test methods";**

Part 2: "Harmonized EN covering essential requirements of article 3.2 of the R&TTE Directive".

### National transposition dates

Date of adoption of this EN:	29 June 2007
Date of latest announcement of this EN (doa):	30 September 2007
Date of latest publication of new National Standard or endorsement of this EN (dop/e):	31 March 2008
Date of withdrawal of any conflicting National Standard (dow):	31 March 2008



---

## Introduction

ULP-AMI/ULP-AMI-P equipment in the MICS service is a unique new technology, available world wide in the medical field, that will provide high speed communications capability between individuals with AIMDs and medical practitioners engaged in utilizing these AIMDs for the purposes of diagnosing and delivering therapy to individuals with various illnesses. Equipment in the MICS service consists of active medical implants that communicate to other active medical implants and/or to ULP-AMI-P as e.g. external programmer/control transmitters.

The present document includes methods of measurement for Ultra Low Power Active Medical Implants (ULP-AMI) and Peripherals (ULP-AMI-P), fitted with antenna connector and/or integral antenna. Equipment designed for use with an integral antenna may be supplied with a temporary or permanent internal connector for the purpose of testing, providing the characteristics being measured are not expected to be affected.

If equipment, which is available on the market, is required to be checked it should be tested in accordance with the methods of measurement specified in the present document.

Clauses 1 through 3 provide a general description on the types of equipment covered by the present document and the definitions, symbols and abbreviations used.

Clause 4 provides a guide to requirements, the number of samples required in order that tests may be carried out and any markings on the equipment that the provider has to supply.

Clauses 5 and 6 provide general test conditions to be used.

Clause 7 gives the maximum measurement uncertainty values.

Clauses 8, 9 and 10 specify spectrum utilization parameters and the measurement methods that are required for the protection of the spectrum and patient. Clause 10 describes channel access requirements and methods. In particular clause 10.1 describes the monitoring system performance specifications that have been chosen to minimize harmful interference to other equipment or services, reduce the potential for disturbance to this equipment from ambient sources or other medical device users in the band and provide a high degree of link reliability in the interest of the patient.

Annex A (normative) provides specifications concerning radiated measurements.

Annex B (normative) provides specifications for test equipment.

Annex C (informative) Strategy for ULP-AMI/ULP-AMI-P in the band 402 MHz to 405 MHz (source ERM\_TG30 see URL: <http://www.etsi.org>).

Annex D (informative) bibliography; provides additional information.

---

# 1 Scope

The present document covers, for Ultra Low Power Active Medical Implants (ULP-AMI) and Peripherals (ULP-AMI-P) used in a Medical Implant Communications Service (MICS), the required characteristics considered necessary to efficiently use the available spectrum and serve the interests of implant users. The specifications contained in the present document were developed to ensure that the health and safety of the patients that are using this equipment under the direction of medical practitioners is protected. Of particular importance is the inclusion of spectrum monitoring and access requirements (listen before talk protocol) designed to significantly reduce any interference potential between MICS systems operating in the band or between a MICS system and the primary users of the band. Also included in the present document is the capability of Low Duty Cycle/Low Power Access in the 403,5 MHz to 403,8 MHz frequency band.

An AIMD is regulated under the AIMD Directive 90/385/EEC: radio parts contained therein (referred to herein as ULP-AMI and ULP-AMI-P for peripheral devices) are regulated under the Directive 1999/5/EC (R&TTE Directive) [1].

It is intended that the present document applies to operation in the band 402 MHz to 405 MHz only and that devices that can also operate in spectrum outside this band also meet any applicable requirements for operation in such bands.

The present document contains the technical characteristics for ULP-AMI radio equipment which is also addressed by CEPT/ERC/REC 70-03 [2] and annex 12 to that document. It does not necessarily include all the characteristics, which may be required by a user, nor does it necessarily represent the optimum performance achievable.

It applies to ULP-AMI devices and accessories operating in the band 402 MHz to 405 MHz:

- for telecommand and telemetry to/from an AIMD in a patient's body to an ULP-AMI-P; or
- for telecommand and telemetry to/from an AIMD to another AIMD within the human body;
- with or without an integral antenna; and/or
- with an antenna connection provided only for the purpose of connecting an external dedicated antenna.

The present document covers requirements for radiated emissions above 25 MHz.

---

# 2 References

The following documents contain provisions which, through reference in this text, constitute provisions of the present document.

- References are either specific (identified by date of publication and/or edition number or version number) or non-specific.
- For a specific reference, subsequent revisions do not apply.
- For a non-specific reference, the latest version applies.

Referenced documents which are not found to be publicly available in the expected location might be found at <http://docbox.etsi.org/Reference>.

NOTE: While any hyperlinks included in this clause were valid at the time of publication ETSI cannot guarantee their long term validity.

- [1] Directive 1999/5/EC of the European Parliament and of the Council of 9 March 1999 on radio equipment and telecommunications terminal equipment and the mutual recognition of their conformity (R&TTE Directive).
- [2] CEPT/ERC/REC 70-03 (02-2007): "Relating to the use of Short Range Devices (SRD)".
- [3] ETSI TR 100 028 (all parts) (V1.4.1): "Electromagnetic compatibility and Radio spectrum Matters (ERM); Uncertainties in the measurement of mobile radio equipment characteristics".

- [4] ITU-T Recommendation O.153: "Basic parameters for the measurement of error performance at bit rates below the primary rate".
- [5] ITU-R Recommendation RS.1346: "Sharing between the meteorological aids service and medical implant communication systems (MICS) operating in the mobile service in the frequency band 401-406 MHz".
- [6] CISPR 16-2-3 (2003): "Specification for radio disturbance and immunity measuring apparatus and methods - Part 2-3: Methods of measurement of disturbances and immunity - Radiated disturbance measurements".
- [7] Radiofrequency Radiation Dosimetry Handbook (October 1986), USAF School of Aerospace Medicine, Aerospace Medical Division (AFSC), Brooks Air Force Base, TX 78235-5301.
- NOTE: See <http://niremf.ifac.cnr.it/docs/HANDBOOK/home.htm>.
- [8] ANSI C63.17 (1998): "American National Standard for Methods of Measurement of the Electromagnetic and Operational Compatibility of Unlicensed Personal Communications Services (PCS) Devices".
- NOTE: See [http://www.techstreet.com/cgi-bin/detail?product\\_id=43013](http://www.techstreet.com/cgi-bin/detail?product_id=43013).
- [9] CEPT/ERC/REC 74-01: "Unwanted Emissions in the Spurious Domain".

---

## 3 Definitions, symbols and abbreviations

### 3.1 Definitions

iTech STANDARD PREVIEW  
(standards.iteh.ai)

For the purposes of the present document, the following terms and definitions apply:

**access protocol:** specification for measuring natural and man-made ambient background levels for the purpose of providing a technique for spectrum access that reduces the potential for harmful interference to/from other users of the spectrum

**Active Implantable Medical Device (AIMD):** any active medical device (AMD) which is intended to be totally or partially introduced, surgically or medically, into the human body or by medical intervention into a natural orifice, and which is intended to remain after the procedure

**Adaptive Frequency Agility (AFA):** ability to determine an unoccupied sub-band or channel of operation in order to maximize spectrum utilization

**artificial antenna:** reduced-radiating dummy load equal to the nominal impedance specified by the applicant

**conducted measurements:** measurements that are made using a direct connection to the equipment under test

**dedicated antenna:** removable antenna supplied and tested with the radio equipment, designed as an indispensable part of the equipment

**duty cycle:** the ratio, expressed as a percentage, of the maximum transmitter "on" time monitored over one hour, relative to a one hour period

NOTE: See clause 8.6.1.

**effective radiated power:** power radiated within the emission bandwidth of the EUT in the direction of the maximum level under specified conditions of measurements in the presence of modulation or without modulation as appropriate

NOTE: See clause 8.3.1.

**emission bandwidth:** measured as the width of the signal between the points on either side of carrier centre frequency that are 20 dB down relative to the maximum level of the modulated carrier

NOTE: See clause 8.2.1.1 for details on how to determine compliance.

**frequency error:** difference between the nominal frequency as measured on the devices under test and under normal test conditions and the frequency under extreme conditions

NOTE: See also clause 8.1.1.

**frequency stability under low voltage condition:** ability of the equipment to remain on the nominal operating frequency when the battery voltage falls below the lower extreme voltage level

NOTE: See also clause 8.5.1.

**integral antenna:** permanent fixed antenna, which may be built-in, designed as an indispensable part of the equipment

**LBT threshold power level:** ambient signal power level above which the monitoring system shall select spectrum for use in a MICS communication session according to the next available channel with the lowest level of ambient signal power or least interfered channel (LIC)

**Least Interfered Channel (LIC):** channel, among the available channels that has the lowest potential for causing disturbance to or receiving disturbance from other users of the band

**listen mode:** action taken by an interrogator to detect an unoccupied sub-band or channel

**Listen Before Talk (LBT):** combination of the listen mode followed by the talk mode

**Medical Device (MD):** any instrument, apparatus, appliance, material or other article, whether used alone or in combination, together with any accessories or software for its proper functioning, intended by the manufacturer to be used for human beings in the:

- diagnosis, prevention, monitoring, treatment or alleviation of disease or injury;
- investigation, replacement or modification of the anatomy or of a physiological process;
- control of conception.

and which does not achieve its principal intended action by pharmacological, chemical, immunological or metabolic means, but which may be assisted in its function by such means

**medical implant communication channel:** any continuous segment of spectrum that is equal to the emission bandwidth of the device with the largest bandwidth that is to participate in a MICS session

NOTE 1: As stated in CEPT/ERC/REC 70-03 [2], annex 12 Band a, it is permitted to aggregate 25 kHz segments up to a maximum of 300 kHz for each channel bandwidth.

NOTE 2: Two types of devices for Medical Implant Communications Systems are covered by the present document:

- (i) Frequency agile devices designed to access a minimum of nine channels evenly distributed across the 402 MHz to 405 MHz band.
- (ii) Single frequency devices restricted to the 403,5 MHz to 403,8 MHz centre channel.

**Medical Implant Communication Link (MICS):** collection of transmissions that may or may not be continuous, between ULP-AMIs and ULP-AMI-Ps, including programmer/controllers, transferring patient related information in a communications service

**Medical Implant Communications System (MICS):** system specifically for the purpose of providing non-voice digital communications between one or several ULP-AMI and one ULP-AMI-P or between ULP-AMI

**Medical Implant Communications System (MICS) session:** collection of transmissions that may or may not be continuous, between co-operating ULP-AMI and ULP-AMI-P

**medical implant event:** occurrence or the lack of an occurrence recognized by a medical implant device or duly authorized health care professional that requires the immediate transmission of data from a medical implant transmitter in order to protect the safety of the person in whom the medical implant transmitter has been placed

NOTE: It is not permitted that this is the only mechanism a medical implant transmitter can use to access spectrum.

**monitoring system:** circuitry in an ULP-AMI and/or ULP-AMI-P that assures conformity with the spectrum access protocol requirements based on Listen before Talk, Adaptive Frequency Agility and selection of the least interfered channel for operation (LIC)

**radiated measurements:** measurements, which involve the absolute measurement of a radiated field

**spurious emissions:** emissions at frequencies other than those of the carrier and sidebands associated with normal test modulation

NOTE: See clause 8.4.1.

**spurious radiations from the receiver:** components at any frequency, generated and radiated by active receiver circuitry and the antenna

NOTE: See clause 9.1.1.

**talk mode:** transmission of intentional radiation by a transmitter

**telecommand:** use of radio communication for the transmission of signals to initiate, modify or terminate functions of equipment at a distance

**telemetry:** use of radio communication for indicating or recording data at a distance

**time-critical data:** data which if not transferred immediately will result in compromising the health and/or safety of the patient

**Ultra Low Power Active Medical Implant (ULP-AMI):** the radio part of an AIMD

**Ultra Low Power Active Medical Implant Peripheral (ULP-AMI-P) device:** the radio part of equipment outside the human body that communicates with an ULP-AMI to establish a MICL

PRELIMINARY STANDARD DRAFT  
(standards.iteh.ai)

## 3.2 Symbols

SIST EN 301 839-1 V1.2.1:2007

For the purposes of the present document, the following symbols apply:

B	bandwidth
dB	decibel
E	electrical field strength
E <sub>0</sub>	reference electrical field strength (see annex A)
f	frequency
f <sub>c</sub>	channel centre frequency
f <sub>e</sub>	frequency under extreme conditions
G	Antenna Gain
NaCl	sodium chloride
P	power
R	distance
R <sub>0</sub>	Reference distance (see annex A)
P <sub>Th</sub>	maximum threshold power level (see clause 10)
t	time
λ	wavelength

## 3.3 Abbreviations

For the purposes of the present document, the following abbreviations apply:

AFA	Adaptive Frequency Agility
AIMD	Active Implantable Medical Device
CW	Continuous Wave
e.r.p.	effective radiated power
EUT	Equipment Under Test
LBT	Listen Before Talk
LDC	Low Duty Cycle

LIC	Least Interfered Channel (see definitions)
LP	Low Power
MD	Medical Device
MICL	Medical Implant Communication Link
MICS	Medical Implant Communications System
RF	Radio Frequency
r.m.s.	root mean square
SRD	Short Range Device
ULP-AMI	Ultra Low Power Active Medical Implant
ULP-AMI-P	Ultra Low Power Active Medical Implant Peripheral
VSWR	Voltage Standing Wave Ratio

## 4 Technical requirements and specifications

### 4.1 General requirements

#### 4.1.1 Transmitter requirements

See clause 8 for transmitter requirements.

#### 4.1.2 Receiver requirements

See clause 9.

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

### 4.2 Presentation of equipment for testing purposes

Each equipment submitted for testing shall fulfil the requirements of the present document on all frequencies over which it is intended to operate. Compliance with this requirement should be shown by testing each unit on a frequency near 403,5 MHz.

The provider shall complete the appropriate application form when submitting the equipment for testing. In addition, the provider shall declare the range of operating conditions and power requirements, as applicable; to establish the appropriate test conditions.

Additionally, technical documentation and operating manuals, sufficient to make the test, shall be supplied for all ULP-AMI-P and ULP-AMI.

A human torso simulator and tissue substitute material for testing ULP-AMI shall be used (see clause 6.5).

Measurements shall be performed, according to the present document, on samples of equipment defined in clauses 4.2.1 to 4.2.3.3.

#### 4.2.1 Choice of model for testing

The provider shall supply one or more samples of each model or type of transmitter (ULP-AMI and/or ULP-AMI-P), as appropriate for testing. Any ancillary equipment needed for testing shall be provided as requested by the testing laboratory.

If an equipment has several optional features, considered not to affect the RF parameters, then the tests need only to be performed on the equipment configured with that combination of features considered to be the most complex or most likely to affect the RF parameters, as proposed by the provider and agreed to by the test laboratory.

#### 4.2.2 Testing of equipment with alternative power levels

Equipment designed to operate with different carrier powers shall have each transmitter parameter tested on samples of equipment defined in clause 4.2.1. See clause 8 for details on testing. Spurious emissions tests shall be performed in accordance with requirements in clause 8.4.1.1.

## 4.2.3 Presentation of equipment that does not have an external RF connector (integral antenna equipment)

### 4.2.3.1 Equipment with an internal permanent or temporary antenna connector

The means to access and/or implement the internal permanent or temporary connector shall be stated by the provider with the aid of a diagram. The fact that use has been made of the internal antenna connection, or of a temporary connection, to facilitate measurements shall be recorded in the test report.

### 4.2.3.2 Equipment with a temporary antenna connector

The provider may submit one set of equipment with the normal antenna connected, to enable the radiated measurements to be made. He shall attend the test laboratory at the conclusion of the radiated measurements, to disconnect the antenna and fit the temporary connector. The testing laboratory staff shall not connect or disconnect any temporary antenna connector.

Alternatively, the provider may submit two sets of equipment to the test laboratory, one fitted with a temporary antenna connector with the antenna disconnected and the other with the antenna connected. Each equipment shall be used for the appropriate tests. The provider shall declare that two sets of equipment are identical in all respects.

### 4.2.3.3 Equipment intended to be implanted in a human body

The provider shall submit the equipment, a human torso simulator as described in clause 6.5 and annex A, and a sufficient quantity of tissue substitute material to fill the test fixture. The provider and/or test laboratory shall determine and agree on the arrangement of the equipment antenna and any additional device leads on the AIMD holding grid within the fixture as prescribed in annex A.

## 4.3 Mechanical and electrical design

### 4.3.1 General

(standards.iteh.ai)  
SIST EN 301 839-1 V1.2.1:2007  
<https://standards.iteh.ai/catalog/standards/sist/3429f0c3-778a-4dbd-bb33-c8dd68b44dde/sist-en-301-839-1-v1-2-1-2007>

The equipment submitted by the provider should be designed, constructed and manufactured in accordance with sound engineering practice and with the aim of minimizing harmful interference to other equipment and services.

Transmitters and receivers may be individual or combination units.

### 4.3.2 Controls

Those controls that, if maladjusted, might increase the interference potentialities of the equipment shall not be easily accessible to the user.

### 4.3.3 Transmitter shut-off facility

Transmitters may be equipped with an automatic transmitter shut-off facility or battery-saving feature. In this case please refer to clause 8 for specific regarding testing requirements.

### 4.3.4 Marking

The equipment shall be marked in a visible place. This marking shall be legible and durable. If impossible the marking shall be in the users' manual, according to the provisions of the Directive 1999/5/EC (R&TTE Directive) [1].

### 4.3.5 Equipment identification

AIMD may also have a unique electronic identification that prevents unauthorized access to the telecommand and telemetry functions of the ULP-AMI.