



SLOVENSKI STANDARD

SIST EN 14546:2005

01-september-2005

Prehrana - Določanje slednih elementov - Določanje celotnega arsena s hidridno atomsko absorpcijsko spektrometrijo (HGAAS) po sušenju

Foodstuffs - Determination of trace elements - Determination of total arsenic by hydride generation atomic absorption spectrometry (HGAAS) after dry ashing

Lebensmittel - Bestimmung von Elementspuren - Bestimmung von Gesamtarsen mit Atomabsorptionsspektrometrie-Hydridtechnik (HGAAS) nach Trockenveraschung

Produits alimentaires - Détermination des éléments traces - Détermination de l'arsenic total par spectrométrie d'absorption atomique par génération d'hydrures (SAAGH) après calcination par voie sèche

Ta slovenski standard je istoveten z: EN 14546:2005

ICS:

67.050	Splošne preskusne in analizne metode za živilske proizvode	General methods of tests and analysis for food products
--------	--	---

SIST EN 14546:2005

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 14546:2005](https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005)

<https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005>

EUROPEAN STANDARD

EN 14546

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2005

ICS 67.050

English version

Foodstuffs - Determination of trace elements - Determination of total arsenic by hydride generation atomic absorption spectrometry (HGAAS) after dry ashing

Produits alimentaires - Détermination des éléments traces -
Détermination de l'arsenic total par spectrométrie
d'absorption atomique par génération d'hydrures (SAAGH)
après calcination par voie sèche

Lebensmittel - Bestimmung von Elementspuren -
Bestimmung von Gesamtarsen mit
Atomabsorptionsspektrometrie-Hydridtechnik (HGAAS)
nach Trockenveraschung

This European Standard was approved by CEN on 15 March 2005.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents	page
Foreword	3
1 Scope	4
2 Normative references	4
3 Principle.....	4
4 Reagents.....	4
5 Apparatus and equipment.....	6
6 Procedure	6
7 Calculation.....	8
8 Precision.....	8
9 Test report	9
Annex A (informative) Results of the inter-laboratory test.....	10
Bibliography.....	12

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 14546:2005](https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005)

<https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005>

Foreword

This document (EN 14546:2005) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2005, and conflicting national standards shall be withdrawn at the latest by October 2005.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 14546:2005](https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005)

<https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005>

EN 14546:2005 (E)

1 Scope

This European Standard specifies a method for the determination of total arsenic in foodstuffs by hydride generation atomic absorption spectrometry (HGAAS) after dry ashing.

Specific foodstuffs for which European Standards exist are excluded from the scope of this horizontal European Standard. It is the task of the analyst to review if vertical standards exist.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13804, *Foodstuffs – Determination of trace elements – Performance criteria, general considerations and sample preparation.*

3 Principle

The samples treated with nitric acid and ashing aid are evaporated to dryness and then mineralized at 425 °C with a gradual increase in temperature. The ash is dissolved in hydrochloric acid and arsenic is quantified by HGAAS-procedure at the arsenic line at 193,7 nm [1].

WARNING — The use of this standard may involve hazardous materials, operations and equipment. This standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate health and safety practices and determine the applicability of regulatory limitations prior to use.

4 Reagents

4.1 General

The concentration of arsenic in the reagents and water used shall be low enough not to affect results of the determination.

4.2 Hydrochloric acid

4.2.1 Hydrochloric acid, not less than 30 % of approximately ρ (HCl) = 1,15 g/ml.

4.2.2 Diluted hydrochloric acid, c (≈ 6 mol/l).

Dilute 500 ml hydrochloric acid (4.2.1) with water in a proportion of 1+1 parts by volume.

4.3 Nitric acid, not less than 65 % (mass fraction) of approximately $\rho(\text{HNO}_3) = 1,4$ g/ml.

4.3.1 Diluted nitric acid 1, $w = 32\%$ ¹⁾

To be prepared by mixing mix nitric acid (4.3) and water in a proportion of 1 + 1 parts by volume.

4.3.2 Diluted nitric acid 2, $w = 6,5\%$ ¹⁾

Mix nitric acid (4.3) and water in a proportion of 1 + 9 parts by volume.

4.3.3 Diluted nitric acid 3, $w = 3,3\%$ ¹⁾

Mix nitric acid (4.3) and water in a proportion of 1 + 19 parts by volume.

4.4 Magnesium oxide, $w =$ not less than 90 %¹⁾

4.5 Magnesium nitrate hexahydrate, $w =$ not less than 98 %¹⁾

4.6 Ashing aid

Suspend 20 g of magnesium nitrate hexahydrate (4.5) and 2 g of magnesium oxide (4.4) in 100 ml of water.

4.7 Sodium hydroxide, $w =$ not less than 95 %¹⁾

4.8 Potassium iodide, $w =$ not less than 99,5 %¹⁾

4.9 Ascorbic acid, $w =$ not less than 99,7 %¹⁾

4.10 Pre-reducing solution

<https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-7a41fab1306/sist-en-14546-2005>

Dissolve 5 g of potassium iodide (4.8) and 5 g of ascorbic acid (4.9) in 100 ml of water.

4.11 Arsenic stock solution, in nitric acid, with a mass concentration $\rho(\text{As (V)}) = 1\,000$ mg/l

The use of a stock solution accompanied by a certificate is advisable.

4.12 Arsenic calibration solutions

4.12.1 Intermediate standard solution

Prepare a standard of 25 $\mu\text{g/l}$ in HCl (4.2.2) from the stock solution (4.11)

4.12.2 Calibration standards

Take an aliquot of the intermediate standard (e.g. 3 ml for a final concentration of 3 $\mu\text{g/l}$) and make up to 5 ml with HCl (4.2.2). Add 5 ml of pre-reducing solution (4.10). Wait 30 min and make up to 25 ml with HCl (4.2.2). Following the same procedure prepare at least three calibration standards at different concentrations plus the zero point of the curve. The concentrations selected shall not exceed the linear range of the calibration curve of arsenic.

1) w is the mass fraction.

EN 14546:2005 (E)

4.13 Sodium borohydride, w = not less than 96%¹⁾

4.14 Sodium hydroxide, w = not less than 98%¹⁾

4.15 Sodium borohydride solution

The concentration of this reducing solution may vary according to the system of hydride generation employed and the recommendations given by the manufacturer in the manual for the equipment. Filter if necessary. This solution shall be prepared daily.

5 Apparatus and equipment**5.1 General**

To minimize the contamination, all apparatus which come into direct contact with the sample and the solutions shall be carefully pre-treated according to EN 13804.

5.2 Programmable muffle furnace

Equipped with a microprocessor or muffle furnace with thermostat.

5.3 Atomic absorption spectrometer

Equipped with an electrically heated quartz cell and an accessory for hydride generation with or without flow injection.

5.4 Lamp for arsenic

Hollow-cathode or electrodeless discharge lamp.
<https://standards.iteh.ai/catalog/standards/sist/84a10a0d-2c03-4795-ad7c-1306/sist-en-14546-2005>

5.5 Hot plate or sand bath

With stepwise heating control.

6 Procedure**6.1 Pre-treatment**

Homogenise the sample in accordance with EN 13804.

6.2 Test portion

Weigh an amount of 250 mg to 1000 mg dry sample or an equivalent quantity of fresh sample into a tall heat resistant glass beaker with a minimum capacity of 200 ml.

6.3 Dry ashing**6.3.1 Ashing procedure**

Add 2,5 ml of ashing aid (4.6) per 250 mg of dry sample. In order to avoid precipitation add the ashing aid while shaking the suspension. Then add 5 ml of nitric acid (4.3.1) and evaporate in a sand bath (5.5) to a dry

residue. Increase the temperature gradually and slowly to avoid boiling, the production of foam, and splashing or spattering.

Cover the glass beaker with a watch glass and place in the muffle furnace at an initial temperature not higher than 150 °C. Increase the temperature to (425 ± 25) °C at a maximum rate of 50 °C/h and maintain for 12 h. The rate of increase in temperature shall be maintained at the level recommended to avoid a loss of material. Take the crucible out of the furnace and let it cool.

Repeat the ashing procedure until the sample is completely incinerated. For this purpose add 5 ml of nitric acid (4.3.2), evaporate in the sand bath, and place again the ashes in the muffle furnace, i.e. the ashes should be white/grey or slightly coloured (the number of repetitions necessary varies depending on the type of sample).

6.3.2 Dissolution of the sample ash

Add 1 ml of water to wet the ash and then add 5 ml of hydrochloric acid (4.2.2) to the crucible, ensuring that all the ash comes into contact with the acid. Then add 5 ml of pre-reducing agent (4.10). Wait 30 min and make up to 25 ml with diluted hydrochloric acid (4.2.2).

If necessary, filter the samples before diluting to volume. The filter shall be decontaminated by an initial washing with nitric acid (4.3.3) and two subsequent washings with deionized water. After filtering the samples and in order to avoid losses, wash the filter with hydrochloric acid (4.2.2) and add the washing liquids to the solution before diluting to volume, as above.

Treat three reagent blanks in the same way as the samples.

The sample solutions are stable for 24 h at 4 °C.

6.4 Hydride generation atomic absorption spectrometry (HGAAS)

[SIST EN 14546:2005](#)

6.4.1 Spectrometer conditions

Optimize the instrument in accordance with the recommendations described in the manual provided with the instrument. The readings shall be made at 193,7 nm.

6.4.2 Hydride generation technique

6.4.2.1 General

Certain elements (e.g. Co, Cu, Cr, Fe, Mg, Ni, Sb, Sn and Se) may cause interference in hydride generation. This interference is limited, or eliminated, when the sample is diluted in diluted hydrochloric acid (4.2.2). If the sample requires further dilution, do so with diluted hydrochloric acid (4.2.2). Confirm that there are no matrix effects or inter-element interference by verifying that the differences between the slopes of the calibration curves and the standard addition curves are not significant. If the matrix effect persists, use the method of standard addition.

6.4.2.2 Standard addition method

Determine the linear range of the standard calibration function. It is important that the measurements are made in the linear range when the method of standard addition is used. A standard addition curve should consist of at least three points of which at least two are standard additions. The concentration of the highest standard should be 3 to 5 times the concentration in the sample solution. The concentration of the lower standard should be half of the highest standard.

Plot a graph of the absorbances obtained in this way against the added concentrations and extrapolate the resulting straight line until it intercepts the concentration axis.