



SLOVENSKI STANDARD

SIST EN 62244:2011

01-oktober-2011

Oprema za zaščito pred sevanjem - Vgrajeni monitorji sevanja za odkrivanje radioaktivnih in posebnih nuklearnih materialov na državnih mejah

Radiation protection instrumentation - Installed radiation monitors for the detection of radioactive and special nuclear materials at national borders

Strahlenschutz-Messgeräte - Fest installierte Strahlungsmonitore für den Nachweis von radioaktiven Stoffen und spaltbarem Nuklearmaterial an Staatsgrenzen

Instrumentation pour la radioprotection - Moniteurs de rayonnement installés pour la détection des matériaux nucléaires radioactifs et spéciaux aux frontières nationales

[https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-](https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011)

Ta slovenski standard je istoveten z: **EN 62244:2011**

ICS:

13.280 Varstvo pred sevanjem Radiation protection

SIST EN 62244:2011 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 62244:2011

<https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 62244

July 2011

ICS 13.280

English version

**Radiation protection instrumentation -
Installed radiation monitors for the detection of radioactive and special
nuclear materials at national borders
(IEC 62244:2006, modified)**

Instrumentation pour la radioprotection -
Moniteurs de rayonnement installés pour
la détection des matériaux nucléaires
radioactifs et spéciaux aux frontières
nationales
(CEI 62244:2006, modifiée)

Strahlenschutz-Messgeräte -
Fest installierte Strahlungsmonitore für
den Nachweis von radioaktiven Stoffen
und spaltbarem Nuklearmaterial an
Staatsgrenzen
(IEC 62244:2006, modifiziert)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

This European Standard was approved by CENELEC on 2011-06-27. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of the International Standard IEC 62244:2006, prepared by SC 45B, "Radiation protection instrumentation", of IEC TC 45, "Nuclear instrumentation", together with the common modifications prepared by the Technical Committee CENELEC TC 45B, Radiation protection instrumentation, was submitted to the CENELEC formal vote and was approved by CENELEC as EN 62244 on 2011-06-27.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

- | | | |
|--|-------|------------|
| – latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement | (dop) | 2012-06-27 |
| – latest date by which the national standards conflicting with the EN have to be withdrawn | (dow) | 2014-06-27 |

Annex ZA has been added by CENELEC.

Subclauses which are additional to those in IEC 62244 are prefixed "Z".

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 62244:2011

<https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011>

Endorsement notice

The text of the International Standard IEC 62244:2006 was approved by CENELEC as a European Standard with agreed common modifications as given below.

COMMON MODIFICATIONS

3 Terms and definitions

3.1 alarm assembly

Change text to “component that provides audio or visual signals, or both, when an alarm level is exceeded or a malfunction occurs”.

3.2 detection assembly

Change text to “component of the installed radiation monitor system that contains the detector(s) and associated electronic devices”.

3.3 detection zone

Insert “being monitored” between “person” and “may”.

3.6 reference point of the detector zone

Delete “that” at start of definition.

iteh STANDARD PREVIEW
(standards.iteh.ai)

Add a new definition

“3.Z1 special nuclear materials

[SIST EN 62244:2011](https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-275738802710/iec-62244-2006)

[https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-](https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-275738802710/iec-62244-2006)

[275738802710/iec-62244-2006](https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-275738802710/iec-62244-2006)

plutonium except that with isotopic concentration exceeding 80 % in plutonium-238; uranium-233; uranium enriched in the isotope 235 or 233; uranium containing the mixture of isotopes as occurring in nature other than in the form of ore or ore-residue; any material containing one or more of the foregoing.

(IAEA-TECDOC-1311)”

4 General Characteristics

4.1 Overview

Replace the last paragraph with “Figure 1 shows an example of the detection zone dimensions for a monitor that uses two detection assemblies. The following detection zone widths, W , are applicable to the above four types of monitors.

- pedestrian: Single detector assembly, 1,0 m.
Two detector assemblies, 1,0 m;
- road vehicle: not exceeding 5,0 m;
- rail vehicles: not exceeding 6,0 m;
- conveyor: See 4.5.”.

4.2 Pedestrian

Delete last sentence of paragraph 1.

2nd paragraph, delete 2nd sentence.

Replace last sentence with “For dynamic monitors, the range of passage speeds shall be specified by the manufacturer, this range shall include 1,2 m/s for pedestrian monitors.”

4.3 Road vehicles (includes road transported containers)

Replace last sentence with "For dynamic monitors, the range of passage speeds shall be specified by the manufacturer, this range shall include 8 km/h for road vehicle monitors."

4.4 Rail vehicles (includes rail transported containers)

Replace last sentence with "For dynamic monitors, the range of passage speeds shall be specified by the manufacturer, this range shall include 8 km/h for rail vehicle monitors."

4.5 Conveyor

Replace last sentence with "For dynamic monitors, the range of passage speeds shall be specified by the manufacturer, this range shall include 1 m/s for conveyor monitors."

4.6 Configuration

Replace 2nd paragraph with "A detection assembly may consist of a single detector or multiple detectors contained within a protective cabinet. A monitor may consist of a single detector assembly or multiple detector assemblies."

Start the last two paragraphs with "The monitor shall ...".

4.8 Speed control

Replace the section with "The performance of the monitor is dependent on the speed at which the object being monitored passes through the detection zone. Therefore the monitors, with the exception of the pedestrian and conveyor monitors, shall include a device for measuring the average speed of the object passing through the detection zone. An excessive speed alarm shall be provided that would indicate if the passage speed exceeds the upper speed of the range stated by the manufacturer. The speed alarm set point shall be adjustable up to a manufacturer-defined maximum speed."

[SIST EN 62244:2011](https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011)

5 General test procedures

<https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011>

5.6 Alarm probability for gamma and neutron radiation

Change in line two "detection probability" to "alarm probability"

5.7 Reference radiation**5.7.1 Gamma**

First NOTE: Change "NOTE" to "NOTE 1".

First line of new NOTE 1: Change "0,1 µSv/h" to "0,1 µGy/h".

Line 3 of third paragraph: After "if not encapsulated" insert "in 0,25 mm stainless steel".

Last NOTE: Change "NOTE" to "NOTE 2".

5.7.2 Neutron

Replace whole section with "²⁵²Cf shall be used for neutron tests. The source should have an activity that would produce an unscattered neutron fluence rate of 0,04 cm⁻² s⁻¹ at the closest point of the detector. This fluence rate is approximately equivalent to a neutron emission rate of 20 000 s⁻¹ at 2 m source to detection surface distance."

5.7.3 SNM

Replace heading with “**Special nuclear material (SNM)**”

Replace whole section with “Standardized SNM test sources given in Annex A may be used by agreement between manufacturer and user. Test should be performed using the acceptance requirements given in 6.3 or 6.4 (i.e. 49 successes out of 50 trials).”.

6 Radiation characteristics

Change title to “Radiation characteristics (applicable to all types of monitors)”.

6.1 False alarm test

6.1.1 Requirements

Replace whole section with “No more than one alarm shall occur over a period of 100 h with the monitor operating in a stable background with a level below the set alarm level.”.

6.1.2 Method of tests

Replace first sentence with “With the monitor operating normally in a stable background initiate a monitoring cycle as per the manufacturer’s specification at 5 min intervals for a period of 100 h.”.

6.2 Background effects

iTeh STANDARD PREVIEW
(standards.iteh.ai)

6.2.1 Requirement

Insert new sentence between first and second paragraph. “The instrument shall alarm when exposed at 1,5 times the manufacturer’s stated instrument’s maximum background level.”.

<https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011>

6.2.2 Method of tests

Number first paragraph a).

Replace 2nd paragraph with the following two paragraphs:

“b) Place the monitor for 100 s in a background level equal to the manufacturer’s stated maximum background plus 50 %. During this 100 s the monitor shall indicate “non operational”.

c) Place the monitor in a background level equal to the manufacturer’s stated maximum background. Then perform a gamma detection test as stated in step 6.3.2 for ¹³⁷Cs only. The monitor shall meet the requirements given in 6.3.1.”.

6.3 Gamma radiation detection

6.3.2 Method of test

Replace first sentence with “For a dynamic monitor, the source shall be moved through the detection zone at the passage velocity stated for that particular monitor type, 1,2 m/s for pedestrian monitors, 8 km/h for road and rail vehicle monitors and 1 m/s for conveyor monitors.”.

6.4 Neutron radiation detection

6.4.2 Method of test

Change NOTE to "This fluence rate is approximately equivalent to a neutron emission rate of $20\,000\text{ s}^{-1}$ at 2 m source to detection surface distance."

Add at the end of second paragraph ", 1,2 m/s for pedestrian monitors, 8 km/h for road and rail vehicle monitors and 1 m/s for conveyor monitors."

6.5 Overload characteristics

6.5.2 Method of test

Remove in line 2 "ambient".

Add at the end of this section "This shall be demonstrated using a ^{137}Cs source with the test method described in 6.3 but with a reduced number of trials."

7 Electrical characteristics

7.1 Mains operation

7.1.2 Method of test

Replace in the first line "defined in 6.3" with "defined in 5.7".

ITeH STANDARD PREVIEW
(standards.iteh.ai)

8 Electromagnetic compatibility

[SIST EN 62244:2011](https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011)

8.1 External magnetic fields

<https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011>

8.1.1 Requirements

Insert at the end of the paragraph "When exposed to external magnetic fields the count rate shall not vary by more than $\pm 15\%$ from the rate under standard test conditions."

8.1.2 Method of test

Add at the end of the last sentence "due to the magnetic field."

8.2 Radiated electromagnetic fields

8.2.2 Method of test

Add at the end of the penultimate sentence "with maximum of $\pm 15\%$ of count rate change"

Change in the last line "setting" to "status".

8.3 Conducted disturbances by bursts and radio frequencies

8.3.2 Method of test

Change in test c) "decades" to "order of magnitude".

Add in the last sentence at the end "due to the conducted disturbances".

9 Mechanical characteristics

9.3 Vibration test

9.3.2 Method of test

Add at the end of section a) "(IEC 60068-2-6)".

9.4 Microphonics/impact

9.4.1 Requirements

Insert between "unaffected" and "by" "(remain within $\pm 15\%$ of the pre-test values)".

9.4.2 Method of test

Delete second paragraph.

10 Environmental characteristics

10.1 Ambient temperature

10.1.2 Method of test

Add at the end of second paragraph "and in 10.1.1".

10.2 Relative humidity

10.2.2 Method of test

Add in line 5 between "Table 3" and "is" and in 10.2.1".

10.3 Sealing

10.3.2 Method of test

Change in line 2 "10 l/m" to "10 l/m²".

11 Documentation

11.2 Certificate

Add at the end another bullet "– maximum operational background, see 6.2.1".

Table 1 – Reference conditions and standard test conditions

Add a new row below "Reference Gamma radiation source": "Reference Neutron radiation source" and add in both new columns 2 and 3 "²⁵²Cf".

Table 3 – Tests performed with variations of influence quantities

Add in row 3 (External magnetic fields) in column 3 at the end "by more than $\pm 15\%$ of the rate under standard test conditions".

Insert in row 8 (Mechanical characteristics) in column 2, "Shock" before first paragraph and "Vibration" before second paragraph.

Replace in the last row the content of column 2 with "Rain of 10 l/m²".

Annex ZA
(normative)
Normative references to international publications
with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60038	-	IEC standard voltages	EN 60038 ¹	-
IEC 60050-393	-	International Electrotechnology Vocabulary - Part 393: Nuclear instrumentation - Physical phenomena and basic concepts	-	-
IEC 60050-394	1995	International Electrotechnical Vocabulary - Chapter 394: Nuclear instrumentation:	-	-
+ A1	1996	Instruments	-	-
+ A2	2000		-	-
IEC 60068-2-18	-	Environmental testing - Part 2-18: Tests - Tests R and guidance: Water	EN 60068-2-18	-
IEC 60068-2-27	-	Environmental testing - Part 2-27: Tests - Test Ea and guidance: Shock	EN 60068-2-27	-
IEC 60068-2-75	-	Environmental testing - Part 2-75: Tests - Test Eh: Hammer tests	EN 60068-2-75	-
IEC 60359	-	Electrical and electronic measurement equipment - Expression of performance	EN 60359	-
IEC 61000-4-2	-	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2	-
IEC 61000-4-3	-	Electromagnetic compatibility (EMC) - Part 4-3: Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test	EN 61000-4-3	-
IEC 61000-4-4	-	Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques - Electrical fast transient/burst immunity test	EN 61000-4-4	-
IEC 61000-4-5	-	Electromagnetic compatibility (EMC) - Part 4-5: Testing and measurement techniques - Surge immunity test	EN 61000-4-5	-
IEC 61000-4-6	-	Electromagnetic compatibility (EMC) - Part 4-6: Testing and measurement techniques - Immunity to conducted disturbances, induced by radio-frequency fields	EN 61000-4-6	-
IEC 61000-4-12	-	Electromagnetic compatibility (EMC) - Part 4-12: Testing and measurement techniques - Ring wave immunity test	EN 61000-4-12	-
IEC 61187	-	Electrical and electronic measuring equipment - Documentation	EN 61187	-

¹ At draft stage.

NORME
INTERNATIONALE
INTERNATIONAL
STANDARD

CEI
IEC

62244

Première édition
First edition
2006-06

**Instrumentation pour la radioprotection –
Moniteurs de rayonnement installés pour la
détection des matériaux nucléaires radioactifs
et spéciaux aux frontières nationales**

iTeh STANDARD PREVIEW

**Radiation protection instrumentation –
Installed radiation monitors for the detection
of radioactive and special nuclear materials
at national borders**

<https://standards.iteh.ai/catalog/standards/sist/75580a3c-97fd-4cbb-95cd-c27f5338e8a3/sist-en-62244-2011>

© IEC 2006 Droits de reproduction réservés — Copyright - all rights reserved

Aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'éditeur.

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembe, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

CODE PRIX
PRICE CODE

U

*Pour prix, voir catalogue en vigueur
For price, see current catalogue*

CONTENTS

FOREWORD.....	7
1 Scope and object.....	11
2 Normative references	11
3 Terms and definitions	13
4 General characteristics.....	17
4.1 Overview	17
4.2 Pedestrian.....	17
4.3 Road vehicles (includes road transported containers).....	19
4.4 Rail vehicles (includes rail transported containers).....	19
4.5 Conveyor.....	19
4.6 Configuration.....	19
4.7 Indication features.....	21
4.8 Speed control.....	21
5 General test procedures	21
5.1 Nature of tests.....	21
5.2 Reference conditions and standard test conditions	23
5.3 Tests performed under standard test conditions	23
5.4 Tests performed with variation of influence quantities.....	23
5.5 Statistical fluctuations	23
5.6 Alarm probability for gamma and neutron radiation.....	23
5.7 Reference radiation.....	23
6 Radiation characteristics	25
6.1 False alarm test	25
6.2 Background effects.....	25
6.3 Gamma radiation detection.....	27
6.4 Neutron radiation detection	27
6.5 Overload characteristics	29
7 Electrical characteristics.....	29
7.1 Mains operation.....	29
7.2 Occupancy sensor.....	31
8 Electromagnetic compatibility	31
8.1 External magnetic fields	31
8.2 Radiated electromagnetic fields.....	31
8.3 Conducted disturbances induced by bursts and radio frequencies	33
8.4 Surges and oscillatory waves	33
8.5 Electrostatic discharge	35
9 Mechanical characteristics.....	37
9.1 Areas of application.....	37
9.2 Mechanical shocks	37
9.3 Vibration test.....	37
9.4 Microphonics/impact.....	39