



SLOVENSKI STANDARD
SIST EN 62467-1:2015
01-december-2015

Medicinska električna oprema - Dozimetrični instrumenti, ki se uporabljajo pri brahiterapiji - 1. del: Instrumenti na osnovi jaškastih ionizacijskih komor

Medical electrical equipment - Dosimetric instruments as used in brachytherapy - Part 1: Instruments based on well-type ionization chambers

iTeh STANDARD PREVIEW

Appareils électromédicaux - Instruments de dosimétrie utilisés en curiethérapie - Partie 1: Instruments conçus pour les chambres d'ionisation à puits

[SIST EN 62467-1:2015](https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-6b4cc91807f9/sist-en-62467-1-2015)

Ta slovenski standard je istoveten z: EN 62467-1:2015

ICS:

| | | |
|-----------|------------------------------|------------------------------|
| 11.040.01 | Medicinska oprema na splošno | Medical equipment in general |
|-----------|------------------------------|------------------------------|

SIST EN 62467-1:2015 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 62467-1:2015](https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015)

<https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015>

EUROPEAN STANDARD

EN 62467-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2015

ICS 11.040.50; 11.040.60

English Version

Medical electrical equipment - Dosimetric instruments as used in
brachytherapy - Part 1: Instruments based on well-type
ionization chambers
(IEC 62467-1:2009)

Appareils électromédicaux - Instruments de dosimétrie
utilisés en curiethérapie - Partie 1: Instruments conçus pour
les chambres d'ionisation à puits
(IEC 62467-1:2009)

Medizinische elektrische Geräte - Dosimetrieeräte zur
Anwendung in der Brachytherapie - Teil 1: Messgeräte mit
Schachtionisationskammern
(IEC 62467-1:2009)

This European Standard was approved by CENELEC on 2015-09-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 62467-1:2015**European foreword**

The text of document 62C/460/FDIS, future edition 1 of IEC 62467-1, prepared by SC 62C "Equipment for radiotherapy, nuclear medicine and radiation dosimetry", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 62467-1:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-06-15
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-09-15

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document.

<https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015>

Endorsement notice

The text of the International Standard IEC 62467-1:2009 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

| | | |
|--------------------|------|---|
| IEC 60601-1-3:2008 | NOTE | Harmonized as EN 60601-1-3:2008 (not modified). |
| IEC 61010-1 | NOTE | Harmonized as EN 61010-1. |
| IEC 61676:2002 | NOTE | Harmonized as EN 61676:2002 (not modified). |

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

| <u>Publication</u> | <u>Year</u> | <u>Title</u> | <u>EN/HD</u> | <u>Year</u> |
|--------------------|-------------|---|--------------------|-------------|
| IEC 60050-393 | 2003 | International Electrotechnical Vocabulary - Part 393: Nuclear instrumentation - Physical phenomena and basic concepts | - | - |
| IEC 60417 | - | Graphical symbols for use on equipment | - | - |
| IEC 60580 | 2000 | Medical electrical equipment - Dose area product meters | EN 60580 | 2000 |
| IEC 60601-1 | 2005 | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance | EN 60601-1 | 2006 |
| - | - | | + corrigendum Mar. | 2010 |
| - | - | | + A12 | 2014 |
| IEC 60731 | 1997 | Medical electrical equipment - Dosimeters with ionization chambers as used in radiotherapy | - | - |
| IEC/TR 60788 | 2004 | Medical electrical equipment - Glossary of defined terms | - | - |
| IEC 61187 | - | Electrical and electronic measuring equipment - Documentation | EN 61187 | - |
| IEC 61674 | 1997 | Medical electrical equipment - Dosimeters with ionization chambers and/or semiconductor detectors as used in X-ray diagnostic imaging | EN 61674 | 1997 |
| ISO/IEC Guide 99 | 2007 | International vocabulary of metrology - Basic and general concepts and associated terms (VIM) | - | - |

EN 62467-1:2015

Annex ZZ
(informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 62467-1:2015

<https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015>



IEC 62467-1

Edition 1.0 2009-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment – Dosimetric instruments as used in
brachytherapy –
Part 1: Instruments based on well-type ionization chambers**

**Appareils électromédicaux – Instruments de dosimétrie utilisés en
curiethérapie –
Partie 1: Instruments conçus pour les chambres d'ionisation à puits**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

T

ICS 11.040.50; 11.040.60

ISBN 978-2-88910-735-3

CONTENTS

| | |
|--|----|
| FOREWORD..... | 4 |
| INTRODUCTION..... | 6 |
| 1 Scope and object..... | 7 |
| 2 Normative references..... | 7 |
| 3 Terms and definitions..... | 8 |
| 4 General requirements..... | 12 |
| 4.1 PERFORMANCE REQUIREMENTS..... | 12 |
| 4.2 MEASURING ASSEMBLY..... | 12 |
| 4.3 Source types..... | 12 |
| 4.3.1 General..... | 12 |
| 4.3.2 Beta particle-emitting sources..... | 13 |
| 4.3.3 Low-energy-photon-emitting sources..... | 13 |
| 4.4 Quantity to be measured..... | 13 |
| 4.5 Reference and STANDARD TEST CONDITIONS..... | 13 |
| 4.6 General test conditions..... | 13 |
| 4.6.1 STANDARD TEST CONDITIONS..... | 13 |
| 4.6.2 STABILIZATION TIME..... | 13 |
| 4.6.3 Adjustments during test..... | 14 |
| 4.6.4 Batteries..... | 14 |
| 4.7 Constructional requirements as related to performance..... | 14 |
| 4.7.1 General..... | 14 |
| 4.7.2 Components..... | 14 |
| 4.7.3 Display..... | 14 |
| 4.7.4 Inserts..... | 14 |
| 4.7.5 STABILIZATION TIME..... | 14 |
| 4.8 Test of components..... | 15 |
| 5 Limits of performance characteristics..... | 15 |
| 5.1 Position of source in insert and repeatability..... | 15 |
| 5.2 USABLE LENGTH..... | 15 |
| 5.3 RESOLUTION OF THE DISPLAY..... | 15 |
| 5.4 STABILIZATION TIME..... | 15 |
| 5.5 LEAKAGE CURRENT..... | 16 |
| 5.5.1 In AIR KERMA STRENGTH measuring mode..... | 16 |
| 5.5.2 In charge measuring mode..... | 16 |
| 5.6 Stability..... | 16 |
| 5.6.1 Long term stability..... | 16 |
| 5.6.2 MANUFACTURER method to check long term stability..... | 16 |
| 6 LIMITS OF VARIATION for effects of influence quantities..... | 16 |
| 6.1 General..... | 16 |
| 6.2 IONIZATION CHAMBER – recombination losses..... | 17 |
| 6.3 Operating voltage..... | 17 |
| 6.3.1 Mains operated MEASURING ASSEMBLY..... | 17 |
| 6.3.2 Battery operated MEASURING ASSEMBLY..... | 17 |
| 6.3.3 Rechargeable MEASURING ASSEMBLY..... | 18 |
| 6.4 Air pressure..... | 18 |
| 6.5 Change of air pressure and EQUILIBRATION TIME of the radiation detector..... | 18 |

| | | |
|-------|---|----|
| 6.5.1 | VENTED WELL TYPE IONIZATION CHAMBERS | 18 |
| 6.5.2 | SEALED WELL TYPE IONIZATION CHAMBERS..... | 19 |
| 6.6 | Temperature and humidity | 19 |
| 6.7 | Length RESPONSE | 19 |
| 6.8 | Electromagnetic immunity..... | 20 |
| 7 | Marking | 20 |
| 7.1 | WELL-TYPE IONIZATION CHAMBER ASSEMBLY | 20 |
| 7.2 | MEASURING ASSEMBLY | 20 |
| 8 | ACCOMPANYING DOCUMENTS | 20 |
| 8.1 | General | 20 |
| 8.2 | Use of the instrument | 20 |
| 8.3 | Documentation | 21 |
| | Bibliography..... | 22 |
| | Index of defined terms | 23 |
| | Table 1 – REFERENCE and STANDARD TEST CONDITIONS | 13 |
| | Table 2 – LIMITS OF VARIATION for the effects of INFLUENCE QUANTITIES | 17 |

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 62467-1:2015

<https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –
DOSIMETRIC INSTRUMENTS AS USED IN BRACHYTHERAPY –**

Part 1: Instruments based on well-type ionization chambers

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62467-1 has been prepared by subcommittee 62C, Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62, Electrical equipment in medical practice.

The text of this standard is based on the following documents:

| | |
|--------------|------------------|
| FDIS | Report on voting |
| 62C/460/FDIS | 62C/468/RVD |

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 62467 series, published under the general title *Medical electrical equipment – Dosimetric instruments as used in brachytherapy*, can be found on the IEC website.

In this standard the following print types are used: Requirements, compliance with which can be tested, and definitions: in roman type;

- notes, explanations, advice, general statements and exceptions: in small roman type;
- *test specifications: in italic type;*
- TERMS USED THROUGHOUT THIS STANDARD THAT HAVE BEEN DEFINED IN CLAUSE 3 OR IN THE PUBLICATIONS INDICATED IN THE INDEX OF DEFINED TERMS: IN SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 62467-1:2015

<https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015>

INTRODUCTION

The wide range of WELL-TYPE IONIZATION CHAMBER instruments currently being used for BRACHYTHERAPY sources indicates the need for a standard for uniformity in measurement and test techniques for WELL-TYPE IONIZATION CHAMBER instruments. Measurements of the output of BRACHYTHERAPY sources have distinct requirements that differ from the assay of sources used in diagnostic nuclear medicine. This translates into the requirements for the measurement devices. Many times similar instrumentation is used for both applications; however, there are tighter requirements for those instruments used for BRACHYTHERAPY sources. Such devices are composite systems consisting of an IONIZATION CHAMBER, either integrally coupled or connected to appropriate electronic circuitry that converts the ionization current to a readout, which can be converted to a quantity appropriate to the source being measured. The ionization current produced can be either read directly or as accumulated charge (current integrated over time) and then converted manually to the appropriate quantity, AIR KERMA STRENGTH (REFERENCE AIR KERMA RATE) or ABSORBED DOSE TO WATER. The principles of operation of the IONIZATION CHAMBER are well known and are not repeated here. In addition, the readout device many times also has application to therapy uses and is well known. Although this standard is written using the quantity AIR KERMA STRENGTH, the principles are the same for other quantities such as REFERENCE AIR KERMA RATE.

In principle the quantity measured is the dose volume integral from which under specified conditions the dose quantities AIR KERMA STRENGTH, REFERENCE AIR KERMA RATE, or ABSORBED DOSE TO WATER at a depth can be deduced. The signal produced by the chamber is the electrical current or charge, which is to be measured with an electrometer meeting criteria according to IEC 60731. The current or charge is converted to the dosimetric quantity of interest by means of a source type specific CALIBRATION FACTOR.

(standards.iteh.ai)

SIST EN 62467-1:2015

<https://standards.iteh.ai/catalog/standards/sist/71668fed-523c-4af4-b1fb-bb4ec91807f9/sist-en-62467-1-2015>