

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

Electroacoustics – Octave-band and fractional-octave-band filters –  
Part 3: Periodic tests

[standards.iteh.ai](https://standards.iteh.ai)

Électroacoustique – Filtres de bande d'octave et de bande d'une fraction  
d'octave –

Partie 3: Essais périodiques

<https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-6bf896797e3a/iec-61260-3-2016>



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2016 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

#### IEC Catalogue - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

#### IEC publications search - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [csc@iec.ch](mailto:csc@iec.ch).

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Catalogue IEC - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

#### Recherche de publications IEC - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 15 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [csc@iec.ch](mailto:csc@iec.ch).

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

---

**Electroacoustics – Octave-band and fractional-octave-band filters –  
Part 3: Periodic tests** (standards.iteh.ai)

**Électroacoustique – Filtres de bande d'octave et de bande d'une fraction  
d'octave –** <https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-6bf896797e3a/iec-61260-3-2016>  
**Partie 3: Essais périodiques**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

---

ICS 17.140.50

ISBN 978-2-8322-3246-0

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions .....	7
4 Submission for testing .....	7
5 Conformance.....	7
6 Preliminary inspection.....	8
7 Power supply.....	8
8 Environmental conditions .....	8
9 Mandatory facilities and general requirements.....	8
9.1 General.....	8
9.2 Test instruments .....	10
10 Test of relative attenuation at midband frequency or effective bandwidth deviation.....	10
10.1 General.....	10
10.2 Tests of relative attenuation at midband frequency.....	10
10.3 Test of effective bandwidth deviation .....	10
11 Linear operating range, measurement range, level range control and overload indicator .....	11
12 Test of lower limit of linear operating range.....	12
13 Measurement of relative attenuation.....	12
14 Documentation .....	13
Annex A (informative) Uncertainty related to test by sinusoidal sweeps .....	16
A.1 General.....	16
A.2 Digitally generated signal .....	16
A.3 Test signal from a signal generator.....	17
A.4 Comparing measurements.....	18
Annex B (informative) Test of effective bandwidth deviation with the use of an exponential sweep – Example .....	19
B.1 General.....	19
B.2 Example.....	19
Annex C (informative) Normalized frequencies for test of one-third-octave-band filters .....	21
C.1 General.....	21
C.2 Example calculation .....	21
Bibliography .....	23
Table 1 – Frequency parameter $R$ and acceptance limits on relative attenuation for fractional-octave-band filters .....	13
Table C.1 – Normalized test frequencies and acceptance limits on relative attenuation for one-third-octave-band filters .....	22

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ELECTROACOUSTICS – OCTAVE-BAND  
AND FRACTIONAL-OCTAVE-BAND FILTERS –****Part 3: Periodic tests****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.  
<https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9->
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61260-3 has been prepared by IEC technical committee 29: Electroacoustics.

This first edition of IEC 61260-3 (together with IEC 61260-1:2014 and IEC 61260-2:2016), cancels and replaces the first edition of IEC 61260 published in 1995 and its Amendment 1 published in 2001. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to IEC 61260.

- a) The single document in the first edition of IEC 61260:1995 is now separated into three parts of IEC 61260 covering: specifications, pattern evaluation tests and periodic tests;
- b) IEC 61260:1995 specified three performance categories: class 0, 1 and 2 while the IEC 61260 series specifies requirements for class 1 and 2;
- c) In IEC 61260:1995, the design goals for the specification can be based on base-2 or base-10 design. In the IEC 61260 series only base-10 is specified;

- d) The reference environmental conditions have been changed from 20 °C/65 % RH to 23 °C/50 % RH;
- e) IEC 61260:1995 specified tolerance limits without considering the uncertainty of measurement for verification of the specifications while the IEC 61260 series specifies acceptance limits for the observed values and maximum-permitted uncertainty of measurements for laboratories testing conformance to specifications in the standard.

The text of this standard is based on the following documents:

CDV	Report on voting
29/846/CDV	29/882A/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all the parts of the IEC 61260 series, published under the general title *Electroacoustics – Octave-band and fractional-octave-band filters* can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or [IEC 61260-3:2016](https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-6bf896797e3a/iec-61260-3-2016)
- amended. <https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-6bf896797e3a/iec-61260-3-2016>

(standards.iteh.ai)

## INTRODUCTION

IEC 61260:1995 and IEC 61260:1995/AMD 1:2001 are now separated into the following three parts of IEC 61260 series:

- Part 1: Specifications
- Part 2: Pattern evaluation tests
- Part 3: Periodic tests

For assessments of conformance to performance specifications, IEC 61260-1 uses different criteria than were used for the IEC 61260:1995 edition.

IEC 61260:1995 did not provide any requirements or recommendations to account for the uncertainty of measurement in assessments of conformance to specifications. This absence of requirements or recommendations to account for uncertainty of measurement created ambiguity in determinations of conformance to specifications for situations where a measured deviation from a design goal was close to the limit of the allowed deviation. If conformance was determined based on whether a measured deviation did or did not exceed the limits, the end-user of the octave-band and fractional-octave-band filters incurred the risk that the true deviation from a design goal exceeded the limits.

To remove this ambiguity, IEC Technical Committee 29, at its meeting in 1996, adopted a policy to account for measurement uncertainty in assessments of conformance in International Standards that it prepares.

This edition of IEC 61260-3 uses an amended criterion for assessing conformance to a specification. Conformance is demonstrated when (a) measured deviations from design goals do not exceed the applicable *acceptance limits* and (b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty. Acceptance limits are analogous to the tolerance limits allowances for design and manufacturing implied in the IEC 61260:1995.

Actual and maximum-permitted uncertainties of measurement are determined for a coverage probability of 95 %. Unless more specific information is available, the evaluation of the contribution of a specific filter or filter set to a total measurement uncertainty can be based on the acceptance limits and maximum-permitted uncertainties specified in this standard.

# ELECTROACOUSTICS – OCTAVE-BAND AND FRACTIONAL-OCTAVE-BAND FILTERS –

## Part 3: Periodic tests

### 1 Scope

**1.1** This part of IEC 61260 describes procedures for periodic testing of octave-band and fractional-octave-band filters that were designed to conform to the class 1 or class 2 specifications given in IEC 61260-1:2014. The aim of this standard is to ensure that periodic testing is performed in a consistent manner by all laboratories.

**1.2** The purpose of periodic testing is to assure the user that the performance of an octave-band and fractional-octave-band filter conforms to the applicable specifications of IEC 61260-1 for a limited set of key tests and for the environmental conditions under which the tests were performed.

**1.3** The extent of the tests in this standard is deliberately restricted to the minimum considered necessary for periodic tests.

**1.4** Periodic tests described in this standard apply to filters for which the manufacturer claims conformance to the specifications in IEC 61260-1:2014. Periodic tests in this standard apply to filters for which the model has been, or has not been, pattern approved by an independent testing organization responsible for pattern approvals in accordance with the test procedures of IEC 61260-2.

<https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-61896707e36e/iec-61260-3:2016>

**1.5** Because of the limited extent of the periodic tests, if evidence of pattern approval is not publicly available, no general conclusion about conformance to the specifications of IEC 61260-1 can be made, even if the results of the periodic tests conform to all applicable requirements of this standard.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61260-1:2014, *Electroacoustics – Octave-band and fractional-octave-band filters – Part 1: Specifications*

IEC 61260-2:2016, *Electroacoustics – Octave-band and fractional-octave-band filters – Part 2: Pattern-evaluation tests*

IEC 61672-1, *Electroacoustics – Sound level meters – Part 1: Specifications*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 98-4, *Uncertainty of measurement – Part 4: Role of measurement uncertainty in conformity assessment*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 61260-1, ISO/IEC Guide 98-3 and ISO/IEC Guide 98-4 apply.

### 4 Submission for testing

**4.1** An instruction manual applicable to the model and version of the filter shall be available in order to perform periodic tests of a filter. If an applicable instruction manual is not submitted along with the filter, nor available at the calibration laboratory, nor publicly accessible from the Internet web site of the manufacturer or supplier of the filter, then no periodic tests shall be performed.

**4.2** The source for the instruction manual shall be described in the documentation for the periodic tests.

**4.3** All items or accessories for the filter that are necessary for periodic testing shall accompany the filter when it is submitted for testing.

**4.4** Periodic tests as described in this standard shall not be performed unless the markings on the filter are as required by IEC 61260-1 or there is evidence that the filter was originally so marked. At least the serial number and the model designation shall be visible on the filter or instrument containing the filter.

**4.5** Data required to perform the periodic tests shall be available, and the source of the data shall be recorded and reported by the laboratory. The data shall include all relevant information required by IEC 61260-1.

**IEC STANDARD PREVIEW**  
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-6bf896797e3a/iec-61260-3-2016>

### 5 Conformance

**5.1** Conformance to a performance specification is demonstrated when the following criteria are both satisfied:

- a) a measured deviation from a design goal does not exceed the applicable acceptance limit and;
- b) the corresponding uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty of measurement given in IEC 61260-1 for the same coverage probability of 95 %.

IEC 61260-1:2014 gives example assessments of conformance using these criteria.

**5.2** Laboratories performing periodic tests shall calculate all uncertainties of measurements in accordance with the guidelines of ISO/IEC Guide 98-3. Actual measurement uncertainties shall be calculated for a coverage probability of 95 %. Calculation of the actual measurement uncertainty for a particular test should consider at least the following components, as applicable:

- the uncertainty attributed to calibration of the individual instruments and equipment used to perform the test;
- the uncertainty resulting from environmental effects or adjustments;
- the uncertainty resulting from errors that may be present in the applied signals;
- the uncertainty attributed to effects associated with the repeatability of the results of the measurements. When a laboratory is only required to perform a single measurement, it is necessary for the laboratory to make an estimate of the contribution of random effects to the total uncertainty. The estimate should be determined from an evaluation of several measurement results previously obtained for a similar filter and parameter;

- the uncertainty associated with the resolution of the applied display device. For digital display devices that indicate signal levels with a resolution of 0,1 dB, the uncertainty component should be taken as a rectangular distribution with semi-range of 0,05 dB; and
- the uncertainty associated with each correction applied to the measurement data.

**5.3** If an actual uncertainty of measurement for a test performed by the laboratory exceeds the corresponding maximum-permitted uncertainty, the result of the test shall not be used to evaluate conformance to this standard for periodic testing.

## 6 Preliminary inspection

Prior to any measurements, the filter and all accessories shall be visually inspected to ensure that the filter is in normal working order. All relevant controls shall be operated to ensure that they are in working order. If the controls or other essential elements are not in proper working order, no periodic tests shall be performed.

## 7 Power supply

For all tests, the filter shall be powered from its preferred supply or a suitable alternative. Before and after conducting the set of tests, the power supply for the filter shall be checked by the method stated in the instruction manual to ensure that it is within the specified operating limits. If the voltage or the equivalent indication of the status of the power supply is not within the operating limits and the reason cannot be attributed to partially discharged batteries or an incorrect selection of the voltage of the public power supply, then no periodic tests shall be performed as a malfunction is indicated.

## 8 Environmental conditions IEC 61260-3:2016

[https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-](https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-61f967053a1c/iec-61260-3-2016)

**8.1** Periodic tests shall be performed within the following ranges of environmental conditions: 20 °C to 26 °C for air temperature, and 25 % to 70 % for relative humidity.

**8.2** As a minimum, the air temperature and the relative humidity shall be measured and recorded at the start and end of periodic testing. The recorded values expanded with the actual expanded uncertainty of measurement shall not exceed the limits in 8.1.

## 9 Mandatory facilities and general requirements

### 9.1 General

**9.1.1** No test specified in this standard shall be omitted unless the bandpass filter does not possess the feature described for the test.

**9.1.2** If the filter does not possess the mandatory features listed in IEC 61260-1, including an overload indicator and means to check that the power supply is adequate for battery powered instruments which contain the filter, the filter does not conform to the specifications of IEC 61260-1, and no periodic tests shall be performed.

**9.1.3** If, for an instrument containing filters with more than one bandwidth, the supplier claims conformance to IEC 61260-1, each bandwidth for which the supplier claims conformance shall demonstrate conformance to the specifications in this standard, otherwise the instrument does not conform to the requirement in this standard.

**9.1.4** For all periodic tests, the configuration of the filter shall be as specified in the instruction manual for one of the normal modes of operation, including required accessories. The input and output terminals shall be terminated with the impedances specified by the supplier, if appropriate.

**9.1.5** For filters enclosed in a sound level meter with detachable preamplifier, the signal input to the filter may be, as specified by the supplier, the input of the preamplifier through a suitable input device replacing the microphone, or the terminal where the signal from the preamplifier normally is connected.

**9.1.6** For filters with digital readout devices, or with an output that is available in a manufacturer-specified digital format (for example over a digital interface connection), the level of the output should be determined from the numeric readout or via the digital output to a suitable display or recording device.

**9.1.7** If the filter is enclosed in an instrument containing a level detector and a display device for displaying the level of the filtered signal with a resolution of at least 0,1 dB, the displayed value from this display device shall be used for testing. If an electrical output is provided corresponding to the displayed value and the testing laboratory intends to utilize the electrical output instead of the display device, the laboratory shall verify that changes in the levels of applied electrical input signals produce corresponding changes in the signal levels indicated on the display device and at the electrical output that are in accordance with the specifications of IEC 61260-1. Where multiple outputs are present, if an output is specified in the instruction manual for testing, this output should be used for the periodic tests.

**9.1.8** For filters that are designed to operate with measuring devices that comply with the requirements for sound level meters as specified in IEC 61672-1, the display indicator of this device shall be used to measure the level of the output signal from the filter.

**9.1.9** If the instruction manual specifies a procedure for adjusting the filter, e.g. sensitivity adjustment, this procedure shall be followed before any measurements are performed.

**9.1.10** The filter shall be allowed to reach equilibrium with the prevailing environmental conditions before switching on the power to perform a test.

<https://standards.iteh.ai/catalog/standards/sist/46677819-075f-4a95-96a9-b0896797c341-61260-3-2016>

**9.1.11** As appropriate, the laboratory shall utilize the recommendations given in the instruction manual for performing the periodic tests.

**9.1.12** If the filter has more than one signal-processing channel, periodic tests shall be performed for each channel that utilizes unique signal processing techniques. For multi-channel systems with the same functional equivalence in all channels, the number of channels to be tested may be less than the total number of channels, at the discretion of the testing laboratory.

**9.1.13** The number of channels tested may also be limited by the customer. In this case, the test report, and possibly any calibration marks on the instrument, shall clearly indicate that the reported results are only valid for the channels tested.

**9.1.14** During periodic tests, as described in this standard, extensive tests are only performed on a limited number of filters in a set of filters. Should any test not meet the requirements of this standard, the set of filters does not conform to the specifications of IEC 61260-1.

**9.1.15** If, during testing, the test laboratory uncovers evidence that the set of filters do not comply with the requirements of IEC 61260-1 in general, even if all tests carried out as specified in this standard do meet the requirements of IEC 61260-1, the filter set shall be reported as having passed all tests specified in this standard, but with a note detailing where the filter set does not conform to IEC 61260-1 and making it clear the filter set overall does not comply with IEC 61260-1.

## 9.2 Test instruments

**9.2.1** The laboratory shall use instruments that have been calibrated for the appropriate quantities at appropriate intervals. As required, the calibrations shall be traceable to national measurement standards.

**9.2.2** Most of the required tests utilize steady sinusoidal signals of various frequencies and signal levels. Sinusoidal signals for testing filter attenuation shall have a total distortion including noise of not more than 0,01 % for class 1 filters and not more than 0,03 % for class 2 filters. The total distortion for other sinusoidal signals shall not exceed 0,3 %.

**9.2.3** Measurement of the effective bandwidth deviation may use a constant amplitude sinusoidal signal the frequency of which is varied, or swept, at an exponential rate. The uncertainty of the measured deviation due to the uncertainty in the amplitude and sweep-rate shall be calculated. The expanded uncertainty shall not exceed the limits given in IEC 61260-1 for test of time invariant operation.

NOTE See Annex B for an example of measuring effective bandwidth using an exponential swept sine. See Annex A for information regarding the computation of uncertainty due to using a swept sine as input.

**9.2.4** Instruments for measuring the environmental conditions during the tests shall have an uncertainty appropriate to ensure that the requirements in 8.1 are maintained.

## 10 Test of relative attenuation at midband frequency or effective bandwidth deviation

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

### 10.1 General

**10.1.1** The purpose of the tests in Clause 10 is to demonstrate that every filter in a set of filters is in working order since more extensive tests are only performed on a limited number of filters in a set of filters.

**10.1.2** The test may either be performed as the measurement of the relative attenuation at the midband frequency of every filter in a set, or alternatively, by measurement of the response to an exponential, sinusoidal sweep covering all filters in a set. For time invariant filters, the response to an exponential sweep corresponds to a measurement of the effective bandwidth deviation. The tests are described in 10.2 and 10.3 respectively. For time invariant filters, testing may be made using either the tests in 10.2 or 10.3, at the choice of the testing laboratory. For filters not being time invariant, only tests in 10.2 apply.

### 10.2 Tests of relative attenuation at midband frequency

**10.2.1** The relative attenuation at the exact midband frequency shall be measured for every filter in a set of filters. The relative attenuation  $\Delta A(\Omega)$  at any midband frequency is determined from Formula (8) given in IEC 61260-1:2014. The reference level range shall be selected for the test. The level of the test signal shall be equal to the reference input signal level.

**10.2.2** The measured relative attenuation shall not exceed the acceptance limits  $\pm 0,4$  dB for Class 1 filters or  $\pm 0,6$  dB for class 2 filters as specified in 5.10 in IEC 61260-1:2014.

### 10.3 Test of effective bandwidth deviation

**10.3.1** The effective bandwidth deviation of each filter in a set of filters shall be measured by a swept-frequency test as described in 5.14 in IEC 61260-1:2014 for the test for time-invariant operation. The test shall be conducted on the reference level range. The level of the input signals shall be  $(3 \pm 0,1)$  dB less than the upper boundary of the linear operating range on the reference level range.

**10.3.2** The sweep shall start at the frequency,  $f_{\text{start}}$ , being lower than the lowest bandedge frequency for the filter with the lowest midband frequency in the filter set and where the relative attenuation for this filter is at least 55 dB. The sweep shall end at a frequency,  $f_{\text{end}}$ , higher than the highest bandedge frequency for the filter with the highest midband frequency in the filter set, and where the relative attenuation is at least 55 dB.

**10.3.3** The sweep shall have constant amplitude, and the frequency of the signal shall be increased at a constant exponential rate as described in Annex G of IEC 61260-1:2014. The sweep rate shall correspond to one decade in frequency in not less than 2 s.

**10.3.4** The time-averaged level of the output signal is measured for an averaging time,  $T_{\text{avg}}$ , which starts no later than the time when the sweep frequency is less than the lowest midband frequency and where the relative attenuation of a filter is at least 55 dB, and ends at a time not less than when the sweep frequency is greater than the highest midband frequency where the relative attenuation of the filter is again at least 55 dB.

The averaging time shall be sufficiently long to also contain parts of the output signal delayed by the operation of the filter. See Annex G in IEC 61260-1:2014 for more information. Annex B gives an example of how the test may be performed.

**10.3.5** The measured time-average or equivalent-continuous output signal level for each filter in the set shall be compared with the calculated value,  $L_c$ , given in Formula (17) in IEC 61260-1:2014.

**10.3.6** For each filter in a filter set, the acceptance limits for the deviation of a measured time-averaged output signal level,  $L_{\text{out}}$ , from the corresponding constant theoretical time-averaged output signal level,  $L_c$ , as determined according to Formula (17) in IEC 61260-1:2014, are  $\pm 0,4$  dB for class 1 filters and  $\pm 0,6$  dB for class 2 filters.

**10.3.7** Both the amplitude and the sweep-rate shall be considered when the uncertainty of measurement is calculated.

NOTE See Annex A for further information about uncertainties related to tests by sinusoidal sweeps.

## 11 Linear operating range, measurement range, level range control and overload indicator

**11.1** Linearity of the response of a filter resulting from changes in the level of the signal at the input shall be tested with steady sinusoidal signals with specified level and frequency. The linearity shall be measured at the exact midband frequency. The level linearity deviation shall be determined in accordance with 5.13 in IEC 61260-1:2014.

**11.2** For an input signal at midband frequency and reference input signal level, the level linearity deviation is zero on the reference level range.

**11.3** The level linearity shall be tested for three filters in a set of filters. The filters for the test shall be selected by the laboratory performing the test, if not required otherwise. The filters selected shall represent filters in the lower, in the middle and in the higher range of midband frequencies for the set of filters. For a set of filters covering the audible range of frequencies, it is recommended to test filters with frequencies close to 31,5 Hz, 1 kHz and 16 kHz.

**11.4** The level range control shall be set to select the reference level range. The level of the input signal shall first be set to the specified reference input signal level. The corresponding output level shall be used for calculating the level linearity deviation for all input levels at any level range for the particular filter.

**11.5** The test shall be performed on the reference level range for levels from the specified lower boundary of the specified linear operating range up to a level where the overload indicator displays an overload. Adjust the level of the input signal with steps that are not greater than 5 dB. The difference between successive steps of the input signal level shall be reduced to 1 dB when the distance to the lower or upper boundaries of a linear operating range is less than 5 dB and when the level is above the upper boundary. The boundaries are as stated in the instruction manual for the filter. If no overload is displayed, the filter does not conform to the requirements.

**11.6** The averaging time during a measurement shall be long enough to establish a stable indication considering the actual frequency and the influence of internally generated noise at low input signal levels.

**11.7** The measured level linearity deviation shall not exceed the acceptance limits given in 5.13 in IEC 61260-1:2014 for all measured levels between the lower boundary of the linear operating range, as stated in the instruction manual for the filter, and up to the highest level, measured as described above, without an overload indication.

**11.8** An overload shall not be indicated if the level of the input signal is below the stated upper boundary of each appropriate linear operating range.

**11.9** For the same three filters as selected above, test each available level range in the following way: based on the same reference level, adjust the input level to be 30 dB below upper boundary of the linear operating range for each of the selected range settings. The measured level linearity deviation shall not exceed the acceptance limits given in 5.13.3 and 5.13.4 of IEC 61260-1:2014.

(standards.iteh.ai)

## 12 Test of lower limit of linear operating range

**12.1** The test in Clause 12 is an abbreviated test to verify that the self-generated noise in the filter is lower than the lower limit of the linear operating range. The test shall be performed on the reference level range and on the level range with the highest sensitivity.

**12.2** Short-circuit the input terminal or use similar means to ensure that the level of the input signal is below the lower limit of the specified linear operating range. Record the output level from each filter in the set. The output level shall not exceed the specified lower limit for the appropriate filter and range.

## 13 Measurement of relative attenuation

**13.1** The relative attenuation on the reference level range shall be tested for the same three filters as selected in Clause 11.

**13.2** The measurements of relative attenuation are made as the response to constant amplitude sinusoidal signals at various frequencies. The level of the input signals shall be  $(1 \pm 0,1)$  dB below the specified upper boundary of the linear operating range.

**13.3** The normalized frequency  $\Omega_k = f_k/f_m$ , of the sinusoidal test signal for each filter with midband frequency,  $f_m$ , shall be calculated by the following formula:

$$\Omega_k = 1 + \frac{G^{1/(2b)} - 1}{G^{1/2} - 1} (R_k - 1) \quad (1)$$

where

$G$  is the octave frequency ratio;

$b$  is the inverse of the bandwidth designator;  
 $R_k$  is a frequency parameter defined in Table 1;  
 $k$  is a whole number in the range 0, 1,...7

The list of normalized frequencies for test shall be extended by:

$$\Omega_{-k} = 1/\Omega_k \quad (2)$$

Where  $\Omega_k$  and  $\Omega_{-k}$  have the same acceptance limits on relative attenuation.

NOTE 1 The specifications in this clause are an abbreviation of the general requirements in 5.10 and Table 1 of IEC 61260-1:2014.

NOTE 2 For octave-band filters,  $\Omega_k = R_k$ .

NOTE 3 Annex C shows an example calculation for one-third-octave-band filters.

**Table 1 – Frequency parameter  $R$  and acceptance limits on relative attenuation for fractional-octave-band filters**

Index $k$	Frequency parameter $R_k$	Minimum; maximum acceptance limits on relative attenuation dB	
		Class 1	Class 2
0	$G^0 = 1$	-0,4; +0,4	-0,6; +0,6
1	$G^{1/8}$	-0,4; +0,5	-0,6; +0,7
2	$G^{1/4}$	-0,4; +0,7	-0,6; +0,9
3	$G^{3/8}$	-0,4; +1,4	-0,6; +1,7
4	$G$	+16,6; +∞	+15,6; +∞
5	$G^2$	+40,5; +∞	+39,5; +∞
6	$G^3$	+60; +∞	+54; +∞
7	$G^4$	+70; +∞	+60; +∞

**13.4** Each of the filters selected for test of relative attenuation shall be tested with the normalized frequency as specified in 13.3 for  $k$  in the range -7, -6...7 as long as the frequencies applied are above 0,5 times the exact midband frequency of the filter in the set with the lowest midband frequency, and below 1,5 times the midband frequency of the filter in the set with the highest midband frequency for all filters in the filter set.

**13.5** Deviation between actual and requested frequency shall be considered when stating the uncertainty for testing of relative attenuation. The expanded uncertainty of measurement shall not exceed the corresponding maximum-permitted uncertainty of measurement given in Annex B of IEC 61260-1:2014.

**13.6** The measured relative attenuation shall not exceed the acceptance limits given in Table 1 for the appropriate class of filter.

## 14 Documentation

The documentation of the periodic test shall contain at least the following information, as applicable, unless national regulations require otherwise:

a) the date(s) when the periodic tests were performed;