

SLOVENSKI STANDARD

SIST EN 60601-2-1:2015

01-december-2015

Nadomešča:

SIST EN 60601-2-1:2002

SIST EN 60601-2-1:2002/A1:2003

Medicinska električna oprema - 2-1. del: Posebne zahteve za osnovno varnost in bistvene lastnosti elektronskih pospeševalnikov v območju od 1 MeV do 50 MeV

Medical electrical equipment -- Part 2-1: Particular requirements for basic safety and essential performance of electron accelerators in the range 1 MeV to 50 MeV

iTeh STANDARD PREVIEW

Medizinische elektrische Geräte - Teil 2-1: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Elektronenbeschleunigern im Bereich von 1 MeV bis 50 MeV

[SIST EN 60601-2-1:2015](https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-cccb51721b6f/sist-en-60601-2-1-2015)

[https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-](https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-cccb51721b6f/sist-en-60601-2-1-2015)

Appareils électromédicaux -- Partie 2-1: Règles particulières pour la sécurité de base et les performances essentielles des accélérateurs d'électrons dans la gamme de 1 MeV à 50 MeV

Ta slovenski standard je istoveten z: EN 60601-2-1:2015

ICS:

11.040.50 Radiografska oprema Radiographic equipment

SIST EN 60601-2-1:2015 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 60601-2-1:2015](https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-ccb51721b6f/sist-en-60601-2-1-2015)

<https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-ccb51721b6f/sist-en-60601-2-1-2015>

EUROPEAN STANDARD

EN 60601-2-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2015

ICS 11.040.60

Supersedes EN 60601-2-1:1998

English Version

**Medical electrical equipment - Part 2-1: Particular requirements
for the basic safety and essential performance of electron
accelerators in the range 1 MeV to 50 MeV
(IEC 60601-2-1:2009 + A1:2014)**

Appareils électromédicaux - Partie 2-1: Exigences
particulières de sécurité de base et de performances
essentielle pour les accélérateurs d'électrons dans la
gamme de 1 MeV à 50 MeV
(IEC 60601-2-1:2009 + A1:2014)

Medizinische elektrische Geräte - Teil 2-1: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von
Elektronenbeschleunigern
im Bereich von 1 MeV bis 50 MeV
(IEC 60601-2-1:2009 + A1:2014)

This European Standard was approved by CENELEC on 2015-09-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 60601-2-1:2015**European foreword**

The text of document 62C/474/FDIS, future edition 3 of IEC 60601-2-1, and the text of document 62C/532/CDV, future IEC 60601-2-1/A1, prepared by SC 62C "Equipment for radiotherapy, nuclear medicine and radiation dosimetry", of IEC/TC 62 "Electrical equipment in medical practice" were submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-1:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-06-15
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-09-15

This document supersedes EN 60601-2-1:1998.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document.

Endorsement notice

The text of the International Standards IEC 60601-2-1:2009 and IEC 60601-2-1:2009/A1:2014 were approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60976:2007	NOTE	Harmonized as EN 60976:2007 (not modified).
IEC 62366	NOTE	Harmonized as EN 62366.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Addition to Annex ZA of EN 60601-1:2006:				
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 61217	1996	Radiotherapy equipment - Coordinates, movements and scales	EN 61217	1996

EN 60601-2-1:2015

Annex ZZ
(informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 60601-2-1:2015](https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-ccb51721b6f/sist-en-60601-2-1-2015)

<https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-ccb51721b6f/sist-en-60601-2-1-2015>



FINAL VERSION

VERSION FINALE

Medical electrical equipment –
Part 2-1: Particular requirements for the basic safety and essential performance
of electron accelerators in the range 1 MeV to 50 MeV

Appareils électromédicaux –
Partie 2-1: Exigences particulières de sécurité de base et de performances
essentielle pour les accélérateurs d'électrons dans la gamme de 1 MeV à
50 MeV

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	11
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	15
201.7 ME EQUIPMENT identification, marking and documents	15
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	21
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	22
201.10 Protection against unwanted and excessive radiation HAZARDS	25
201.11 Protection against excessive temperatures and other HAZARDS	47
201.12 Accuracy of controls and instruments and protection against hazardous outputs	47
201.13 HAZARDOUS SITUATIONS and fault conditions	48
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	48
201.15 Construction of ME EQUIPMENT.....	49
201.16 ME SYSTEMS	49
201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS	49
206 Usability.....	50
Annexes	60
Annex B (informative) Sequence of testing	60
Annex I (informative) ME SYSTEMS aspects.....	60
Bibliography.....	61
Index of defined terms	62
Figure 201.101 – Limits of STRAY X-RADIATION during ELECTRON IRRADIATION (201.10.1.2.102.1)	52
Figure 201.102 – Limits of RELATIVE SURFACE DOSE during X-IRRADIATION (201.10.1.2.102.2)	53
Figure 201.103 – Elevation view – Application of LEAKAGE RADIATION requirements (201.10.1.2.103 and 201.10.1.2.104)	54
Figure 201.104 – 24 measurement points for averaging LEAKAGE RADIATION during X-RADIATION (201.10.1.2.103.2.1)	55
Figure 201.105 – Limits of LEAKAGE RADIATION through the BEAM LIMITING DEVICES during ELECTRON IRRADIATION (201.10.1.2.103.2.2).....	56
Figure 201.106 – Measurement points for averaging LEAKAGE RADIATION during ELECTRON IRRADIATION (201.10.1.2.103.2.2).....	57
Figure 201.107 – 24 measurement points for averaging LEAKAGE RADIATION outside area <i>M</i> (201.10.1.2.103.3).....	58
Figure 201.108 – ME EQUIPMENT movements and scales	59

IEC 60601-2-1:2009

– 3 –

+AMD1:2014 CSV © IEC 2014

Table 201.101 – Colours of indicator lights and their meaning for ME EQUIPMENT.....	16
Table 201.102 – Data required in the technical description to support Clause 201.10 SITE TEST compliance	18
Table 201.103 – Clauses and subclauses in this particular standard that require the provision of information in the ACCOMPANYING DOCUMENTS, INSTRUCTIONS FOR USE and the technical description	20
Table 201.104 – Limits of STRAY X-RADIATION during ELECTRON IRRADIATION (see Figure 201.101)	40
Table 201.105 – Limits of RELATIVE SURFACE DOSE during X-IRRADIATION (see Figure 201.102)	40

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

[SIST EN 60601-2-1:2015](https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-cccb51721b6f/sist-en-60601-2-1-2015)

<https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-cccb51721b6f/sist-en-60601-2-1-2015>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-1: Particular requirements for the basic safety
and essential performance of electron accelerators
in the range 1 MeV to 50 MeV**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This Consolidated version of IEC 60601-2-1 bears the edition number 3.1. It consists of the third edition (2009-10) [documents 62C/474/FDIS and 62C/480/RVD] and its amendment 1 (2014-07) [documents 62C/532/CDV and 62C/562/RVC]. The technical content is identical to the base edition and its amendment.

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

This publication has been prepared for user convenience.

IEC 60601-2-1:2009
+AMD1:2014 CSV © IEC 2014

– 5 –

International Standard IEC 60601-2-1 has been prepared by IEC subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition addresses the following issues not covered in previous editions:

- alignment with the new relevant collateral standards;
- new technologies in radiotherapy, including:
 - stereotactic radiosurgery (SRS) and stereotactic radiotherapy (SRT);
 - intensity modulated radiotherapy (IMRT);
 - electronic imaging devices (e.g. EPID);
 - moving beam radiotherapy (dynamic therapy).

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

(standards.iteh.ai)

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment* can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

[SIST EN 60601-2-1:2015](https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-cccb51721b6f/sist-en-60601-2-1-2015)

<https://standards.iteh.ai/catalog/standards/sist/8bc2411d-4c1b-4922-8c39-cccb51721b6f/sist-en-60601-2-1-2015>

INTRODUCTION

The use of ELECTRON ACCELERATORS for RADIOTHERAPY purposes may expose PATIENTS to danger if the ME EQUIPMENT fails to deliver the required dose to the PATIENT, or if the ME EQUIPMENT design does not satisfy standards of electrical and mechanical safety. The ME EQUIPMENT may also cause danger to persons in the vicinity if the ME EQUIPMENT itself fails to contain the RADIATION adequately and/or if there are inadequacies in the design of the TREATMENT ROOM.

This particular standard establishes requirements to be complied with by MANUFACTURERS in the design and construction of ELECTRON ACCELERATORS for use in RADIOTHERAPY; it does not attempt to define their optimum performance requirements. Its purpose is to identify those features of design that are regarded, at the present time, as essential for the safe operation of such ME EQUIPMENT. It places limits on the degradation of ME EQUIPMENT performance beyond which it can be presumed that a fault condition exists and where an INTERLOCK then operates to prevent continued operation of the ME EQUIPMENT.

Clause 201.10 contains limits beyond which INTERLOCKS prevent, INTERRUPT or TERMINATE IRRADIATION in order to insure ESSENTIAL PERFORMANCE is maintained to avoid an unsafe condition. TYPE TESTS that are performed by the MANUFACTURER, and/or SITE TESTS, which are not necessarily performed by the MANUFACTURER, are SPECIFIED for each requirement. It is understood that SITE TESTS may or may not be required of the MANUFACTURER, per the agreement between the MANUFACTURER and end user.

Given that before installation a MANUFACTURER cannot provide SITE TEST data, data available from SITE TESTS should be incorporated in the ACCOMPANYING DOCUMENTS, in the form of a SITE TEST report, by those who test the ME EQUIPMENT at installation.

This International Standard was first published in 1981. It was amended in 1984 and 1990. A second edition was published in 1998 and amended in 2002. This third edition is prompted by the need to align this particular standard with the third edition of the general standard, IEC 60601-1:2005.

IEC 60976 and IEC/TR 60977 are closely related to this standard. The former specifies test methods and reporting formats for performance tests of ELECTRON ACCELERATORS for use in RADIOTHERAPY, with the aim of providing uniform methods for conducting such tests. The latter is not a standard per se, but suggests performance values, measured per the methods specified in IEC 60976, that could be achievable with present technology.

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-1: Particular requirements for the basic safety
and essential performance of electron accelerators
in the range 1 MeV to 50 MeV****201.1 Scope, object and related standards**

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTRON ACCELERATORS, hereafter referred to as ME EQUIPMENT, in the range 1 MeV to 50 MeV, used for treatment of PATIENTS.

This particular standard, with the inclusion of TYPE TESTS and SITE TESTS, applies respectively to the manufacture and some installation aspects of ELECTRON ACCELERATORS

- intended for RADIOTHERAPY in human medical practice, including those in which the selection and DISPLAY of operating parameters can be controlled automatically by PROGRAMMABLE ELECTRONIC SUBSYSTEMS (PESS),
- that, under NORMAL CONDITIONS and in NORMAL USE, deliver a RADIATION BEAM of X-RADIATION and/or ELECTRON RADIATION having
 - NOMINAL ENERGY in the range 1 MeV to 50 MeV,
 - maximum ABSORBED DOSE RATES between $0,001 \text{ Gy} \times \text{s}^{-1}$ and $1 \text{ Gy} \times \text{s}^{-1}$ at 1 m from the RADIATION SOURCE,
 - NORMAL TREATMENT DISTANCES (NTDs) between 0,5 m and 2 m from the RADIATION SOURCE,

and

- intended to be
 - for NORMAL USE, operated under the authority of appropriately licensed or QUALIFIED PERSONS by OPERATORS having the required skills for a particular medical application, for particular specified clinical purposes, e.g. STATIONARY RADIOTHERAPY or MOVING BEAM RADIOTHERAPY,
 - maintained in accordance with the recommendations given in the INSTRUCTIONS FOR USE,
 - subject to regular quality assurance performance and calibration checks by a QUALIFIED PERSON.

NOTE 1 In this particular standard, all references to installation refer to installation in the RESPONSIBLE ORGANIZATION'S premises.

NOTE 2 In this particular standard, all references to ABSORBED DOSE refer to ABSORBED DOSE in water.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

IEC 60601-2-1:2009

– 9 –

+AMD1:2014 CSV © IEC 2014

IEC 61271 gives guidance on the designation of ME EQUIPMENT movements; the marking of scales, their zero positions and the direction of movement with increasing value (see 201.7.4.101).

IEC 60676 specifies methods of testing and disclosure of functional performance of medical ELECTRON ACCELERATORS. The standard is intended to facilitate comparisons of accelerator-based ME EQUIPMENTS of different manufacture. IEC 60676 contains no safety requirements, and is therefore not required for compliance with this particular standard. It should also be noted (as stated in the Introduction to IEC 60976:2007) that tests specified in IEC 60976 are not necessarily appropriate for ensuring that any individual medical ELECTRON ACCELERATOR conforms to the declared functional performance during the course of its working lifetime.

NOTE 3 IEC/TR 60977, *Medical electrical equipment – Medical electron accelerators – Guidelines for functional performance characteristics*, is a related technical report that provides performance guidelines. It shall not be construed as a standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ELECTRON ACCELERATORS in the range 1 MeV to 50 MeV and to specify tests to check compliance to those requirements.

NOTE The adoption of this standard helps to ensure that the ME EQUIPMENT

- maintains PATIENT safety during ME EQUIPMENT movements and failure of the SUPPLY MAINS,
- delivers the pre-selected RADIATION TYPE, NOMINAL ENERGY, and ABSORBED DOSE,
- delivers the RADIATION in accordance with the pre-selected relationship of the RADIATION BEAM to the PATIENT, by utilizing STATIONARY RADIOTHERAPY, MOVING BEAM RADIOTHERAPY, RADIATION BEAM modifying devices, etc., without causing unnecessary risk to the PATIENT, the OPERATOR, other persons or the environment.

201.1.3 Collateral standards

Addition:

Collateral standards published after the date of publication of this standard shall only apply subject to further amendment to this standard.

This particular standard refers to those applicable collateral standards that are listed in clause 2 of the general standard and clause 201.2 of this particular standard.

IEC 60601-1-6 apply as modified in Clauses 206. IEC 60601-1-3, IEC 60601-1-8 and 60601-1-10² do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

² IEC 60601-1-10, Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers