



SLOVENSKI STANDARD SIST EN 60601-2-18:2015

01-december-2015

Nadomešča:

SIST EN 60601-2-18:1998

SIST EN 60601-2-18:1998/A1:2002

Medicinska električna oprema - 2-18. del: Posebne zahteve za osnovno varnost in bistvene lastnosti endoskopske opreme

Medical electrical equipment -- Part 2-18: Particular requirements for basic safety and essential performance of endoscopic equipment

iTeh STANDARD PREVIEW

Medizinische elektrische Geräte - Teil 2-18: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von endoskopischen Geräten

[SIST EN 60601-2-18:2015](https://standards.iteh.ai/catalog/standards/sist/0a23b829-01b3-4176-9f5d-9380ad3580e/sist-en-60601-2-18-2015)

Appareils électromédicaux -- Partie 2-18: Exigences particulières pour la sécurité de base et les performances essentielles des appareils d'endoscopie

Ta slovenski standard je istoveten z: EN 60601-2-18:2015

ICS:

11.040.50 Radiografska oprema Radiographic equipment

SIST EN 60601-2-18:2015 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 60601-2-18:2015](#)

<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015>

EUROPEAN STANDARD

EN 60601-2-18

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2015

ICS 11.040.50

Supersedes EN 60601-2-18:1996

English Version

**Medical electrical equipment - Part 2-18: Particular requirements
for the basic safety and essential performance of endoscopic
equipment
(IEC 60601-2-18:2009)**

Appareils électromédicaux - Partie 2-18: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils d'endoscopie
(IEC 60601-2-18:2009)

Medizinische elektrische Geräte - Teil 2-18: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von endoskopischen
Geräten
(IEC 60601-2-18:2009)

This European Standard was approved by CENELEC on 2015-09-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 60601-2-18:2015**European foreword**

The text of document 62D/682/CDV, future edition 3 of IEC 60601-2-18, prepared by SC 62D "Electromedical equipment", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-18:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-06-15
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-09-15

This document supersedes EN 60601-2-18:1996.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

iTeh STANDARD PREVIEW

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document.

<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015>

Endorsement notice

The text of the International Standard IEC 60601-2-18:2009 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated :

IEC 60601-2-57 NOTE Harmonized as EN 60601-2-57.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Replacement in Annex ZA of EN 60601-1:2006:				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corrigendum Mar.	2007 2010
-	-			
Addition to Annex ZA of EN 60601-1:2006:				
IEC 60601-2-2	2009	Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	EN 60601-2-2 + A11	2009 2011
-	-			
IEC 60601-2-37	-	Medical electrical equipment - Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment	EN 60601-2-37	-
ISO 8600-1	-	Optics and photonics - Medical endoscopes and endotherapy devices - Part 1: General requirements	-	-

EN 60601-2-18:2015

Annex ZZ
(informative)**Coverage of Essential Requirements of EU Directives**

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 60601-2-18:2015](https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015)

<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015>



IEC 60601-2-18

Edition 3.0 2009-08

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Medical electrical equipment –
Part 2-18: Particular requirements for the basic safety and essential performance of endoscopic equipment

Appareils électromédicaux –
Partie 2-18: Exigences particulières pour la sécurité de base et les performances essentielles des appareils d'endoscopie

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX



ICS 11.040.50

ISBN 978-2-88910-216-7

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
201.1 Scope, object and related standards	8
201.2 Normative references	9
201.3 Terms and definitions.....	10
201.4 General requirements.....	12
201.5 General requirements for testing of ME EQUIPMENT.....	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	14
201.7 ME EQUIPMENT identification, marking and documents.....	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	18
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	19
201.10 Protection against unwanted and excessive radiation HAZARDS.....	20
201.11 Protection against excessive temperatures and other HAZARDS.....	21
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	24
201.13 HAZARDOUS SITUATIONS and fault conditions.....	25
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	25
201.15 Construction of ME EQUIPMENT.....	25
201.16 ME SYSTEMS	26
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	26
202 Electromagnetic compatibility – Requirements and tests	26
Annexes	27
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	27
Annex D (informative) Symbols on marking.....	29
Annex J (informative) Survey of insulation paths.....	31
Annex AA (informative) Particular guidance and rationale.....	33
Annex BB (informative) Clauses of this standard addressing essential principles of safety and performance of medical devices (GHTF/SG1/N41R9:2005).....	43
Index of defined terms used in this particular standard.....	45
Figure 201.101 – Identification of LIGHT EMISSION PART	12
Figure 201.102 – Measurement of CAPACITIVELY-COUPLED HF CURRENT from conductive parts of an ENDOSCOPE.....	24
Figure 201.J.101 – Insulation example 101	31
Figure 201.J.102 – Insulation example 102	32
Figure 201.J.103 – Insulation example 103	32
Figure AA.101 – Illustration of typical CONFIGURATION FOR ENDOSCOPIC APPLICATION	34
Table 201.101 – List of ESSENTIAL PERFORMANCE requirements	13

Table 201.C.101 – Marking on the outside of ENDOSCOPIC EQUIPMENT or its parts	27
Table 201.C.102 – Marking on the inside of ENDOSCOPIC EQUIPMENT or its parts	27
Table 201.C.104 – ACCOMPANYING DOCUMENTS, general	28
Table 201.C.105 – ACCOMPANYING DOCUMENTS, instructions for use	28
Table 201.D.101 – Symbols for marking ENDOSCOPIC EQUIPMENT or its parts.....	29
Table BB.1 – Correspondence between this standard and GHTF/SG1/N41R9:2005	43

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 60601-2-18:2015](https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015)

<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-18: Particular requirements for the basic safety
and essential performance of endoscopic equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d->
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-18 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62, Electrical equipment in medical practice.

This third edition cancels and replaces the second edition, published in 1996, and its Amendment 1 (2000). This edition constitutes a technical revision and has been aligned or harmonized with IEC 60601-1:2005.

The main changes with respect to the previous edition include:

- a) alignment of requirements with IEC 60601-1:2005;
- b) inclusion of essential performance requirements;
- c) the inclusion of energized endoscopes and energized endotherapy devices used through second and subsequent punctures within the scope of the standard;
- d) reference to IEC 60601-2-2 for the dielectric strength testing of HF energized endotherapy devices, rather than defining different tests.

The text of this particular standard is based on the following documents:

Enquiry draft	Report on voting
62D/682/CDV	62D/743/RVC

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 60601-2-18:2015](https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015)

<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015>

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of endoscopic equipment.

This particular standard amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as 'the general standard'.

The requirements are followed by specifications for the relevant tests.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 60601-2-18:2015](https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015)

<https://standards.iteh.ai/catalog/standards/sist/0a23b829-0acb-4176-9f5d-f9380dd3580e/sist-en-60601-2-18-2015>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-18: Particular requirements for the basic safety and essential performance of endoscopic equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ENDOSCOPIC EQUIPMENT together with its INTERCONNECTION CONDITIONS and INTERFACE CONDITIONS.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ENDOSCOPIC EQUIPMENT [as defined in 201.3.204].

NOTE This object includes endoscopic intense light source equipment which is part of the ENDOSCOPIC EQUIPMENT including its supply unit, therefore IEC 60601-2-57 does not apply.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

Amendment:

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

Addition:

IEC 60601-2-2:2009, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-37, *Medical electrical equipment – Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment*