



# SLOVENSKI STANDARD SIST EN 60601-2-54:2009

01-november-2009

A YX]W]bg\_UYY\_f] bUcdfYa U!'&! ("XY. 'DcgYVbY'nU hYj Y'nUcgbcj bc`j Ufbcgh]b  
V]glj YbY`Uglbcgh]fYbh[ Ybg\_YcdfYa Y'nUfUX]c[ fUZ]c`]b'fUX]cg\_cd]c`f#97`\* \$\* \$%&! )(.&\$-\$-Ł

Medical electrical equipment - Part 2-54: Particular requirements for basic safety and essential performance of X-ray equipment for radiography and radioscopy (IEC 60601-2-54:2009)

**iTeh STANDARD PREVIEW**

Medizinische elektrische Geräte - Teil 2-54: Besondere Festlegungen für die Sicherheit und die wesentlichen Leistungsmerkmale von Röntgeneinrichtungen für Radiographie und Radioskopie (IEC 60601-2-54:2009)

[SIST EN 60601-2-54:2009](https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-11eb-35b4-20c000000000)

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-11eb-35b4-20c000000000>

Appareils électromédicaux - Partie 2-54: Exigences particulières relatives à la sécurité de base et aux performances essentielles des appareils à rayonnement X utilisés pour la radiographie et la radioscopie (CEI 60601-2-54:2009)

**Ta slovenski standard je istoveten z: EN 60601-2-54:2009**

## ICS:

11.040.50 Radiografska oprema Radiographic equipment

**SIST EN 60601-2-54:2009 en,fr**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 60601-2-54:2009

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN 60601-2-54**

September 2009

ICS 11.040.50

Supersedes EN 60601-2-7:1998, EN 60601-2-28:1993 (partially) and EN 60601-2-32:1994

English version

**Medical electrical equipment -  
Part 2-54: Particular requirements  
for the basic safety and essential performance of X-ray equipment  
for radiography and radioscopy  
(IEC 60601-2-54:2009)**

Appareils électromédicaux -  
Partie 2-54: Exigences particulières  
pour la sécurité de base  
et les performances essentielles  
des appareils à rayonnement X utilisés  
pour la radiographie et la radioscopie  
(CEI 60601-2-54:2009)

Medizinische elektrische Geräte -  
Teil 2-54: Besondere Festlegungen  
für die Sicherheit und die wesentlichen  
Leistungsmerkmale  
von Röntgeneinrichtungen  
für Radiographie und Radioskopie  
(IEC 60601-2-54:2009)

[standards.iteh.ai](https://standards.iteh.ai/)

[SIST EN 60601-2-54:2009](https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-)

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8->

This European Standard was approved by CENELEC on 2009-08-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

**CENELEC**

European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**Central Secretariat: Avenue Marnix 17, B - 1000 Brussels**

## Foreword

The text of document 62B/735/FDIS, future edition 1 of IEC 60601-2-54, prepared by SC 62B, Diagnostic imaging equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-54 on 2009-08-01.

EN 60601-2-54 was developed for use with EN 60601-1:2006.

This European Standard supersedes EN 60601-2-7:1998, EN 60601-2-32:1994 and EN 60601-2-28:1993 (partially).

The following dates were fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2010-05-01
- latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 2012-08-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive MDD (93/42/EEC). See Annex ZZ.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

Annexes ZA and ZZ have been added by CENELEC.

---

### Endorsement notice

The text of the International Standard IEC 60601-2-54:2009 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

[1] IEC 60627	NOTE Harmonized as EN 60627:2001 (not modified).
[2] IEC 61267	NOTE Harmonized as EN 61267:2006 (not modified).
[3] ISO 4090	NOTE Harmonized as EN ISO 4090:2004 (not modified).
[10] IEC 60601-2-7	NOTE Harmonized as EN 60601-2-7:1998 (not modified).
[11] IEC 60601-2-28	NOTE Harmonized as EN 60601-2-28:1993 (not modified).
[12] IEC 60601-2-32	NOTE Harmonized as EN 60601-2-32:1994 (not modified).
[13] IEC 60601-1-8	NOTE Harmonized as EN 60601-1-8:2007 (not modified).
[14] IEC 60601-1-10	NOTE Harmonized as EN 60601-1-10:2008 (not modified).
[15] IEC 60601-2-43	NOTE Harmonized as EN 60601-2-43:2000 (not modified).

**STANDARD PREVIEW**  
**(standards.iteh.ai)**

---

[SIST EN 60601-2-54:2009](https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009)  
<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009>

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

*Annex ZA of EN 60601-1:2006 applies, except as follows:*

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<b>Replace the reference to IEC 60601-1-2 by:</b>				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2	2007
<b>Replace the reference to IEC 60601-1-3 by:</b>				
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3	2008
<b>Addition:</b>				
IEC 60336	- <sup>1)</sup>	Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Characteristics of focal spots	EN 60336	2005 <sup>2)</sup>
IEC 60580	2000	Medical electrical equipment - Dose area product meters	EN 60580	2000
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	—	—
IEC 60806	- <sup>1)</sup>	Determination of the maximum symmetrical radiation field from a rotating anode X-ray tube for medical diagnosis	EN 60806	2004 <sup>2)</sup>
IEC 62220-1	2003	Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1: Determination of the detective quantum efficiency	EN 62220-1	2004

<sup>1)</sup> Undated reference.

<sup>2)</sup> Valid edition at date of issue.

## **Annex ZZ** (informative)

### **Coverage of Essential Requirements of EC Directives**

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

**WARNING:** Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

---

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 60601-2-54:2009](https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009)

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 60601-2-54:2009

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009>





IEC 60601-2-54

Edition 1.0 2009-06

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical electrical equipment –**  
**Part 2-54: Particular requirements for the basic safety and essential performance**  
**of X-ray equipment for radiography and radioscopy**

**Appareils électromédicaux –**  
**Partie 2-54: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des appareils à rayonnement X utilisés pour la radiographie et la**  
**radioscopie**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

PRICE CODE  
CODE PRIX

**XB**

ICS 11.040.50

ISBN 2-8318-1047-5

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
201.1 Scope, object and related standards .....	7
201.2 Normative references.....	8
201.3 Terms and definitions.....	9
201.4 General requirements .....	10
201.5 General requirements for testing of ME EQUIPMENT .....	11
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	11
201.7 ME EQUIPMENT identification, marking and documents .....	11
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	14
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	17
201.10 Protection against unwanted and excessive radiation HAZARDS .....	21
201.11 Protection against excessive temperatures and other HAZARDS .....	21
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	22
201.13 HAZARDOUS SITUATIONS and fault conditions .....	22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	22
201.15 Construction of ME EQUIPMENT.....	22
201.16 ME SYSTEMS .....	22
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	22
202 Electromagnetic compatibility – Requirements and tests .....	22
203 Radiation protection in diagnostic X-ray equipment.....	23
Annexes .....	58
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	59
Annex AA (informative) Particular guidance and rationale.....	61
Bibliography.....	64
Index of defined terms used in this particular standard.....	65
Figure 203.101 – Zone of EXTRA-FOCAL RADIATION .....	41
Figure 203.102 – Discrepancies in covering the IMAGE RECEPTION AREA.....	43
Figure 203.103 – Discrepancies in visual indication of the X-RAY FIELD .....	47
Figure 203.104 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY SOURCE ASSEMBLY below the PATIENT SUPPORT) .....	55
Figure 203.105 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE ASSEMBLY below the PATIENT SUPPORT) .....	55
Figure 203.106 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY SOURCE ASSEMBLY above the PATIENT SUPPORT) .....	56
Figure 203.107 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE ASSEMBLY above the PATIENT SUPPORT).....	57
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	10
Table 203.101 – Tests for verifying reproducibility and linearity .....	29

Table 203.102 – Loadings for testing AUTOMATIC EXPOSURE CONTROLS.....	31
Table 203.103 – ATTENUATION for the measurement of AIR KERMA .....	33
Table 203.104 – ATTENUATION EQUIVALENT of items in the X-RAY BEAM.....	49
Table 203.105 – Application categories .....	51
Table 203.106 – Requirements for PRIMARY PROTECTIVE SHIELDING .....	51
Table 203.107 – STRAY RADIATION in SIGNIFICANT ZONES OF OCCUPANCY .....	53
Table 201.C.101 – Marking on the outside of ME EQUIPMENT or its parts .....	59
Table 201.C.102 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS.....	59

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 60601-2-54:2009](https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009)

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009>

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-54 has been prepared by IEC subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

IEC 60601-2-54 has been developed for use with the third edition of IEC 60601-1 (2005). It replaces and supersedes IEC 60601-2-7 and IEC 60601-2-32, as well as IEC 60601-2-28:1993 (currently under revision), all of which were developed to amend earlier editions of IEC 60601-1 and consequently no longer apply to this particular standard.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62B/735/FDIS	62B/750/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## INTRODUCTION

This particular standard has been prepared to provide, based on IEC 60601-1:2005 (third edition) and its collaterals, a complete set of safety requirements for ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. While the previously existing standards for such equipment were dedicated to components and subsystems, this particular standard addresses the system level of X-RAY EQUIPMENT, which consists of a combination of an X-RAY GENERATOR, ASSOCIATED EQUIPMENT and ACCESSORIES. Component functions are addressed as far as necessary.

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. Requirements for additional provisions for ME EQUIPMENT for interventional applications are covered by IEC 60601-2-43.

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

[SIST EN 60601-2-54:2009](https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009)

<https://standards.iteh.ai/catalog/standards/sist/a6471d56-da5a-4608-81e8-0fd4b7d4b64f/sist-en-60601-2-54-2009>

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1)</sup> applies, except as follows:

##### 201.1.1 Scope

*Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ME EQUIPMENT and ME SYSTEMS intended to be used for projection RADIOGRAPHY and RADIOSCOPY. IEC 60601-2-43 applies to ME EQUIPMENT and ME SYSTEMS intended to be used for interventional applications and refers to applicable requirements in this particular standard.

ME EQUIPMENT and ME SYSTEMS intended to be used for bone or tissue absorption densitometry, computed tomography, mammography or dental applications are excluded from the scope of this International Standard. The scope of this International Standard also excludes radiotherapy simulators.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE Taking into account economic and social factors, the scope of this particular standard includes ME EQUIPMENT intended to be used for DIRECT RADIOSCOPY. In some countries examinations performed with DIRECT RADIOSCOPY are prohibited.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ME EQUIPMENT and ME SYSTEMS for RADIOGRAPHY and RADIOSCOPY.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 and IEC 60601-1-3 apply as modified in Clauses 202 and 203 respectively. IEC 60601-1-8 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

<sup>1)</sup> The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*