

SLOVENSKI STANDARD

SIST EN 60601-2-28:2010

01-september-2010

Nadomešča:

SIST EN 60601-2-28:1995

Medicinska električna oprema - 2-28. del: Posebne zahteve za osnovno varnost in bistvene lastnosti rentgenskih naprav za zdravniške preglede (IEC 60601-2-28:2010)

Medical electrical equipment - Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis (IEC 60601-2-28:2010)

iTeh STANDARD PREVIEW

(standards.iteh.ai)

Medizinische elektrische Geräte - Teil 2-28: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Röntgenstrahlern für die medizinische Diagnostik (IEC 60601-2-28:2010)

<https://standards.iteh.ai/catalog/standards/sist/c449283b-4b7e-4b0e-a8c0-faac5d8f617/sist-en-60601-2-28-2010>

Appareils électromédicaux - Partie 2-28: Exigences particulières pour la sécurité de base et les performances essentielles des gaines équipées pour diagnostic médical (CEI 60601-2-28:2010)

Ta slovenski standard je istoveten z: EN 60601-2-28:2010

ICS:

11.040.50 Radiografska oprema Radiographic equipment

SIST EN 60601-2-28:2010

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 60601-2-28:2010

<https://standards.iteh.ai/catalog/standards/sist/c449283b-4b7e-4b0e-a8c0-faaac5d8f617/sist-en-60601-2-28-2010>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 60601-2-28

May 2010

ICS 11.040.55

Supersedes EN 60601-2-28:1993

English version

**Medical electrical equipment -
Part 2-28: Particular requirements for the basic safety and essential
performance of X-ray tube assemblies for medical diagnosis
(IEC 60601-2-28:2010)**

Appareils électromédicaux -
Partie 2-28: Exigences particulières
pour la sécurité de base
et les performances essentielles
des gaines équipées pour diagnostic
médical
(CEI 60601-2-28:2010)

Medizinische elektrische Geräte -
Teil 2-28: Besondere Festlegungen
für die Sicherheit einschließlich
der wesentlichen Leistungsmerkmale
von Röntgenstrahlern für die medizinische
Diagnostik
(IEC 60601-2-28:2010)

iteh STANDARD PREVIEW
(standards.iteh.ai)

This European Standard was approved by CENELEC on 2010-04-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 62B/778/FDIS, future edition 2 of IEC 60601-2-28, prepared by SC 62B, Diagnostic imaging equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-28 on 2010-04-01.

This European Standard supersedes EN 60601-2-28:1993.

The second edition of this particular standard has been prepared to fit EN 60601-1:2006, which is referred to as the general standard.

When the first edition was developed, mainly X-RAY TUBE ASSEMBLIES holding a glass insert were considered and EN 60601-1:1990 was in place. While the variety of modern X-RAY TUBE ASSEMBLIES and technologies has increased, the third edition of the general standard requires the MANUFACTURER to perform RISK MANAGEMENT. The technical modifications versus the first edition of EN 60601-2-28 account for these changes.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2011-01-01
- latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 2013-04-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive 93/42/EEC. See Annex ZZ.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-2-28:2010 was approved by CENELEC as a European Standard without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 60601-2-28:2010](https://standards.iteh.ai/catalog/standards/sist/c449283b-4b7e-4b0e-a8c0-faaac5d8f617/sist-en-60601-2-28-2010)

<https://standards.iteh.ai/catalog/standards/sist/c449283b-4b7e-4b0e-a8c0-faaac5d8f617/sist-en-60601-2-28-2010>

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement:</i>				
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3 + corr. March	2008 2010
<i>Addition:</i>				
IEC 60336	-	Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Characteristics of focal spots	EN 60336	-
IEC 60522	-	Determination of the permanent filtration of X-ray tube assemblies	EN 60522	-
IEC 60613	2010	Electrical and loading characteristics of X-ray tube assemblies for medical diagnosis	EN 60613	2010
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-

Annex ZZ (informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 60601-2-28:2010

<https://standards.iteh.ai/catalog/standards/sist/c449283b-4b7e-4b0e-a8c0-faaac5d8f617/sist-en-60601-2-28-2010>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 60601-2-28:2010

<https://standards.iteh.ai/catalog/standards/sist/c449283b-4b7e-4b0e-a8c0-faaac5d8f617/sist-en-60601-2-28-2010>



IEC 60601-2-28

Edition 2.0 2010-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
Part 2-28: Particular requirements for the basic safety and essential performance
of X-ray tube assemblies for medical diagnosis

Appareils électromédicaux –
Partie 2-28: Exigences particulières pour la sécurité de base et les performances
essentiels des gaines équipées pour diagnostic médical

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

R

ICS 11.040.55

ISBN 2-8318-1082-2

CONTENTS

FOREWORD.....	3
201.1 Scope, object and related standards.....	5
201.2 Normative references	6
201.3 Terms and definitions	7
201.4 General requirements.....	7
201.5 General requirements for testing ME EQUIPMENT.....	7
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	8
201.7 ME EQUIPMENT identification, marking and documents.....	8
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	10
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	11
201.10 Protection against unwanted and excessive radiation HAZARDS.....	12
201.11 Protection against excessive temperatures and other HAZARDS.....	12
201.12 Accuracy of controls and instruments and protection against hazardous outputs	13
201.13 HAZARDOUS SITUATIONS and fault conditions	13
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	13
201.15 Construction of ME EQUIPMENT	13
201.16 ME SYSTEMS.....	13
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	13
203 RADIATION protection in diagnostic X-RAY EQUIPMENT	13
Annexes	14
Annex AA (informative) Test of X-RAY TUBE ASSEMBLIES for pressure-related RISKS.....	15
Index of defined terms used in this particular standard.....	17

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-28 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 1993. This edition constitutes a technical revision.

The second edition of this particular standard has been prepared to fit IEC 60601-1:2005 (the third edition of IEC 60601-1), which is referred to as the general standard.

When the first edition was developed, mainly X-RAY TUBE ASSEMBLIES holding a glass insert were considered and IEC 60601-1:1988 (the second edition of the general standard) was in place. While the variety of modern X-RAY TUBE ASSEMBLIES and technologies has increased, the third edition of the general standard requires the MANUFACTURER to perform RISK MANAGEMENT. The technical modifications versus the first edition of IEC 60601-2-28 account for these changes.