



SLOVENSKI STANDARD
SIST EN ISO 8536-8:2005
01-marec-2005

Infusion equipment for medical use - Part 8: Infusion equipment for use with pressure infusion apparatus (ISO 8536-8:2004)

Infusionsgeräte zur medizinischen Verwendung - Teil 8: Infusionsgeräte zur Verwendung mit Druckinfusionsapparaten (ISO 8536-8:2004)

Matériel de perfusion a usage médical - Partie 8: Matériel de perfusion pour utilisation avec des appareils de perfusion sous pression (ISO 8536-8:2004)

Ta slovenski standard je istoveten z: EN ISO 8536-8:2004

ICS:

11.040.20	Transfuzijska, infuzijska in injekcijska oprema	Transfusion, infusion and injection equipment
-----------	---	---

SIST EN ISO 8536-8:2005

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8536-8:2005](#)

<https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 8536-8

August 2004

ICS 11.040.20

English version

**Infusion equipment for medical use - Part 8: Infusion equipment
for use with pressure infusion apparatus (ISO 8536-8:2004)**

Matériel de perfusion à usage médical - Partie 8: Matériel
de perfusion pour utilisation avec des appareils de
perfusion sous pression (ISO 8536-8:2004)

Infusionsgeräte zur medizinischen Verwendung - Teil 8:
Infusionsgeräte zur Verwendung mit
Druckinfusionsapparaten (ISO 8536-8:2004)

This European Standard was approved by CEN on 29 July 2004.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN ISO 8536-8:2005](https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

EN ISO 8536-8:2004 (E)**Foreword**

This document (EN ISO 8536-8:2004) has been prepared by Technical Committee ISO/TC 76 "Transfusion, infusion and injection equipment for medical and pharmaceutical use" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2005, and conflicting national standards shall be withdrawn at the latest by February 2005.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 8536-8:2004 has been approved by CEN as EN ISO 8536-8:2004 without any modifications.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8536-8:2005](https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005>

INTERNATIONAL STANDARD

ISO
8536-8

First edition
2004-08-15

Infusion equipment for medical use — Part 8: Infusion equipment for use with pressure infusion apparatus

*Matériel de perfusion à usage médical —
Partie 8: Matériel de perfusion pour utilisation avec des appareils de
perfusion sous pression*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8536-8:2005](https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005>



Reference number
ISO 8536-8:2004(E)

© ISO 2004

ISO 8536-8:2004(E)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8536-8:2005](https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed62e12a-d961-400d-82df-68ff1a28ab48/sist-en-iso-8536-8-2005>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope.....	1
2 Normative references	1
3 General requirements	1
4 Designation.....	4
4.1 Infusion set.....	4
4.2 Air-inlet device	4
5 Materials.....	4
6 Physical requirements.....	5
6.1 Particulate contamination	5
6.2 Tensile strength	5
6.3 Leakage	5
6.4 Male conical fitting.....	5
6.5 Injection site	5
6.6 Fluid filter	5
6.7 Flow rate of infusion fluid	5
6.8 Closure-piercing device	5
6.9 Air-inlet device	5
6.10 Drip chamber and drip tube	5
6.11 Tubing	5
6.12 Flow regulator	5
6.13 Protective caps.....	6
6.14 Storage volume	6
7 Chemical requirements	6
8 Biological requirements	6
9 Packaging	6
10 Labelling.....	6
10.1 Unit container	6
10.2 Shelf or multi-unit container	7
Annex A (normative) Physical tests	8
Annex B (normative) Chemical tests.....	9
Annex C (normative) Biological tests.....	10
Bibliography	11

ISO 8536-8:2004(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8536-8 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection equipment for medical and pharmaceutical use*.

ISO 8536 consists of the following parts, under the general title *Infusion equipment for medical use*:

- iTeh STANDARD PREVIEW**
(standards.iteh.ai)
- Part 1: *Infusion glass bottles*
 - Part 2: *Closures for infusion bottles*
 - Part 3: *Aluminium caps for infusion bottles*
 - Part 4: *Infusion sets for single use, gravity feed*
 - Part 5: *Burette infusion sets for single use, gravity feed*
 - Part 6: *Freeze drying closures for infusion bottles*
 - Part 7: *Caps made of aluminium-plastics combinations for infusion bottles*
 - Part 8: *Infusion equipment for use with pressure infusion apparatus*
 - Part 9: *Fluid lines for use with pressure infusion equipment*
 - Part 10: *Accessories for fluid lines for use with pressure infusion equipment*
 - Part 11: *Infusion filters for use with pressure infusion equipment*

Infusion equipment for medical use —

Part 8: Infusion equipment for use with pressure infusion apparatus

1 Scope

This part of ISO 8536 gives users information on sterilized infusion sets for single use with pressure infusion equipment up to a maximum of 200 kPa (2 bar).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-2:1998, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*

ISO 8536-4:2004, *Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

IEC 60601-2-24, *Medical electrical equipment — Part 2-24: Particular requirements for the safety of infusion pumps and controllers*

3 General requirements

3.1 The nomenclature to be used for components of infusion sets and of a separate air-inlet device is given in Figures 1, 2 and 3. These figures illustrate examples of the configuration of infusion sets and air-inlet devices; other configurations may be used provided they lead to the same results. Infusion sets as illustrated in Figure 2 should only be used for collapsible plastics containers. Infusion sets as illustrated in Figure 2 used with separate air-inlet devices as illustrated in Figure 3, or infusion sets as illustrated in Figure 1 shall be used for rigid containers.