



SLOVENSKI STANDARD
SIST EN ISO 8536-10:2005
01-marec-2005

Infusion equipment for medical use - Part 10: Accessories for fluid lines for use with pressure infusion equipment (ISO 8536-10:2004)

Infusionsgeräte zur medizinischen Verwendung - Teil 10: Zubehörteile für Übertragungsleitungen zur Verwendung mit Druckinfusionsapparaten (ISO 8536-10:2004)

Matériel de perfusion a usage médical - Partie 10: Accessoires de tubulures pour utilisation avec des appareils de perfusion sous pression (ISO 8536-10:2004)

Ta slovenski standard je istoveten z: EN ISO 8536-10:2004

ICS:

11.040.20	Transfuzijska, infuzijska in injekcijska oprema	Transfusion, infusion and injection equipment
-----------	---	---

SIST EN ISO 8536-10:2005 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8536-10:2005](#)

<https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ee953400cb2/sist-en-iso-8536-10-2005>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 8536-10

December 2004

ICS 11.040.20

English version

**Infusion equipment for medical use - Part 10: Accessories for
fluid lines for use with pressure infusion equipment (ISO 8536-
10:2004)**

Matériel de perfusion à usage médical - Partie 10:
Accessoires de tubulures pour utilisation avec des
appareils de perfusion sous pression (ISO 8536-10:2004)

This European Standard was approved by CEN on 30 September 2004.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
<https://standards.cen.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ef953400cb2/sist-en-iso-8536-10-2005>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

EN ISO 8536-10:2004 (E)**Foreword**

This document (EN ISO 8536-10:2004) has been prepared by Technical Committee ISO/TC 76 "Transfusion, infusion and injection equipment for medical and pharmaceutical use" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2005, and conflicting national standards shall be withdrawn at the latest by June 2005.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 8536-10:2004 has been approved by CEN as EN ISO 8536-10:2004 without any modifications.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8536-10:2005](https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ee953400cb2/sist-en-iso-8536-10-2005)

<https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ee953400cb2/sist-en-iso-8536-10-2005>

INTERNATIONAL
STANDARD

ISO
8536-10

First edition
2004-10-01

**Infusion equipment for medical use —
Part 10:
Accessories for fluid lines for use with
pressure infusion equipment**

Matériel de perfusion à usage médical —

*Partie 10: Accessoires de tubulures pour utilisation avec des appareils
de perfusion sous pression*

iTeh STANDARD-PREVIEW
(standards.iteh.ai)

SIST EN ISO 8536-10:2005

<https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ee953400cb2/sist-en-iso-8536-10-2005>



Reference number
ISO 8536-10:2004(E)

© ISO 2004

ISO 8536-10:2004(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8536-10:2005](https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ee953400cb2/sist-en-iso-8536-10-2005)

<https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-ee953400cb2/sist-en-iso-8536-10-2005>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

1	Scope	1
2	Normative references	1
3	Designation	2
4	Materials	2
5	Physical requirements	2
5.1	Avoidance of air bubbles	2
5.2	Particulate contamination	3
5.3	Tensile strength	3
5.4	Leakage	3
5.5	Adapters with female and/or male conical fittings	3
5.6	Protective caps	3
5.7	Manipulation of stopcocks	3
5.8	Unit with injection site	3
5.9	Unit with check valve	3
6	Chemical requirements	3
7	Biological requirements	4
7.1	Sterility	4
7.2	Pyrogens	4
7.3	Haemolysis	4
8	Packaging	4
9	Labelling	4
9.1	Unit container	4
9.2	Shelf or multi-unit container	5
Annex A (normative)	Physical tests	6
Annex B (normative)	Chemical tests	8
Annex C (normative)	Biological tests	9
Bibliography	10

ISO 8536-10:2004(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8536-10 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection equipment for medical and pharmaceutical use*.

ISO 8536 consists of the following parts, under the general title *Infusion equipment for medical use*:

- Part 1: *Infusion glass bottles*
- Part 2: *Closures for infusion bottles*
- Part 3: *Aluminium caps for infusion bottles*
- Part 4: *Infusion sets for single use, gravity feed*
- Part 5: *Burette infusion sets for single use, gravity feed*
- Part 6: *Freeze drying closures for infusion bottles*
- Part 7: *Caps made of aluminium-plastics combinations for infusion bottles*
- Part 8: *Infusion equipment for use with pressure infusion apparatus*
- Part 9: *Fluid lines for use with pressure infusion equipment*
- Part 10: *Accessories for fluid lines for use with pressure infusion equipment*
- Part 11: *Infusion filters for use with pressure infusion equipment*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 8536-10:2005

<https://standards.iteh.ai/catalog/standards/sist/71fa6c30-f18a-4dcc-9be0-71499cb2/sist-en-iso-8536-10-2005>

Infusion equipment for medical use —

Part 10:

Accessories for fluid lines for use with pressure infusion equipment

1 Scope

This part of ISO 8536 applies to sterilized accessories for single use in fluid lines and pressure infusion equipment as specified in ISO 8536-8.

This part of ISO 8536 includes:

- a) Two-way stopcocks (2SC), three-way stopcocks (3SC), four-way stopcocks (4SC) and stopcocks manifold (SM);

NOTE Designation of a stopcock depends on the number of connections. The number of possible functional positions can be expressed by addition of a complementary note, using a diagonal stroke and a numeral indicating the number of possible stopcock positions, e.g. 3/4-way stopcock for three-way stopcock with four possible positions.

- b) units with injection site (UIS) or check valve (UCV);
- c) stoppers (S) or adapters (A).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-2, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*

ISO 7886-1, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use*

ISO 8536-4, *Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed*

ISO 8536-8, *Infusion equipment for medical use — Part 8: Infusion equipment for use with pressure infusion apparatus*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*