



# SLOVENSKI STANDARD SIST EN 60601-2-23:2015

01-december-2015

Nadomešča:

SIST EN 60601-2-23:2002

---

**Medicinska električna oprema - 2-23. del: Posebne zahteve za osnovno varnost in bistvene lastnosti opreme za skozikožno nadzorovanje delnega (parcialnega) krvnega tlaka**

Medical electrical equipment - Part 2-23: Particular requirements for basic safety and essential performance of transcutaneous partial pressure monitoring equipment

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

Appareils électromédicaux - Partie 2-23: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de surveillance de la pression partielle transcutanée

**Ta slovenski standard je istoveten z: EN 60601-2-23:2015**

---

**ICS:**

11.040.55      Diagnostična oprema      Diagnostic equipment

**SIST EN 60601-2-23:2015**      en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 60601-2-23:2015](https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015)

<https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015>

EUROPEAN STANDARD

**EN 60601-2-23**

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2015

ICS 11.040.55

Supersedes EN 60601-2-23:2000

English Version

**Medical electrical equipment - Part 2-23: Particular requirements  
for the basic safety and essential performance of transcutaneous  
partial pressure monitoring equipment  
(IEC 60601-2-23:2011)**

Appareils électromédicaux - Partie 2-23: Exigences  
particulières pour la sécurité de base et les performances  
essentielles des appareils de surveillance de la pression  
partielle transcutanée  
(IEC 60601-2-23:2011)

Medizinische elektrische Geräte - Teil 2-23: Besondere  
Festlegungen für die Sicherheit einschließlich der  
wesentlichen Leistungsmerkmale von Geräten für die  
transkutane Partialdrucküberwachung  
(IEC 60601-2-23:2011)

This European Standard was approved by CENELEC on 2015-09-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

**EN 60601-2-23:2015****European foreword**

The text of document 62D/885/FDIS, future edition 3 of IEC 60601-2-23, prepared by SC 62D "Electromedical equipment", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-23:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-06-15
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-09-15

This document supersedes EN 60601-2-23:2000.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

**iTeh STANDARD PREVIEW**

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document. <https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015>

**Endorsement notice**

The text of the International Standard IEC 60601-2-23:2011 was approved by CENELEC as a European Standard without any modification.

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cenelec.eu](http://www.cenelec.eu).

#### **Annex ZA of EN 60601-1:2006 applies, except as follows:**

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<b>Replacement in Annex ZA of EN 60601-1:2006:</b>				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corrigendum Mar.	2007 2010
-	-			
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8 + corrigendum Mar.	2007 2010
-	-			

#### **Addition to Annex ZA of EN 60601-1:2006:**

IEC 60601-2-49	2011	Medical electrical equipment - Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitoring equipment	EN 60601-2-49	2015
----------------	------	--	---------------	------

EN 60601-2-23:2015

**Annex ZZ**  
(informative)**Coverage of Essential Requirements of EU Directives**

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

**WARNING:** Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 60601-2-23:2015](https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015)

<https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015>



IEC 60601-2-23

Edition 3.0 2011-02

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Medical electrical equipment –**  
**Part 2-23: Particular requirements for the basic safety and essential performance**  
**of transcutaneous partial pressure monitoring equipment**

**Appareils électromédicaux –**  
**Partie 2-23: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des appareils de surveillance de la pression partielle transcutanée**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

PRICE CODE  
CODE PRIX



ICS 11.040.55

ISBN 978-2-88912-370-4

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards .....	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	10
201.4 General requirements .....	11
201.5 General requirements for testing of ME EQUIPMENT .....	11
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7 ME EQUIPMENT identification, marking and documents .....	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	13
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	14
201.10 Protection against unwanted and excessive radiation HAZARDS .....	14
201.11 Protection against excessive temperatures and other HAZARDS .....	14
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	18
201.13 HAZARDOUS SITUATIONS and fault conditions .....	23
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	23
201.15 Construction of ME EQUIPMENT.....	23
201.16 ME SYSTEMS .....	23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	24
202 Electromagnetic compatibility – Requirements and tests.....	24
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems .....	28
Annexes .....	34
Annex AA (informative) Particular guidance and rationale.....	35
Annex BB (informative) Alarm diagrams of Clause 208/IEC 60601-1-8:2006 .....	43
Index of defined terms used in this particular standard.....	46
Figure 201.101 – TRANSDUCER cable strain relief.....	14
Figure 201.102 – Foam block test (see 201.11.1.2.2.104 and 201.11.1.2.2.105).....	15
Figure 201.103 – Linearity and hysteresis test set-up – Gas mix chamber, assembled.....	20
Figure 201.104 – Linearity and hysteresis test set-up – Gas mix chamber, manufacturing dimensions .....	21
Figure 201.105 – Linearity and hysteresis test set-up – Gas mix chamber, dimensions of hose connector .....	22
Figure 202.101 – Set-up for radiated and conducted EMISSIONS testing according to 202.6.1.1.2 a) .....	25
Figure 202.102 – Set-up for radiated immunity test according to 202.6.2.3.2 .....	27
Figure BB.1 – NON-LATCHING ALARM SIGNALS without ALARM RESET.....	43
Figure BB.2 – NON-LATCHING ALARM SIGNALS with ALARM RESET .....	43
Figure BB.3 – LATCHING ALARM SIGNALS with ALARM RESET .....	44
Figure BB.4 – Two ALARM CONDITIONS with ALARM RESET.....	44



Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	11
Table 201.102 – Required readings and tolerances .....	19
Table 201.103 – Calibration test gases .....	19
Table 208.101 – ALARM CONDITION priorities .....	29
Table 208.102 – Characteristics of the burst of auditory ALARM SIGNALS .....	30

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 60601-2-23:2015](https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015)

<https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015>

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-23 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 1999 and constitutes a technical revision. This edition of IEC 60601-2-23 was revised to align structurally with the 2005 edition of IEC 60601-1.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/885/FDIS	62D/907/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## **iTeh STANDARD PREVIEW (standards.iteh.ai)**

[SIST EN 60601-2-23:2015](https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015)

<https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015>

## INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of TRANSCUTANEOUS PARTIAL PRESSURE MONITORING EQUIPMENT. It amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this Standard.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 60601-2-23:2015](https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015)

<https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-dab99b16fe66/sist-en-60601-2-23-2015>

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-23: Particular requirements for the basic safety and essential performance of transcutaneous partial pressure monitoring equipment

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1</sup> applies, except as follows:

##### 201.1.1 \* Scope

*Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of TRANSCUTANEOUS PARTIAL PRESSURE MONITORING EQUIPMENT as defined in 201.3.63 and hereinafter referred to as ME EQUIPMENT, whether this ME EQUIPMENT is stand alone or part of a system.

This standard applies to transcutaneous monitors used with adults, children and neonates, and it includes the use of these devices in foetal monitoring during birth.

This standard does not apply to haemoglobin saturation oximeters or to devices applied to surfaces of the body other than the skin (for example conjunctiva, mucosa).

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the General Standard.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for TRANSCUTANEOUS PARTIAL PRESSURE MONITORING EQUIPMENT as defined in 201.3.63.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

<sup>1</sup> The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2:2007 and IEC 60601-1-8:2006 apply as modified in Clauses 202 and 208 respectively. IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published

#### 201.1.4 Particular standards

##### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

(standards.iteh.ai)

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

[https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-](https://standards.iteh.ai/catalog/standards/sist/827c098c-9ea1-475d-9722-d2b99b16f666/sist-en-60601-2-23-2015)

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.