

INTERNATIONAL ELECTROTECHNICAL COMMISSION

IEC TR 61000-1-6
Edition 1.0 2012-07

Electromagnetic compatibility (EMC) –

Part 1-6: General – Guide to the assessment of measurement uncertainty

C O R R I G E N D U M 1

Figure 17 – Situations, where and how an instrument is suitable for tests and/or measurements as specified in the applicable basic standard with tolerances

Replace the text of the existing footnote ¹⁾ with the following new text:

- ¹⁾ This approach deviates from other concepts found in literature, because the tolerances given in the standards of the 61000-4-x series once were defined without taking into account uncertainties. Therefore uncertainties should not be used to demonstrate compliance with tolerances.

[IEC TR 61000-1-6:2012/COR1:2014](https://standards.iteh.ai/catalog/standards/iec/a7c20e2d-d11b-4ad8-b6fa-794d42a8de2c/iec-tr-61000-1-6-2012-cor1-2014)

<https://standards.iteh.ai/catalog/standards/iec/a7c20e2d-d11b-4ad8-b6fa-794d42a8de2c/iec-tr-61000-1-6-2012-cor1-2014>