

SLOVENSKI STANDARD SIST EN 60601-2-47:2015

01-september-2015

Nadomešča: SIST EN 60601-2-47:2002

Medicinska električna oprema - 2-47. del: Posebne zahteve za osnovno varnost in bistvene lastnosti ambulantnih elektrokardiografskih sistemov

Medical electrical equipment - Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

Medizinische elektrische Geräte - Teil 2-47: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von ambulanten elektrokardiogrphischen Systemen

SIST EN 60601-2-47:2015

Appareils électromédicaux de Partie 2-47 Exigences particulières pour la sécurité de base et les performances essentielles des systèmes d'électrocardiographie ambulatoires

Ta slovenski standard je istoveten z: EN 60601-2-47:2015

<u>ICS:</u>

11.040.55 Diagnostična oprema

Diagnostic equipment

SIST EN 60601-2-47:2015

en

SIST EN 60601-2-47:2015

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 60601-2-47:2015

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 60601-2-47

May 2015

ICS 11.040.55

Supersedes EN 60601-2-47:2001

English Version

Medical electrical equipment - Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems (IEC 60601-2-47:2012)

Appareils électromédicaux - Partie 2-47: Exigences particulières pour la sécurité de base et les performances essentielles des systèmes d'électrocardiographie ambulatoires (IEC 60601-2-47:2012) Medizinische elektrische Geräte - Teil 2-47: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von ambulanten elektrokardiographischen Systemen (IEC 60601-2-47:2012)

This European Standard was approved by CENELEC on 2015-04-14. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member. **ICLASS.ICEN.21**)

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions. dards. iteh ai/catalog/standards/sist/11d9743e-ca64-41e3-9314-9909fd823967/sist-en-60601-2-47-2015

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

© 2015 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC Members.

Foreword

The text of document 62D/963/FDIS, future edition 2 of IEC 60601-2-47, prepared by SC 62D "Electromedical equipment", of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-47:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at (dop) 2016-01-14 national level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with (dow) 2018-04-14 the document have to be withdrawn

This document supersedes EN 60601-2-47:2001.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive 93/42/EEC. See informative Annex ZZ, which is an integral part of this document. https://standards.iteh.ai/catalog/standards/sist/11d9743e-ca64-41e3-9314-9909fd823967/sist-en-60601-2-47-2015

Endorsement notice

The text of the International Standard IEC 60601-2-47:2012 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-25	NOTE	Harmonized as EN 60601-2-25.
IFC 60601-2-27	NOTE	Harmonized as FN 60601-2-27

EN 60601-2-47:2015

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: <u>www.cenelec.eu</u>.

Annex ZA of EN 60601-1:2006 applies, except as follows:

Publication	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
In Annex ZA of EN	60601-	1:2006 replace IEC 60601-1-2 by:	W	
IEC 60601-1-2 (mod)	2007	Medical electrical equipment -	EN 60601-1-2	2007
-	-	Part 1-2: General requirements for basic safety and essential performance -	+ corrigendum Mar.	2010
		Collateral standard: Electromagnetic		
http	ps://stand	compatibility - Requirements and tests ards.itch.ar/catalog/stantlards/sist/11d9/43e-ca64-41c	e3-9314-	
		000063922067/aint are 60601.2.47.2015		

9909fd823967/sist-en-60601-2-47-2015

Annex ZZ

(informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW (standards.iteh.ai)





Edition 2.0 2012-02

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment ANDARD PREVIEW Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

SIST EN 60601-2-47:2015

Appareils électromédicauxetrai/catalog/standards/sist/11d9743e-ca64-41e3-9314-Partie 2-47: Exigences particulières/pour/la0sécurité5de base et les performances essentielles des systèmes d'électrocardiographie ambulatoires

INTERNATIONAL ELECTROTECHNICAL COMMISSION

COMMISSION ELECTROTECHNIQUE INTERNATIONALE

PRICE CODE CODE PRIX

ICS 11.040.55

ISBN 978-2-88912-911-9

Warning! Make sure that you obtained this publication from an authorized distributor. Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

 Registered trademark of the International Electrotechnical Commission Marque déposée de la Commission Electrotechnique Internationale

CONTENTS

FOREWO	PRD	4		
INTRODU	ICTION	6		
201.1	Scope, object and related standards	7		
201.2	Normative references	9		
201.3	Terms and definitions	9		
201.4	General requirements	11		
201.5	General requirements for testing of ME EQUIPMENT	12		
201.6	Classification of ME EQUIPMENT and ME SYSTEMS	12		
201.7	ME EQUIPMENT identification, marking and documents	12		
201.8	Protection against electrical HAZARDS from ME EQUIPMENT	13		
201.9	Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	14		
201.10	Protection against unwanted and excessive radiation HAZARDS	14		
201.11	Protection against excessive temperatures and other HAZARDS	14		
201.12	Accuracy of controls and instruments and protection against hazardous outputs	14		
201.13	HAZARDOUS SITUATIONS and fault conditions	38		
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	38		
201.15	Construction of ME EQUIPMENT	38		
201.16	ME SYSTEMS	39		
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	39		
202	Electromagnetic compatibility - Requirements and tests 41e3-9314-	40		
Annexes	9909fd823967/sist-en-60601-2-47-2015	42		
Annex AA	(informative) Particular guidance and rationale	43		
Bibliograp	bhy	64		
Index of d	lefined terms used in this particular standard	65		
Figure 20	1.101 – General test circuit for 201.12.4.4	29		
Figure 20	1.102 – Test signal for input dynamic range test according to 201.12.4.4.101	30		
Figure 20	1.103 – Test circuit for common mode rejection according to 201.12.4.4.103	33		
Figure 20	1.104 – Test circuit for pacemaker pulse tolerance according to			
201.12.4.	4.109	37		
Figure 20 radiated e	2.101 – Test set-up for conductive emission test according to 202.6.1.1.2 and emission and radiated immunity test according to 202.6.1.1.2 and 202.6.2.3.2	41		
Table 201	1.101 – Distributed additional ESSENTIAL PERFORMANCE requirements	11		
Table 201	1.102 – LEAD WIRE colour codes	13		
Table 201	.103 – Requirements for all arrhythmia algorithms	17		
Table 201	1.104 – Requirements for algorithms with optional capabilities	18		
Table 201.105 – Beat label classifications				
Table 201.106 – Example of noise floor calculation results 24				
Table 201.107 – Example of HRV test results				
Table 201	.108 – Run sensitivity summary matrix	25		
Table 201	1.109 – Run positive predictivity summary matrix	26		

- 3 -

Table AA.1 – Records to be included in a complete test	44
Table AA.2 – Example of a line-format, beat-by-beat performance report	48
Table AA.3 – Condensed beat-by-beat summary matrix containing 11 elements	49
Table AA.4 – Summary table (matrix format) of beat-by-beat comparison	49
Table AA.5 – Example of a line-format SHUTDOWN report	50
Table AA.6 – Example of a line-format report	51
Table AA.7 – Example of VF performance report	51
Table AA.8 – Example of false VF performance report	51
Table AA.9 – Example of a line-format couplet and run performance report	52
Table AA.10 – Example of device measurements of synthetic test patterns	53
Table AA.11 – Example of predicted ideal values for synthetic test patterns	54
Table AA.12 – Example of choice of test patterns	54
Table AA.13 – Example of RMS interval differences	57
Table AA.14 – Example of summary of frequency components	58

iTeh STANDARD PREVIEW (standards.iteh.ai)

- 4 -

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user. (standards.iteh.al)
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national for regional publication shall be clearly indicated in the latter. https://standards.iteh.ai/catalog/standards/sist/11d9743e-ca64-41e3-9314-
- 5) IEC itself does not provide any attestation of conformity independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-47 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2001. It constitutes a technical revision. This edition was revised to align structurally with the 2005 edition of IEC 60601-1.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/963/FDIS	62D/980/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

60601-2-47 © IEC:2012

- 5 -

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- Test specifications: italic type.
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type.
 Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this standard;
- SIST EN 60601-2-47:2015
 "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard; 60601-2-47-2015
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard concerns the basic safety and essential performance of AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS. It amends and supplements IEC 60601-1 (third edition 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard. The requirements of this particular standard take priority over those of the general standard.

A "General guidance and rationale" for the requirements of this particular standard is included in Annex AA.

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

iTeh STANDARD PREVIEW (standards.iteh.ai)

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS, hereafter referred to as ME SYSTEMS

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard. <u>SIST EN 60601-2-47:2015</u> https://standards.iteh.ai/catalog/standards/sist/11d9743e-ca64-41e3-9314-

Within the scope of this standard are systems of the following types:

- a) systems that provide continuous recording and continuous analysis of the ECG allowing full re-analysis giving essentially similar results. The systems may first record and store the ECG and analyse it later on a separate unit, or record and analyse the ECG simultaneously. The type of storage media used is irrelevant with regard to this standard;
- b) systems that provide continuous analysis and only partial or limited recording not allowing a full re-analysis of the ECG.

The safety aspects of this standard apply to all types of systems falling in one of the abovementioned categories.

If the AMBULATORY ELECTROCARDIOGRAPHIC SYSTEM offers automatic ECG analysis, minimal performance requirements for measurement and analysis functions apply. MEDICAL ELECTRICAL EQUIPMENT covered by IEC 60601-2-25 and IEC 60601-2-27 are excluded from the scope of this standard.

This standard does not apply to systems that do not continuously record and analyse the ECG (for example, 'intermittent event recorders').

201.1.2 Object

Replacement:

¹ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.*

- 8 -

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS.

201.1.3 Collateral Standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3, IEC 60601-1-8 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the standard addresses the content of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

60601-2-47 © IEC:2012

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

Amendment:

IEC 60601-1-2:2007, Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility -Requirements and tests.

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, apply, except as follows:

NOTE An index of defined terms is found beginning on page 64.

Additional definitions: iTeh STANDARD PREVIEW

(standards.iteh.ai)

201.3.201

AF

SIST EN 60601-2-47:2015

ATRIAL FIBRILLATION https://standards.iteh.ai/catalog/standards/sist/11d9743e-ca64-41e3-9314-

ECG rhythm involving either no P-waves and irregular RR intervals (atrial fibrillation) or high frequency flutter waves and regular or irregular RR intervals (atrial flutter)

201.3.202

AMBULATORY ELECTROCARDIOGRAPHIC SYSTEM

ME SYSTEM, AMBULATORY RECORDER and a PLAYBACK EQUIPMENT, both of which may contain an analysis function

Note 1 to entry: This ME SYSTEM is often referred to as Holter monitoring system after its inventor Dr. Norman Holter.

201.3.203

AMBULATORY RECORDER

recording ME EQUIPMENT worn or carried by the PATIENT including associated ELECTRODES and cables for recording heart action potentials

Note 1 to entry: An AMBULATORY RECORDER may also analyse the heart action potentials. It may record selectively when significant events are detected, or continuously.

201.3.204

CONTINUOUS RECORDER

ME EQUIPMENT, which performs continuous recording of the ECG

201.3.205 DATABASE

DB

sampled ECGs or artificial signals of one or more channels together with descriptive (clinical) information